
Živila - Določevanje aflatoksina v začimbah, razen v papriki, z IAC-čiščenjem in HPLC-FLD s postkolonsko derivatizacijo

Foodstuffs - Determination of aflatoxins in spices other than paprika by IAC clean-up and HPLC-FLD with post-column derivatization

Lebensmittel - Bestimmung von Aflatoxinen in Gewürzen außer Paprika mit IAC Reinigung und HPLC-FLD mit Nachsäulenderivatisierung

Produits alimentaires - Détermination des aflatoxines dans les épices (pour lesquelles un niveau maximal a été établi par l'UE) autres que le paprika

<https://standards.iteh.ai/catalog/standards/sist/af07dfc6-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021>

Ta slovenski standard je istoveten z: EN 17424:2020

ICS:

67.050	Splošne preskusne in analizne metode za živilske proizvode	General methods of tests and analysis for food products
67.220.10	Začimbe	Spices and condiments

SIST EN 17424:2021**en,fr,de**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 17424:2021

<https://standards.iteh.ai/catalog/standards/sist/af07df66-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021>

EUROPEAN STANDARD

EN 17424

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2020

ICS 67.050

English Version

Foodstuffs - Determination of aflatoxins in spices other than paprika by IAC clean-up and HPLC-FLD with post-column derivatization

Produits alimentaires - Dosage des aflatoxines dans les épices autres que le paprika par purification sur colonne d'immunoaffinité et CLHP-FLD avec dérivation post-colonne

Lebensmittel - Bestimmung von Aflatoxinen in Gewürzen außer Paprika mit IAC Reinigung und HPLC-FLD mit Nachsäulenderivatisierung

This European Standard was approved by CEN on 14 September 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Introduction	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions.....	5
4 Principle.....	5
5 Reagents.....	6
6 Apparatus and equipment.....	10
7 Procedure	11
7.1 Extraction of aflatoxins from the sample	11
7.2 Determination of recovery	12
7.3 IAC clean-up	12
7.4 HPLC-FLD analysis.....	12
7.4.1 General.....	12
7.4.2 Confirmation.....	12
7.4.3 Post-column derivatization.....	13
7.4.4 Identification.....	13
7.4.5 Calibration.....	13
8 Calculation.....	13
9 Precision	14
9.1 General.....	14
9.2 Repeatability	14
9.3 Reproducibility.....	15
10 Test report.....	16
Annex A (informative) Example conditions for suitable HPLC-FLD systems	17
A.1 General.....	17
A.2 Example HPLC Conditions.....	17
A.3 Other examples of suitable columns and mobile phase conditions	17
Annex B (informative) Typical chromatograms	19
Annex C (informative) Precision data	24
Annex D (informative) Supplementary information.....	27
Bibliography	31

European foreword

This document (EN 17424:2020) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2021, and conflicting national standards shall be withdrawn at the latest by May 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 17424:2021](https://standards.iteh.ai/catalog/standards/sist/af07dfe6-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021)

<https://standards.iteh.ai/catalog/standards/sist/af07dfe6-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021>

Introduction

Aflatoxins consist of a group of approximately twenty related fungal metabolites, although only aflatoxins B₁, B₂, G₁ and G₂ are normally found in foods. Aflatoxins B₂ and G₂ are the dihydro derivatives of the parent compounds. Aflatoxins are produced by at least three species of *Aspergillus*, *A. flavus*, *A. parasiticus* and *A. nomius*, and can occur in a wide range of important raw food commodities, including cereals, nuts, spices, figs and dried fruit.

WARNING 1 — Suitable precaution and protection measures need to be taken when carrying out working steps with harmful chemicals. The latest version of the hazardous substances ordinance, Regulation (EC) No 1907/2006 [3], should be taken into account as well as appropriate national statements e.g. such as in [4].

WARNING 2 — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

WARNING 3 — Aflatoxins are known to have carcinogenic effects and to be both acutely and chronically toxic.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 17424:2021](https://standards.iteh.ai/catalog/standards/sist/af07dfe6-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021)

<https://standards.iteh.ai/catalog/standards/sist/af07dfe6-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021>

1 Scope

This document describes a procedure for the determination of aflatoxins B₁, B₂, G₁ and G₂ and total aflatoxins (sum of B₁, B₂, G₁ and G₂) in spices for which EU maximum levels are established, other than paprika, by high performance liquid chromatography (HPLC) with post-column derivatization (PCD) and fluorescence detection (FLD) after immunoaffinity column (IAC) clean-up.

The method is applicable to the spices capsicum (excluding paprika), pepper, nutmeg, ginger, turmeric and mixtures thereof.

The method has been validated for aflatoxins B₁, B₂, G₁ and G₂ and total aflatoxins in a range of test samples that comprised: ginger, pepper, nutmeg, chilli, turmeric as individual spices and mixed pepper + chilli + nutmeg (90 + 5 + 5, $m + m + m$), mixed spice+ginger (6 + 4, $m + m$) mixed spice, mixed turmeric+ginger (2 + 8, $m + m$).

The validation was carried out over the following concentration ranges: aflatoxin B₁ = 1 µg/kg to 16 µg/kg and total aflatoxins = 2,46 µg/kg to 36,1 µg/kg.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

3 Terms and definitions (standards.iteh.ai)

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

4 Principle

Aflatoxins are extracted from the spices with a mixture of methanol, acetonitrile and water. The sample extract is filtered, diluted with phosphate buffered saline (PBS) and applied to an immunoaffinity column containing antibodies specific to aflatoxins B₁, B₂, G₁ and G₂. The aflatoxins are eluted from the IAC and are quantified by reversed-phase HPLC with PCD followed by FLD.

5 Reagents

Use only reagents of recognized analytical grade, p.a. (pro analysis) and water complying with grade 1 of EN ISO 3696, unless otherwise specified. Solvents shall be of quality for HPLC analysis, unless otherwise specified.

WARNING 1 — Decontamination procedures for laboratory wastes of aflatoxins were developed by the International Agency for Research on Cancer (IARC) [5], [6].

WARNING 2 — Aflatoxins are subject to light degradation. Protect the laboratory, where the analyses are done, adequately from daylight. This can be achieved effectively by using Ultraviolet (UV) absorbing foil on the windows in combination with subdued light (no direct sunlight) or curtains or blinds in combination with artificial light (fluorescent tubes are acceptable). Protect aflatoxin containing solutions from light as much as possible (keep in the dark, use aluminium foil or amber-coloured glassware) and store at the temperature recommended by the manufacturer (e.g. $-18\text{ }^{\circ}\text{C}$).

5.1 Methanol (CH_3OH)

5.2 Acetonitrile (CH_3CN)

5.3 Potassium bromide (KBr)

5.4 Sodium hydroxide (NaOH)

5.5 Sodium hydroxide solution, substance concentration $c(\text{NaOH}) = 2\text{ mol/l}$

Dissolve 8,0 g of sodium hydroxide (5.4) in 100 ml of water.

5.6 Sodium hydroxide solution, $c(\text{NaOH}) = 0,1\text{ mol/l}$

Dissolve 0,4 g of sodium hydroxide (5.4) in 100 ml of water.

5.7 Hydrochloric acid solution (HCl), volume fraction $\varphi(\text{HCl}) = 37\%$ (acidimetric)

5.8 Hydrochloric acid solution, $c(\text{HCl}) = 0,1\text{ mol/l}$

Dilute 8,28 ml of hydrochloric acid solution (5.7) to 1 l with water.

5.9 Sulfuric acid solution, for rinsing glassware, $c(\text{H}_2\text{SO}_4) = 2\text{ mol/l}$

5.10 Sodium chloride (NaCl)

5.11 Potassium chloride (KCl)

5.12 Potassium dihydrogen phosphate (KH_2PO_4)

5.13 Disodium hydrogen orthophosphate (Na_2HPO_4)

5.14 Phosphate buffered saline (PBS), pH = 7,4

Weigh 0,20 g of potassium chloride (5.11), 0,20 g of potassium dihydrogen phosphate (5.12), 1,16 g of disodium hydrogen orthophosphate (5.13) and 8,00 g of sodium chloride (5.10) to the nearest 0,01 g and transfer into a 1 l volumetric flask. Dissolve in water and add 900 ml of water.

After dissolution adjust the pH to 7,4 with hydrochloric acid solution (5.8) or sodium hydroxide solution (5.6) as appropriate, then fill up to the mark with water.

Alternatively, a PBS solution with equivalent properties may be prepared from commercially available PBS material.

5.15 Polysorbate 20, e.g. Tween® 20¹, lauric acid ≥ 40 %

5.16 PBS/Polysorbate 20 solution

Mix Polysorbate 20 (5.15) with PBS (5.14) (1 + 9, V + V).

5.17 Extraction solution

Mix acetonitrile (5.2), methanol (5.1) and water (40 + 35 + 25, V + V + V).

5.18 Nitric acid solution (HNO₃), c(HNO₃) = 4 mol/l

Slowly add 25,5 ml of concentrated HNO₃ (70 %, mass concentration $\rho(\text{HNO}_3) = 1,413 \text{ g/ml}$ at 25 °C) to 25 ml water, adjust the final volume to 100 ml with water.

5.19 Pyridinium bromide perbromide solution (PBPB), $\rho(\text{PBPB}) = 50 \text{ mg/l}$

Weigh 50 mg ± 10 mg of PBPB into a 20 ml glass vial. Dissolve in approximately 20 ml water, use an ultra-sonic bath if required, and transfer quantitatively into a 1 l amber glass bottle. Fill up to 1 l with water. The solution can be stored in a dark place at room temperature for four days.

5.20 Aflatoxin B₁ e.g. crystalline, as a film or as certified standard solution

5.21 Aflatoxin B₂ e.g. crystalline, as a film or as certified standard solution

5.22 Aflatoxin G₁ e.g. crystalline, as a film or as certified standard solution

5.23 Aflatoxin G₂ e.g. crystalline, as a film or as certified standard solution

5.24 Aflatoxins (B₁, B₂, G₁ and G₂) stock solutions

Dissolve aflatoxin B₁ (5.20), B₂ (5.21), G₁ (5.22) and G₂ (5.23) separately in acetonitrile (5.2) to give separate solutions with a mass concentration of 10 µg/ml for each aflatoxin. Transfer the stock solutions to amber vials and store them below 4 °C when not in use.

¹ Tween 20 is a trade name of a polysorbate 20-type nonionic surfactant available from various suppliers. This information is given for the convenience of users of this European standard and does not constitute an endorsement by CEN of this product. Equivalent products may be used if they can be shown to lead to the same results.

EN 17424:2020 (E)

To determine the exact mass concentration of aflatoxins in each stock solution, record the absorption curve between a wavelength of 330 nm and 370 nm in 1 cm quartz glass cells using a UV-spectrometer (6.5) with acetonitrile (5.2) as reference. Calculate the mass concentration of each aflatoxin, ρ , in $\mu\text{g/ml}$, according to Formula (1):

$$\rho = \frac{A_{\max} \times M \times 100}{\delta \times \varepsilon} \quad (1)$$

where

A_{\max} is the maximum extinction value determined from the absorption curve (here determined at a wavelength of 350 nm);

M is the molar mass of each aflatoxin, in g/mol;

δ is the path length of the quartz cell, in cm;

ε is the molar absorption coefficient of each aflatoxin in acetonitrile (5.2), in m^2/mol .

M and ε of aflatoxins B₁, B₂, G₁ and G₂ are given in Table 1 [7].

Table 1 — Molar mass and molar absorption coefficient of aflatoxins B₁, B₂, G₁ and G₂ in acetonitrile (5.2)

Aflatoxin	M g/mol	ε m^2/mol
B ₁	312	2 070
B ₂	314	2 250
G ₁	328	1 760
G ₂	330	1 890

5.25 Mixed stock solution

Prepare a mixed stock solution containing 1 000 ng/ml of aflatoxin B₁ and G₁, 500 ng/ml of aflatoxin B₂ and G₂ in acetonitrile (5.2) by appropriate dilution of aflatoxins (B₁, B₂, G₁ and G₂) stock solutions (5.24).

A certified mixed aflatoxins stock standard solution which is ready to use may be used as an alternative.

Protect the solution from light and store below 4 °C when not in use. This solution can be stored under these conditions for two months.

5.26 Mixed intermediate solution

Prepare a mixed intermediate solution by pipetting 2,0 ml of the mixed stock solution (5.25) into a 10 ml volumetric flask. Dilute to the mark with acetonitrile (5.2) and shake well. The concentration of this mixed intermediate solution is 200 ng/ml of aflatoxin B₁ and G₁, and 100 ng/ml of aflatoxin B₂ and G₂.

Protect the solution from light and store below 4 °C when not in use. This solution can be stored under these conditions for two months.

5.27 Mixed standard solution

Pipette 1,0 ml of the mixed intermediate solution (5.26) into a 10 ml volumetric flask, fill to the mark with the acetonitrile (5.2) and mix well to give a solution containing 20 ng/ml of aflatoxin B₁ and G₁, 10 ng/ml of aflatoxin B₂ and G₂ for preparation of calibration solutions.

Protect the solution from light and store below 4 °C when not in use. This solution can be stored under these conditions for two months.

5.28 Calibration solutions

Add different volumes of the mixed standard solution (5.27) to five 10 ml volumetric flasks to obtain five calibration levels across the calibration range. The values in Table 2 were used in the validation study.

Evaporate the acetonitrile until dryness under a stream of nitrogen at room temperature. To each flask, add 4 ml of methanol (5.1), dissolve the aflatoxins by mixing, then dilute to 10 ml with water, and shake well. Methanol and water are subject to volume contraction when mixed, so adjust the volume with water again to the given volume.

These calibration solutions cover the range from 0,48 µg/kg to 9,6 µg/kg for aflatoxins B₁ and G₁ and the range from 0,24 µg/kg to 4,8 µg/kg for B₂ and G₂.

Table 2 — Example preparation of HPLC calibration solutions

Calibration solution	Mixed standard solution (5.27) µl	Targeted mass concentration per analyte			
		B ₁ ng/ml	B ₂ ng/ml	G ₁ ng/ml	G ₂ ng/ml
1	10	0,02	0,01	0,02	0,01
2	20	0,04	0,02	0,04	0,02
3	50	0,10	0,05	0,10	0,05
4	100	0,20	0,10	0,20	0,10
5	200	0,40	0,20	0,40	0,20

These solutions should be freshly made on each day of analysis.

5.29 HPLC mobile phase solvent A

Mix water, acetonitrile (5.2) and methanol (5.1) (56 + 30 + 14, V + V + V). Degas the solution before use if an online system is not available on the HPLC (6.13).

5.30 HPLC mobile phase solvent B

Mix water, acetonitrile (5.2) and methanol (5.1) (56 + 30 + 14, V + V + V). Add 350 µl of nitric acid solution (5.18) and 119 mg of potassium bromide (5.3) per litre of mobile phase. Degas the solution before use if an online system is not available on the HPLC (6.13).

5.31 Immunoaffinity column

The IAC contains antibodies raised against aflatoxins B₁, B₂, G₁ and G₂. The column shall have a capacity of not less than 100 ng of aflatoxin B₁.

6 Apparatus and equipment

The use of non-acid washed glassware may cause losses of aflatoxins. It is recommended that all glassware coming into contact with aqueous solutions of aflatoxins should be washed with acid solution before use. Many laboratory washing machines do this as part of the washing program. Otherwise soak such laboratory glassware in sulfuric acid (5.9) for several h (e.g. 15 h overnight), then rinse well (e.g. three times) with water to remove all traces of acid. Check the absence of acid with pH paper.

In practice, the treatment is necessary for round bottomed flasks, volumetric flasks, measuring cylinders, vials or tubes used for calibration solutions and final extracts (particularly autosampler vials), and Pasteur pipettes, if these are used to transfer calibration solutions or extracts.

Usual laboratory apparatus and, in particular, the following:

- 6.1 **Laboratory balance**, accuracy of 0,01 g.
- 6.2 **Analytical balance**, accuracy of 0,1 mg.
- 6.3 **Centrifugation bottle**, e.g. 250 ml.
- 6.4 **Centrifuge**, suitable for relative centrifugal force of 2 000 *g* to 2 500 *g*.

NOTE $g = 9,81 \text{ m} \cdot \text{s}^{-2}$.

- 6.5 **UV-spectrometer with quartz cuvettes**.

- 6.6 **Laboratory blender**, e.g. Ultra Turrax^{®2}.

- 6.7 **Laboratory shaker**, adjustable.

- 6.8 **pH meter**.

SIST EN 17424:2021
<https://standards.iteh.ai/catalog/standards/sist/af07dfe6-6d89-4fde-aba6-6348d6ad48f8/sist-en-17424-2021>

- 6.9 **Filter paper**, e.g. 190 mm diameter, pre-folded, wet strengthened, nominal particle retention rating of 30 μm .

- 6.10 **Pipettes**, 25 μl , 50 μl , 250 μl , 1 ml, 10 ml capacity.

- 6.11 **Syringe**, plastic disposable, 5 ml capacity.

- 6.12 **Syringe filter**, 0,45 μm polytetrafluoroethylene (PTFE).

- 6.13 **HPLC-FLD system**, with the following components:

- 6.13.1 **HPLC pump**, suitable for flow rate at 1,0 ml/min.

- 6.13.2 **Injection system**, capable for total loop injection (a 100 μl loop is recommended).

- 6.13.3 **HPLC column**, e.g. C18 or ODS-1 (length of 250 mm, inner diameter of 4,6 mm and particle size of 5 μm), which ensures a baseline resolution of the aflatoxin B₁, B₂, G₁ and G₂ peaks from all other peaks. A suitable pre-column shall be used.

² Ultra Turrax[®] is a trade name of a product commercially available from various suppliers. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of the product named. Equivalent products may be used if they can be shown to lead to the same results.