



**SLOVENSKI STANDARD**  
**oSIST prEN 17425:2019**

**01-oktober-2019**

---

**Živila - Določevanje alkaloidov rženih rožičkov (ergot) v žitu in žitnih proizvodih s  
čiščenjem dSPE in LC-MS/MS**

Foodstuffs - Determination of ergot alkaloids in cereals and cereal products by dSPE  
clean-up and LC-MS/MS

Lebensmittel - Bestimmung von Ergotalkaloiden in Getreiden und Getreideerzeugnissen  
mit dSPE-Reinigung und LC-MS/MS

**ITEH STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN 17425:2019](https://standards.iteh.ai/catalog/standards/sist/17425-2019/c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019)

**Ta slovenski standard je istoveten z: prEN 17425**

---

**ICS:**

67.050	Splošne preskusne in analizne metode za živilske proizvode	General methods of tests and analysis for food products
67.060	Žita, stročnice in proizvodi iz njih	Cereals, pulses and derived products

**oSIST prEN 17425:2019**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN 17425:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**DRAFT**  
**prEN 17425**

August 2019

ICS 67.050; 67.060

English Version

## Foodstuffs - Determination of ergot alkaloids in cereals and cereal products by dSPE clean-up and LC-MS/MS

Lebensmittel - Bestimmung von Ergotalkaloiden in  
Getreiden und Getreideerzeugnissen mit dSPE-  
Reinigung und LC-MS/MS

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 275.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

**Warning** : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	<b>Page</b>
European foreword .....	4
Introduction .....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions.....	6
4 Principle.....	6
5 Reagents.....	6
6 Apparatus and equipment.....	8
7 Procedure .....	10
7.1 Preparation of the test sample.....	10
7.2 Extraction of ergot alkaloids.....	10
7.2.1 Precautions .....	10
7.2.2 Test sample .....	10
7.2.3 Sample extraction.....	10
7.2.4 Spiking procedure .....	10
7.2.5 Clean-up.....	10
8 LC-MS/MS analysis.....	11
8.1 General.....	11
8.2 Batch composition and analytical sequence.....	11
8.3 Identification.....	11
8.4 Calibration.....	11
9 Calculation.....	12
9.1 Calculation of individual ergot alkaloid concentration .....	12
9.2 Calculation of ergot alkaloid sum concentration .....	12
10 Precision .....	12
10.1 General.....	12
10.2 Repeatability .....	12
10.3 Reproducibility.....	12
11 Test report.....	13
Annex A (informative) Example conditions for suitable LC-MS/MS systems .....	14
A.1 System settings for Waters® Xevo® TQ-S Mass Spectrometer.....	14
A.1.1 Settings for chromatography.....	14
A.1.2 Detector parameters.....	15
A.2 System settings for Micromass® Quattro Ultima Pt Mass Spectrometer .....	17
A.2.1 Settings for chromatography.....	17
A.2.2 Detector parameters.....	17
A.3 System settings for SCIEX API 5500 Mass Spectrometer <sup>5</sup> .....	19
A.3.1 Settings for chromatography.....	19

<b>A.3.2</b>	<b>Detector parameters</b> .....	<b>20</b>
<b>A.4</b>	<b>Other LC conditions</b> .....	<b>21</b>
<b>Annex B</b>	<b>(informative) Typical chromatograms</b> .....	<b>23</b>
<b>Annex C</b>	<b>(informative) Precision data</b> .....	<b>29</b>
<b>Annex D</b>	<b>(informative) Data comparison</b> .....	<b>43</b>
<b>Bibliography</b>	.....	<b>44</b>

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

[oSIST prEN 17425:2019](https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019)

<https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019>

**prEN 17425:2019 (E)**

## **European foreword**

This document (prEN 17425:2019) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This document is currently submitted to the CEN Enquiry.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN 17425:2019](https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019)

<https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019>

## Introduction

Ergot alkaloids are a group of mycotoxins produced by several species of *Claviceps* fungi growing on cereals and forage grass. These toxins are a risk for consumers as they can enter the food chain. All ergot alkaloids share a common structure, the ergoline system, and are divided into several classes, based on the presence of functional groups. The chiral carbon atom C-8 is responsible for the epimerization.

The isomers of each of these compounds are nominally known as the 'inines'.

**WARNING 1 — Suitable precaution and protection measures need to be taken when carrying out working steps with harmful chemicals. The latest version of the hazardous substances ordinance, Regulation (EC) No 1907/2006 [3], should be taken into account as well as appropriate national statements.**

**WARNING 2 — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.**

**WARNING 3 - Ergot alkaloids can cause vasoconstrictive, neurotoxic, reproductive and developmental adverse effects, and can be acutely and chronically toxic [4].**

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN 17425:2019](https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019)

<https://standards.iteh.ai/catalog/standards/sist/4097e30c-75a1-4b73-94d9-34e0d6d7a4b1/osist-pren-17425-2019>

**prEN 17425:2019 (E)****1 Scope**

This document describes a method for the determination of the sum total of six ergot alkaloids (ergocornine, ergometrine, ergocristine, ergotamine, ergosine and ergocryptine) and their -inine epimer pairs by liquid chromatography coupled with tandem mass spectrometry (LC-MS/MS) after clean-up by dispersive solid phase extraction (SPE).

The method has been validated for cereals and cereal-based food products.

The method has been validated in the range 13,2 µg/kg to 168 µg/kg for the sum of the twelve ergot alkaloids, in rye flour, rye bread and cereal products (breakfast cereal, infant breakfast cereal, and crispbread) that contained rye as an ingredient, as well as seeded wholemeal flour and a barley and rye flour mixture.

Method performance was satisfactory in the range 24,1 µg/kg to 168 µg/kg, however at lower concentrations RSD<sub>R</sub> values were greater than 44 %, and HorRat values exceeded 2,0, indicating the method may not be fully suitable at concentrations below 24 µg/kg for sum of ergot alkaloids, although it is suitable for screening at these concentrations. Method performance may be improved by inclusion of an isotopically labelled internal standard, but this was not available at the time of the method validation study.

**2 Normative references**

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

oSIST prEN 17425:2019

**3 Terms and definitions**

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

**4 Principle**

Ergot alkaloids are extracted from cereals and cereal-based foods with buffer at pH 8,9 and cleaned up with a dispersive solid phase material prior to filtering. The ergot alkaloids are quantified by liquid chromatography coupled to tandem mass spectrometry (LC-MS/MS).

**5 Reagents**

Use only reagents of recognized analytical grade and water complying with grade 1 of EN ISO 3696, unless otherwise specified. Solvents shall be of quality for LC analysis, unless otherwise specified.

NOTE Ergometrine and ergotamine are listed as Category 1 scheduled substances in Regulation (EC) No 273/2004 [5] on drug precursors. It is a requirement to have an appropriate licence in order to purchase and store these compounds (and their related -inine epimers).

**5.1 Ergocornine, e.g. crystalline, as a film or as certified standard solution.**



- 5.2 **Ergocorninine**, e.g. crystalline, as a film or as certified standard solution.
- 5.3 **Ergocristine**, e.g. crystalline, as a film or as certified standard solution.
- 5.4 **Ergocristinine**, e.g. crystalline, as a film or as certified standard solution.
- 5.5  **$\alpha$ -Ergocryptine**, e.g. crystalline, as a film or as certified standard solution.
- 5.6  **$\alpha$ -Ergocryptinine**, e.g. crystalline, as a film or as certified standard solution.
- 5.7 **Ergometrine (maleate)**, e.g. crystalline, as a film or as certified standard solution.
- 5.8 **Ergometrinine**, e.g. crystalline, as a film or as certified standard solution.
- 5.9 **Ergosine**, e.g. crystalline, as a film or as certified standard solution.
- 5.10 **Ergosinine**, e.g. crystalline, as a film or as certified standard solution.
- 5.11 **Ergotamine (tartrate)**, e.g. crystalline, as a film or as certified standard solution.
- 5.12 **Ergotaminine**, e.g. crystalline, as a film or as certified standard solution.
- 5.13 **Acetonitrile**, LC-MS grade.
- 5.14 **Methanol**, LC grade.
- 5.15 **Water**, suitable for LC-MS/MS.
- 5.16 **Solid phase extraction material, primary secondary amine (PSA) bulk sorbent**, 40  $\mu\text{m}$ , 10 g; e.g. Bondesil™, 1.
- 5.17 **Sodium hydroxide (NaOH)**, analytical reagent grade.
- 5.18 **Sodium hydroxide solution**, substance concentration  $c(\text{NaOH}) = 1,0 \text{ mol/l}$ .
- Dissolve 4 g NaOH (5.17) in water (5.15) to a final volume of 100 ml.
- 5.19 **Hydrochloric acid (HCl)**, analytical reagent grade, volume fraction  $\varphi(\text{HCl}) = 37 \%$  (acidimetric), density: 1,18 g/cm<sup>3</sup> (20 °C).
- 5.20 **Hydrochloric acid solution**,  $c(\text{HCl}) = 1,0 \text{ mol/l}$ .
- Dilute 8,3 ml HCl (5.19) in water (5.15) to a final volume of 100 ml.
- 5.21 **Ammonium carbonate ((NH<sub>4</sub>)<sub>2</sub>CO<sub>3</sub>)**, LC-MS grade.
- 5.22 **Ammonium carbonate solution**, mass concentration  $\rho((\text{NH}_4)_2\text{CO}_3) = 200 \text{ mg/l}$ ; pH = 8,9 ± 0,3.
- Weigh 200 mg of ammonium carbonate (5.21) to the nearest 2 mg and transfer into a 1 l glass laboratory screw top flask. Add 500 ml of water (5.15). Shake the flask vigorously to ensure all solid has been dissolved.

<sup>1</sup> Bondesil™ is a trade name of a product commercially available from various suppliers. This information is given for the convenience of users of this European standard and does not constitute an endorsement by CEN of the products named. Equivalent products may be used if they can be shown to lead to the same results.

**prEN 17425:2019 (E)**

After dissolution adjust the pH  $8,9 \pm 0,3$  with sodium hydroxide solution (5.18) or hydrochloric acid solution (5.20) as appropriate, then fill up to the mark with water (5.15).

This solution can be stored at room temperature for 3 months.

Alternatively, a ready to use ammonium carbonate solution ( $c((\text{NH}_4)_2\text{CO}_3) = 3,03 \text{ mmol/l}$ ) of the correct analytical grade may be used ensuring the pH  $8,9 \pm 0,3$ .

**5.23 Extraction solvent.**

Mix acetonitrile (5.13) and ammonium carbonate solution (5.22) (84 + 16, v + v). Shake vigorously.

**5.24 Individual stock solutions,  $\rho = 100 \mu\text{g/ml}$ , in acetonitrile.**

Follow any specific manufacturers' instructions to re-dissolve the films of individual ergot alkaloids [5.1 to 5.12]. If crystalline material is used, weigh 5 mg to the nearest 0,2 mg, of each of the solid standards (5.1 to 5.12) individually into glass weighing boats and transfer quantitatively into individual 50 ml volumetric flasks, then fill up to the mark with acetonitrile (5.13).

**5.25 Mixed standard solution,  $\rho = 0,5 \mu\text{g/ml}$ .**

Using a pipette (6.5), transfer 50  $\mu\text{l}$  of each of the individual stock solutions (5.24) into a 10 ml volumetric flask and make up to volume with acetonitrile (5.13).

When using an individual stock solution with a mass concentration other than  $\rho = 100 \mu\text{g/ml}$  calculate the appropriate volume required to prepare the 0,5  $\mu\text{g/ml}$  standard solution mixture.

**5.26 Calibration solutions:**

Prepare e.g. the following calibration solutions as outlined in Table 1. Dispense volumes of mixed standard solution (5.25) into volumetric flasks and fill up to the mark with acetonitrile (5.13).

**Table 1 — Examples of suitable calibration solutions**

Calibration solution	Mixed standard solution (5.25)	Final volume	Mass concentration of alkaloids	Equivalent to mass fraction of alkaloids
	$\mu\text{l}$	ml	ng/ml	$\mu\text{g/kg}$
1	10	50	0,1	0,5
2	20	50	0,2	1
3	10	5	1	5
4	20	5	2	10
5	40	5	4	20
6	100	5	10	50

**6 Apparatus and equipment**

Usual laboratory apparatus and, in particular, the following. Unless otherwise stated volumetric glassware shall be of grade 'A' quality. Ergot alkaloids are sensitive to epimerisation by light and amber glass shall be used where possible.

**6.1 Laboratory balance**, accuracy of 0,01 g.

**6.2 Analytical balance**, accuracy of 0,1 mg.

**6.3 Single or multiple grinding mill.**

**6.4 Extraction flasks with cap**, of sufficient volume to contain 50 ml extraction solvent and 10 g sample, e.g. 100 ml.

**6.5 Pipette, adjustable, e.g.** 25 µl, 50 µl, 250 µl, 1 000 µl, suitable for organic solvents, with disposable tips.

**6.6 Laboratory shaker**, for solvent extraction with suitable 100 ml flasks.

**6.7 Folded filter paper**, diameter 12,5 cm, hardened, grade 54.

**6.8 Vials with caps**, 40 ml amber vials.

**6.9 Vials with caps**, 4,0 ml amber vials.

**6.10 Plastic Luer-lock syringe**, 1 ml.

**6.11 Polytetrafluoroethylene (PTFE) plastic syringe filters**, 13 mm × 0,22 µm.

**6.12 Vials with caps**, 2,0 ml amber vials.

**6.13 Autosampler vials suitable for LC-MS/MS analysis**, e.g. 200 µl.

**6.14 LC-MS/MS system with the following components:**

**6.14.1 LC pump**, capable of delivering a binary gradient at flow rates appropriate for the analytical column in use with sufficient accuracy.

**6.14.2 Injection system**, capable of injecting an appropriate volume of injection solution with sufficient accuracy.

**6.14.3 LC column**, capable of retaining the ergot alkaloids, preferably with a retention factor of at least two. The column shall be suitable for use with a mobile phase at pH > 7.

**NOTE** Under the conditions given it can be possible to separate α- and β-ergocryptine. This is not critical for the application of the method as results are reported as total ergocryptine using α-ergocryptine for quantification. If it is not possible to separate α- and β-ergocryptine using these conditions, a single peak will be measured that should be quantified with α-ergocryptine standard.

**6.14.4 Column filter**, in-line filter suitable for the LC column used (6.14.3).

**6.14.5 Column oven**, capable of maintaining a constant temperature.

**6.14.6 Tandem mass spectrometer (MS/MS)**, capable of ionization of the ergot alkaloids (resulting in positive ions) and selected reaction monitoring (SRM) with a sufficiently wide dynamic range.

Any ionization source providing sufficient yield may be used.

**6.14.7 Data evaluation system.**

**prEN 17425:2019 (E)****7 Procedure****7.1 Preparation of the test sample**

Grind and homogenize the sample with a grinding mill with a mesh or sieve size of 0,5 mm or smaller (6.3) before analysis.

Depending on the starting material (ground or unground material), it is advisable to first grind the sample through a sieve of 1 mm to prevent excessive heat formation during milling, which could lead to partial decomposition of the analytes. Then grind the sample through a sieve of 0,5 mm.

Mix samples well before taking a test portion for analysis. Store the samples at room temperature.

**7.2 Extraction of ergot alkaloids****7.2.1 Precautions**

Ergot alkaloids are sensitive to epimerization by light and amber glass shall be used where possible. Analyse the samples immediately after extraction. Only if absolutely necessary, extracts may be stored overnight at 4 °C.

**7.2.2 Test sample**

Weigh 10 g of the sample to the nearest 0,05 g into a flask (6.4).

A larger test sample size may be used. In that case the amount of extraction solvent shall be adjusted accordingly (7.2.3).

**7.2.3 Sample extraction**

Add 50 ml of the extraction solvent (5.23) to the flask (6.4) using a measuring cylinder.

Place the sample flasks in the shaker (6.6). Shake the samples for approximately 30 min at a moderate speed.

Whilst the samples are shaking, prepare sufficient glass funnels containing folded filter paper (6.7) ready to filter the samples into 40 ml amber vials (6.8).

When the samples have finished shaking, shake each flask individually by hand for approximately 10 s prior to pouring through funnels and filter paper (6.7) into 40 ml amber vials (6.8).

**7.2.4 Spiking procedure**

Add 200 µl of the mixed standard solution (5.25) to 10 g ± 0,05 g of a 'blank' sample. Allow the spiked samples to dry. At least one blank and a spike of the appropriate sample shall be included with each batch. Extract the samples as described in 7.2.3. If a larger test sample size has been used adjust the spiking volume accordingly.

**7.2.5 Clean-up**

Transfer 1 ml of sample filtrate into a 4 ml amber vial (6.9) containing 50 mg ± 5 mg of the solid phase extraction material (5.16).

Shake each tightly sealed sample vial with a laboratory shaker (6.6) at high speed for approximately 45 s.

Using a plastic Luer-lock syringe (6.10), take up as much as possible of the sample. Fit a 13 mm PTFE 0,22 µm syringe filter (6.11) and holding the syringe vertically, allow any solid phase material to rest on the bottom of the syringe. Push the liquid through the filter into a 2 ml amber vial (6.12).

## 8 LC-MS/MS analysis

### 8.1 General

Optimize analytical parameters (i.e. selection of masses of precursor and product ions, cone voltages and collision energies) by infusion and injection of standard solutions of individual ergot alkaloids. Use a tandem mass spectrometer or equivalent in positive electrospray ionization (ESI<sup>+</sup>) mode. Set the acquisition mode to multiple reaction monitoring (MRM), monitoring at least two product ions [6]. Examples are given in Table A.3.

Satisfactory separation of ergot alkaloid epimers can be achieved using a mobile phase with a pH > 7. The use of a mobile phase with pH > 7, requires an analytical column that contains a stationary phase that is resistant to high pH.

Examples of measurement conditions and transitions are given in Annex A. Examples of typical chromatograms are shown in Annex B.

### 8.2 Batch composition and analytical sequence

An example is as follows: The first injection of every run sequence (following any test or priming injections) is usually an aliquot of sample extraction solvent (5.23). Then inject the calibration solutions, followed by an extraction solvent to check for possible carry over. Subsequently inject the sample test solutions, inject one calibration solution at periodic intervals, e.g. one calibration solution for every 5 to 10 sample test solutions injected. At the end of the batch, re-inject the calibration solutions.

### 8.3 Identification

Identify each mycotoxin by comparing the retention times of the calibration solutions with that of the sample test solution. Identify the analyte on the basis of at least two mass transitions. In addition the retention times (peaks in both mass traces) and the area ratio of the two peaks shall match that of the standard substance [6].

As long as standards for  $\beta$ -ergocryptine and  $\beta$ -ergocryptinine are not available, use standards for  $\alpha$ -ergocryptine and  $\alpha$ -ergocryptinine to quantify both the  $\alpha$ - and  $\beta$ -ergocryptine and  $\alpha$ - and  $\beta$ -ergocryptinine and report a sum for  $\alpha$ - and  $\beta$ -ergocryptine and  $\alpha$ - and  $\beta$ -ergocryptinine.

### 8.4 Calibration

For each ergot alkaloid, plot the peak areas of the quantifier ion (y-axis) of all individual calibration solutions (5.26, calibration solutions 1 to 6) against the corresponding mass fraction ( $\mu\text{g}/\text{kg}$ ) (x-axis). The quantifier is the transition which overall gives the best S/N (signal/noise) ratio. Construct a calibration curve using (possibly weighted) regression with all individual data points obtained, estimate the slope and possible intercept of each of the calibration curves. If higher deviations or nonlinearity is observed, identify the cause and, if necessary, re-run the analyses.