

# SLOVENSKI STANDARD

## SIST EN 13016-3:2018

01-junij-2018

Nadomešča:

SIST EN 13016-1:2007

---

**Tekoči naftni proizvodi - Parni tlak - 3. del: Določevanje parnega tlaka in enakovrednega parnega tlaka suhega zraka (DVPE) (z uporabo trojne ekspanzijske metode)**

Liquid petroleum products - Vapour pressure - Part 3: Determination of vapour pressure and calculated dry vapour pressure equivalent (DVPE) (Triple Expansion Method)

**iTeh STANDARD PREVIEW**

Flüssige Mineralölerzeugnisse - Dampfdruck - Teil 3: Bestimmung des Dampfdruckes und des berechneten Trockendampfdruckes (BTDD) (Dreifache Expansions Methodik)

SIST EN 13016-3:2018

Produits pétroliers liquides - Pression de vapeur - Partie 3 - Détermination de la pression de vapeur saturée et de la pression de vapeur sèche calculée (PVSC) (Méthode Expansion Triplique)

**Ta slovenski standard je istoveten z: EN 13016-3:2018**

---

**ICS:**

17.100	Merjenje sile, teže in tlaka	Measurement of force, weight and pressure
75.160.20	Tekoča goriva	Liquid fuels

**SIST EN 13016-3:2018**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 13016-3:2018

<https://standards.iteh.ai/catalog/standards/sist/b7dfbc66-ae3c-4977-8c2b-d21f4fb0b0e6/sist-en-13016-3-2018>

EUROPEAN STANDARD

EN 13016-3

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2018

ICS 75.160.20

English Version

## Liquid petroleum products - Vapour pressure - Part 3: Determination of vapour pressure and calculated dry vapour pressure equivalent (DVPE) (Triple Expansion Method)

Produits pétroliers liquides - Pression de vapeur -  
Partie 3 : Détermination de la pression de vapeur et de  
la pression de vapeur sèche équivalente calculée  
(PVSE) (Méthode triple expansion)

Flüssige Mineralölzeugnisse - Dampfdruck - Teil 3:  
Bestimmung des Dampfdruckes und des berechneten  
dem trockenen Dampfdruck entsprechenden Druckes  
(DVPE) (Dreifach-Expansionsmethode)

This European Standard was approved by CEN on 27 November 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

<b>Contents</b>	<b>Page</b>
European foreword.....	3
Introduction .....	4
1 Scope .....	5
2 Normative references .....	5
3 Terms and definitions .....	5
4 Principle .....	6
5 Reagents and materials.....	6
6 Apparatus.....	7
7 Sampling.....	8
8 Sample preparation.....	8
9 Preparation of apparatus .....	9
10 Calibration of apparatus.....	9
10.1 Pressure transducer.....	9
10.2 Temperature measuring device.....	10
11 Verification of apparatus.....	10
11.1 Reference fluid check.....	10
11.2 Quality control checks .....	11
12 Procedure.....	11
13 Calculation.....	12
14 Expression of results.....	12
15 Precision.....	13
15.1 General.....	13
15.2 Repeatability, $r$ .....	13
15.3 Reproducibility, $R$ .....	13
16 Test report.....	13
Annex A (normative) Accepted reference values .....	15
Bibliography.....	16

## European foreword

This document (EN 13016-3:2018) has been prepared by Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2018, and conflicting national standards shall be withdrawn at the latest by October 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

EN 13016 consists of the following parts, under the general title *Liquid petroleum products — Vapour pressure*:

- *Part 1: Determination of air saturated vapour pressure (ASVP) and calculated dry vapour pressure equivalent (DVPE);*
- *Part 2: Determination of absolute pressure (AVP) between 40 C and 100 C;*
- *Part 3: Determination of vapour pressure and calculated dry vapour pressure equivalent (DVPE) (Triple Expansion Method).*

This part is based on ASTM D6378 [3].

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Introduction

Vapour pressure is used as a classification criterion for the safe handling and carriage of petroleum products, feedstocks and components; it has a relationship to the potential for hydrocarbon emissions, under uncontrolled conditions, and thus is the subject of environmental scrutiny.

Vapour pressure limitations are often imposed to prevent pump cavitation during transfer operations.

Vapour pressure is one measure of the volatility characteristics of fuels used in many differing types of engines with large variations in operating temperatures. Fuels having a high vapour pressure may vaporize too readily in the fuel handling systems, resulting in decreased flow to the engine and possible stoppage by vapour lock. Conversely, fuels of low vapour pressure may not vaporize readily enough, resulting in difficult starting, slow warm-up and poor acceleration.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 13016-3:2018](https://standards.iteh.ai/catalog/standards/sist/b7dfbc66-ae3c-4977-8c2b-d21f4fb0b0e6/sist-en-13016-3-2018)

<https://standards.iteh.ai/catalog/standards/sist/b7dfbc66-ae3c-4977-8c2b-d21f4fb0b0e6/sist-en-13016-3-2018>

## 1 Scope

This European Standard specifies a method for the determination of the vapour pressure, exerted *in vacuo*, by volatile, low viscosity petroleum products, components, ethanol blends up to 85 % (V/V), and feedstocks using a variable volume chamber. A dry vapour pressure equivalent (DVPE) is calculated from the vapour pressure.

The conditions used in the test described in this standard are a vapour-to-liquid ratio of 4:1 and a test temperature of 37,8 °C.

The equipment is not wetted with water during the test, and the method described is therefore suitable for testing samples with or without oxygenates; no account is taken of dissolved water in the sample.

This procedure calculates the partial pressure of the air dissolved in the test portion during the triple expansion process. It is suitable for samples with a DVPE between 13,7 kPa and 98,3 kPa; vapour pressures outside this range can be measured but the precision has not been determined.

This document is applicable to fuels containing oxygenated compounds up to the limits stated in the relevant Council Directive 85/536/EEC [6], and for ethanol-fuel blends up to 85 % (V/V) ethanol.

NOTE For the purposes of this European Standard, the terms “% (m/m)” and “% (V/V)” are used to represent the mass and volume fractions respectively.

**WARNING** — The use of this Standard can involve hazardous materials, operations and equipment. This Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of users of this standard to take appropriate measures to ensure the safety and health of personnel prior to application of the standard, and fulfil statutory and regulatory requirements for this purpose.

SIST EN 13016-3:2018

## 2 Normative references

<https://standards.iteh.ai/catalog/standards/sist/b7dfbc66-ae3c-4977-8c2b-d214fb0b0e6/sist-en-13016-3-2018>

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13016-1, *Liquid petroleum products — Vapour pressure — Part 1: Determination of air saturated vapour pressure (ASVP) and calculated dry vapour pressure equivalent (DVPE)*

EN ISO 3170, *Petroleum liquids — Manual sampling (ISO 3170)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

**vapour pressure**

**VP**

total pressure minus the partial pressure of the dissolved air in the liquid at a vapour to liquid ratio of 4:1 and at 37,8 °C

**EN 13016-3:2018 (E)****3.2****dry vapour pressure equivalent****DVPE**

vapour pressure as calculated via EN 13016-1 by a correlation formula from the measured vapour pressure (VP)

Note 1 to entry: This Part of the European Standard calculates an equivalent DVPE result by applying a correlation formula. DVPE is only applicable to a vapour to liquid ratio of 4:1 and at a test temperature of 37,8 °C

**3.3****partial pressure from dissolved air****PPA**

pressure exerted in vacuum from dissolved air that escapes from the liquid phase into the vapour phase at a vapour to liquid ratio of 4:1 and at 37,8 °C

**3.4****total vapour pressure****TP**

pressure exerted in vacuum by air and gas containing petroleum products, components and feedstocks, and other liquids, in the absence of undissolved water at 37,8 °C

Note 1 to entry: The total vapour pressure reported at the end of the test is at a vapour to liquid ratio of 4:1.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

**4 Principle**

Employing a measuring chamber with a built-in piston, a test portion of known volume is drawn into the measuring chamber that is temperature controlled at approximately 20 °C. After sealing the measuring chamber, the temperature of the measuring chamber is increased to the 37,8 °C test temperature simultaneously with the first expansion using the piston. Two further expansions are performed to a final test volume of five times that of the test specimen. After each expansion, the total vapour pressure is measured. The PPA from the test portion is calculated from the three resulting pressures. The vapour pressure of the test portion is calculated by subtracting the PPA from the final total vapour pressure. The calculated vapour pressure is converted to a dry vapour pressure equivalent (DVPE) by applying a correlation formula.

NOTE For liquids containing very low levels of high vapour pressure contaminants, which behave like a gas, this test method that determines the PPA and gases, can lead to wrong results since the partial pressure of the contaminants will be included in the PPA.

**5 Reagents and materials**

Use chemicals of 99 % (m/m) minimum purity for check samples for verification of apparatus.

**5.1 Pentane**<sup>1)</sup>.

**5.2 2,2-Dimethylbutane**<sup>1)</sup>.

**5.3 2,3-Dimethylbutane**<sup>1)</sup>.

---

1) Annex A provides an accepted reference value (ARV) for this material.



## 6 Apparatus

### 6.1 Instrument:

NOTE This European Standard does not give full details of instruments suitable for carrying out this test. Details on the installation, operation and maintenance of each instrument can be found in the manufacturer's manual.

**6.1.1 General**, apparatus suitable for this test method employs a small volume, cylindrically shaped measuring chamber with associated equipment to control the chamber temperature at 37,8 °C. The measuring chamber assembly contains a movable piston to allow sample introduction into the measuring chamber, and expansion in three predefined volume steps, to a final 4:1 vapour to liquid ratio. A pressure transducer is incorporated in the piston. The measuring chamber contains an inlet/outlet valve combination for test portion introduction and expulsion. The piston and the valve combination are at the same temperature as the measuring chamber to avoid any condensation or excessive evaporation. The apparatus rinses the measuring chamber with the next sample to be tested, or a solvent if required.

**6.1.2 Measuring chamber assembly**, approximately 5 ml total test volume (liquid and vapour) with a moveable piston to provide a vapour to liquid ratio of 4:1, at the end of the test.

**6.1.2.1** The accuracy of this 4:1 vapour to liquid ratio shall be within 3,95:1 and 4,05:1.

**6.1.2.2** The volumes used in the calculations for vapour and partial pressures shall be accurate to within 1 % of the total test volume of the measuring chamber.

**6.1.2.3** The chamber shall be capable of retaining a vacuum within 1 kPa for 300 s at the total test volume position of the piston.

**6.1.2.4** The maximum dead volume with the piston at its lowest position shall be less than 1 % of the total test volume.

The measuring chamber and piston used in generating the precision and DVPE correlation formula were constructed of nickel plated aluminium, and stainless steel and/or brass respectively with a total volume of 5 ml. Measuring chambers with different capacities and materials may be used, but the precision and DVPE correlation formula have not been determined.

**6.1.2.5** The test chamber shall be vacuum-tight, with a provision for introducing the sample

The test chamber shall be capable of controlling the temperature of the sample to achieve the specified test temperature to within  $\pm 0,1$  °C.

**6.1.2.6 Temperature measuring device**, platinum resistance thermometer and electronic temperature control shall be used to maintain the measuring chamber at 37,8 °C  $\pm 0,1$  °C for the duration of the vapour pressure measurements.

**6.1.3** The apparatus shall be capable of measuring the vapour pressure of small samples of petroleum products, components and feedstocks up to, at least, 200 kPa by means of a pressure transducer.

**6.1.3.1 Pressure transducer**, having a minimum measuring range from 0 kPa to 200 kPa with an accuracy of  $\leq 0,2$  kPa and a resolution of  $\leq 0,1$  kPa with calibration or verification traceable to national measurement standards.

**6.2 Vacuum pump**, capable of reducing the pressure in the measuring chamber to less than 0,01 kPa absolute, for use during calibration checks.