

---

---

**Non-destructive testing — Qualification  
of radiographic film digitisation  
systems —**

**Part 2:  
Minimum requirements**

**iTeh STANDARD PREVIEW**  
*Essais non destructifs — Qualification des systèmes de numérisation  
des films radiographiques —  
(standards.iteh.ai)  
Partie 2: Exigences minimales*

[ISO 14096-2:2005](https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005)

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 14096-2:2005](https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005)

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 14096-2 was prepared by the European Committee for Standardization (CEN) (as EN 14096-2:2003) and was adopted, under a special “fast-track procedure”, by Technical Committee ISO/TC 135, *Non-destructive testing*, Subcommittee SC 5, *Radiation methods*, in parallel with its approval by the ISO member bodies.

ISO 14096 consists of the following parts, under the general title *Non-destructive testing — Qualification of radiographic film digitisation systems*:

- *Part 1: Definitions, quantitative measurements of image quality parameters, standard reference film and qualitative control*
- *Part 2: Minimum requirements*

<b>Contents</b>		Page
Foreword .....		3
Introduction.....		4
1	Scope .....	5
2	Normative references .....	5
3	Terms and definitions.....	5
4	Digitisation quality classes.....	7
5	Minimum requirements for film digitisation quality classes .....	7

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 14096-2:2005](https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005)  
<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

## Foreword

This document (EN 14096-2:2003) has been prepared by Technical Committee CEN/TC 138, "Non-destructive testing", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2003, and conflicting national standards shall be withdrawn at the latest by October 2003.

EN 14096 comprises a series of European Standards for radiographic film digitisation systems which is made up of the following:

EN 14096-1, *Non-destructive testing – Qualification of radiographic film digitisation systems – Part 1: Definitions, quantitative measurements of image quality parameters, standard reference film and qualitative control*

EN 14096-2, *Non-destructive testing – Qualification of radiographic film digitisation systems – Part 2: Minimum requirements*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Slovakia, Spain, Sweden, Switzerland and the United Kingdom.

[ISO 14096-2:2005](https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005)

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

## Introduction

Radiographic film systems are used for industrial inspection by X- and gamma rays. To apply modern means of computer support for analysis, transmission and storage the information stored in the radiographic film should be converted into digital data (digitisation). This European Standard defines minimum requirements to ensure that the relevant information for evaluation of the digital data is preserved during the film digitisation process.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 14096-2:2005](https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005)

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

## 1 Scope

This European Standard specifies three film-digitisation quality classes for the requirements of non-destructive testing. The selected class depends on the radiation energy, penetrated material thickness and the quality level of the original radiographic film. This European Standard does not address signal processing, display and storage of the digitised data.

## 2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text, and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN 444, *Non-destructive testing - General principles for radiographic examination of metallic materials by X- and gamma-rays.*

EN 1435, *Non-destructive examination of welds — Radiographic examination of welded joints.*

EN 12681, *Founding — Radiographic examination.*

EN 14096-1, *Non-destructive testing — Qualification of radiographic film digitisation systems — Part 1: Definitions, quantitative measurements of image quality parameters, standard reference film and qualitative control.*

ISO 5579, *Non-destructive testing — Radiographic examination of metallic materials by X- and gamma rays — Basic rules.*

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

## 3 Terms and definitions

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

For the purposes of this European Standard, the following terms and definitions apply.

### 3.1

#### radiographic film digitisation system

digitiser

sequential application of the two functions below:

- a) detection of the diffuse transmittance of a small unit area of the film (pixel, picture element) by means of an optical detector, giving an electric output signal (geometrical digitisation);
- b) conversion of the above electrical signal into a numerical value (densitometrical digitisation)

### 3.2

#### pixel size

$P$

geometrical centre-to-centre distance between adjacent pixels in a row (horizontal pitch) or column (vertical pitch) of the scanned image

**3.3**  
**optical density**  
 $D$   
logarithmic value to the base 10 of the diffuse light intensity ratio in front of ( $I_0$ ) and behind ( $I_D$ ) the radiographic film according to equation (1):

$$D = \lg \frac{I_0}{I_D} \quad (1)$$

**3.4**  
**spatial frequency**  
 $f$   
described by a sinusoidal intensity variation along a geometrical axis

The period of this function is measured in number of line pairs per millimetre (lp/mm).

**3.5**  
**modulation transfer function**  
**MTF**  
normalised magnitude of the Fourier-transform (FT) of the differentiated spatial optical density edge spread function (ESF) (see EN 14096-1:2003, Figure 1)

It describes the unsharpness function of the digitiser (contrast transmission as a function of the object size).

NOTE This MTF calculation is based on optical densities, which correspond to the X-ray dose.

**3.6**  
**digital resolution** in bit  
number of bits provided by the analogue-to-digital converter of the digitiser used for densitometrical digitisation

NOTE A digital resolution of  $N$  bits corresponds to  $2^N$  digital values.  
<https://standards.iteh.ai/catalog/standards/sis/e9722b23-943d-4e16-8057-7855f4d14a6d/iso-14096-2-2005>

**3.7**  
**density contrast sensitivity**  
 $\Delta D_{CS}$   
minimum density variation of the film, which is resolved by the digitiser

This is mostly determined by the digitisation noise of the digitise (quantum noise of the light detector).

NOTE The determination of this value is described in EN 14096-1:2003, 4.1.5.

**3.8**  
**density range**  
 $D_R$   
range of maximum and minimum optical densities, which can be measured by the digitiser

Depending on the construction of the digitiser, this density range can be split into several working ranges (e.g. by a different illumination power and/or a different detector integration time).

**3.9**  
**working range**  
 $D_{WR}$   
range of optical densities, where the digitiser guarantees a minimum density contrast sensitivity in one single acquisition



Only in this working range the digitised data can be used for evaluation. Depending on the digitiser construction there can be more than one working range, e.g. for brighter or darker films.

## 4 Digitisation quality classes

All radiographic film digitisation systems are subdivided into 3 quality classes DS, DB and DA :

- DS** - the enhanced technique which performs the digitisation with an insignificant reduction of signal-to-noise-ratio and spatial resolution.  
**Application field:** digital archiving of films (digital storage).
- DB** - the enhanced technique, which permits some reduction of image quality.  
**Application field:** digital analysis of films, original radiographic films have to be archived.
- DA** - the basic technique, which permits some reduction of image quality and further reduced spatial resolution.  
**Application field:** digital analysis of films, original radiographic films have to be archived.

Each radiographic film digitisation system for NDT applications shall be identified with all working ranges of optical densities. It shall be classified by the quality class corresponding to Table 1 and the maximum MTF 20 % value, corresponding to EN 14096-1, which can be performed by this system.

**EXAMPLE** A digitisation system of class DS 4,2 (scanner class DS (see Table 1), maximum MTF 20 % = 4,2 lp/mm (see Table 2)) can be applied for archiving of radiographs taken with X-rays above 200 keV or  $\gamma$ -rays and can be applied for all class DB and DA digitisation tasks.

iTeh STANDARD PREVIEW

## 5 Minimum requirements for film digitisation quality classes

### 5.1 Density range and working ranges of the film digitisation system

<https://standards.iteh.ai/catalog/standards/sist/e9722b23-943d-4e16-8057-7850f40c7b0d/iso-14096-2:2003>

Table 1 defines the minimum density range of the radiographic digitisation system. In this density range the digitiser shall provide a density contrast sensitivity  $\Delta D_{CS}$  with  $\Delta D_{CS} \leq 0,02$ . Depending on the construction of the digitiser, this density range can be split into several working ranges.

The minimum digital resolution is given for devices converting the digital value proportional to the optical density. If the digital value is converted proportional to the light intensity, the digital resolution has to be increased at least by 2 additional bits.

**Table 1 — Minimum density range of the radiographic digitisation system with a minimum density contrast sensitivity**

Parameter	Class DS	Class DB	Class DA
Density range <sup>a</sup> $D_R$	0,5 to 4,5	0,5 to 4,0	0,5 to 3,5
Digital resolution in bit	$\geq 12$	$\geq 10$	$\geq 10$
Density contrast sensitivity $\Delta D_{CS}$ within $D_R$	$\leq 0,02$	$\leq 0,02$	$\leq 0,02$

<sup>a</sup> This density range may be splitted into separated working ranges.

### 5.2 Minimum spatial resolution of film digitisation systems

Due to the energy dependence of the inherent unsharpness of industrial X-ray film systems the following parameters (see Table 2) shall be observed: