



SLOVENSKI STANDARD

oSIST ISO 22036:2019

01-maj-2019

Kakovost tal - Določevanje elementov v sledovih z atomsko emisijsko spektrometrijo z induktivno sklopljeno plazmo (ICP - AES)

Soil quality - Determination of trace elements in extracts of soil by inductively coupled plasma - atomic emission spectrometry (ICP - AES)

iTeh Standards

Qualité du sol - Dosage des éléments traces dans des extraits de sol par spectrométrie d'émission atomique avec plasma induit par haute fréquence (ICP-AES)

Document Preview

Ta slovenski standard je istoveten z: ISO 22036:2008

SIST ISO 22036:2019

<https://standards.iteh.ai/catalog/standards/sist/b1fc3db3-ccbc-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019>

ICS:

13.080.10	Kemijske značilnosti tal	Chemical characteristics of soils
71.040.50	Fizikalnokemijske analitske metode	Physicochemical methods of analysis

oSIST ISO 22036:2019

en

INTERNATIONAL STANDARD

ISO
22036

First edition
2008-12-15

Soil quality — Determination of trace elements in extracts of soil by inductively coupled plasma - atomic emission spectrometry (ICP-AES)

*Qualité du sol — Dosage des éléments traces dans des extraits de sol
par spectrométrie d'émission atomique avec plasma induit par haute
fréquence (ICP-AES)*

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST ISO 22036:2019](https://standards.iteh.ai/catalog/standards/sist/b1fc3db3-ccb3-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019)

<https://standards.iteh.ai/catalog/standards/sist/b1fc3db3-ccb3-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019>



Reference number
ISO 22036:2008(E)

© ISO 2008

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[SIST ISO 22036:2019](https://standards.iteh.ai/catalog/standards/sist/b1fc3db3-ccbc-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019)

<https://standards.iteh.ai/catalog/standards/sist/b1fc3db3-ccbc-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle.....	3
5 Interferences	6
5.1 General	6
5.2 Spectral interferences.....	6
5.3 Non-spectral interferences	7
6 Reagents.....	8
7 Instrumentation	9
8 Procedure	10
8.1 Cleaning of glassware.....	10
8.2 Instrument performance parameters.....	10
8.3 Instrument optimization.....	11
8.4 Alignment of the spectrometer	11
8.5 Calibration methods	12
8.6 Solutions to be prepared	12
8.7 Measurement procedure.....	13
9 Calculation of results	14
10 Precision.....	14
11 Expression of results	14
12 Test report.....	15
Annex A (informative) Repeatability and precision results	16
Annex B (informative) Interferences	19
Bibliography.....	30

ISO 22036:2008(E)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 22036 was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[SIST ISO 22036:2019](https://standards.itih.ai/catalog/standards/sist/b1fc3db3-ccbc-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019)

<https://standards.itih.ai/catalog/standards/sist/b1fc3db3-ccbc-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019>

Soil quality — Determination of trace elements in extracts of soil by inductively coupled plasma - atomic emission spectrometry (ICP-AES)

WARNING — The procedures in this International Standard should be carried out by competent, trained persons. Some of the techniques and reagents, including the use of equipment, are potentially very dangerous. Users of this International Standard who are not thoroughly familiar with the potential dangers and related safe practices should take professional advice before commencing any operation.

1 Scope

This International Standard describes the determination of trace elements in digests or extraction solutions from soil by inductively coupled plasma - atomic emission spectrometry (ICP-AES) for 34 elements (see Table 1).

This multi-element determination method is applicable to soil extracts obtained with aqua regia in accordance with ISO 11466, with DTPA in accordance with ISO 14870 or other weak extractants, or soil extracts for the determination of total element contents using the acid digestion method of ISO 14869-1 or the fusion method of ISO 14869-2.

The choice of calibration method depends on the extractant and can be adapted to the extractant concentration.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO Guide 32, *Calibration in analytical chemistry and use of certified reference materials*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5725-1, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*

ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

ISO 11465, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*

ISO 11466, *Soil quality — Extraction of trace elements soluble in aqua regia*

ISO 14869-1, *Soil quality — Dissolution for the determination of total element content — Part 1: Dissolution with hydrofluoric and perchloric acids*

ISO 14869-2, *Soil quality — Dissolution for the determination of total element content — Part 2: Dissolution by alkaline fusion*

ISO 14870, *Soil quality — Extraction of trace elements by buffered DTPA solution*

ISO 22036:2008(E)**3 Terms and definitions**

For the purposes of this document, the terms and definitions given in ISO 5725-1, ISO 5725-2, ISO Guide 32 and the following apply.

- 3.1 analyte**
element to be determined
- 3.2 blank calibration solution**
solution prepared in the same way as the calibration solution but leaving out the analytes
- 3.3 blank test solution**
solution prepared in the same way as the test sample solution but omitting the test portion
- 3.4 calibration solution**
solution used to calibrate the instrument, prepared from stock solutions by adding acids, buffer, reference element and salts as needed
- 3.5 instrument detection limit**
lowest concentration that can be detected with a defined statistical probability using a clean instrument and a clean solution
- NOTE The clean solution is usually dilute nitric acid.
- 3.6 laboratory sample**
sample sent to the laboratory for analysis
- 3.7 linearity**
straight-line relationship between the mean result of measurement and the quantity (concentration) of the analyte <https://standards.iteh.ai/catalog/standards/sist/b1fc3db3-ccb3-4b92-8cb3-e6b2c2fea458/sist-iso-22036-2019>
- 3.8 method detection limit**
lowest concentration that can be detected using a specific analytical method with a defined statistical probability for defined maximum matrix element concentrations
- 3.9 pure chemical**
chemical with the highest available purity and known stoichiometry
- NOTE The content of analyte and contaminants should be known with an established degree of certainty.
- 3.10 stock solution**
solution with accurately known analyte concentration(s), prepared from pure chemicals (3.9)
- NOTE Stock solutions are reference materials within the meaning of ISO Guide 30.
- 3.11 test sample**
portion taken from the laboratory sample after homogenizing, grinding, dividing, etc.

3.12

test sample solution

solution prepared after extraction or dissolution of the test sample according to appropriate specifications

NOTE The test sample solution is intended for use for measurement.

4 Principle

Inductively coupled plasma - atomic emission spectrometry (ICP-AES) can be used to determine trace elements in solution. The solution is dispersed by a suitable nebulizer and the resulting aerosol is transported into the plasma torch. In a radio-frequency inductively coupled plasma the solvent is evaporated, the dried salts are then vaporized, dissociated, atomized and ionized. The atoms or ions are excited thermally and the number of photons emitted during transition to a lower energy level are measured with optical emission spectrometry. The spectra are dispersed by a grating spectrometer, and the intensities of the emission lines are monitored by photosensitive devices. The identification of the element takes place by means of the wavelength of the radiation (energy of photons), while the concentration of the element is proportional to the intensity of the radiation (number of photons). The ICP-AES method can be used to perform multi-element determinations using sequential or simultaneous optical systems and axial or radial viewing of the plasma.

Table 1 shows examples of recommended wavelengths, and detection limits for one particular instrument. Data given are valid for water acidified with nitric acid with an optimized instrument. Using other instruments can lead to different detection limits. Adoption of other wavelengths is possible.

Table 1 — Recommended wavelengths and estimated detection limits for selected elements and wavelengths obtained using ICP-AES Varian, Vista-MPX megapixel (CD detector features) [9]

Element wavelengths and analytical lines			Axial viewing		Radial viewing	
Element	Wavelength nm	Lines I = atom II = ion	Detection limit $\mu\text{g/l}^a$	Detection limit mg/kg^b	Detection limit $\mu\text{g/l}^a$	Detection limit mg/kg^b
Aluminium	396,068		1	0,10	4	0,4
	308,215	I	2,6	0,26		
	309,271	I				
	396,152	I	0,1	0,01	4	0,4
	167,078	I	0,3	0,03		
Antimony	206,833	I	0,5	0,5	16	1,6
	217,581	I	1,8	0,18	5	0,5
	231,146	I	2	0,2		
Arsenic	188,979		2	0,2	12	1,2
	193,696		1	0,1		
	197,198	I	5	0,5		
	189,042	I				
	188,979	I	1,5	0,15	5	0,5
Barium	233,527	II	0,06	0,006	0,7	0,07
	455,403	II	0,01	0,001	0,15	0,02
	493,409	II	0,04	0,004	0,15	0,02
Beryllium	313,107	II	0,03	0,003	0,15	0,02
	313,402	II	0,01	0,001	0,15	0,02
	234,861	II	0,01	0,001	0,05	0,005
Bismuth	223,061	I	1,8	0,18	6	0,6
	306,771	I	17	1,7		
	315,887					
Boron	208,959	I	0,7	0,07	1,2	0,12
	249,678	I	1,1	0,11	1,5	0,15
	249,772	I	0,5	0,05	1	0,1

Table 1 (continued)

Element wavelengths and analytical lines			Axial viewing		Radial viewing	
Element	Wavelength nm	Lines I = atom II = ion	Detection limit $\mu\text{g/l}^a$	Detection limit mg/kg^b	Detection limit $\mu\text{g/l}^a$	Detection limit mg/kg^b
Cadmium	214,438	II	0,1	0,01	0,5	0,05
	226,502	II	0,11	0,011	0,6	0,06
	228,802	II	0,20	0,02	0,5	0,05
Calcium	396,847	II	0,5	0,05	0,3	0,03
	317,933	II	0,3	0,03	6,5	0,7
	393,366	II	0,5	0,05		
Chromium	267,716	II	0,1	0,01	1	0,1
	205,552	II	0,3	0,03		
	206,149	II				
	283,563	II	0,2	0,02		
	284,325	II				
Cobalt	238,892	II	0,4	0,04	1,2	0,1
	228,616	II	0,4	0,04	1	0,1
	230,786	II				
Copper	327,396	I	0,3	0,03	1,5	0,1
	224,700	II				
	324,754	I	0,6	0,06		
Iron	238,204	II	0,3	0,03	0,9	0,09
	239,562	II				
	259,940	II	0,5	0,05	0,7	0,07
Lead	220,353	II	0,4	0,04	8	0,8
	216,999	I				
	224,688	I				
	261,418	I				
	283,306	I	1,8	0,18		
Lithium	670,783	I	1,7	0,17	1	0,1
	460,286	I	67	6,7		
Magnesium	279,553	II	0,02	0,002	0,1	0,01
	279,079	II	1	0,1	4	0,4
	285,213	I	0,06	0,006	0,25	0,025
	279,806	II	1,5	0,15	10	1
Manganese	257,610	II	0,10	0,01	0,13	0,01
	260,569	II				
	279,482	II				
	293,306	II	0,4	0,04	1	0,1
	403,076	I	0,8	0,08		
	259,372	II	0,05	0,005		
Mercury	194,227	II	1,2	0,12	2,5	0,25
	253,652	I	1	0,1	2	0,20
	184,890	I				
Molybdenum	202,030	II	0,2	0,02	2	0,2
	204,598	II	0,6	0,06	3	0,3
Nickel	231,604	II	0,4	0,04	2,1	0,2
	221,647	II	0,3	0,03	1,4	0,14
	216,555	I	0,15	0,015		
	232,003	II				
Phosphorus	177,428	I	1,5	0,15	25	2,5
	178,222	I	7	0,7		
	213,618	I	1,3	0,13	5,3	0,53
	214,914	I	1	0,1	11	1,1

Table 1 (continued)

Element wavelengths and analytical lines			Axial viewing		Radial viewing	
Element	Wavelength nm	Lines I = atom II = ion	Detection limit $\mu\text{g/l}^a$	Detection limit mg/kg^b	Detection limit $\mu\text{g/l}^a$	Detection limit mg/kg^b
Potassium	766,491 769,896	I I	0,2 23	0,02 2,3	4 12	0,4 1,2
Rubidium	780,03	I	1	0,1	5	0,5
Selenium	196,026 203,985	I I	0,8 2,8	0,08 0,28	16	1,6
Silicon	251,611 212,412 288,158	I I I	0,9 1,3 1	0,09 0,13 0,1	2,2 5	0,22 0,5
Silver	328,068 338,289	I I	0,4 1	0,04 0,1	1 2	0,1 0,2
Sodium	589,592 588,995 330,237	I I I	0,6 12 69	0,06 1,2 6,9	1,5 15	0,2 0,15
Strontium	407,771 421,552 460,733	II II I	0,01 0,01 0,3	0,001 0,001 0,03	0,1 0,1	0,01 0,01
Sulfur	181,962 182,036	I	4	0,4	13	1,3
Thallium	190,800 190,864	II II	2	0,2	13	0,1
Tin	189,933 235,484 283,998	II I I	6 23 11	0,6 2,3	8 20	0,8 2,0
Titanium	336,121 334,941 337,280	II II II	0,15 0,2 0,2	0,015 0,02 0,02	1 0,25 1	0,1 0,25 0,1
Vanadium	292,402 309,310 311,837 290,882 310,230	II II II II II	0,3 0,08 0,1	0,03 0,008 0,01	2	0,2
Zinc	213,856 202,548 206,200	I II II	0,05 0,03 0,15	0,005 0,003 0,015	0,8 0,7 2	0,08 0,07 0,02

^a Typical 3-sigma detection limits using 30 s integration time.

^b The detection limit (LOD), as a mass fraction of the soil sample in mg/kg dry matter, is given assuming that a test sample of 1 g is extracted and diluted to 100 ml. The LOD shown in Table 1 are only examples of a given equipment and laboratory conditions. Each laboratory shall select appropriate wavelengths and determine LOD under its specific laboratory conditions.

NOTE The wavelengths given in Table 1 are often used, but they are given here only as an example. Adoption of other wavelengths is possible. The limit of detection and the linear range vary for each element with the wavelength, spectrometer, operating conditions and matrix load in the sample solution. If solutions with high salt concentrations (typical for soil extract solutions) are measured, the LOD is substantially increased compared with water samples.