



SLOVENSKI STANDARD SIST EN ISO 7886-1:2018

01-junij-2018

Nadomešča:
SIST EN ISO 7886-1:2000

Sterilne podkožne injekcijske brizge za enkratno uporabo - 1. del: Injekcijske brizge za ročno injiciranje (ISO 7886-1:2017)

Sterile hypodermic syringes for single use - Part 1: Syringes for manual use (ISO 7886-1:2017)

Sterile Einmalspritzen für medizinische Zwecke - Teil 1: Spritzen zum manuellen Gebrauch (ISO 7886-1:2017)

Seringues hypodermiques stériles, non réutilisables - Partie 1: Seringues pour utilisation manuelle (ISO 7886-1:2017)

Ta slovenski standard je istoveten z: EN ISO 7886-1:2018

ICS:

11.040.25	Injekcijske brizge, igle in katetri	Syringes, needles and catheters
-----------	-------------------------------------	---------------------------------

SIST EN ISO 7886-1:2018

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 7886-1:2018

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>

EUROPEAN STANDARD

EN ISO 7886-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2018

ICS 11.040.25

Supersedes EN ISO 7886-1:1997

English Version

Sterile hypodermic syringes for single use - Part 1: Syringes for manual use (ISO 7886-1:2017)

Seringues hypodermiques stériles, non réutilisables -
Partie 1: Seringues pour utilisation manuelle (ISO
7886-1:2017)

Sterile Einmalspritzen für medizinische Zwecke - Teil
1: Spritzen zum manuellen Gebrauch (ISO 7886-
1:2017)

This European Standard was approved by CEN on 28 February 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Endorsement notice	4
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered	5

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 7886-1:2018](https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018)
<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>

European foreword

This document (EN ISO 7886-1:2017) has been prepared by Technical Committee ISO/TC 84 " Devices for administration of medicinal products and catheters " in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by *September 2018*, and conflicting national standards shall be withdrawn at the latest by *September 2018*.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 7886-1:1997.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard 'within the meaning of Annex ZA', the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table 1 — Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 15223-1:2016	EN ISO 15223-1:2016	ISO 15223-1:2016
ISO 23908	EN ISO 23908:2013	ISO 23908:2011
ISO 80369-7	EN ISO 80369-7:2017	ISO 80369-7:2016

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

EN ISO 7886-1:2018 (E)

Endorsement notice

The text of ISO 7886-1:2017 has been approved by CEN as EN ISO 7886-1:2018 without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 7886-1:2018](https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered

This European standard has been prepared under a Commission's standardisation request M/295 concerning the development of European Standards related to medical devices to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with essential requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Directive 93/42/EEC [OJ L 169]

Essential Requirements of Directive 93/42/EEC	Clause(s)/subclause(s) of this EN	Remarks/Notes
7.3	6.2	Standard Clause 6.2 meets ER 7.3 in respect of the device altering the pH of the contents of the device.
7.5	13.2	Standard Clause 13.2 covers ER 7.5 only in respect of leakage past the plunger.
7.6	14.1.1, 14.1.2	Standard Clause 14.1.1 meets ER 7.6 in respect of packaging only. Standard Clause 14.1.2 meets the ER 7.6 up to the point of use.

EN ISO 7886-1:2018 (E)

Essential Requirements of Directive 93/42/EEC	Clause(s)/subclause(s) of this EN	Remarks/Notes
8.3	14.1.1, 14.1.2	
9.2	5, 6, 10, 11	Standard Clauses 5, 6, 10 and 11 meet ER 9.2 for the aspects detailed in the Standard Clauses.
10.1	8, 9.4, 11.2	Standard Clause 8 meets ER 10.1 except for the last sentence. Standard Clauses 9.4 and 11.2 meet the requirements of ER 10.1 as they relates to the relationship between the zero graduation line of the scale and the fiducial line on the plunger stopper only.
10.2	9.1, 9.2	
10.3	15.2.1 b)	
13.1	15	
13.3 (a)	15.2.2 b), 15.3 f), 15.4.1 b), 15.5 e)	
13.3 (b)	15.3 e), 15.4.1 e), 15.5 f), 15.6 a)	
13.3 (c)	15.3 a), 15.4.2 a), 15.5 a), 15.5 b), 15.6 c)	
13.3 (d)	15.3 c), 15.4.1 c), 15.5 d), 15.6 b)	
13.3 (e)	15.3 g), 15.4.1 f), 15.5 g), 15.6 f)	
13.3 (f)	15.2.2 a), 15.3 b), 15.4.1 a), 15.5 c)	
13.3 (i)	15.6 e)	
13.3 (k)	15.4.2 b)	

WARNING 1: Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2: Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

INTERNATIONAL
STANDARD

ISO
7886-1

Second edition
2017-05

**Sterile hypodermic syringes for
single use —**

**Part 1:
Syringes for manual use**

Seringues hypodermiques stériles, non réutilisables —

Partie 1: Seringues pour utilisation manuelle

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 7886-1:2018](https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>



Reference number
ISO 7886-1:2017(E)

© ISO 2017

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 7886-1:2018

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Nomenclature	3
5 General requirements	5
6 Extraneous matter	5
6.1 General.....	5
6.2 Limits for acidity or alkalinity.....	5
6.3 Limits for extractable metals.....	5
7 Lubricant	6
8 Tolerance on graduated capacity	6
9 Graduated scale	7
9.1 Scale.....	7
9.2 Numbering of scales.....	8
9.3 Overall length of scale to nominal capacity line.....	8
9.4 Position of scale.....	9
10 Barrel	9
10.1 Dimensions.....	9
10.2 Barrel flanges.....	9
11 Plunger stopper/plunger assembly	9
11.1 Design.....	9
12 Nozzle	10
12.1 Conical fitting.....	10
12.2 Position of nozzle on end of barrel.....	10
12.3 Nozzle lumen.....	10
13 Performance	10
13.1 Dead space.....	10
13.2 Freedom from air and liquid leakage past plunger stopper.....	10
13.3 Force to operate the piston.....	10
13.4 Fit of plunger stopper/plunger in barrel.....	10
14 Packaging	11
14.1 Unit packaging and self-contained syringe units.....	11
14.1.1 Unit packaging.....	11
14.1.2 Self-contained syringe units.....	11
14.2 Multiple unit pack.....	11
14.3 User packaging.....	11
15 Information supplied by the manufacturer	12
15.1 General.....	12
15.2 Syringes.....	12
15.2.1 General.....	12
15.2.2 Additional marking for self-contained syringe units.....	12
15.3 Unit packaging.....	12
15.4 Multiple unit packs.....	13
15.4.1 General.....	13
15.4.2 Multiple unit packs with self-contained syringes.....	13
15.5 User packaging.....	13

ISO 7886-1:2017(E)

15.6	Storage container.....	14
15.7	Transport wrapping.....	14
Annex A (normative) Method for preparation of extracts.....		15
Annex B (normative) Test method for air leakage past syringe plunger stopper during aspiration, and for separation of plunger stopper and plunger.....		16
Annex C (normative) Method for determination of dead space.....		18
Annex D (normative) Test method for liquid leakage at syringe plunger stopper under compression.....		19
Annex E (informative) Test method for the determination of forces required to operate the piston.....		21
Annex F (informative) Test method for the quantity of silicone.....		25
Bibliography.....		28

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 7886-1:2018](https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2514321a3bcb/sist-en-iso-7886-1-2018>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical committee ISO/TC 84, *Devices for administration of medicinal products and catheters*.

<https://standards.iteh.ai/catalog/standards/sist/65e8a60b-599b-4b4c-9652-2511a2b34b6f/iso-7886-1:2018>

This second edition cancels and replaces the first edition (ISO 7886-1:1993), which has been technically revised. It also incorporates the Technical corrigendum ISO 7886-1:1993/Cor.1:1995.

The main changes to the previous edition are the following:

- a) clarified the Scope, e.g. excluding single-use syringes made of glass;
- b) added new Normative references;
- c) added new terms and definitions;
- d) clarified the drawing to illustrate the component of the syringe;
- e) included general requirements;
- f) revised test methods for syringes;
- g) revised the labelling requirement;
- h) clarified the type of lubricant for the different types of syringes;
- i) replaced Annex E (informative): Examples of test methods for incompatibility between syringes and injection fluids with [Annex E](#) (informative): Test method for the determination of forces required to operate the piston;
- j) added [Annex F](#) (informative): Test method for the quantity of silicone;
- k) informative annex on materials has been deleted.

A list of all parts in the ISO 7886 series can be found on the ISO website.