



Designation: D 2572 – 97

AMERICAN SOCIETY FOR TESTING AND MATERIALS
100 Barr Harbor Dr., West Conshohocken, PA 19428
Reprinted from the Annual Book of ASTM Standards. Copyright ASTM

Standard Test Method for Isocyanate Groups in Urethane Materials or Prepolymers¹

This standard is issued under the fixed designation D 2572; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope

1.1 This test method covers the determination of the isocyanate group (NCO) content of a urethane intermediate or prepolymer.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Section 8.

2. Referenced Documents

2.1 *ASTM Standards:*

D 841 Specification for Nitration Grade Toluene²

3. Summary of Test Method

3.1 The urethane prepolymer is allowed to react with an excess of di-*n*-butylamine in toluene (Note 1). After the reaction is complete, the excess of di-*n*-butylamine is determined by back titration with standard hydrochloric acid.

NOTE 1—Solvents other than toluene may be needed in some cases, for example, anhydrous dimethylformamide (DMF), but the procedure has not been evaluated using them.

4. Significance and Use

4.1 The percent NCO is used to establish unit ratios for stoichiometric reactions with co-reactants.

5. Interferences

5.1 Phosgene, the carbamyl chloride of isocyanate, hydrogen chloride, or any other acidic or basic impurities of sufficient strength will interfere. The concentration of these compounds is usually sufficiently low that their effect on the determination is negligible.

6. Apparatus

6.1 *Polyethylene Stoppers (or corks)* covered with aluminum foil to fit 250-mL Erlenmeyer flasks.

6.2 *Magnetic Stirrer* with TFE-fluorocarbon-covered stirring bar.

6.3 *Pipet*,³ 25-mL.

6.4 *Buret*,³ 50 or 100-mL.

6.5 *Analytical Balance*.

7. Reagents

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests unless otherwise specified. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁴ Other grades may be used, provided that it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 *Bromphenol Blue Indicator*, prepare solution 0.1% by mixing 0.10 g of acid, nonwater soluble bromphenol blue with 1.5 mL of 0.1 *N* sodium hydroxide solution and diluting to 100 mL with distilled water.

7.3 *Di-*n*-butylamine Solution*, 0.1 *N* in dry toluene. Not standardized.

7.4 *Hydrochloric Acid Solution*, 0.1 *N* standardized using bromphenol blue indicator solution.

7.5 *Isopropyl Alcohol*.

7.6 *Toluene*, anhydrous, conforming to Specification D 841 or equivalent dried with molecular sieve pellets.

8. Hazards

8.1 Diisocyanates are strong skin irritants and sensitizers. They are also hazardous to the lungs. (TLV for toluene 2,4-diisocyanate is 0.005 ppm.) Avoid contact with skin, eyes, and clothing, and avoid breathing these substances. Use with adequate ventilation. In case of contact, wash with plenty of water and flush eyes for 15 min. Call a physician before reuse. See Supplier's Material Safety Data Sheet for additional information.

³ Burets and pipets shall conform to National Institute of Standards and Technology tolerances.

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

¹ This test method is under the jurisdiction of ASTM Committee D-1 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.33 on Polymers and Resins.

Current edition approved Nov. 10, 1997. Published June 1998. Originally published as D 2572 – 67 T. Last previous edition D 2572 – 91.

² *Annual Book of ASTM Standards*, Vol 06.04.