

SLOVENSKI STANDARD SIST EN ISO 6507-2:2018

01-julij-2018

Nadomešča:

SIST EN ISO 6507-2:2006

Kovinski materiali - Preskus trdote po Vickersu - 2. del: Preverjanje in umerjanje naprav za preskušanje (ISO 6507-2:2018)

Metallic materials - Vickers hardness test - Part 2: Verification and calibration of testing machines (ISO 6507-2:2018)

Metallische Werkstoffe - Härteprüfung nach Vickers - Teil 2./Überprüfung und Kalibrierung der Prüfmaschinen (ISO 6507-2:2018)

Matériaux métalliques - Essai de dureté: <u>Vickerso7 Partie</u> 2: Vérification et étalonnage des machines d'essai (ISO 6507-2:2018) atalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018

Ta slovenski standard je istoveten z: EN ISO 6507-2:2018

ICS:

77.040.10 Mehansko preskušanje kovin Mechanical testing of metals

SIST EN ISO 6507-2:2018 en,fr,de

SIST EN ISO 6507-2:2018

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6507-2:2018

https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM EN ISO 6507-2

March 2018

ICS 77.040.10

Supersedes EN ISO 6507-2:2005

English Version

Metallic materials - Vickers hardness test - Part 2: Verification and calibration of testing machines (ISO 6507-2:2018)

Matériaux métalliques - Essai de dureté Vickers - Partie 2: Vérification et étalonnage des machines d'essai (ISO 6507-2:2018) Metallische Werkstoffe - Härteprüfung nach Vickers -Teil 2: Überprüfung und Kalibrierung der Prüfmaschinen (ISO 6507-2:2018)

This European Standard was approved by CEN on 18 January 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

iTeh STANDARD PREVIEW

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Roland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 6507-2:2018 (E)

Contents	Page
Euronean foreword	3

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6507-2:2018 https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018

European foreword

This document (EN ISO 6507-2:2018) has been prepared by Technical Committee ISO/TC 164 "Mechanical testing of metals" in collaboration with Technical Committee ECISS/TC 101 "Test methods for steel (other than chemical analysis)" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2018, and conflicting national standards shall be withdrawn at the latest by September 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 6507-2:2005.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW

Endorsement notice (standards.Iten.al)

The text of ISO 6507-2:2018 has been approved by CEN as EN ISO 6507-2:2018 without any modification.

https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018

SIST EN ISO 6507-2:2018

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6507-2:2018

https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018

SIST EN ISO 6507-2:2018

INTERNATIONAL STANDARD

ISO 6507-2

Fourth edition 2018-01

Metallic materials — Vickers hardness test —

Part 2: **Verification and calibration of testing machines**

iTeh STMatériaux métalliques — Essai de durété Vickers —
Partie 2: Vérification et étalonnage des machines d'essai

SIST EN ISO 6507-2:2018 https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018



ISO 6507-2:2018(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6507-2:2018 https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-c7ac9f0670f8/sist-en-iso-6507-2-2018



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva, Switzerland Tel. +41 22 749 01 11 Fax +41 22 749 09 47 copyright@iso.org www.iso.org

Published in Switzerland

ISO 6507-2:2018(E)

Co	ntents	Page
Fore	eword	iv
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	General conditions	1
5	Direct verification 5.1 General 5.2 Calibration of the test force 5.3 Verification of the indenter 5.4 Calibration and verification of the diagonal measuring system 5.5 Verification of the testing cycle 5.6 Uncertainty of calibration/verification	2 3 4
6	Indirect verification 6.1 General 6.2 Test force and hardness levels 6.3 Measurement of reference indentation 6.4 Number of indentations 6.5 Verification result 6.6 Repeatability 6.7 Bias 6.8 Uncertainty of calibration/verification	5
7	Intervals between verifications	7
8	Verification report/calibration certificate 7-2:2018 8.1 Vicker's testing machine atalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb- 8.2 Vickers indenter c7ac9f0670f8/sist-en-iso-6507-2-2018	
	nex A (informative) Uncertainty of the calibration results of the hardness testing system	
Bibl	liography	18

ISO 6507-2:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html. (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 164, *Mechanical testing of metals*, Subcommittee SC 3, *Hardness testing*. SIST EN ISO 6507-2:2018
https://standards.iteh.ai/catalog/standards/sist/a7a1e11b-dfda-4e0e-b1bb-

This fourth edition cancels and replaces the third edition (ISO 6507-2:2005), which has been technically revised.

The main changes compared to the previous edition are as follows:

- all references of indentation diagonals, <0,020 mm, have been removed;
- requirements for the calibration and verification of the measuring system have been revised;
- requirements for the maximum permissible error in measuring a reference indentation have been revised:
- recommendations for inspection and monitoring of the indenter have been moved to ISO 6507-1;
- requirements have been added to the test report for reporting the hardness values of reference blocks used;
- Annex A has been revised.

A list of all parts in the ISO 6507 series can be found on the ISO website.

Metallic materials — Vickers hardness test —

Part 2:

Verification and calibration of testing machines

1 Scope

This document specifies a method of verification and calibration of testing machines and diagonal measuring system for determining Vickers hardness in accordance with ISO 6507-1.

A direct method of verification and calibration is specified for the testing machine, indenter and the diagonal length measuring system. An indirect verification method using reference blocks is specified for the overall checking of the machine.

If a testing machine is also to be used for other methods of hardness testing, it shall be verified independently for each method.

This document is also applicable to portable hardness testing machines but not applicable to hardness testing machines based on different measurement principles, e.g. ultrasonic impedance method.

iTeh STANDARD PREVIEW

2 Normative references (standards.iteh.ai)

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

c7ac9f0670f8/sist-en-iso-6507-2-2018
ISO 376, Metallic materials — Calibration of force-proving instruments used for the verification of uniaxial testing machines

ISO 6507-1, Metallic materials — Vickers hardness test — Part 1: Test method

ISO 6507-3, Metallic materials — Vickers hardness test — Part 3: Calibration of reference blocks

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at https://www.iso.org/obp

4 General conditions

Before a Vickers hardness testing machine is verified, the machine shall be checked to ensure that it is properly set up in accordance with the manufacturer's instructions.

Especially, it should be checked that:

- a) the plunger holding the indenter is capable of sliding in its guide without any friction or excessive side play;
- b) the indenter-holder is firmly mounted in the plunger;