



**SLOVENSKI STANDARD
SIST EN ISO 14607:2018**

01-julij-2018

**Nadomešča:
SIST EN ISO 14607:2009**

Neaktivni kirurški vsadki (implantati) - Prsni vsadki - Posebne zahteve (ISO 14607:2018)

Non-active surgical implants - Mammary implants - Particular requirements (ISO 14607:2018)

Nichtaktive chirurgische Implantate - Mammaimplantate - Besondere Anforderungen (ISO 14607:2018)

**ITeC STANDARD PREVIEW
(standards.iteh.ai)**

Implants chirurgicaux non actifs - Implants mammaires - Exigences particulières (ISO 14607:2018)

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>

Ta slovenski standard je istoveten z: EN ISO 14607:2018

ICS:

11.040.40	Implantanti za kirurgijo, protetiko in ortetiko	Implants for surgery, prosthetics and orthotics
-----------	---	---

SIST EN ISO 14607:2018

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 14607:2018

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>

EUROPEAN STANDARD

EN ISO 14607

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2018

ICS 11.040.40

Supersedes EN ISO 14607:2009

English Version

Non-active surgical implants - Mammary implants - Particular requirements (ISO 14607:2018)

Implants chirurgicaux non actifs - Implants
mammaires - Exigences particulières (ISO
14607:2018)

Nichtaktive chirurgische Implantate -
Mammaimplantate - Besondere Anforderungen (ISO
14607:2018)

This European Standard was approved by CEN on 26 February 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC [OJ L 169] aimed to be covered.....	5

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 14607:2018
<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>

European foreword

This document (EN ISO 14607:2018) has been prepared by Technical Committee ISO/TC 150 “Implants for surgery” in collaboration with Technical Committee CEN/TC 285 “Non-active surgical implants” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2018, and conflicting national standards shall be withdrawn at the latest by November 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 14607:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table — Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 34-1:2015	–	ISO 34-1:2015
ISO 37:2017	–	ISO 37:2017
ISO 4287	EN ISO 4287:1998, EN ISO 4287:1998/AC:2008 and EN ISO 4287:1998/A1:2009	ISO 4287:1997, ISO 4287:1997/Cor 1:1998/Cor 2:2005 and ISO 4287:1997/Amd:2009

EN ISO 14607:2018 (E)

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 7619-1	–	ISO 7619-1:2010
ISO 10993-1	EN ISO 10993-1:2009 and EN ISO 10993-1:2009/AC:2010	ISO 10993-1:2009 and ISO 10993-1:2009/Cor 1:2010
ISO 10993-5	EN ISO 10993-5:2009	ISO 10993-5:2009
ISO 10993-18	EN ISO 10993-18:2009	ISO 10993-18:2005
ISO 11607-1	EN ISO 11607-1:2017	ISO 11607-1:2006 and ISO 11607-1:2006/Amd 1:2014
ISO 14155	EN ISO 14155:2011 and EN ISO 14155:2011/AC:2011	ISO 14155:2011 and ISO 14155:2011/Cor. 1:2011
ISO 14630:2012	EN ISO 14630:2012	ISO 14630:2012

Endorsement notice

The text of ISO 14607:2018 has been approved by CEN as EN ISO 14607:2018 without any modification.

ITh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14607:2018](https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018)

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC [OJ L 169] aimed to be covered

This European Standard has been prepared under a Commission's standardization request [M/023 concerning the development of European Standards related to medical devices] to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Directive 93/42/EEC [OJ L 169]

Essential Requirements of directive 93/42/EEC	Clause(s)/subclause(s) of this EN	Remarks/Notes
7.5	6.2, 6.3, 6.4, 7.2.2.1	<p>The Essential Requirement is covered with respect to the risk posed by silicone gel leaking from the implant.</p> <p>The standard requires several tests to establish safe composition and properties of silicone gel. Beyond that the standard also requires several tests to establish the integrity of the implant shell to prevent leaks. Additionally, an evaluation of diffusion of gel through the implant shell is required in 7.3.4.</p> <p>Standard Clause 6.2 covers Directive Annex I, ER 7.5 in respect of cytotoxicity only.</p> <p>Standard Clause 6.3 covers Directive</p>

EN ISO 14607:2018 (E)

Essential Requirements of directive 93/42/EEC	Clause(s)/subclause(s) of this EN	Remarks/Notes
		<p>Annex I, ER 7.5 in respect of oligomers, cyclotetrasiloxane (D4) and cyclopentasiloxane (D5) only.</p> <p>Standard Clause 6.4 covers Directive Annex I, ER 7.5 in respect of the specified metals and in production raw materials only.</p> <p>Standard Clause 7.2.2.1 covers Directive Annex I, ER 7.5 in respect of the integrity of the implant shell only.</p>
9.2, second indent	7.2.2.2.2	<p>The Essential Requirement is covered with respect to a potential external impact on the implant, e.g. due to the safety belt being triggered in a car accident.</p> <p>The standard provides an impact resistance test to make sure implants withstand such impacts.</p> <p>Standard Clause 7.2.2.2.2 covers Directive Annex I, ER 9.2 second dash in respect of impact resistance of the device only.</p>
13.1	11.3, Annex I, Annex J	<p>While the references to ISO 14630:2012 in 11.3 point to useful information, they are not considered relevant within the context of this Annex ZA.</p> <p>Only the requirements provided specifically in this standard should be considered when claiming coverage of the Essential Requirement with this standard.</p>
13.3 a)	11.7.2 a), c), Annex J a)	<p>Directive Annex I, ER 13.3 a) is covered in respect of name or trade name and address of the manufacturer only. The requirement regarding the name and address of authorized representative is not covered.</p>

Essential Requirements of directive 93/42/EEC	Clause(s)/subclause(s) of this EN	Remarks/Notes
13.3 b)	11.2	While the references to ISO 14630:2012 in 11.2 point to useful information, they are not considered relevant within the context of this Annex ZA. Only the requirements provided specifically in this standard, which only partly cover the Essential Requirement 13.3 b), should be considered when claiming coverage of the Essential Requirement with this standard.
13.3 d)	11.7.2 b)	—
13.3 j)	11.5, Annex I	Standard Clause 11.5 covers Directive Annex I, ER 13.3 i) in respect of filling materials only. Standard Annex I covers Directive Annex I, ER 13.3 i) only in respect of the aspects detailed in the standard.
13.3 k)	Annex I	The ER is covered only in respect of the aspects detailed in the standard.
13.6.d)	Annex I h)	Standard Annex I h) covers Directive Annex I, 13.6 d) provided the information is in the instructions for use.
13.6 e)	11.6, Annex J c), d), h)	Standard Clause 11.6 and Annex J c) and d) cover Directive Annex I, ER 13.6 e) only in respect of expected lifetime. Standard Annex J h) covers Directive Annex I, ER 13.6 e) only in respect of possible effects on breast feeding.
13.6. f)	11.3.3, Annex J j), k), l),	The ER is covered only in respect of the aspects detailed in the standard.
13.6 l)	Annex J l)	Standard Annex J l) covers Directive Annex I, 13.6 l) provided the information is in the instructions for use and is given to the patient.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14607:2018](#)

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>

INTERNATIONAL
STANDARD

ISO
14607

Third edition
2018-04

**Non-active surgical implants —
Mammary implants — Particular
requirements**

*Implants chirurgicaux non actifs — Implants mammaires —
Exigences particulières*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14607:2018](https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018)

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>



Reference number
ISO 14607:2018(E)

© ISO 2018

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 14607:2018

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9acc5/sist-en-iso-14607-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Intended performance	3
5 Design attributes	3
6 Materials	3
6.1 General.....	3
6.2 Cytotoxicity.....	4
6.3 Residual low molecular weight oligomers.....	4
6.4 Trace elements.....	4
6.5 Physico-mechanical properties and characterization.....	4
6.6 Documentation of materials.....	4
7 Design evaluation	5
7.1 General.....	5
7.2 Pre-clinical evaluation.....	5
7.2.1 General.....	5
7.2.2 Mechanical tests.....	5
7.2.3 Physical evaluation.....	6
7.2.4 Chemical evaluation.....	7
7.2.5 Biological evaluation.....	7
7.3 Clinical evaluation.....	7
7.4 Post-market surveillance.....	7
8 Manufacturing	7
9 Sterilization	7
10 Packaging	7
11 Information supplied by the manufacturer	8
11.1 General.....	8
11.2 Product labelling.....	8
11.3 Information for the user.....	8
11.3.1 General.....	8
11.3.2 Resterilization.....	8
11.3.3 Effects on diagnostic techniques.....	8
11.4 Marking on implants.....	8
11.5 Filling materials.....	8
11.6 Information on expected lifetime.....	8
11.7 Information for the patient.....	9
11.7.1 General.....	9
11.7.2 Patient record label.....	9
11.7.3 Patient card.....	9
Annex A (normative) Determination of octamethylcyclotetrasiloxane (D4) and decamethylcyclopentasiloxane (D5) in silicone gels	10
Annex B (normative) Tests for shell integrity	14
Annex C (normative) Mechanical tests on a mammary implant in its implantable state	17
Annex D (normative) Test method for valve competence and injection site competence	23
Annex E (normative) Test for silicone gel cohesion (silicone filling materials only)	25

ISO 14607:2018(E)

Annex F (normative) Test for silicone gel penetration (silicone filling materials only)	27
Annex G (informative) Assessment of silicone diffusion from mammary implants using an in vitro method	32
Annex H (informative) Test for surface characteristics	36
Annex I (normative) Information for the user	39
Annex J (normative) Information for the patient	40
Bibliography	41

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14607:2018](https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9ace5/sist-en-iso-14607-2018)

<https://standards.iteh.ai/catalog/standards/sist/709774e2-4649-4cc7-bf7a-f33b1ea9ace5/sist-en-iso-14607-2018>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*,

This third edition cancels and replaces the second edition (ISO 14607:2007), which has been technically revised.

The main changes compared to the previous edition are as follows:

- limit values for trace elements have been added ([6.4](#));
- determination of octamethylcyclotetrasiloxane (D4) and decamethylcyclopentasiloxane (D5) in silicone gels (new [Annex A](#)) has been included;
- mechanical test on a mammary implant in its implantable state (new [Annex C](#), previously [Annex E](#)), specifically the fatigue test ([C.1](#)), has undergone major revision;
- test for silicone gel penetration (silicone filling materials only) (new [Annex F](#)) has been included;
- silicone diffusion assessment from mammary implants by an *in vitro* method (new [Annex G](#), previously [Annex H](#)) has undergone major revision;
- test for surface characteristics (new [Annex H](#), previously [Annex A](#)) has undergone major revision.