



Designation: D 1209 – 05

Standard Test Method for Color of Clear Liquids (Platinum-Cobalt Scale)¹

This standard is issued under the fixed designation D 1209; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last approval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope*

1.1 This test method describes a procedure for the visual measurement of the color of essentially light colored liquids (Note 1). It is applicable only to materials in which the color-producing bodies present have light absorption characteristics nearly identical with those of the platinum-cobalt color standards used.

NOTE 1—A procedure for estimating color of darker liquids, described for soluble nitrocellulose base solutions, is given in Guide D 365.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 For purposes of determining conformance of an observed or a calculated value using this test method to relevant specifications, test result(s) shall be rounded off “to the nearest unit” in the last right-hand digit used in expressing the specification limit, in accordance with the rounding-off method of Practice E 29.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements see Section 6.

1.5 For specific hazard information, see the Material Safety Data Sheet.

2. Referenced Documents

2.1 ASTM Standards:²

D 156 Test Method for Saybolt Color of Petroleum Products (Saybolt Chromometer Method)

D 365 Guide for Soluble Nitrocellulose Base Solutions

D 1193 Specification for Reagent Water

E 29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

E 202 Test Methods for Analysis of Ethylene Glycols and Propylene Glycols

E 346 Test Methods for Analysis of Methanol

3. Significance and Use

3.1 The property of color of a solvent varies in importance with the application for which it is intended, the amount of color that can be tolerated being dependent on the color characteristics of the material in which it is used. The paint, varnish, and lacquer solvents, or diluents commercially available on today’s market normally have little or no color. The presence or absence of color in such material is an indication of the degree of refinement to which the solvent has been subjected or of the cleanliness of the shipping or storage container in which it is handled, or both.

3.2 For a number of years the term “water-white” was considered sufficient as a measurement of solvent color. Several expressions for defining “water-white” gradually appeared and it became evident that a more precise color standard was needed. This was accomplished in 1952 with the adoption of Test Method D 1209 using the platinum-cobalt scale. This test method is similar to the description given in *Standard Methods for the Examination of Water and Waste Water*³ and is referred to by many as “APHA Color.” The preparation of these platinum-cobalt color standards was originally described by A. Hazen in the *American Chemical Journal*⁴ in which he assigned the number 5 (parts per ten thousand) to his platinum-cobalt stock solution. Subsequently, in their first edition (1905) of *Standard Methods for the Examination of Water*, the American Public Health Association, using exactly the same concentration of reagents, assigned the color designation 500 (parts per million) which is the same ratio. The parts per million nomenclature is not used since color is not referred directly to a weight relationship. It is therefore recommended that the incorrect term “Hazen Color” should not be used. Also, because it refers primarily to water, the term “APHA Color” is

¹ This test method is under the jurisdiction of ASTM Committee D01 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.35 on Solvents, Plasticizers, and Chemical Intermediates.

Current edition approved May 15, 2005. Published May 2005. Originally approved in 1952. Last previous edition approved in 2000 as D 1209 – 00.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.

³ *Standard Methods for the Examination of Water and Waste Water*, M. Franson, Ed., American Public Health Assoc., 14th ed., 1975, p. 65.

⁴ Hazen, A., “New Color Standard for Natural Waters,” *American Chemical Journal*, Vol XIV, 1892, p. 300–310.

*A Summary of Changes section appears at the end of this standard.

TABLE 1 Absorbance Tolerance Limits For No. 500 Platinum-Cobalt Stock Solution

Wavelength, nm	Absorbance
430	0.110 to 0.120
455	0.130 to 0.145
480	0.105 to 0.120
510	0.055 to 0.065

TABLE 2 Platinum-Cobalt Color Standards

Color Standard Number	Stock Solution, mL	Color Standard Number	Stock Solution, mL
5	1	70	14
10	2	100	20
15	3	150	30
20	4	200	40
25	5	250	50
30	6	300	60
35	7	350	70
40	8	400	80
50	10	450	90
60	12	500	100 ^A

^A This is platinum-cobalt color No. 10 in Guide **D 365**.

TABLE 3 Platinum-Cobalt Color Standards for Very Light Colors

Color Standard Number	Stock Solution, mL	Color Standard Number	Stock Solution, mL
1	0.20	9	1.80
2	0.40	10	2.00
3	0.60	11	2.20
4	0.80	12	2.40
5	1.00	13	2.60
6	1.20	14	2.80
7	1.40	15	3.00
8	1.60		

undesirable. The recommended nomenclature for referring to the color of organic liquids is “Platinum-Cobalt Color, Test Method D 1209.”

3.3 The petroleum industry uses the Saybolt colorimeter Test Method **D 156** for measuring and defining the color of hydrocarbon solvents; however, this system of color measurement is not commonly employed outside of the petroleum industry. It has been reported by various sources that a Saybolt color of +25 is equivalent to 25 in the platinum-cobalt system or to colors produced by masses of potassium dichromate ranging between 4.8 and 5.6 mg dissolved in 1 L of distilled water. Because of the differences in the spectral characteristics of the several color systems being compared and the subjective manner in which the measurements are made, exact equivalencies are difficult to obtain.

4. Apparatus

4.1 *Spectrophotometer*, equipped for liquid samples and for measurements in the visible region.

NOTE 2—The spectrophotometer used must be clean and in first-class operating condition. The instrument should be calibrated in accordance with the instructions given in the Standards for Checking the Calibration of Spectrophotometers (200 to 1000 nm).⁵

⁵ See NIST Letter Circular LC-1017.

4.2 *Spectrophotometer Cells*, matched having a 10-mm light path.

4.3 *Color Comparison Tubes*—Matched 100-mL, tall-form Nessler tubes, provided with ground-on, optically clear, glass caps. Tubes should be selected so that the height of the 100-mL graduation mark is 275 to 295 mm above the bottom of the tube.

4.4 *Color Comparator*—A color comparator constructed to permit visual comparison of light transmitted through tall-form, 100-mL Nessler tubes in the direction of their longitudinal axes. The comparator should be constructed so that white light is passed through or reflected off a white glass plate and directed with equal intensity through the tubes, and should be shielded so that no light enters the tubes from the side.⁶

5. Reagents

5.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁷ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Type IV of Specification **D 1193**.

5.3 *Cobalt Chloride* (CoCl₂·6H₂O).

5.4 *Hydrochloric Acid* (sp gr 1.19)—Concentrated hydrochloric acid (HCl).

5.5 *Potassium Chloroplatinate* (K₂PtCl₆).

6. Platinum-Cobalt Reference Standards

6.1 *Platinum-Cobalt Stock Solution*— Dissolve 1.245 g of potassium chloroplatinate (K₂PtCl₆) and 1.00 g of cobalt chloride (CoCl₂·6H₂O) in water. Carefully add 100 mL of hydrochloric acid (HCl, sp gr 1.19) and dilute to 1 L with water. The absorbance of the 500 platinum-cobalt stock solution in a cell having a 10-mm light path, with reagent water in a matched cell as the reference solution,⁸ must fall within the limits given in **Table 1**.

NOTE 3—This stock solution is commercially available from reputable chemical suppliers.

6.2 *Platinum-Cobalt Standards*—From the stock solution, prepare color standards in accordance with **Table 2** by diluting the required volumes to 100 mL with water in the Nessler

⁶ The sole source of supply of the unit known to the committee at this time is Scientific Glass and Instruments, Inc., P.O. Box 6, Houston, TX 77001. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,¹ which you may attend.

⁷ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

⁸ See the manufacturer’s instruction manual for complete details for operating the spectrophotometer.