



**SLOVENSKI STANDARD**  
**kSIST FprEN 1434-3:2015**

**01-julij-2015**

---

**Toplotni števci - 3. del: Izmenjava podatkov in vmesniki**

Heat meters - Part 3: Data exchange and interfaces

Wärmezähler - Teil 3: Datenaustausch und Schnittstellen

Compteurs d'énergie thermique - Partie 3 : Échange de données et interfaces

**Ta slovenski standard je istoveten z: FprEN 1434-3**

---

**ICS:**

17.200.10      Toplota. Kalorimetrija      Heat. Calorimetry

**kSIST FprEN 1434-3:2015**      **en,fr,de**



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**FINAL DRAFT**  
**FprEN 1434-3**

March 2015

ICS 17.200.10

Will supersede EN 1434-3:2008

English Version

## Heat meters - Part 3: Data exchange and interfaces

Compteurs d'énergie thermique - Partie 3 : Échange de données et interfaces

Wärmezähler - Teil 3: Datenaustausch und Schnittstellen

This draft European Standard is submitted to CEN members for unique acceptance procedure. It has been drawn up by the Technical Committee CEN/TC 294.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

**Warning** : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

	Page
Foreword.....	4
<b>1 Scope .....</b>	<b>5</b>
<b>2 Normative references .....</b>	<b>5</b>
<b>3 Meter interfaces and protocols overview .....</b>	<b>5</b>
<b>4 Physical layer .....</b>	<b>6</b>
4.1 General.....	6
4.2 Physical layer optical interface .....	6
4.3 Physical layer M-Bus .....	6
4.4 Physical layer wireless interface.....	6
4.5 Physical layer current loop interface.....	6
4.6 Physical layer Local Bus.....	6
<b>5 Link layer .....</b>	<b>7</b>
5.1 Link layer optical interface .....	7
5.2 Link layer of M-Bus and Local Bus .....	7
5.3 Link layer wireless interface.....	7
5.4 Link layer current-loop interface.....	7
<b>6 Application layer .....</b>	<b>7</b>
6.1 Application layer optical interface .....	7
6.2 Application layer M-Bus and Local Bus .....	8
<b>7 Application .....</b>	<b>9</b>
7.1 General.....	9
7.2 Physical layer .....	9
7.3 Link layer .....	9
7.4 Application layer.....	9
7.5 Control applications .....	9
<b>Annex A (informative) Recommendation for heat meter test interface .....</b>	<b>10</b>
<b>Annex B (informative) Additional information for heat meters .....</b>	<b>11</b>
B.1 Additional information regarding the EN 62056-21 protocol .....	11
B.2 Data set .....	11
B.3 Coding of the data set identification number .....	12
<b>Annex C (informative) Automatic protocol detection and wake-up for the optical interface .....</b>	<b>21</b>
C.1 Introduction .....	21
C.2 Trying EN 13757-2 protocol .....	21
C.3 Trying the EN 62056-21 protocol.....	21
<b>Annex D (informative) Usage of heat meters in control applications .....</b>	<b>23</b>
D.1 Heat meter .....	23
D.2 Controller.....	24
<b>Annex E (informative) Protection techniques for M-Bus meters against surge/lightning.....</b>	<b>26</b>
<b>Annex F (informative) Additional information about the master-unit for the M-Bus.....</b>	<b>30</b>
F.1 Master side interface to the M-Bus .....	30

<b>F.2</b>	<b>Master side interface for local data read out.....</b>	<b>30</b>
<b>F.3</b>	<b>Full size level converter.....</b>	<b>31</b>
	<b>Bibliography.....</b>	<b>33</b>

**FprEN 1434-3:2015 (E)****Foreword**

This document (FprEN 1434-3:2015) has been prepared by Technical Committee CEN/TC 294 "Communication systems for meters and remote reading of meters", the secretariat of which is held by DIN.

This document is currently submitted to the Unique Acceptance Procedure.

This document will supersede EN 1434-3:2008.

The following significant editorial changes compared to the previous edition have been incorporated in this European Standard:

- a) update of normative references;
- b) update of Table 1 "Possible combinations of interfaces and standards";
- c) addition of explanations to Table B.1 "Values for "UU", register codes".

EN 1434 consists of the following parts, under the general title "Heat meters":

- *Part 1: General requirements*
- *Part 2: Constructional requirements*
- *Part 4: Pattern approval tests*
- *Part 5: Initial verification tests*
- *Part 6: Installation, commissioning, operational monitoring and maintenance*

## 1 Scope

This European Standard specifies the general requirements and applies to heat meters. Heat meters are instruments intended for measuring the energy which in a heat-exchange circuit is absorbed (cooling) or given up (heating) by a liquid called the heat-conveying liquid. The meter indicates heat in legal units.

Part 3 specifies the data exchange between a meter and a readout device (POINT / POINT communication). For these applications using the optical readout head, the EN 62056-21 protocol is recommended.

For direct or remote local readout of a single or a few meters via a battery driven readout device, the physical layer of EN 13757-6 (local bus) is recommended.

For bigger networks with up to 250 meters, a master unit with AC mains supply according to EN 13757-2 is necessary to control the M-Bus. For these applications the physical and link layer of EN 13757-2 and the application layer of EN 13757-3 is required.

For wireless meter communications, EN 13757-4 describes several alternatives of walk/drive-by readout via a mobile station or by using stationary receivers or a network. Both unidirectionally and bidirectionally transmitting meters are supported by this standard.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13757-2, *Communication systems for and remote reading of meters - Part 2: Physical and link layer*

EN 13757-3:2013, *Communication systems for and remote reading of meters - Part 3: Dedicated application layer*

EN 13757-4, *Communication systems for meters and remote reading of meters - Part 4: Wireless meter readout (Radio meter reading for operation in SRD bands)*

EN 13757-6, *Communication systems for meters and remote reading of meters - Part 6: Local Bus*

EN 62056-21:2002, *Electricity metering - Data exchange for meter reading, tariff and load control - Part 21: Direct local data exchange (IEC 62056-21:2002)*

## 3 Meter interfaces and protocols overview

Table 1 — Possible combinations of interfaces and standards

Hardware interface type	Recommended standard	Alternative standards
Optical EN 62056-21:2002, 3.2	EN 13757-2	EN 62056-21:2002, Subclause 4.1
M-Bus	EN 13757-2	No alternative
Wireless	EN 13757-4 and EN 13757-5	No alternative
Current loop	EN 62056-21:2002, 3.1	No alternative
Local Bus	EN 13757-6	No alternative
Application layer (All interfaces)	EN 13757-3	EN 13757-1

**FprEN 1434-3:2015 (E)****4 Physical layer****4.1 General**

A meter can have either none or a number of interfaces to communicate with the outside world. If a meter has an interface in accordance with this standard, it shall fulfil at least one of the following requirements for the physical layer.

**4.2 Physical layer optical interface**

The optical interface is used for local data readout. A hand held unit, equipped with an optical readout head, is temporarily connected to one heat meter and the data is read out, one heat meter at a time. The physical properties of the optical interface are defined in EN 62056-21.

**4.3 Physical layer M-Bus**

The physical layer of the M-Bus is described in EN 13757-2. It can be used for "point to point" or for "multi-point" communication in bus systems. If a heat meter presents more than one unit load to the bus, the number of unit loads has to be shown on the meter documentation as "xUL" where x is the number of unit loads. Only integer values are allowed. Especially in extended installation, meters with an M-Bus interface might need additional protection against surge and lightning. Annex E shows various techniques for either constructing meters with an M-bus interface and integrated enhanced protection elements. In addition, it shows how to construct external protection elements for meters with a standard (unprotected) M-Bus interface. Two variants are given: one (preferred) for situation where a ground connection is available and a variant with weaker protection if no ground connection is available. An enhanced version of the protection additionally protects the meter and its interface from destruction if mains power is connected to the M-Bus terminals of the meter. If the readout frequency of the meter is limited either by software or by the battery capacity, the meter documentation shall signal the readout frequency as "x per day", "y per h" or "z per min" where x, y or z are the number of readouts within the corresponding period allowed by the software without impairing the battery lifetime. Heat meters with unlimited readout frequency do not need such information.

**4.4 Physical layer wireless interface**

The physical layer wireless interface shall be according to EN 13757-4.

**4.5 Physical layer current loop interface**

Type of signal: 20 mA (CL interface in accordance with EN 62056-21:2002, 4.1 with galvanic separation).

Power supply: on the heat meter side, the interface shall be passive. The readout device supplies the necessary power.

Connections: via terminals or suitable connectors.

**4.6 Physical layer Local Bus**

The Local Bus is an alternative to the M-Bus. It is restricted to small installations (Mini installation/ meter cluster according to EN 13757-2) and optimized for special battery-driven masters. It does not support meter power supply from the bus. Note that this interface is not compatible with M-Bus masters according to EN 13757-2. Its physical layer is described in EN 13757-6.