

# TECHNICAL SPECIFICATION **ISO/TS 11073-92001**

First edition  
2007-09-01

---

---

## **Health informatics — Medical waveform format —**

### **Part 92001: Encoding rules**

*Informatique de santé — Forme d'onde médicale —*

**iTeh STANDARD PREVIEW**  
*Partie 92001: Règles d'encodage*  
**(standards.iteh.ai)**

[ISO/TS 11073-92001:2007](https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007)

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>



Reference number  
ISO/TS 11073-92001:2007(E)

© ISO 2007

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO/TS 11073-92001:2007](https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007)

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword.....	iv
Introduction .....	v
<b>1 Scope .....</b>	<b>1</b>
<b>2 Terms and definitions .....</b>	<b>1</b>
<b>3 Abbreviated terms .....</b>	<b>1</b>
<b>4 Basic specifications .....</b>	<b>2</b>
<b>4.1 Basic attributes .....</b>	<b>2</b>
4.1.1 General.....	2
4.1.2 Sampling attributes .....	2
4.1.3 Frame attributes .....	3
<b>4.2 Encoding rule .....</b>	<b>4</b>
4.2.1 General.....	4
4.2.2 Tag (T).....	4
4.2.3 Data length (L).....	5
4.2.4 Value (V).....	6
<b>4.3 Encoding principle.....</b>	<b>6</b>
4.3.1 General.....	6
4.3.2 Definition levels .....	6
4.3.3 General principles in interpretation, scope and priority of definitions.....	6
<b>5 Basic rules (Level 1) .....</b>	<b>8</b>
<b>5.1 Primary description .....</b>	<b>8</b>
5.1.1 Sampling attributes .....	8
5.1.2 Frame attributes .....	9
5.1.3 Waveform.....	10
5.1.4 Channel.....	13
<b>5.2 Auxiliary rule .....</b>	<b>14</b>
5.2.1 Data description .....	14
5.2.2 Other definition .....	16
5.2.3 Information description.....	17
<b>6 Supplemental description (Level 2) .....</b>	<b>21</b>
<b>7 Extended description (Level 3).....</b>	<b>24</b>
<b>Annex A (informative) MFER conformance statement .....</b>	<b>27</b>
<b>Annex B (informative) Description example .....</b>	<b>28</b>
<b>Annex C (informative) Event information description .....</b>	<b>35</b>
<b>Annex D (informative) Example of standard encoding.....</b>	<b>36</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In other circumstances, particularly when there is an urgent market requirement for such documents, a technical committee may decide to publish other types of normative document:

- an ISO Publicly Available Specification (ISO/PAS) represents an agreement between technical experts in an ISO working group and is accepted for publication if it is approved by more than 50 % of the members of the parent committee casting a vote;
- an ISO Technical Specification (ISO/TS) represents an agreement between the members of a technical committee and is accepted for publication if it is approved by 2/3 of the members of the committee casting a vote.

<https://standards.iteh.ai/catalog/standards/sist/c5590d41-2543-4ee8-9cca-110689e47e68/iso-ts-11073-92001-2007>

An ISO/PAS or ISO/TS is reviewed after three years in order to decide whether it will be confirmed for a further three years, revised to become an International Standard, or withdrawn. If the ISO/PAS or ISO/TS is confirmed, it is reviewed again after a further three years, at which time it must either be transformed into an International Standard or be withdrawn.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TS 11073-92001 was prepared by Technical Committee ISO/TC 215, *Health informatics*.

## Introduction

Medical waveform data such as an electrocardiogram (ECG) or an electroencephalogram (EEG) are widely utilized in physiological examinations, physiological research, electronic medical records, healthcare information and other areas in the clinical field. Medical waveform data can be used for many medical and research purposes if digital signal processing technology is applied to standardize the data in a digital format. For medical waveforms it is essential to standardize the data format to expedite the mutual application of the standard so that the data can be processed electronically and used in a variety of ways.

**Simple and easy implementation:** application of Medical waveform Format Encoding Rules (MFER) is very simple and is designed to facilitate understanding, easy installation, trouble-shooting and low implementation cost.

**Harmonization with other standards:** MFER is specially utilized to describe the medical waveform data. Other information than waveform data, such as patient demographic data and finding information, etc. should be written using other healthcare standards, such as HL7, DICOM, ISO/IEEE 11073.

In addition, experts in each field should independently develop relevant standards for medical specifications; for example MFER for ECG is developed by cardiologists and EEG is developed by neurologists.

**Combination with coded information and text information:** MFER policy is that both machine and human readable manner are used. Namely coded information is for computer processable and text data are for human readable information. ART (arterial blood pressure) is coded as 129 and information description fields indicate "Right radial artery pressure", for example. As the description of MFER is quite flexible, MFER neither hinders the features of each system nor impedes the development of technologies.

[ISO/TS 11073-92001:2007](https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007)

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO/TS 11073-92001:2007](https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007)

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>

# Health informatics — Medical waveform format —

## Part 92001: Encoding rules

### 1 Scope

This Technical Specification specifies how medical waveforms, such as electrocardiogram, electroencephalogram, spirometry waveform, etc., are described for interoperability among healthcare information systems.

This Technical Specification may be used with other relevant protocols, such as HL7, DICOM, ISO/IEEE 11073, and database management systems for each purpose.

This is a general specification, so specifications for particular waveform types and for harmonization with DICOM, SCP-ECG, X73, etc. are not given.

This Technical Specification does not include lower layer protocols for message exchange. For example, a critical real-time application like a patient monitoring system is out of scope and this is an implementation issue.

### 2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

#### 2.1

##### **frame**

waveform encoding unit consisting of data blocks, channels and sequences

#### 2.2

##### **medical waveform**

time sequential data that are sampled by A/D converter or transmitted from medical equipment

#### 2.3

##### **sampling**

data that are converted at a fixed time interval

#### 2.4

##### **channel**

individual waveform data group

### 3 Abbreviated terms

AAMI	Association for the Advancement of Medical Instrumentation
A/D	Analog to Digital
CSE	Common Standards for Quantitative Electrocardiography

CEN	Comité Européen de Normalisation/European Committee for Standardization
ECG	Electrocardiogram
EEG	Electroencephalogram
GPS	Global Positioning System
HL7	Health Level Seven
DICOM	Digital Imaging and Communications in Medicine
IEEE	Institute of Electrical and Electronic Engineers
IEC	International Electrotechnical Commission
JIS	Japanese Industrial Standard
LSB	Least significant bit
MFER	Medical waveform Format Encoding Rules
MSB	Most significant bit
SCP-ECG	Standard Communications Protocol for Computerized Electrocardiography (EN 1064)
SPO2	Saturation of Peripheral Oxygen
VCG	Vectorcardiogram

iteh STANDARD PREVIEW  
(standards.iteh.ai)  
<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>

## **4 Basic specifications**

### **4.1 Basic attributes**

#### **4.1.1 General**

Medical waveform data described in accordance with the MFER consists of Sampling attributes (Figure 1), Frame attributes (Figure 2) and other supplemental information.

#### **4.1.2 Sampling attributes**

Sampling information has two attributes, sampling rate and sampling resolution.


##### **a) Sampling rate**

The sampling rate is described with sampling interval or sampling frequency. The sampling interval stands for the time or distance interval of each sampled data as distributed sampled waveform data.

##### **b) Sampling resolution**

Sampling resolution represents a minimum sampling value per least significant bit (LSB).



**Key**

T sampling interval (or frequency)

R resolution

iTeh STANDARD PREVIEW

(standards.iteh.ai) Figure 1 — Sampling attributes

ISO/TS 11073-92001:2007

**4.1.3 Frame attributes**

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>

The frame is a waveform encoding unit consisting of data blocks, channels and sequences. A configuration example of a frame is shown as Figure 2.

## a) Data block


The data block is the waveform data array for each channel.

## b) Channels

The channels indicate different waveform groups, e.g. if three waveform groups exist, the number of channels is 3.

## c) Sequence

The sequence represents the repetition of the group with the data block and channel.



- Key**
- 1 frame
  - 2 data block
  - 3 channel
  - 4 sequence

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

**Figure 2 — Frame attributes**


[ISO/TS 11073-92001:2007](https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007)

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>

**4.2 Encoding rule**

**4.2.1 General**

The header and waveform data should be encoded based on the encoding rules which are composed of the tag, length and value (TLV), as shown in Figure 3.



**Figure 3 — Data unit**

- The tag (T) consists of one or more octets and indicates the attribute of the data value.
- The data length (L) is the length of data values indicated in one or more octets.
- The value (V) is contents which are indicated by tag (T); e.g. attribute definition, waveform data, etc.

**4.2.2 Tag (T)**

The tag is composed of a class, primitive/context (P/C) and tag number. The tag is classified into four classes (Table 1). Classes 0 to 2 are MFER standard coding and class 3 is for private use. The private definition is intended for special purposes but should be included within any updated future version.

Table 1 — Tag

8	7	6	5	4	3	2	1
<b>Class</b>		<b>P/C</b>	<b>Tag number</b>				
0	0	0/1	MFER				
0	1						
1	0		Private				
1	1						

a) Primitive type (P/C = 0)

P/C = 0 indicates a primitive description.

b) Context type (P/C = 1)

This has only two Tags which are Group and Channel definition on current MFER. Figure 4 gives an example of a group definition.

8	7	6	5	4	3	2	1
0	1	1	0	0	1	1	1

Figure 4 — Group definition  
(standards.iteh.ai)

4.2.3 Data length (L)

The data length indicates the number of octets used for data values in the value (V) section (i.e. the length excluding octets used for tag and data length sections). The data length encoding method differs depending on whether the number of octets used for data is less than 127 or more than 128 octets.

a) In case the data value section uses 127 octets or less

The length is encoded in one octet, as shown in Figure 5.

8	7	6	5	4	3	2	1
0	Data length						

Figure 5 — Data length ≤ 127 octets

b) In case the data value section uses 128 octets or more

The long data length can be encoded using multiple octets. The first octet indicates the number of octets used to represent the total data length. For example, two subsequent octets are used to indicate the waveform data length from 128 to 65 535 and thus three octets are used to encode the data length as in Figure 6. However, MFER allows representation of a data length using multiple octets even if the length is less than 127 octets. For example, four octets can describe up to 4 294 967 295 bytes length as a data part.

8	7	6	5	4	3	2	1	8	7	6	5	4	3	2	1	8	7	6	5	4	3	2	1					
1	Length number (e.g. 3 octets)							Most significant octet							The second octet							The third octet						

Figure 6 — Data length

c) Designation of indefinite data length

MFER allows designation of an indefinite data length by encoding 80 h on the top of the data length field (Figure 7). This indefinite length designation is terminated by encoding the end-of-contents (tag = 00, data length = 00).




Figure 7 — Indefinite length designation and end-of-contents designation

d) In case the data length is 0

MFER indicates that the definition indicated by tag resets to the default value. Namely, on the root definition the concerned items re-initialize to default values and in case of the channel definition, the channel definition is re-initialized to the root definition.

4.2.4 Value (V)

The header or waveform data values are encoded in the value section according to descriptors specified by the tag.

4.3 Encoding principle

iTech STANDARD PREVIEW  
(standards.iteh.ai)

4.3.1 General

All definitions in MFER have default values, so any additional or amended definitions are optional. Thus the definition corresponding to each tag has a default value, so re-definition is not necessary if the default value is retained. It is expected that default definitions will suffice for most purposes.

ISO 11073-92001-2007  
110689e47a68/iso-ts-11073-92001-2007

4.3.2 Definition levels

4.3.2.1 Level 1 — basic definitions

Definitions at level 1 are basic definitions, which are ordinary rules (marked with an asterisk) and ensure precise encoding.

4.3.2.2 Level 2 — supplementary definitions

Definitions at level 2 are supplementary definitions. They may be used as required but it is desirable to associate the supplementary definitions with a host protocol where they can be defined with the host protocol.

4.3.2.3 Level 3 — extended definitions

Definitions at level 3 are extended definitions, which should be used as little as possible. Items of these extended definitions may considerably affect the system with regard to privacy, security, etc.; thus, great care should be taken in using them.

4.3.3 General principles in interpretation, scope and priority of definitions

4.3.3.1 Initial values (default value)

All definitions in MFER have initial values that are applied until redefined by any subsequent definition.

#### 4.3.3.2 Multiple definitions

Multiple definitions may be made for any item. Depending on items, a new definition, an old definition or all definitions (such as for events), may be used multiple times.

For example, setting the sampling frequency to 250 Hz overrides the initial value of 1 kHz.

If multiple events occur, they are interpreted in definition order.

#### 4.3.3.3 Later definition priority

Each definition is interpreted in definition order. If an item has related definitions, definition should be made in due order. The default endianness is big-endian, so to use little-endian endianness the definition for little-endian must be designated.

For example, before defining each channel, the number of channels should be defined.

#### 4.3.3.4 Channel attributes definition order

Before defining the attributes of a channel, the number of channels should be defined. If the number of channels is defined later, previous channel definitions are reset to the root definition including default values.

#### 4.3.3.5 Root definition (general definition) and channel definition (definition per channel)

The root definition is effective for all channels. The channel definition is effective only for the relevant channel and overrides the root definition. However, care should be taken because if a subsequent change to the root definition is made, it will override the default content of the relevant channel for subsequent channel definitions.

For example, if EEG is designated in the root definition, ECG designated for a channel in the channel definition overrides EEG.

<https://standards.iteh.ai/catalog/standards/sist/c559f0a4-2543-4ee8-9cca-110689e47a68/iso-ts-11073-92001-2007>

#### 4.3.3.6 Definition reset

If the data length is defined as zero (no data) in the definition of an item, the content in the definition is reset to default value. If the data length is designated as zero in a channel definition, the definition follows the root definition including the default value. If the number of channels is defined, contents defined for the channel attribute are all reset to the root definition including the default value.

#### 4.3.3.7 Incomplete definition ignored

If a definition is made without an adequate preceding definition, the definition is ignored.

In the absence of any complete definition, the default root definition will be applied.

For example, if the number of channels is undefined, any dependent channel definition is ignored.

#### 4.3.3.8 Succession of definitions

Unless redefined, each definition applies to all succeeding frames, in the effective range, except for the data pointer which is succeeding renewed. Thus, contents defined in the root definition apply to all frames unless overridden by channel definition(s), so it suffices to define common items in the root definition.

For examples, to use little-endian for all encodings with MFER, define little-endian once, then it is effective over the whole region irrespective of frames.