



SLOVENSKI STANDARD
oSIST ISO 9696:2009
01-junij-2009

Kakovost vode - Merjenje celotne alfa aktivnosti v neslanih vodah - Metoda robustnega vira

Water quality - Measurement of gross alpha activity in non-saline water - Thick source method

Qualité de l'eau - Mesurage de l'activité alpha globale des eaux non salines - Méthode par source concentrée

Ta slovenski standard je istoveten z: ISO 9696:2007

ICS:

13.060.60	Preiskava fizikalnih lastnosti vode	Examination of physical properties of water
17.240	Merjenje sevanja	Radiation measurements

oSIST ISO 9696:2009

en,fr

INTERNATIONAL STANDARD

ISO
9696

Second edition
2007-11-15

Water quality — Measurement of gross alpha activity in non-saline water — Thick source method

*Qualité de l'eau — Mesurage de l'activité alpha globale des eaux non
salines — Méthode par source concentrée*



Reference number
ISO 9696:2007(E)

© ISO 2007

ISO 9696:2007(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Symbols, definitions and units	2
4 Principle	2
5 Reagents and equipment	3
6 Procedure	4
7 Contamination check	7
8 Expression of results	8
9 Test report	10
Bibliography	11

ISO 9696:2007(E)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9696 was prepared by Technical Committee ISO/TC 147, *Water quality*.

This second edition cancels and replaces the first edition (ISO 9696:1992), which has been technically revised.

Water quality — Measurement of gross alpha activity in non-saline water — Thick source method

WARNING — Persons using this International Standard should be familiar with normal laboratory practice. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted in accordance with this International Standard be carried out by suitably trained staff.

1 Scope

This International Standard specifies a method for the determination of gross alpha activity in non-saline waters for alpha-emitting radionuclides which are not volatile at 350 °C. It is possible to determine supported volatile radionuclides measured to an extent determined by half-life, matrix retention (of the volatile species) and the duration of measurement (counting time).

The method is applicable to raw and potable waters.

The range of application depends on the amount of suspended matter in the water and on the performance characteristics (background count rate and counting efficiency) of the counter.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 31-9, *Quantities and units — Part 9: Atomic and nuclear physics*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of water samples*

ISO 5667-14, *Water quality — Sampling — Part 14: Guidance on quality assurance of environmental water sampling and handling*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

Guide to the expression of uncertainty in measurement (GUM), BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, OIML

3 Symbols, definitions and units

For the purposes of this document, the definitions, symbols and abbreviations defined in ISO 31-9, as well as the following symbols, apply.

V_t	Volume of the water sample, in litres
V	Volume of test sample, in litres, equivalent to the mass of solid on the planchet
m	Mass, in milligrams, of ignited residue from volume V
m_r	Mass of the residue deposited on the planchet sample, in milligrams
A	Alpha activity of the calibration source, in becquerels
c_A	Alpha activity concentration, in becquerels per litre
t_0	Background counting time, in seconds
t_g	Sample counting time, in seconds
r_0	Background count rate, per second
r_g	Sample gross count rate, per second
r_s	Calibration count rate, per second
ε	Counting efficiency of the specified calibration source
S	Area of the planchet, in square millimetres
$u(c_A)$	Standard uncertainty associated with the measurement result, in becquerels per litre
U	Expanded uncertainty calculated by $U = k \cdot u(a)$ with $k = 1, 2, \dots$, in becquerels per litre
c_A^*	Decision threshold, in becquerels per litre
$c_A^\#$	Detection limit, in becquerels per litre
$c_A^<, c_A^>$	Lower and upper limits of the confidence interval, in becquerels per litre

4 Principle

Gross alpha radioactivity is determined by using proportional counting or solid scintillation counting on water residue deposited on a planchet. Due to the strong absorption of the residue deposit, it is considered that the alpha emission from the surface is proportional to the alpha activity of the deposit.

Gross alpha determination is not an absolute determination of the sample alpha radioactive content, but a relative determination referring to a specific alpha emitter which constitutes the standard calibration source. This type of determination is also known as alpha index.

The sample is acidified to stabilize it, evaporated almost to dryness, converted to the sulfate form and then ignited at 350 °C. A portion of the residue is transferred to a planchet and the alpha activity measured by counting in an alpha-particle detector or counting system previously calibrated against an alpha-emitting standard and the alpha activity concentration calculated.