



Designation: E1380 – 90 (Reapproved 2005)

Standard Specification for Color Coding of Laboratory Pipets with Multiple Graduations, 0.1 mL and Larger, But Excluding Disposable Prothrombin and Disposable Micropipets¹

This standard is issued under the fixed designation E1380; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification provides a system for color coding disposable pipets (glass and plastic) with multiple graduations. The color will indicate capacity and subdivisions for identification purposes.

1.2 The intent of this specification is to ensure that when color coding is utilized, all manufacturers will be encouraged to use the same color for nominal value identification.

2. Referenced Documents

2.1 *ASTM Standards*:²

E1273 Specification for Color Coding of Reusable Laboratory Pipets

3. General Requirements

3.1 Colors shall be as specified in Table 1.

3.2 *Durability*—Disposable pipet coding shall be as durable as the graduations of the pipet. (Any institution or individual who reuses a disposable product must bear full responsibility for its safety and effectiveness).

4. Style

4.1 Each pipet may be color coded by one or more of the following methods:

4.1.1 A single or double color band. The single color band shall be 6 to 10 mm wide and the double color bands shall be 3 to 5 mm wide and separated by a clear space of 2 to 3 mm. Color bands need not be continuous, but may be broken as long as 40 % of the circumference of the pipet is covered.

4.1.2 Colored plugging material.

4.1.3 Permanent color constituent (ingredient) of the pipet material(s).

TABLE 1 Color Coding

NOTE 1—The above mentioned color codes are consistent with Specification E1273.

Nominal Volume mL	Sub-division mL	Color Code
0.1	0.001	2 green
	0.005	1 red
	0.01	1 white
	0.05	2orange
0.125	0.0125	2 yellow
	0.025	2 blue
0.2	0.001	2 white
	0.002	1 black
	0.01	1orange
	0.1	1 green
0.5	0.005	2 yellow
	0.01	2 red
	0.02	2 black
	0.05	2 green
1	0.01	1 yellow
	0.05	2 green
	0.1	1 red
	0.5,1.0,1.1	1 purple
1.5	0.01	2 red
	0.01	2 white
2	0.01	1 black
	0.02	2orange
	0.05	1 green
	0.1	1 black
2.2	1.0,2.0,2.1,2.2	1 green
	0.1	1 black
3	0.01	2 blue
	0.05	1 red
5	0.1	1 blue
	0.1	1orange
10	0.2	2orange
	0.1	2 green
15	0.1	2 green
	0.1	2 yellow
20	0.1	1 white
	0.1	1 green
25	0.2	2orange
	0.1	1 black
50	0.5	1 purple
	0.2	1 red
100	0.5	1 purple
	0.2	1 red

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Apparatus.

Current edition approved May 1, 2005. Published October 2005. Originally approved in 1990. Last previous edition approved in 1999 as E1380-90(1999)^{e1}. DOI: 10.1520/E1380-90R05.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

4.1.4 Coloring of the standard identification marking(s) of the pipet.

4.2 A color blow out band or bands shall not be the only method used to signify size.

4.3 *Location*—Color coding shall be near the top end of the pipet. In the instances where a separate band or bands are utilized to indicate “blow-out” or “blow-out” style pipets, there