



Designation: E1048 – 88 (Reapproved 2005)

# Standard Specification for Color-Coding Pipets or Containers Coated With Anticoagulants<sup>1</sup>

This standard is issued under the fixed designation E1048; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers a system to identify the anticoagulants used in coating pipets or containers not exceeding 1 mL in volume.

1.2 The purpose of this specification is to ensure that if a color code is used with an anticoagulant, all manufacturers will be encouraged to use the same code; it is not intended to require color coding.

## 2. Color Code

2.1 The color coding used to identify the anticoagulant used in coating the pipets shall be in accordance with Table 1.

2.2 Variations in the enamels used and in the methods of application appropriate for pipets made from different types of glass inevitably result in minor variations of color, and it is therefore not appropriate to specify more closely the colors mentioned in the table.

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Apparatus.

Current edition approved May 1, 2005. Published October 2005. Originally approved in 1985. Last previous edition approved in 1999 as E1048 – 88(1999)<sup>e1</sup>. DOI: 10.1520/E1048-88R05.

TABLE 1 Anticoagulant Color Code Identification

Anticoagulant	Color Code Band
Ammonium Heparin	Red
Sodium Heparin	Green
Lithium Heparin	

## 3. Dimensions and Permissible Variations

3.1 The color code shall consist of a color band 1-3 mm wide which shall completely encircle the pipet in a plane perpendicular to its longitudinal axis.

3.2 The location of the color code shall not be more than 20 mm from the top of the pipet.

3.3 If other color code bands are used to indicate capacity, they shall be 2-6 mm wide. The anticoagulant color code band should be separated with a space of at least 5 mm. The color code band indicating the anticoagulant should be closest to the top of the pipet.

## 4. Keywords

4.1 anticoagulants; colorcoding; pipets

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).