
**Workplace atmospheres — Determination
of inorganic acids by ion
chromatography —**

Part 2:

**Volatile acids, except hydrofluoric acid
(hydrochloric acid, hydrobromic acid and
nitric acid)**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

*Air des lieux de travail — Détermination des acides inorganiques par
chromatographie ionique —*

*Partie 2: Acides volatils, sauf acide fluorhydrique (acide chlorhydrique,
acide bromhydrique et acide nitrique)*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 21438-2:2009

<https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-c446ac019c2a/iso-21438-2-2009>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction.....	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
3.1 General definitions	2
3.2 Particle size fraction definitions	3
3.3 Sampling definitions	4
3.4 Analytical definitions	4
3.5 Statistical terms	5
4 Principle.....	6
5 Requirement.....	7
6 Reagents.....	7
7 Apparatus	9
7.1 Sampling equipment.....	9
7.2 Laboratory apparatus.....	11
8 Occupational exposure assessment	13
8.1 General	13
8.2 Personal sampling.....	13
8.3 Static sampling.....	13
8.4 Selection of measurement conditions and measurement pattern	13
9 Sampling.....	14
9.1 Preliminary considerations	14
9.2 Preparation for sampling.....	16
9.3 Sampling position	17
9.4 Collection of samples	17
9.5 Transportation	17
10 Analysis.....	18
10.1 Preparation of test and calibration solutions	18
10.2 Instrumental analysis	19
10.3 Estimation of detection and quantification limits	20
10.4 Quality control	20
10.5 Measurement uncertainty	21
11 Expression of results	22
12 Method performance	22
12.1 Sampling efficiency and sample storage.....	22
12.2 Quantification limits	22
12.3 Upper limits of the working range	22
12.4 Bias and precision.....	23
12.5 Uncertainty of sampling and analysis method.....	23
12.6 Interferences	23
13 Test report.....	23
13.1 Test record	23
13.2 Laboratory report	24
Annex A (informative) Temperature and pressure correction.....	25

Annex B (normative) Filter materials	27
Bibliography	28

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 21438-2:2009

<https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-c446ac019c2a/iso-21438-2-2009>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21438-2 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*.

ISO 21438 consists of the following parts, under the general title *Workplace atmospheres — Determination of inorganic acids by ion chromatography*:

- Part 1: *Non volatile acids (sulfuric acid and phosphoric acid)*
- Part 2: *Volatile acids, except hydrofluoric acid (hydrochloric acid, hydrobromic acid and nitric acid)*
- Part 3: *Hydrofluoric acid and particulate fluorides*

Introduction

The health of workers in many industries is at risk through exposure by inhalation of volatile inorganic acids. Industrial hygienists and other public health professionals need to determine the effectiveness of measures taken to control workers' exposure, and this is generally achieved by making workplace air measurements. This part of ISO 21438 has been published in order to make available a method for making valid exposure measurements for volatile inorganic acids in use in industry, such as hydrochloric acid, hydrobromic acid and nitric acid, but excluding hydrofluoric acid. It is intended to be of benefit to: agencies concerned with health and safety at work; industrial hygienists and other public health professionals; analytical laboratories; industrial users of hydrochloric acid, hydrobromic acid and nitric acid and their workers, etc.

It has been assumed in the drafting of this part of ISO 21438 that the execution of its provisions and the interpretation of the results obtained are entrusted to appropriately qualified and experienced people.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 21438-2:2009](https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-c446ac019c2a/iso-21438-2-2009)

<https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-c446ac019c2a/iso-21438-2-2009>

Workplace atmospheres — Determination of inorganic acids by ion chromatography —

Part 2: Volatile acids, except hydrofluoric acid (hydrochloric acid, hydrobromic acid and nitric acid)

1 Scope

This part of ISO 21438 specifies a method for the determination of the time-weighted average mass concentration of hydrogen chloride (HCl) gas and hydrochloric acid mist, hydrogen bromide (HBr) vapour and hydrobromic acid mist and nitric acid (HNO₃) vapour and mist in workplace air by collection on an alkali-impregnated quartz fibre filter and analysis by ion chromatography.

For mist sampling, the method is applicable to the personal sampling of the inhalable fraction of airborne particles, as defined in ISO 7708, and to static (area) sampling.

The analytical method is applicable to the determination of masses of 0,01 mg to 2,5 mg of HCl, HBr and HNO₃ per sample.

The range of concentrations of HCl, HBr and HNO₃ in air for which the measuring procedure is applicable is determined by the sampling method selected by the user. For a 240 l air sample, the working range is approximately 0,04 mg·m⁻³ to 10 mg·m⁻³ for HCl, HBr and HNO₃.

The procedure is intended to differentiate between the acids and their corresponding salts. If both are present in the air, particulate salts are trapped on a pre-filter. Co-sampled particulate matter trapped on the pre-filter and/or deposited on the walls of the sampler may be analysed, if desired.

Acids can react with co-sampled particulate matter on the pre-filter, causing interference with the measurement of the acid concentration.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 648, *Laboratory glassware — Single-volume pipettes*

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

ISO 3585, *Borosilicate glass 3.3 — Properties*

ISO 7708:1995, *Air quality — Particle size fraction definitions for health-related sampling*

ISO 8655-1, *Piston-operated volumetric apparatus — Part 1: Terminology, general requirements and user recommendations*

ISO 8655-2, *Piston-operated volumetric apparatus — Part 2: Piston pipettes*

ISO 8655-6, *Piston-operated volumetric apparatus — Part 6: Gravimetric methods for the determination of measurement error*

ISO 8756, *Air quality — Handling of temperature, pressure and humidity data*

EN 13205, *Workplace atmospheres — Assessment of performance of instruments for measurement of airborne particle concentrations*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1 General definitions

3.1.1

chemical agent

any chemical element or compound, on its own or admixed as it occurs in the natural state or as produced, used or released, including release as waste, by any work activity, whether produced intentionally or not and whether placed on the market or not

[EN 1540:1998^[1]]

iTeh STANDARD PREVIEW
(standards.iteh.ai)

3.1.2

breathing zone

⟨general definition⟩ space around the worker's face from where he or she takes his or her breath

ISO 21438-2:2009

3.1.3

breathing zone

⟨technical definition⟩ hemisphere (generally accepted to be 0,3 m in radius) extending in front of the human face, centred on the mid point of a line joining the ears, the base of the hemisphere being a plane through this line, the top of the head and the larynx

[https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-](https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-e446ac019e2a/iso-21438-2-2009)

[e446ac019e2a/iso-21438-2-2009](https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-e446ac019e2a/iso-21438-2-2009)

NOTE 1 The definition is not applicable when respiratory protective equipment is used.

NOTE 2 Adapted from EN 1540:1998^[1].

3.1.4

exposure (by inhalation)

situation in which a chemical agent is present in air which is inhaled by a person

3.1.5

measuring procedure

procedure for sampling and analysing one or more chemical agents in the air and including storage and transportation of the sample

3.1.6

operating time

period during which a sampling pump can be operated at specified flow rate and back pressure without recharging or replacing the battery

[EN 1232:1997^[2]]

3.1.7**time-weighted average concentration
TWA concentration**

concentration of a chemical agent in the atmosphere, averaged over a reference period

NOTE More detailed discussion of TWA concentrations is available in Reference [3].

3.1.8**limit value**

reference figure for concentration of a chemical agent in air

NOTE An example is the Threshold Limit Value® (TLV) for a given substance in workplace air, as established by the ACGIH^[3].

3.1.9**reference period**

specified period of time stated for the limit value of a specific chemical agent

NOTE Examples of limit values for different reference periods are short-term and long-term exposure limits, such as those established by the ACGIH^[3].

3.1.10**workplace**

defined area or areas in which the work activities are carried out

[EN 1540:1998^[1]]

iTeh STANDARD PREVIEW

3.2 Particle size fraction definitions**3.2.1****inhalable convention**

target specification for sampling instruments when the inhalable fraction is of interest

ISO 21438-2:2009

<https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-c446ac019c2a/iso-21438-2-2009>

[ISO 7708:1995]

3.2.2**inhalable fraction**

mass fraction of total airborne particles which is inhaled through the nose and mouth

NOTE The inhalable fraction depends on the speed and direction of air movement, on breathing rate and on other factors.

[ISO 7708:1995]

3.2.3**total airborne particles**

all particles surrounded by air in a given volume of air

NOTE Because all measuring instruments are size-selective to some extent, it is often impossible to measure the total airborne particles concentration.

[ISO 7708:1995]

3.3 Sampling definitions

3.3.1

personal sampler

device attached to a person that samples air in the breathing zone

[EN 1540:1998^[1]]

3.3.2

personal sampling

process of sampling carried out using a personal sampler

[EN 1540:1998^[1]]

3.3.3

sampling instrument sampler

(for the purposes of this part of ISO 21438) device for collecting airborne particles

NOTE Instruments used to collect airborne particles are frequently referred to by a number of other terms, e.g. sampling heads, filter holders, filter cassettes.

3.3.4

static sampling area sampling

process of air sampling carried out in a particular location

ITeH STANDARD PREVIEW
(standards.iteh.ai)

3.4 Analytical definitions

3.4.1

blank solution

solution prepared by taking a reagent blank, laboratory blank or field blank through the same procedure as used for sample dissolution

ISO 21438-2:2009

<https://standards.iteh.ai/catalog/standards/sist/7ec4cf86-7ef6-467f-9302-34ba015ca20-21438-2-2009>

3.4.2

calibration blank solution

calibration solution prepared without the addition of any working standard solution

NOTE The concentration of chloride, nitrate and bromide in the calibration blank solution is taken to be zero.

3.4.3

calibration solution

solution, prepared by dilution of a working standard solution, containing chloride, nitrate and bromide at concentrations that are suitable for use in calibration of an analytical instrument

3.4.4

extraction solution

solvent or solution used to solubilize the analyte(s) of interest

3.4.5

field blank

filter that is taken through the same handling procedure as a sample, except that it is not used for sampling, i.e. it is loaded into a sampler, transported to the sampling site and then returned to the laboratory for analysis

3.4.6

laboratory blank

unused filter, taken from the same batch as used for sampling, that does not leave the laboratory

3.4.7**linear dynamic range**

range of concentrations over which the calibration curve for chloride, nitrate and bromide is linear

NOTE The linear dynamic range extends from the detection limit to the onset of calibration curvature.

3.4.8**reagent blank**

all reagents used in sample dissolution, in the same quantities as used for preparation of laboratory blank, field blank and sample solutions

3.4.9**sample dissolution**

process of obtaining a solution containing chloride, nitrate and bromide from a sample, which might or might not involve complete dissolution of the sample

3.4.10**sample preparation**

all operations carried out on a sample, after transportation and storage, to prepare it for analysis, including transformation of the sample into a measurable state, where necessary

3.4.11**sample solution**

solution prepared from a sample by the process of sample dissolution

NOTE

A sample solution might need to be subjected to further operations, e.g. dilution, in order to produce a test solution that is ready for analysis.

3.4.12**stock standard solution**

solution, used for preparation of the calibration solutions, containing chloride, nitrate and/or bromide at a certified concentration that is traceable to national standards

3.4.13**test solution**

blank solution or sample solution that has been subjected to all operations required to bring it into a state in which it is ready for analysis, e.g. dilution

NOTE

The blank test solution is the blank solution and the sample test solution is the sample solution if these solutions are not subjected to any further operations before analysis.

3.4.14**working standard solution**

solution, prepared by dilution of the stock standard solution(s), that contains chloride, nitrate and bromide at concentrations that are better suited to preparation of calibration solutions than the concentration of chloride, nitrate and bromide in the stock standard solutions

3.5 Statistical terms**3.5.1****analytical recovery**

ratio of the mass of analyte measured when a sample is analysed to the known mass of analyte in that sample, expressed as a percentage

3.5.2**bias**

consistent deviation of the results of a measurement process from the true value of the air quality characteristic itself

**3.5.3
coverage factor**

k
numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

NOTE The coverage factor, k , is typically in the range from 2 to 3.

[ISO/IEC Guide 98-3:2008^[4]]

**3.5.4
combined standard uncertainty**

u_c
standard uncertainty of the result of measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or covariances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[ISO/IEC Guide 98-3:2008^[4]]

**3.5.5
expanded uncertainty**

quantity defining an interval about a result of a measurement, expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

[ISO/IEC Guide 98-3:2008^[4]]

iTeh STANDARD PREVIEW
(standards.iteh.ai)

**3.5.6
precision**

closeness of agreement of results obtained by applying the method several times under prescribed conditions

**3.5.7
true value**

value which characterizes a quantity perfectly defined in the conditions which exist when that quantity is considered

NOTE The true value of a quantity is a theoretical concept and, in general, cannot be known exactly.

**3.5.8
uncertainty (of measurement)**

parameter associated with the result of a measurement that characterizes the dispersion of the values that could reasonably be attributed to the measurand

NOTE 1 The parameter may be, for example, a standard deviation (or a given multiple of it), or the width of a confidence interval.

NOTE 2 Uncertainty of measurement comprises, in general, many components. Some of these components may be evaluated from the statistical distribution of the results of a series of measurements, and can be characterized by standard deviations. The other components, which can also be characterized by standard deviations, are evaluated from assumed probability distributions based on experience or other information. ISO/IEC Guide 98-3^[4] refers to these different cases as Type A and Type B evaluations of uncertainty, respectively.

NOTE 3 Adapted from ISO/IEC Guide 99:2007^[5].

4 Principle

4.1 A known volume of air is drawn through a pre-filter and an alkali-impregnated quartz fibre sampling filter mounted in an inhalable sampler to collect HCl, HBr and HNO₃. The acids are collected on the sampling filter, while particulate salts of the acids are trapped on the pre-filter.