

INTERNATIONAL STANDARD

IEC
60601-1-1

Second edition
2000-12

Medical electrical equipment –

Part 1-1:

General requirements for safety –

Collateral standard: Safety requirements for medical electrical systems

(<https://standards.iteh.ai>)
Document Preview

[IEC 60601-1-1:2000](https://standards.iteh.ai/standards/iec/60601-1-1-1-1-2000)

<https://standards.iteh.ai/standards/iec/60601-1-1-1-1-2000>

*This **English-language** version is derived from the original **bilingual** publication by leaving out all French-language pages. Missing page numbers correspond to the French-language pages.*



Reference number
IEC 60601-1-1:2000(E)

Publication numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series. For example, IEC 34-1 is now referred to as IEC 60034-1.

Consolidated editions

The IEC is now publishing consolidated versions of its publications. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Further information on IEC publications

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology. Information relating to this publication, including its validity, is available in the IEC Catalogue of publications (see below) in addition to new editions, amendments and corrigenda. Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is also available from the following:

- **IEC Web Site (www.iec.ch)**

- **Catalogue of IEC publications**

The on-line catalogue on the IEC web site (www.iec.ch/searchpub) enables you to search by a variety of criteria including text searches, technical committees and date of publication. On-line information is also available on recently issued publications, withdrawn and replaced publications, as well as corrigenda.

- **IEC Just Published**

This summary of recently issued publications (www.iec.ch/online_news/justpub) is also available by email. Please contact the Customer Service Centre (see below) for further information.

- **Customer Service Centre**

If you have any questions regarding this publication or need further assistance, please contact the Customer Service Centre:

Email: custserv@iec.ch

Tel: +41 22 919 02 11

Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

IEC
60601-1-1

Second edition
2000-12

Medical electrical equipment –

Part 1-1:

General requirements for safety –

Collateral standard: Safety requirements for medical electrical systems

(<https://standards.iteh.ai>)
Document Preview

IEC 60601-1-1:2000

<https://standards.iteh.ai/est/standards/iec/d8007407-0172-4d10-80c7-9956ec47b89b/iec-60601-1-1-2000>

© IEC 2000 Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembe, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

U

For price, see current catalogue

CONTENTS

	Page
FOREWORD	7
SECTION ONE — GENERAL	
Clause	
1 Scope and object.....	11
2 Terminology and definitions.....	11
3 General requirements.....	13
6 Identification, marking and documents.....	15
SECTION TWO — ENVIRONMENTAL CONDITIONS	
10 Environmental conditions.....	17
SECTION THREE — PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
16 ENCLOSURES and PROTECTIVE COVERS.....	17
17 Separation.....	17
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS	19
SECTION FOUR — PROTECTION AGAINST MECHANICAL HAZARDS	
22 Moving parts.....	21
SECTION FIVE — PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
SECTION SIX — PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN — PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility.....	21
49 Interruption of the power supply	21
SECTION EIGHT — ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
SECTION NINE — ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS	
52 Abnormal operation and fault conditions.....	23
SECTION TEN — CONSTRUCTIONAL REQUIREMENTS	
56 Components and general assembly	23
57 MAINS PARTS, components and layout	23
58 Protective earthing - Terminals and connections.....	25
59 Construction and layout.....	25

	Pages
Figure 201 Example of PATIENT ENVIRONMENT	27
Annex AAA (informative) General guidance and rationale	29
Annex BBB (informative) Examples of combinations of MEDICAL ELECTRICAL EQUIPMENT and non-medical electrical equipment	41
Annex CCC (normative) Normative references	47
Annex DDD (informative) Bibliography.....	49
Annex EEE (normative) Requirements for MULTIPLE PORTABLE SOCKET-OUTLETS	51
Annex FFF (informative) Examples of application of MULTIPLE PORTABLE SOCKET-OUTLETS.....	55

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC 60601-1-1:2000](https://standards.iteh.ai/standards/iec/d8007407-0172-4d10-80c7-9956ec47b89b/iec-60601-1-1-2000)

<https://standards.iteh.ai/standards/iec/d8007407-0172-4d10-80c7-9956ec47b89b/iec-60601-1-1-2000>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 1-1: General requirements for safety –
Collateral standard:
Safety requirements for medical electrical systems**

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a world-wide organisation for standardisation comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardisation in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organisations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organisation for Standardisation (ISO) in accordance with conditions determined by agreement between the two organisations.
- 2) The formal decisions or agreements of the IEC on technical matters, prepared by technical committees on which all the National Committees having a special interest therein are represented, express, as nearly as possible, an international consensus of opinion on the subjects dealt with.
- 3) They have the form of recommendations for international use published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-1-1 has been prepared by subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition of 60601-1-1 cancels and replaces the first edition published in 1992 and its amendment 1(1995) and constitutes a technical revision.

This second edition is a Collateral Standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety*, hereinafter referred to as the General Standard, and is the first of a series of Collateral Standards amplifying the General Standard.

The text of this Collateral Standard is based on the following documents:

FDIS	Report on voting
62A/312/FDIS	62A/318/RVD

Full information on the voting for the approval of this Collateral Standard can be found in the report on voting indicated in the above table.

In the 60601 series of publications, Collateral Standards specify general requirements for safety applicable to

- a group of MEDICAL ELECTRICAL EQUIPMENT (for example, radiological equipment);
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the General Standard (for example, electromagnetic compatibility).

The numbering of sections, clauses and subclauses of this Collateral Standard corresponds with that of the General Standard.

Subclauses and figures which are additional to those of the General Standard are numbered starting from 201; additional annexes are lettered AAA, BBB, etc.

In this Collateral Standard, the following print types are used:

- requirements, compliance with which can be tested and definitions: in roman type;
- explanations, advice, general statements, exceptions and references: in smaller roman type,
- *test specifications: in italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD OR OF THIS COLLATERAL STANDARD: SMALL CAPITALS.

The requirements are followed by specifications for the relevant tests.

Some provisions or statements in the body of this Collateral Standard require additional information. Such information is presented in the informative annex AAA, General guidance and rationale. An asterisk (*) at the left margin of a clause or subclause indicates the presence of additional information.

Annexes AAA, BBB, DDD and FFF are for information only.

Annexes CCC and EEE form an integral part of this Collateral Standard.

The committee has decided that the contents of this publication will remain unchanged until 2005. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

MEDICAL ELECTRICAL EQUIPMENT –
Part 1-1: General requirements for safety –
Collateral Standard:
Safety requirements for medical electrical systems

SECTION ONE — GENERAL

1 Scope and object

*1.201 Scope

This standard applies to the safety of MEDICAL ELECTRICAL SYSTEMS, as defined in 2.201. It describes the safety requirements necessary to provide protection for the PATIENT, the OPERATOR and surroundings.

2 Terminology and definitions

In this Collateral Standard, terms printed in small capitals are used in accordance with their definitions in IEC 60601-1.

Where the terms "voltage" and "current" are used, they mean the r.m.s. values of an alternating, direct or composite voltage or current.

For the purpose of this standard the following additional definitions apply:

2.201

MEDICAL ELECTRICAL SYSTEM (hereinafter referred to as SYSTEM)
combination of items of equipment, at least one of which must be MEDICAL ELECTRICAL EQUIPMENT and inter-connected by FUNCTIONAL CONNECTION or use of a MULTIPLE PORTABLE SOCKET-OUTLET

NOTE Equipment, when mentioned in connection with a SYSTEM, should be taken to include EQUIPMENT. (See also examples given in annexes BBB and FFF.)

*2.202

PATIENT ENVIRONMENT

any volume in which intentional or unintentional contact can occur between PATIENT and parts of the SYSTEM or between PATIENT and other persons touching parts of the SYSTEM (see figure 201)

*2.203

SEPARATION DEVICE

a component or arrangement of components with input parts and output parts that, for safety reasons, prevent a transfer of unwanted voltage or current between parts of a SYSTEM

*2.204

MULTIPLE PORTABLE SOCKET-OUTLET

a combination of two or more socket-outlets intended to be connected to, or integral with, flexible cables or cords, and which can easily be moved from one place to another while connected to the supply

NOTE A MULTIPLE PORTABLE SOCKET-OUTLET may be a separate item or an integral part of medical or non-medical equipment

2.205*FUNCTIONAL CONNECTION**

connection, electrical or otherwise, including those intended to transfer signals and/or power and/or substances

3 General requirements***3.201 General requirements for the SYSTEM**

After installation or subsequent modification, a SYSTEM shall not cause a SAFETY HAZARD.

A SYSTEM shall provide:

- within the PATIENT ENVIRONMENT, a level of safety comparable to that provided by MEDICAL ELECTRICAL EQUIPMENT complying with IEC 60601-1, and
- outside the PATIENT ENVIRONMENT, the level of safety appropriate for non-medical electrical equipment complying with other IEC or ISO safety standards.

Compliance is considered to exist if the requirements of 3.201.1, 3.201.2, 3.201.3 and 3.201.4 are met. A SYSTEM incorporating equipment or parts, which use materials or have forms of construction different from those detailed in relevant standards as mentioned in 3.201.1 and 3.201.2, shall be accepted if it can be demonstrated that an equivalent degree of safety is obtained.

3.201.1 MEDICAL ELECTRICAL EQUIPMENT

MEDICAL ELECTRICAL EQUIPMENT shall comply with the requirements of IEC 60601-1 and its relevant particular standards.

Compliance is checked by inspection of appropriate documents or certificates.

3.201.2 Non-medical electrical equipment

Non-medical electrical equipment shall comply with IEC and ISO safety standards that are relevant to that equipment. See also annex DDD.

Equipment in which protection against electric shock relies on BASIC INSULATION only shall not be used in a SYSTEM.

Compliance is checked by inspection of appropriate documents or certificates.

***3.201.3. Specified power supply**

A specified power supply according to 10.2.2.201 shall be in accordance with IEC 60601-1 or shall demonstrate an equivalent degree of safety.

NOTE The party assembling or modifying the SYSTEM should calculate the power consumption of the SYSTEM, make sure that this consumption is consistent with the power that the MULTIPLE PORTABLE SOCKET-OUTLET(S) can support and document it.

Compliance is checked by inspection of appropriate documents or certificates.

***3.201.4. SYSTEM**

After installation or subsequent modification, the SYSTEM shall be in compliance with the requirements of this Collateral Standard.

Compliance is checked by inspection, by testing or by analysis, as specified in the relevant subclause.

Only hazards arising from the interconnection of various equipment to constitute a SYSTEM shall be considered.

Safety tests which have already been carried out on individual equipment of the SYSTEM according to relevant standards shall not be repeated.

Tests shall be carried out:

- in NORMAL CONDITION unless otherwise specified in this standard, and
- under the operating conditions specified by the manufacturer of the SYSTEM.

6 Identification, marking and documents***6.8.201 ACCOMPANYING DOCUMENTS of a SYSTEM**

A SYSTEM (including a modified SYSTEM) shall be accompanied by documents containing all the data necessary for safe and intended use.

These documents shall include:

- a) the ACCOMPANYING DOCUMENTS for each item of MEDICAL ELECTRICAL EQUIPMENT (see 6.8 of IEC 60601-1);
- b) the equivalent documents for each item of non-medical electrical equipment;
- c) the following information:
 - instructions for cleaning and, where applicable, sterilizing and disinfecting each item of equipment forming part of the SYSTEM;
 - additional safety measures which should be applied, during installation of the SYSTEM;
 - which parts of the SYSTEM are suitable for use within the PATIENT ENVIRONMENT;
 - additional measures which should be applied during preventive maintenance;
 - a warning that MULTIPLE PORTABLE SOCKET-OUTLETS shall not be placed on the floor;
 - a warning that an additional MULTIPLE PORTABLE SOCKET-OUTLET or extension cord shall not be connected to the system;
 - a warning not to connect items which are not specified as part of the SYSTEM;
 - the maximum permitted load for any MULTIPLE PORTABLE SOCKET-OUTLET(S) used with the SYSTEM;
 - an instruction that MULTIPLE PORTABLE SOCKET-OUTLETS provided with the SYSTEM shall only be used for supplying power to equipment which is intended to form part of the SYSTEM;
 - an explanation of the risks of connecting a non-medical electrical equipment, which has been supplied as a part of the SYSTEM, directly to the wall outlet when the non-medical equipment is intended to be supplied via a MULTIPLE PORTABLE SOCKET-OUTLET with a separating transformer;

- an explanation of the risks of connecting electrical equipment, which has not been supplied as a part of the SYSTEM, to the MULTIPLE PORTABLE SOCKET-OUTLET;
 - any restrictions in the environmental conditions to ensure safety (see clause 10 of the General Standard);
 - instructions to the OPERATOR not to touch parts referred to in 16.201 and the PATIENT simultaneously;
- d) advice to
- the installer, recommending that the SYSTEM be installed in a way that enables the USER to achieve optimal use, and
 - the USER, to carry out all cleaning, adjustment, sterilization and disinfection procedures specified herein.

Compliance is checked by inspection.

SECTION TWO — ENVIRONMENTAL CONDITIONS

10 Environmental conditions

*10.2.2.201 Power supply

A power supply from another equipment for EQUIPMENT in a SYSTEM shall be specified by the manufacturer.

SECTION THREE — PROTECTION AGAINST ELECTRIC SHOCK HAZARDS

16 ENCLOSURES and PROTECTIVE COVERS

16.201 ENCLOSURES

Parts of non-medical electrical equipment in the PATIENT ENVIRONMENT that, after removal of covers, connectors, etc., without the use of a TOOL, may be contacted by the OPERATOR during routine maintenance, calibration, etc., shall operate at a voltage not exceeding 25 V a.c. or 60 V d.c. or peak value supplied from a source which is separated from the SUPPLY MAINS by one of the methods described in 17 g) 1) to 5) of IEC 60601-1.

Compliance is checked by inspection.

17 Separation

*17.201 Electrical separation

If the allowable values of LEAKAGE CURRENTS can be exceeded – caused by FUNCTIONAL CONNECTION between different items of equipment of a SYSTEM and other systems, for example, an emergency calling system or a data processing system – then safety measures incorporating a SEPARATION DEVICE shall be applied.

Such safety measures provide suitable electrical separation between the equipment and/or between the SYSTEM and other systems and shall have the dielectric strength, CREEPAGE