INTERNATIONAL STANDARD

ISO 16628

First edition 2008-11-15

Tracheobronchial tubes — Sizing and marking

Sondes trachéo-bronchiques — Dimensionnement et marquage

iTeh STANDARD PREVIEW (standards.iteh.ai)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 16628:2008 https://standards.iteh.ai/catalog/standards/sist/1797969d-e388-4786-be0b-2259d68dd34e/iso-16628-2008



COPYRIGHT PROTECTED DOCUMENT

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 16628 was prepared by Technical Committee ISO/TC 121, Anaesthetic and respiratory equipment, Subcommittee SC 2, Tracheal tubes and other equipment.

This first edition cancels and replaces the Technical Specification ISO/TS 16628:2003.

(standards.iteh.ai)

Introduction

Originally a Technical Specification, this document is being revised as an International Standard to provide requirements for size designation and labelling of tracheobronchial tubes. Such requirements will have the benefit of enabling clinicians to select the most appropriate device for their patient population and to select accessories such as fiberoptic bronchoscopes based on the measured internal diameter of the bronchial segment of the tracheobronchial tube.

This International Standard provides requirements for manufacturers in establishing a standard method of size designation for tracheobronchial tubes and their parts. A tracheobronchial tube is a double lumen tracheal tube that facilitates selective ventilation to one or both lungs and allows isolation of one lung from the other. It is designed for either right or left mainstem bronchus placement, and has both a tracheal and bronchial cuff.

iTeh STANDARD PREVIEW (standards.iteh.ai)

Tracheobronchial tubes — Sizing and marking

1 Scope

This International Standard provides requirements for size designation and labelling of tracheobronchial tubes, including colour coding of the bronchial cuff and its associated pilot balloon. Tracheobronchial tubes that include bronchus blockers are excluded from the scope of this International Standard.

2 Normative References

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4135, Anaesthetic and respiratory equipment — Vocabulary

ISO 7000, Graphical symbols for use on equipment — Index and synopsis

ISO 15223-1, Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements

<u>ISO 16628:2008</u>

EN 1041, Information supplied by the manufacturer with medical devices 786-be0b-2259d68dd34e/iso-16628-2008

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

tracheobronchial tube


double-lumen tube designed for insertion into the trachea and a main bronchus to separate the right and left lungs

4 General requirements

4.1 Outside diameter of bronchial segment

The outside diameter of the bronchial segment of the tube shall be the outside diameter measured at the midpoint of the deflated bronchial cuff (see Key 9 of Figure 1) in millimetres, determined in accordance with Annex A.

© ISO 2008 – All rights reserved



Key

- 1 machine end of the tracheobronchial tube
- 2 bronchial segment
- 3 cross-section of tracheal lumen (not necessarily circular) ARD PREVIEW
- 4 cross-section of bronchial lumen (not necessarily circular)
- 5 patient end of the bronchial segment (standards.iteh.ai)
- 6 patient end of the tracheal segment
- 7 tracheal cuff <u>ISO 16628:2008</u>
- 8 bronchial cuff https://standards.iteh.ai/catalog/standards/sist/1797969d-e388-4786-be0b-
- 9 point of measurement of the outside diameter of the bronchial segment 2008

Figure 1 — Example of a tracheobronchial tube

4.2 Effective inside diameter

The inside diameter of the lumena of the bronchial and tracheal segments of the tube in millimetres shall be determined in accordance with Annex B.

NOTE The effective inside diameter is intended as a guide to selecting the appropriate diameter bronchoscope or other device to be inserted into the tracheobronchial tube.

4.3 Colour coding

The bronchial cuff and its associated pilot balloon shall be entirely coloured blue.

4.4 Segment differentiation

The tracheal and bronchial segments shall be clearly distinguishable from each other when viewed from the machine end.

5 Marking

5.1 Use of symbols

- **5.1.1** The recommendations of 4.2 can be met by the appropriate symbols as given in ISO 7000 or ISO 15223-1.
- **5.1.2** Marking of tracheobronchial tubes, packages, inserts and information supplied by the manufacturer shall comply with EN 1041.

5.2 Marking of tracheobronchial tube

- **5.2.1** Marking of the tracheobronchial tube shall include the following:
- a) the name and/or trademark of the manufacturer or supplier;
- b) the manufacturer's stated size (e.g. 39 Fr.);
- c) the outside diameter of the bronchial segment, in accordance with 4.1, in millimetres, and prefixed "Br" as shown in the following example:

Br 13,0 OD;

- d) for tracheobronchial tubes for single use, the words "SINGLE-USE" or equivalent;
- e) length marks at 2 cm intervals, measured from the tip of the bronchial segment, and marked on the machine end of the tracheal segment dards.iteh.ai)
- f) the word "right" or "left", as appropriate. ISO 16628:2008
- 5.2.2 Marking should be of a colour that contrasts with the colour of the tube.

5.3 Marking on tracheobronchial tube individual pack and any insert

The following shall be marked on, or visible through, the tracheobronchial tube individual pack and may additionally be given on an insert:

- a) description of contents;
- b) the word "right" or "left" as appropriate;
- c) the manufacturer's stated size (e.g. 39 Fr.);
- d) the outside diameter of the bronchial segment, in accordance with 4.1, in millimetres, and prefixed "Br" as shown in the following example:

Br 13,0 OD;

- e) the effective inside diameter, as determined in accordance with 4.2;
- f) the name and/or trademark of the manufacturer;
- g) the batch number (it is strongly recommended that the "use-by" date be given);
- h) the word "STERILE" if appropriate (it is recommended that the method of sterilization be given);
- i) the words "single-use" or equivalent for tubes not intended for re-use;

ISO 16628:2008(E)

- j) an indication of the presence of natural rubber (latex), if present in the device;
- k) instructions for cleaning and disinfection or sterilization, unless the tube is marked as for single use;
- I) recommendations for an appropriate sterilization method, if delivered non-sterile.

iTeh STANDARD PREVIEW (standards.iteh.ai)

Annex A

(normative)

Outside diameter determination of the bronchial segment

A.1 Principle

The outside diameter (OD) of the bronchial segment of the tracheobronchial tube is measured by wrapping an inelastic wire or thread around the outside of the bronchial segment of the tube and measuring its length. This method determines the equivalent circular diameter (see A.3.3 below).

A.2 Apparatus

- **A.2.1 Length of inelastic wire or thread**, sufficiently long to be wrapped around the outside of the bronchial segment of the tube five times at the mid-point of the deflated bronchial cuff (see Key 9 of Figure 1).
- **A.2.2** Means of measuring inelastic wire or thread (A.2.1) in millimetres.

A.3 Procedure

iTeh STANDARD PREVIEW

- **A.3.1** Wrap a piece of inelastic wire or thread (A.2.1) in a close spiral five times around the outside of the bronchial segment of the tube, at the mid-point of the deflated bronchial cuff (see Key 9 in Figure 1).
- A.3.2 Measure the length of the inelastic wire of thread in millimetres. 4786-be0b-2259d68dd34e/iso-16628-2008
- **A.3.3** Determine the OD by dividing the length of the wrapped thread by $5 \cdot \pi$ (i.e. $5 \times 3,14 = 15,7$).

A.4 Expression of results

Express the results in millimetres, as determined in A.3.3, to one decimal place, rounded up to the nearest 0,5 mm.

EXAMPLE 9,1 mm is expressed as 9,5 mm

9,6 mm is expressed as 10,0 mm

13,1 mm is expressed as 13,5 mm

NOTE The diameter is expressed to the larger 0,5 mm in order to ensure that the selected tube is smaller than the bronchus.

© ISO 2008 – All rights reserved