
**Cereals — Determination of bulk density,
called mass per hectolitre —**

**Part 3:
Routine method**

*Céréales — Détermination de la masse volumique, dite masse
à l'hectolitre —*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Partie 3: Méthode pratique

ISO 7971-3:2009

<https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9fd869cd944/iso-7971-3-2009>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 7971-3:2009

<https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9fd869cd944/iso-7971-3-2009>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 Principle.....	1
5 Apparatus	2
6 Procedure	2
6.1 General.....	2
6.2 Hand-operated instruments.....	2
6.3 Automatic instruments.....	3
6.4 Expression of results	3
7 Precision.....	3
7.1 Interlaboratory test	3
7.2 Repeatability.....	3
7.3 Reproducibility.....	3
7.4 Comparison of two groups of measurements in one laboratory.....	4
7.5 Comparison of two groups of measurements in two laboratories	4
7.6 Uncertainty	4
8 Test report	5
Annex A (informative) Description of dimensions and use of KERN apparatus	6
Annex B (informative) Description of dimensions and use of NILEMA LITRE apparatus	10
Annex C (informative) Results of interlaboratory tests.....	12
Bibliography	14

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7971 (all parts) was prepared by the European Committee for Standardization (CEN) Technical Committee CEN/TC 338, *Cereal and cereal products*, in collaboration with Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

The first edition of ISO 7971-3 cancels and replaces the first edition of ISO 7971-2:1995, which has been technically revised.

ISO 7971 consists of the following parts, under the general title *Cereals — Determination of bulk density, called mass per hectolitre*:

- *Part 1: Reference method*
- *Part 2: Method of traceability for measuring instruments through reference to the international standard instrument*
- *Part 3: Routine method*

Cereals — Determination of bulk density, called mass per hectolitre —

Part 3: Routine method

1 Scope

This part of ISO 7971 specifies a routine method for the determination of bulk density, called “mass per hectolitre” of cereals as grain using manual or automatic, mechanical, electric or electronic mass per hectolitre measuring instruments.

NOTE Further details of the measuring instruments are specified in ISO 7971-2:2009, 6.4.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7971-2, *Cereals — Determination of bulk density, called mass per hectolitre — Part 2: Method of traceability for measuring instruments through reference to the international standard instrument*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

bulk density **“mass per hectolitre”**

⟨cereals⟩ ratio of the mass of a cereal to the volume it occupies after being poured into a container under well-defined conditions

NOTE 1 Bulk density is expressed in kilograms per hectolitre of grains as received.

NOTE 2 The bulk density, as defined in this part of ISO 7971, is different from “packing density” or “intrinsic density” of cereals.

[ISO 7971-1:2009]

4 Principle

The mass per hectolitre of a cereal is obtained from the mass of a volume of cereal determined under controlled sample filling and flow conditions.

The mass per hectolitre can be affected by:

- a) space between the grains, which depends on the grain size and shape;
- b) density of the grains.

5 Apparatus

5.1 General requirement for bulk density apparatus. Any apparatus (5.2 and 5.3) shall be verified according to ISO 7971-2 and shall fulfil the performance demands specified therein.

5.2 Hand-operated measuring instrument. Apparatus consisting of a filling hopper, a measuring container and the accessories necessary for their use.

The manner in which the grain is poured into the measuring container and the way in which it packs into the container can cause the measurements taken by the various instruments to vary and lead to measurement errors.

To minimize such variations, special attention should be given to ensuring that the design of the instruments and their size, material and shape are appropriate.

NOTE Annexes A and B of this part of ISO 7971 contain examples of technical specifications of two hand-operated instruments with a capacity of 1 l.

5.3 Automatic measuring instrument. This category includes various types of devices, some of which can be used on their own or combined with an infrared analyser.

The measurement is based on the application of equations to allow the correcting of the bias and/or the drifts monitored. It does not include manual weighing. The numeric value of the hectolitre mass is directly displayed.

5.4 Analytical balance, capable of being read to the nearest 0,1 g or 0,01 g depending on the volume of the container (see 6.2).

6 Procedure

6.1 General

The measurements shall be taken using grain from which large impurities (straw, stones, husks, etc.) have been discarded, taking environmental conditions into consideration to ensure that there is no difference in temperature between the grain and the room in which the test is performed.

Determine the bulk density in duplicate. For all the devices and for every sample, it is advisable to perform the two measurements on two different grain test portions.

NOTE Repeating the measurement on the same grain test portion changes the friction coefficient which therefore makes it easier for the grains to slide; they are then more tightly packed, which increases the value of the bulk density.

6.2 Hand-operated instruments

Check that the various components of the instrument are clean and that they are working properly.

Make sure that equipment is placed on a firm, flat base, after using a spirit level to check that the base is horizontal.

Take great care to avoid any impact during filling. If the apparatus is jolted, cancel the test and start again.

Each type of apparatus is different; use each according to the manufacturer's instructions.

When using the analytical balance (5.4), weigh to the nearest 1 g for a 1 l container or the nearest 0,1 g for apparatus with a container of smaller volume.

6.3 Automatic instruments

As the operations to be performed prior to the actual measurement differ according to the type of equipment used, reference to the manufacturer's instructions is recommended.

Ensure that the instrument is placed on a horizontal surface in a room protected from extreme temperatures, humidity, dust and vibrations.

Take particular care to:

- a) select the correct cereal to be measured to ensure that the right calibration is used;
- b) use the volume of cereals recommended for the device in question;
- c) empty the collector drawer between samples.

6.4 Expression of results

Take the arithmetic mean of the two determinations as the result if the repeatability conditions are met.

Express the result to the nearest 0,1 kg/hl.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

7 Precision

ISO 7971-3:2009

<https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9fd869cd944/iso-7971-3-2009>

7.1 Interlaboratory test

Details of an interlaboratory test on the precision of the method are summarized in Annex C. The values derived from this interlaboratory test cannot be applied to other bulk density ranges and matrices than those given.

7.2 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, shall not in more that 5 % of cases be greater than the repeatability limit

$$r = 0,4$$

for products whose mass per hectolitre is between 67,5 kg/hl and 84,5 kg/hl (see Tables C.1, C.2, and Figure C.1).

7.3 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, shall not in more that 5 % of cases be greater than the reproducibility limit

$$R = 1,2$$

for products whose mass per hectolitre is between 67,5 kg/hl and 84,5 kg/hl (see Tables C.1, C.2, and Figure C.1).

7.4 Comparison of two groups of measurements in one laboratory

Critical difference, CD_r , is the difference between two averaged values obtained from two test results under repeatability conditions. As the result is a mean of two values (see 6.1), the comparison of two bulk densities shall be made with critical difference.

The critical difference, CD_r , between two averaged values obtained from two test results under repeatability conditions is given by:

$$CD_r = 2,8 s_r \sqrt{\frac{1}{2n_1} + \frac{1}{2n_2}} = 2,8 s_r \sqrt{\frac{1}{2}} = 1,98 s_r = 0,23$$

i.e. 0,2 kg/hl, after rounding, where

s_r is the standard deviation of repeatability;

n_1 and n_2 are the number of test results corresponding to each averaged value (here, $n_1 = n_2 = 2$).

7.5 Comparison of two groups of measurements in two laboratories

The critical difference, CD_R , between two averaged values obtained in two different laboratories from two test results under repeatability conditions is given by:

$$CD_R = 2,8 \sqrt{s_R^2 - s_r^2 \left(\frac{1}{2n_1} + \frac{1}{2n_2} \right)} = 2,8 \sqrt{s_R^2 - 0,5 s_r^2} = 1,18$$

i.e. 1,2 kg/hl, after rounding, where

s_r is the standard deviation of repeatability, <https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9f1869cd944/iso-7971-3-2009>

s_R is the standard deviation of reproducibility, <https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9f1869cd944/iso-7971-3-2009>

n_1 and n_2 are the number of test results corresponding to each of the averaged values (here $n_1 = n_2 = 2$).

7.6 Uncertainty

Uncertainty, U_e , is a parameter representing the distribution of the values which may reasonably be attributed to the result. This uncertainty is given by a statistical distribution of the results from the interlaboratory test and is characterized by the experimental standard deviation.

For mass per hectolitre, the uncertainty is given by

$$U_e = \pm 2s_R$$

$$U_e = \pm 0,43 \times 2 = \pm 0,86$$

i.e. 0,9 kg/hl, after rounding.

8 Test report

The test report shall contain at least the following information:

- a) an indication of the method used, including a reference to this part of ISO 7971;
- b) the result obtained;
- c) all the operating details not specified in this part of ISO 7971 or those regarded as optional, in addition to any possible incidents that might have affected the result;
- d) all the information necessary for a full identification of the sample.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 7971-3:2009

<https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9fd869cd944/iso-7971-3-2009>

Annex A (informative)

Description of dimensions and use of KERN¹⁾ apparatus

A.1 Dimensions of the apparatus

A.1.1 General

The dimensions of the various components of the apparatus should be as specified in A.1.2 to A.1.7. See Figure A.1 for a depiction of the components.

A.1.2 Pre-filling measure

Volume to level mark	1 350 ml ± 10 ml
Internal diameter	86 mm ± 0,2 mm

A.1.3 Filling hopper

Internal diameter	79 mm ± 0,1 mm
Wall thickness	1 mm ± 0,2 mm
Height above piston	280 mm ± 2 mm

<https://standards.iteh.ai/catalog/standards/sist/7c34039e-a8c6-4877-8b04-c9fd869cd944/iso-7971-3-2009>

A.1.4 Piston

Diameter	87,5 mm ± 0,1 mm
Height	40 mm ± 0,2 mm
Mass	450 g ± 2 g

A.1.5 Measuring container

Internal diameter	88,2 mm ± 0,1 mm
Internal height above piston	163,7 mm ± 0,1 mm
Wall thickness	1,2 mm ± 0,5 mm
External reinforcement of upper edge:	
— Thickness	2,5 mm ± 0,5 mm
— Height	6,0 mm ± 1,0 mm

1) Example of a suitable product available commercially. This information is given for the convenience of users of this part of ISO 7971 and does not constitute an endorsement by ISO of this product.

Base thickness	4,5 mm ± 0,1 mm
Diameter of base perforations	3,0 mm ± 0,1 mm
Foot height	9,0 mm ± 0,1 mm
Foot diameter	6,0 mm ± 0,1 mm
Gap between base and base plate	6,0 mm ± 0,1 mm
Number of perforations in base	1 + 4 + 8 + 12 + 16 + 20 + 24 = 85
Measuring ring:	
— internal diameter	88,2 mm ± 0,1 mm
— height	40,5 mm ± 0,1 mm

A.1.6 Base plate

Locating circle diameter	80,0 mm ± 0,1 mm
--------------------------	------------------

A.1.7 Straightedge (levelling blade)

Thickness	1 mm ± 0,05 mm
-----------	----------------

Cut-out angle	90° ± 2°
---------------	----------

Width of cutting edge bevel	3 mm ± 0,5 mm
-----------------------------	---------------

iTech STANDARD PREVIEW
(standards.iteh.ai)
ISO 7971-3:2009
<https://standards.iteh.ai/catalog/standards/sist/7c34059e-a8c6-4877-8b04-c9fd869cd944/iso-7971-3-2009>

A.2 Specifications of the apparatus

A.2.1 Pre-filling measure. The pre-filling measure should be made of metal in the shape of a straight-sided cylinder closed at the bottom end, with a flat base plate. It should have a circular level mark on the inside wall situated no less than 10 mm and no more than 30 mm from the open end.

NOTE The pre-filling measure controls the way in which the filling hopper (A.2.2) is filled with grain and consequently reduces or eliminates possible operator errors.

A.2.2 Filling hopper. The hopper should be made of metal in the shape of a straight-sided cylinder which is open at both ends. At the bottom of the cylinder, an extended projection around the circumference of the cylinder enables the filling hopper to be pushed on to the ring at the top of the measuring container (A.2.3). The hopper receives a quantity of grain of more than 1 l from the pre-filling measure (A.2.1).

A.2.3 Measuring container with measuring ring. The 1 l volume of the measuring container is delimited by the inside surface of the container wall, the upper side of the piston (A.2.4) and the underside of the straightedge (A.2.5), once it is in position. The maximum permissible relative error in the capacity of the container is ± 3/1 000. The wall of the measuring container consists of a seamless drawn brass tube or a stainless steel tube in the shape of a straight-sided cylinder that is open at the top and closed at its base, with a reinforced outer edge. The top edge should be ground flat.

A measuring ring with the same internal diameter as the measuring container should be attached to the edge of the measuring container. The gap between the edge and the measuring ring should be sufficient to allow the straightedge (A.2.5) to pass through easily, but the clearance should not be too great.