

ISO/TC 150

Secretariat: DIN

Voting begins on:  
2008-08-28

Voting terminates on:  
2008-10-28

---

---

## Instrumentation for use in association with non-active surgical implants — General requirements

*Instrumentation à utiliser en association avec les implants chirurgicaux  
non actifs — Exigences générales*

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/ba12a28-6b13-4a16-85d0-3510e7bd25c9/iso-16061-2008>

Please see the administrative notes on page iii

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number  
ISO/FDIS 16061:2008(E)

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/bab2ab38-6b13-4a16-85d0-3510e7bd25c9/iso-16061-2008>

**Copyright notice**

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

## ISO/CEN PARALLEL PROCESSING

The CEN Secretary-General has advised the ISO Secretary-General that this final draft International Standard covers a subject of interest to European standardization. Consultation on the ISO/DIS had the same effect for CEN members as a CEN enquiry on a draft European Standard. In accordance with the ISO-lead mode of collaboration as defined in the Vienna Agreement, this final draft, established on the basis of comments received, is hereby submitted to a parallel two-month FDIS vote in ISO and formal vote in CEN.

**Positive votes shall not be accompanied by comments.**

**Negative votes shall be accompanied by the relevant technical reasons.**

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/bab2ab38-6b13-4a16-85d0-3510e7bd25c9/iso-16061-2008>

**Contents**

Page

Foreword .....	v
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Intended performance</b> .....	<b>2</b>
<b>5 Design attributes</b> .....	<b>2</b>
<b>6 Selection of materials</b> .....	<b>3</b>
<b>7 Design evaluation</b> .....	<b>3</b>
7.1 General .....	3
7.2 Pre-clinical evaluation .....	3
7.3 Clinical evaluation .....	3
<b>8 Manufacture</b> .....	<b>4</b>
<b>9 Sterilization</b> .....	<b>4</b>
9.1 Products supplied sterile .....	4
9.2 Products provided non-sterile .....	4
<b>10 Packaging</b> .....	<b>4</b>
10.1 Protection from damage in storage and transport .....	4
10.2 Maintenance of sterility in transit .....	4
<b>11 Information to be supplied by the manufacturer</b> .....	<b>5</b>
11.1 General .....	5
11.2 Instruments with measuring function .....	5
11.3 Restrictions in combinations .....	5
11.4 Marking on instruments .....	5
11.5 Instructions for use .....	5
11.6 Instruments intended for single use .....	5
<b>Annex A (informative) Examples of typical instrument applications, together with materials found acceptable for instrument manufacture</b> .....	<b>6</b>
<b>Bibliography</b> .....	<b>17</b>
<b>Annex ZA (informative) Relationship between this International Standard and the Essential Requirements of EU Directive 93/42 EEC</b> .....	<b>19</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 16061 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*.

This second edition cancels and replaces the first edition (ISO 16061:2000), which has been technically revised.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/6b13-4a16-85d0-3510e7bd25c9/iso-16061-2008>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/bab2ab38-6b13-4a16-85d0-3510e7bd25c9/iso-16061-2008>

# Instrumentation for use in association with non-active surgical implants — General requirements

## 1 Scope

This International Standard specifies general requirements for instruments to be used in association with non-active surgical implants. These requirements apply to instruments when they are manufactured and when they are resupplied after refurbishment.

This International Standard also applies to instruments which may be connected to power-driven systems, but does not apply to the power-driven systems themselves.

With regard to safety, this International Standard gives requirements for intended performance, design attributes, selection of materials, design evaluation, manufacture, sterilization, packaging and information to be supplied by the manufacturer.

This International Standard is not applicable to instruments associated with dental implants, transendodontic and transradicular implants and ophthalmic implants.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11135-1, *Sterilization of health care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 14155-1, *Clinical investigation of medical devices for human subjects — Part 1: General requirements*

ISO 14155-2, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 17664, *Sterilization of medical devices — Information to be provided by the manufacturer for the processing of resterilizable medical devices*

ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

EN 556-1, *Sterilization of medical devices — Requirements for medical devices to be designated “STERILE” — Part 1: Requirements for terminally sterilized medical devices*

EN 556-2, *Sterilization of medical devices — Requirements for medical devices to be designated “STERILE” — Part 2: Requirements for aseptically processed medical devices*

EN 1041:1998, *Information supplied by the manufacturer with medical devices*

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

**3.1 associated instrument**  
**instrument**  
non-active medical device intended for use during surgical procedures related to a specific non-active surgical implant

**3.2 resupplied instrument**  
instrument or set of instruments that has been returned to the manufacturer and has been re-issued

### 4 Intended performance

The intended performance of an instrument shall be described and documented by addressing the following:

- a) functional characteristics;
- b) intended conditions of use.

NOTE Account should be taken of

- published standards;
- published clinical and scientific literature;
- validated test results.

The extent to which the intended performance of an instrument has been achieved shall be determined (see Clause 7).

### 5 Design attributes

The development of the design attributes of an instrument to meet the performance intended by the manufacturer shall take into account at least the following:

- a) physical, mechanical and chemical properties of the instrument materials (see Clauses 6 and 7);
- b) microbiological and particulate contamination levels (see Clauses 7, 9 and 10);
- c) ease of use, cleaning and maintenance (see Clause 7);



- d) potential deterioration of the material characteristics due to sterilization and storage (see Clauses 6, 7 and 8);
- e) effects of contact between the instrument and body, the implant and other instruments (see Clause 7);
- f) shape and dimensions of the instrument, including their possible effects on the body (see Clause 7);
- g) wear characteristics of materials and the effect of wear and wear products on the instrument and the body (see Clauses 6 and 7);
- h) insertion, removal and interconnection of parts (see Clause 7);
- i) extent of fluid leakage and/or diffusion of substances into or out of instruments (see Clauses 6 and 7);
- j) accuracy and stability of instruments with a measuring function (see Clauses 7 and 8);
- k) ability of the instrument or fragment of instrument to be located by means of an external imaging device (see 11.5).

## 6 Selection of materials

Materials for the manufacture of instruments shall be selected with regard to the properties required for the intended purpose, taking into account the effects of manufacture, handling, sterilization and storage (see Clause 7).

The suitability of a given material for a particular application shall be demonstrated by either:

- a) evaluating in accordance with Clause 7
- or
- b) selection from the materials found suitable by proven clinical use in similar applications.

NOTE Annex A lists some of the materials that have been found acceptable in certain applications.

## 7 Design evaluation

### 7.1 General

Instruments shall be evaluated in association with the implant they are designed for, in order to demonstrate that the intended performance (see Clause 4) is achieved. Safety shall be demonstrated by pre-clinical evaluation and by carrying out a risk analysis in accordance with ISO 14971.

NOTE In certain circumstances, a clinical evaluation can also be required.

### 7.2 Pre-clinical evaluation

If pre-clinical testing of instruments is required, the testing shall simulate conditions of intended use.

### 7.3 Clinical evaluation

If a clinical evaluation is required, it shall be conducted using the associated implant under the intended conditions of use. Where a clinical investigation is carried out, it shall be managed in accordance with ISO 14155-1 and ISO 14155-2.

## 8 Manufacture

Instruments shall be manufactured to specifications in accordance with the required design attributes (see Clause 5).

NOTE 1 The application of quality systems as described in ISO 13485 might be appropriate.

NOTE 2 The design specification for re-supplied instruments need not necessarily be the same as the original, provided that the requirements of this International Standard are met.

## 9 Sterilization

### 9.1 Products supplied sterile

Instruments which are labelled "STERILE" shall comply with EN 556-1 and EN 556-2.

Sterilization processes shall be validated and routinely controlled.

If instruments are to be sterilized by ethylene oxide, ISO 11135-1 applies,

If instruments are to be sterilized by irradiation, ISO 11137-1 applies,

If instruments are to be sterilized by steam, ISO 17665-1 applies.

### 9.2 Products provided non-sterile

For instruments that are supplied non-sterile, the manufacturer shall specify at least one appropriate sterilization method such that the functional safety of the product is not adversely affected. If multiple sterilizations are not allowed, this shall be stated (see 11.6).

For instruments that are supplied non-sterile or claimed to be re-sterilizable, the manufacturer shall provide information on the processing of these instruments in accordance with ISO 17664.

## 10 Packaging

### 10.1 Protection from damage in storage and transport

For each instrument, the packaging shall be designed so that, under conditions specified by the manufacturer for storage, transport and handling (including control of temperature, humidity and ambient pressure, if applicable), it protects against damage and deterioration and does not adversely affect the intended performance of the instrument.

NOTE Possible test methods are specified in IEC 60068-2-27, IEC 60068-2-31 and/or IEC 60068-2-47.

### 10.2 Maintenance of sterility in transit

Instruments labelled "STERILE" shall be packed such that they remain sterile under normal storage, transport and handling conditions, unless the protective package is damaged or opened.

The packaging shall conform to ISO 11607-1 and ISO 11607-2.