



Designation: D 3195 – 90 (Reapproved 1998)<sup>ε1</sup>

## Standard Practice for Rotameter Calibration<sup>1</sup>

This standard is issued under the fixed designation D 3195; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

<sup>ε1</sup> NOTE—Paragraph X1.5 was corrected editorially in October 1998.

### 1. Scope

1.1 This practice covers the calibration of variable-area flowmeters (rotameters) used to determine air sample volumes at or close to ambient conditions of pressure and temperature, in the analysis of atmospheres for pollutant content.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

### 2. Referenced Documents

#### 2.1 ASTM Standards:

- D 1071 Test Methods for Volumetric Measurement of Gaseous Fuel Samples<sup>2</sup>
- D 1356 Terminology Relating to Sampling and Analysis of Atmospheres<sup>3</sup>
- D 3631 Test Methods for Measuring Surface Atmospheric Pressure<sup>3</sup>
- E 1 Specification for ASTM Thermometers<sup>4</sup>
- E 337 Test Method for Measuring Humidity with a Psychrometer (the Measurement of Wet- and Dry-Bulb Temperatures)<sup>3</sup>

### 3. Terminology

#### 3.1 Definitions:

3.1.1 For definitions of terms used in this practice, refer to Terminology D 1356.

3.1.2 *standard conditions* are taken as 25°C and 101.3 kPa (760 mm Hg) at existing ambient humidity. This conforms to most of the ASTM methods for atmospheric sampling and analysis that involve volumetric corrections. Absolute tempera-

ture scales are to be used when substituting values into the formulae used in this procedure.

### 4. Summary of Practice

4.1 Two alternative methods of performing the required volume determinations for rotameter calibration are described:

4.1.1 Using the water-sealed rotating drum meter (wet test meter). See Section 7.

4.1.2 Using the volumetric gasometer (bell prover). See Section 8.

### 5. Significance and Use

5.1 Choice of method depends primarily on which equipment is available. Higher accuracy is possible with the gasometer. The accuracies of the methods of atmospheric analysis, for which the calibration procedure is intended, do not warrant the very highest possible accuracy in flow measurement.

### 6. Apparatus

6.1 *Wet Test Meter, or Volumetric Gasometer*, with water seal and equipped with a water manometer on the inlet.

6.2 *Counter Balance Weights*, for gasometer.

6.3 *Mercury Barometer*—See Test Methods D 3631.

6.4 *Psychrometer*, (if room air is used for calibration gas). See Test Method E 337.

6.5 *Thermometer*, to measure ambient temperature. See Specification E 1.

6.6 *Stopwatch*.

6.7 *Air Supply*, either a cylinder of compressed air or a diaphragm type pump of adequate capacity and a ballast volume or restrictor to eliminate pulsations.

6.8 *Needle Valve*.

### 7. Procedure Using Wet Test Meter

7.1 Unless it was already calibrated within the previous 3 months, calibrate the wet test meter by Test Methods D 1071. The method described in Section 19 is recommended for highest accuracy.

7.2 Set up the apparatus as shown in Fig. 1, making connections as short as possible and large enough inside diameter to avoid any appreciable pressure drops.

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee D22 on Sampling and Analysis of Atmospheres and is the direct responsibility of Subcommittee D22.01 on Quality Control.

Current edition approved Dec. 28, 1990. Published February 1991. Originally published as D 3195 – 73. Last previous edition D 3195 – 73 (1985)<sup>ε1</sup>.

<sup>2</sup> *Annual Book of ASTM Standards*, Vol 05.05.

<sup>3</sup> *Annual Book of ASTM Standards*, Vol 11.03.

<sup>4</sup> *Annual Book of ASTM Standards*, Vol 14.03.




FIG. 1 Calibration Assembly Using Wet Test Meter

7.3 Before and after the complete calibration run, record room temperature, barometric pressure in accordance with Test Methods D 3631, and relative humidity (when room air is used for calibrating gas) in accordance with Test Method E 337. Use average values for subsequent calculations.

7.4 Start air flowing through the rotameter and wet test meter. Adjust the flow to the desired rate with the needle valve. Take a pair of timed readings on the wet test meter, under steady flow, for each of five or more uniformly spaced points on the rotameter scale, going from low values to high values. Repeat, going from high to low. Note the manometer reading and meter water temperature for each meter reading.

### 8. Procedure Using Gasometer

8.1 Unless it was already calibrated within the previous six months, in the same location, calibrate the gasometer by Test Methods D 1071.

8.2 Set up the apparatus as shown in Fig. 2, making connections as short as possible and large enough inside diameter to avoid any appreciable pressure drops.

8.3 Before and after the complete calibration run, record the room temperature, barometric pressure, and relative humidity (when room air is used for calibrating gas). Use average values for subsequent calculations.




FIG. 2 Calibration Assembly Using Gasometer