
Kakovost vode - Merjenje celotne beta aktivnosti v neslanih vodah - Metoda robustnega vira

Water quality - Measurement of gross beta activity in non-saline water - Thick source method

Qualité de l'eau - Mesurage de l'activité bêta globale des eaux non salines - Méthode par source concentrée

Ta slovenski standard je istoveten z: ISO 9697:2008

ICS:

| | | |
|-----------|-------------------------------------|---|
| 13.060.60 | Preiskava fizikalnih lastnosti vode | Examination of physical properties of water |
| 13.280 | Varstvo pred sevanjem | Radiation protection |

oSIST ISO 9697:2009**en,fr**

INTERNATIONAL STANDARD

**ISO
9697**

Second edition
2008-11-01

Water quality — Measurement of gross beta activity in non-saline water — Thick source method

*Qualité de l'eau — Mesurage de l'activité bêta globale des eaux non
salines — Méthode par source concentrée*



Reference number
ISO 9697:2008(E)

© ISO 2008

ISO 9697:2008(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

| | |
|--|----|
| Foreword..... | iv |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Symbols, definitions, and units..... | 2 |
| 4 Principle | 2 |
| 5 Reagents and equipment | 3 |
| 6 Procedure | 4 |
| 7 Source control..... | 7 |
| 8 Expression of results | 7 |
| 9 Test report | 9 |
| Bibliography | 11 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9697 was prepared by Technical Committee ISO/TC 147, *Water quality*.

This second edition cancels and replaces the first edition (ISO 9697:1992), which has been technically revised.

Water quality — Measurement of gross beta activity in non-saline water — Thick source method

WARNING — Persons using this International Standard should be familiar with normal laboratory practice. This International Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this International Standard be carried out by suitably trained staff.

1 Scope

This International Standard specifies a method for the determination of gross beta activity in non-saline waters. The method covers non-volatile radionuclides with maximum beta energies $> 0,3$ MeV. Measurement of very low energy beta emitters, such as ^3H , ^{14}C , ^{35}S and ^{241}Pu , is not included in this International Standard.

The method is applicable to the analysis of raw and potable waters.

2 Normative references

The following referenced documents are indispensable for the application of this documents. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of water samples*

ISO 5667-14, *Water quality — Sampling — Part 14: Guidance on quality assurance of environmental water sampling and handling*

ISO 9696, *Water quality — Measurement of gross alpha activity in non-saline water — Thick source method*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics* ¹⁾

1) To be published. (Revision of ISO 31-9:1992)

ISO 9697:2008(E)

3 Symbols, definitions, and units

For the purposes of this document, the symbols, definitions, and units given in ISO 80000-10, and the following, apply.

| | |
|----------------|---|
| A | beta activity, in becquerels, of the calibration source |
| c_A | beta activity concentration, in becquerels per litre |
| c_A^* | decision threshold, in becquerels per litre |
| $c_A^\#$ | detection limit, in becquerels per litre |
| $c_A^<, c_A^>$ | lower and upper limits of the confidence interval, in becquerels per litre |
| m | mass, in milligrams, of ignited residue from volume, V |
| m_r | mass, in milligrams, of the residue deposited on the planchet sample |
| r_0 | background count rate, in reciprocal seconds |
| $r_{0\alpha}$ | background count rate, in reciprocal seconds, from the alpha window |
| r_g | sample gross count rate, in reciprocal seconds |
| $r_{g\alpha}$ | gross count rate, in reciprocal seconds, from the alpha calibration source |
| r_s | calibration count rate, in reciprocal seconds |
| $r_{s\alpha}$ | calibration count rate, in reciprocal seconds, from the alpha calibration source |
| S | area, in square millimetres, of the planchet |
| t_0 | background counting time, in seconds |
| t_g | sample counting time, in seconds |
| $u(c_A)$ | standard uncertainty, in becquerels per litre, associated with the measurement result |
| U | expanded uncertainty, in becquerels per litre, calculated from $U = ku(c_A)$, with $k = 1, 2 \dots$ |
| V | volume, in litres, of test sample equivalent to the mass of solid on the planchet |
| V_t | volume, in litres, of the water sample |
| ε | counting efficiency for the specified radioactive standard |
| χ | alpha-beta cross-talk, percentage of alpha count going into the beta window from the alpha calibration source |

4 Principle

Gross beta activity determination is not an absolute determination of the radioactivity of all beta-emitting radionuclides in a sample, but a relative determination referred to a specific beta emitter that constitutes the standard calibration source. This type of determination is also known as beta index.

The sample, taken, handled and preserved as specified in ISO 5667-1 and ISO 5667-3, is evaporated almost to dryness, converted to the sulfate form and then ignited at 350 °C. A portion of the residue is transferred to a planchet and the beta activity measured by counting in an appropriate counting assembly calibrated against a suitable beta calibration standard, such as potassium-40 (^{40}K) or strontium-90-yttrium-90 ($^{90}\text{Sr} + ^{90}\text{Y}$).

If simultaneous gross alpha and beta measurements are required on the same water sample, the procedure specified in this International Standard is common to that of ISO 9696, and sample sources can be measured up to 10 mg/cm².