

First edition
2009-08-01

Corrected version
2009-11-15

**Optics and photonics — Designation of
microscope objectives —**

**Part 2:
Chromatic correction**

Optique et photonique — Désignation des objectifs de microscope —

Partie 2: Correction chromatique

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

ISO 19012-2:2009

<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>



Reference number
ISO 19012-2:2009(E)

© ISO 2009

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 19012-2:2009

<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 19012-2 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 5, *Microscopes and endoscopes*.

ISO 19012 consists of the following parts, under the general title *Optics and photonics — Designation of microscope objectives*:

— *Part 1: Flatness of field/Plan*

[ISO 19012-2:2009](#)

— *Part 2: Chromatic correction*

<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>

In this corrected version of ISO 19012-2:2009, on page 3, subclause 4.2.3 has been re-worded and mention of the reference wavelength removed.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 19012-2:2009

<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>

Optics and photonics — Designation of microscope objectives —

Part 2: Chromatic correction

1 Scope

This part of ISO 19012 specifies classes of chromatic correction and defines minimum requirements regarding chromatic correction. The defined marking on the component enables the operator to correctly use the microscope.

The standard application for visual observation refers to the combination of objective and tube lens as specified by the manufacturer. The specifications regarding chromatic correction only refer to axial chromatic aberration.

iTeh STANDARD PREVIEW

2 Normative references (standards.iteh.ai)

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10934-1, *Optics and optical instruments — Vocabulary for microscopy — Part 1: Light microscopy*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10934-1 and the following apply.

3.1

reference wavelength

wavelength of 546,07 nm (e-line)

3.2

blue wavelength

wavelength of 479,99 nm (F'-line)

3.3

red wavelength

wavelength of 643,85 nm (C'-line)

3.4

focus

best focusing point for each wavelength

3.5

focus difference

axial separation of foci for different wavelengths

4 Requirements

4.1 Basic criterion for the depth of field

Equation (1) applies as the basic criterion for the depth of field:

$$\delta_{\text{ob}} = \frac{n\lambda}{2NA^2} \quad (1)$$

where

n is the refractive index of medium in object space;

NA is the numerical aperture of objective;

λ is the wavelength of the reference wave e-line in micrometers.

A table of δ_{ob} depending on NA can be found in Annex A.

4.2 Markings

4.2.1 General

The following markings may be used if the requirements according to 4.3 are met.

The indication of this marking does not apply to objective lenses sold before the year 2011.

This part of ISO 19012 does not apply to the objectives exclusively used on stereo microscopes.

A mixture of a capital letter and a lowercase letter is allowed in marking.

4.2.2 Achromat

Marking is not necessary but possible.

ACH, ACHRO, ACHROMAT

4.2.3 Semiapochromat

Objective lenses shall be marked with one of the following three options:

- a) SEMIAPO, or
- b) FL, or
- c) a naming containing the letter sequence FLU.

4.2.4 Apochromat

APO

4.3 Specifications

4.3.1 General

The specifications of the “Semiapochromat” and “Apochromat” include the criterion of “Achromat”.

4.3.2 Achromat

The absolute value of the focus difference between the red wavelength and the blue wavelength is $\leq 2 \times \delta_{ob}$.

4.3.3 Semiapochromat

The absolute values of the focus differences for the red wavelength and the blue wavelength to the reference wavelength are $\leq 2,5 \times \delta_{ob}$.

4.3.4 Apochromat

The absolute values of the focus differences for the red wavelength and the blue wavelength to the reference wavelength are $\leq \delta_{ob}$.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 19012-2:2009

<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>

Annex A
(informative)

Depth of field, δ_{ob}

Dry		Immersion	
n	1	n	1,518
λ (μm)	0,546	λ (μm)	0,546
NA	δ_{ob} (μm)	NA	δ_{ob} (μm)
0,04	170,63	0,40	2,59
0,07	55,71	0,70	0,85
0,10	27,30	0,90	0,51
0,13	16,15	1,00	0,41
0,15	12,13	1,25	0,27
0,16	10,66	1,30	0,25
0,20	6,83	1,35	0,23
0,22	5,64	1,40	0,21
0,25	4,37		
0,30	3,03		
0,35	2,23		
0,40	1,71		
0,45	1,35		
0,50	1,09		
0,55	0,90		
0,60	0,76		
0,65	0,65		
0,70	0,56		
0,75	0,49		
0,80	0,43		
0,85	0,38		
0,90	0,34		
0,95	0,30		

STANDARD PREVIEW
(standards.iteh.ai)

ISO 19012-2:2009
<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>

Bibliography

- [1] ISO 8578, *Optics and optical instruments — Microscopes — Marking of objectives and eyepieces*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 19012-2:2009](https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009)

<https://standards.iteh.ai/catalog/standards/sist/80ea5011-edd1-4aa0-8755-7259036c844a/iso-19012-2-2009>