
**Non-destructive testing — Infrared
thermography — Vocabulary**

Essais non destructifs — Thermographie infrarouge — Vocabulaire

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 10878:2013

<https://standards.iteh.ai/catalog/standards/iso/30c4fbef-5984-4c07-921a-b2eff275c841/iso-10878-2013>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 10878:2013

<https://standards.iteh.ai/catalog/standards/iso/30c4fbef-5984-4c07-921a-b2eff275c841/iso-10878-2013>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

The reproduction of the terms and definitions contained in this International Standard is permitted in teaching manuals, instruction booklets, technical publications and journals for strictly educational or implementation purposes. The conditions for such reproduction are: that no modifications are made to the terms and definitions; that such reproduction is not permitted for dictionaries or similar publications offered for sale; and that this International Standard is referenced as the source document.

With the sole exceptions noted above, no other part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
Scope	1
1 Terms and definitions	1
Bibliography.....	26
Alphabetical index.....	27

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 10878:2013

<https://standards.iteh.ai/catalog/standards/iso/30c4fbef-5984-4c07-921a-b2eff275c841/iso-10878-2013>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 135, *Non-destructive testing*, Subcommittee SC 8, *Infrared thermography for non-destructive testing*.

iteh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 10878:2013

<https://standards.iteh.ai/catalog/standards/iso/30c4fbef-5984-4c07-921a-b2eff275c841/iso-10878-2013>

Introduction

This International Standard is a compilation of terms and definitions to provide a precise understanding or interpretation of infrared thermography and thermal/infrared non-destructive testing. These serve to secure the foundation of infrared thermography technology growth within the academic and industrial communities.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 10878:2013](https://standards.iteh.ai/catalog/standards/iso/30c4fbef-5984-4c07-921a-b2eff275c841/iso-10878-2013)

<https://standards.iteh.ai/catalog/standards/iso/30c4fbef-5984-4c07-921a-b2eff275c841/iso-10878-2013>

Non-destructive testing — Infrared thermography — Vocabulary

Scope

This International Standard defines terms used in infrared thermography for non-destructive testing and forms a common basis for standard general use.

1 Terms and definitions

1.1

absorptivity

α

absorptance

absorptance coefficient

proportion (as a fraction of 1) of the radiant energy impinging on a material's surface that is absorbed by the material

NOTE 1 Absorptivity is dimensionless.

NOTE 2 For a blackbody, this is unity (1,0). Technically, absorptivity is the internal absorptance per path length. In thermography, the two terms, absorptivity and absorptance, are often used interchangeably.

NOTE 3 Absorptance is the ratio between the radiation energy absorbed by a body and the total radiation incident on the body.

NOTE 4 Absorptivity can vary with wavelength and be quoted for a specified band width or a specific wavelength. See 1.136, Spectral absorption coefficient.

1.2

active thermography

infrared thermographic examination of materials and objects which requires additional thermal stimulation

NOTE The thermal stimulation can be optical, sonic (ultrasonic), inductive, microwave or use any other form of energy.

1.3

ambient operating range

range of ambient temperatures over which an instrument is designed to operate within reported performance specifications

1.4

ambient temperature

temperature of the air in the vicinity of a test object (target)

NOTE "Ambient temperature" is not to be confused with "reflected ambient temperature", which is a term often used to mean "reflected apparent temperature".

1.5

ambient temperature compensation

correction built into infrared instruments to provide automatic compensation of temperature readings affected by the ambient temperature

1.6

angular subtense

angular diameter of an optical system or subsystem

NOTE 1 Angular subtense is expressed in angular degrees or milliradians.

NOTE 2 In infrared thermography, the angle over which a sensing instrument collects radiant energy.

1.7

anomalous thermal image

observed thermal pattern of a structure that is not in accordance with the expected (reference) thermal pattern

1.8

anomaly

irregularity or abnormality in a system

EXAMPLE An irregularity, such as an anomalous thermal pattern or any indication that deviates from what is normally expected in the absence of any anomaly.

1.9

anti-reflectance coating

coating of infrared optical elements (lenses, windows) used to increase the sensitivity of a specified wavelength range through minimization or suppression of reflections causing signal loss

1.10

apparent temperature

uncompensated reading from an infrared thermography camera containing all radiation incident on the detector, regardless of its source

[ISO 18434-1:2008^[6], 3.1]

1.11

area effect

change in infrared radiometer output depending on the area of the measuring target

1.12

artefact

⟨1⟩ product of artificial character due to an extraneous agency

⟨2⟩ error caused by an uncompensated anomaly

EXAMPLE In thermography, an emissivity artefact simulates apparent variation of surface temperature.

1.13

atmospheric absorption

absorption of specific wavelengths of solar radiation, due largely to moisture, atmospheric gases and pollutants

1.14

atmospheric temperature

temperature of the atmosphere between the infrared camera and the object

1.15

atmospheric window

⟨infrared⟩ any spectral interval within the infrared spectrum in which the atmosphere transmits radiant energy well (atmospheric absorption is minimal)

EXAMPLE Atmospheric windows are roughly defined to lie in the wavelength ranges:

- a) 0,78 μm to 2,0 μm in the near infrared (NIR);
- b) 2,0 μm to 5,5 μm in the mid-wave infrared (MWIR);
- c) 7,5 μm to 14,0 μm in the long-wave infrared (LWIR).

1.16**attenuating medium**

material or other medium that attenuates infrared radiation emitted from a source

EXAMPLE Attenuating media include windows, filters, atmospheres, external optics.

1.17**blackbody**

ideal perfect emitter and absorber of thermal radiation at all wavelengths

NOTE A blackbody is described by Planck's law. In its classical form, Planck's law describes the spectral distribution of the radiant energy emitted by a blackbody.

1.17.1**blackbody equivalent temperature**

apparent temperature of a test object that is equal to the temperature of a blackbody emitting the same amount of radiant energy

1.17.2**blackbody radiator**

radiator with the effective emissivity ε close to unity ($\varepsilon \geq 0,98$ across all relevant wavelengths)

1.17.3**blackbody reference**

calibrated, traceable device used to check the calibration of infrared imaging radiometers or infrared thermometers

1.17.4**blackbody simulator**

device whose radiation is close to that of a blackbody at the same temperature

EXAMPLE A cavity or a flat plate with a structured or coated surface characterized by a stable and uniform temperature and with emissivity close to 1.

1.18**centre wavelength**

wavelength in the middle of the spectral sensitivity band of an infrared detector

1.19**cooled sensor**

sensor that needs cooling to improve sensitivity to infrared radiant energy by reducing thermal noise influence

1.20**detecting element**

sensitive part of a detector which is directly affected by the quantity to be measured

EXAMPLE For temperature-sensing devices: a thermocouple junction; resistive element; photoelectric, pyroelectric or quantum sensor.

1.21**differential blackbody**

device for establishing two parallel isothermal planar zones of different temperatures and with effective emissivities close to 1,0

1.22**diffraction limit**

limit of geometric diffraction of optical systems

1.23

diffuse reflector

lambertian reflector

surface that reflects incident radiation equally in all directions

NOTE 1 A lambertian diffuser is a surface that reflects a portion of the incident radiation in such a manner that the reflected radiation is equal in all directions, such as a gold perfect sphere.

NOTE 2 A mirror is not a diffuse reflector.

1.24

edge effect

(1) effect caused by measurement error mainly at the edge due to solid displacement or deformation by variable loading in thermoelastic stress measurement

(2) change in thermal properties at the edge of a target object as a result of different thermal conduction and convection properties

EXAMPLE Effect caused by measurement error at an edge due to solid displacement or deformation by variable loading in thermoelastic stress measurement.

1.25

effective emissivity

ε^*

measured emissivity value of a particular target surface under existing measurement conditions (rather than the generic tabulated value for the same material) that can be used to correct specific temperature readings

NOTE 1 Effective emissivity is also called emittance; however, the latter term is not preferred because it has been used to describe radiant exitance.

NOTE 2 Effective emissivity is context dependent, and is not purely a property of a material.

1.26

effective number of pixels

spatial resolution of a measured infrared image

NOTE The effective number of pixels is determined for a scanning infrared thermographic instrument according to the scanning pitch, and for an infrared thermographic instrument with an array sensor according to the number of pixels of the detector.

1.27

EMI/RFI noise

disturbance to electrical signals caused by electromagnetic interference (EMI) or radio frequency interference (RFI)

NOTE In infrared thermography, EMI/RFI noise can cause patterns to appear on the display and is sometimes due to poor grounding or earthing.

1.28

emissivity

ε

ratio of the radiance of a target surface to that of a blackbody at the same temperature and over the same spectral interval

1.29

emittance

ratio of the radiant flux emitted by a real target and that emitted by a blackbody at the same temperature and under the same conditions

NOTE 1 The total radiance, R^0 , is obtained by an integration of the monochromatic radiance between wavelengths zero and infinity.

$$R^0 = \int_0^{\infty} \frac{2\pi hc^2 \lambda^{-5}}{\exp\left(\frac{hc}{k\lambda T}\right) - 1} d\lambda = \frac{\sigma}{\pi} T^4$$

where

- c is the speed of light in a vacuum;
- h is the Planck constant;
- k is the Boltzmann constant;
- T is the thermodynamic temperature;
- σ is the Stefan-Boltzmann constant, in watts per square metre per kelvin to the power four, given by

$$\sigma = \frac{2\pi^5 k^4}{15c^2 h^3} = 5,67 \times 10^{-8}$$

Radiance and emittance being connected, the total emittance, M^0 , is given by

$$M^0 = \pi R^0 = \sigma T^4$$

NOTE 2 In thermography, the terms “radiance” and “emittance” are technically often used interchangeably.

NOTE 3 Refer to ISO 80000-7

1.30 environmental rating

rating assigned to an operating unit (typically an electrical or mechanical enclosure) to indicate the limits of the environmental conditions under which the unit functions reliably and within reported performance specifications

1.31 extended source

source of infrared radiation whose image fills completely or a larger part of the field of view of the infrared camera

1.32 field of view field of vision FOV

angular subtense over which an instrument integrates total incoming radiant energy

NOTE 1 Angular subtense is expressed in angular degrees or radians per side if rectangular or square and in angular degrees or radians if circular.

NOTE 2 In infrared thermometers, field of view defines the target spot size; in a scanning/staring imager, it defines the scan angle or picture size or a total field of view (TFOV).

NOTE 3 The field of view is the angular extent of the observable world that is seen at any given moment.

See Figure 1.