



SLOVENSKI STANDARD

SIST ISO 21748:2014

01-januar-2014

Nadomešča:

SIST-TS ISO/TS 21748:2006

Napotek za uporabo ocen ponovljivosti, obnovljivosti in pravilnosti pri ocenjevanju merilne negotovosti

Guidance for the use of repeatability, reproducibility and trueness estimates in measurement uncertainty estimation

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Lignes directrices relatives à l'utilisation d'estimations de la répétabilité, de la reproductibilité et de la justesse dans l'évaluation de l'incertitude de mesure

[https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-](https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014)

[1f0647b39c58/sist-iso-21748-2014](https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014)

Ta slovenski standard je istoveten z: ISO 21748:2010

ICS:

17.020

Meroslovje in merjenje na splošno

Metrology and measurement in general

SIST ISO 21748:2014

en,fr

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 21748:2014](#)

<https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014>

**Guidance for the use of repeatability,
reproducibility and trueness estimates in
measurement uncertainty estimation**

*Lignes directrices relatives à l'utilisation d'estimations de la répétabilité,
de la reproductibilité et de la justesse dans l'évaluation de l'incertitude
de mesure*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 21748:2014](https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014)

[https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-
1f0647b39c58/sist-iso-21748-2014](https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014)



ISO 21748:2010(E)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 21748:2014](https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014)

<https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Terms and definitions	1
3 Symbols.....	4
4 Principles	7
4.1 Individual results and measurement process performance	7
4.2 Applicability of reproducibility data	7
4.3 Basic equations for the statistical model	8
4.4 Repeatability data.....	9
5 Evaluating uncertainty using repeatability, reproducibility and trueness estimates	9
5.1 Procedure for evaluating measurement uncertainty	9
5.2 Differences between expected and actual precision.....	9
6 Establishing the relevance of method performance data to measurement results from a particular measurement process.....	10
6.1 General	10
6.2 Demonstrating control of the laboratory component of bias	10
6.3 Verification of repeatability	12
6.4 Continued verification of performance	13
7 Establishing relevance to the test item.....	13
7.1 General	13
7.2 Sampling	13
7.3 Sample preparation and pre-treatment	14
7.4 Changes in test-item type.....	14
7.5 Variation of uncertainty with level of response	14
8 Additional factors	15
9 General expression for combined standard uncertainty	15
10 Uncertainty budgets based on collaborative study data	16
11 Evaluation of uncertainty for a combined result.....	17
12 Expression of uncertainty information	18
12.1 General expression	18
12.2 Choice of coverage factor	18
13 Comparison of method performance figures and uncertainty data.....	18
13.1 Basic assumptions for comparison	18
13.2 Comparison procedure.....	19
13.3 Reasons for differences.....	19
Annex A (informative) Approaches to uncertainty estimation.....	20
Annex B (informative) Experimental uncertainty evaluation.....	25
Annex C (informative) Examples of uncertainty calculations	26
Bibliography.....	37

ISO 21748:2010(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21748 was prepared by Technical Committee ISO/TC 69, *Applications of statistical methods*, Subcommittee SC 6, *Measurement methods and results*.

This first edition cancels and replaces ISO/TS 21748:2004, which has been technically revised.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 21748:2014
<https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014>

Introduction

Knowledge of the uncertainty associated with measurement results is essential to the interpretation of the results. Without quantitative assessments of uncertainty, it is impossible to decide whether observed differences between results reflect more than experimental variability, whether test items comply with specifications, or whether laws based on limits have been broken. Without information on uncertainty, there is a risk of misinterpretation of results. Incorrect decisions taken on such a basis may result in unnecessary expenditure in industry, incorrect prosecution in law, or adverse health or social consequences.

Laboratories operating under ISO/IEC 17025 accreditation and related systems are accordingly required to evaluate measurement uncertainty for measurement and test results and report the uncertainty where relevant. The *Guide to the expression of uncertainty in measurement* (GUM), published by ISO/IEC as ISO/IEC Guide 98-3:2008, is a widely adopted standard approach. However, it applies to situations where a model of the measurement process is available. A very wide range of standard test methods is, however, subjected to collaborative study in accordance with ISO 5725-2:1994. This International Standard provides an appropriate and economic methodology for estimating uncertainty associated with the results of these methods, which complies fully with the relevant principles of the GUM, whilst taking account of method performance data obtained by collaborative study.

The general approach used in this International Standard requires that

- estimates of the repeatability, reproducibility and trueness of the method in use, obtained by collaborative study as described in ISO 5725-2:1994, be available from published information about the test method in use. These provide estimates of the intra- and inter-laboratory components of variance, together with an estimate of uncertainty associated with the trueness of the method;
- the laboratory confirms that its implementation of the test method is consistent with the established performance of the test method by checking its own bias and precision. This confirms that the published data are applicable to the results obtained by the laboratory;
- any influences on the measurement results that were not adequately covered by the collaborative study be identified and the variance associated with the results that could arise from these effects be quantified.

An uncertainty estimate is made by combining the relevant variance estimates in the manner prescribed by the GUM.

The general principle of using reproducibility data in uncertainty evaluation is sometimes called a “top-down” approach.

The dispersion of results obtained in a collaborative study is often also usefully compared with measurement uncertainty estimates obtained using GUM procedures as a test of full understanding of the method. Such comparisons will be more effective given a consistent methodology for estimating the same parameter using collaborative study data.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 21748:2014](#)

<https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8f4f-1f0647b39c58/sist-iso-21748-2014>

Guidance for the use of repeatability, reproducibility and trueness estimates in measurement uncertainty estimation

1 Scope

The International Standard gives guidance for

- evaluation of measurement uncertainties using data obtained from studies conducted in accordance with ISO 5725-2:1994;
- comparison of collaborative study results with measurement uncertainty (MU) obtained using formal principles of uncertainty propagation (see Clause 13).

ISO 5725-3:1994 provides additional models for studies of intermediate precision. However, while the same general approach may be applied to the use of such extended models, uncertainty evaluation using these models is not incorporated in the present International Standard.

This International Standard is applicable in all measurement and test fields where an uncertainty associated with a result has to be determined.

This International Standard does not describe the application of repeatability data in the absence of reproducibility data.

This International Standard assumes that recognized, non-negligible systematic effects are corrected, either by applying a numerical correction as part of the method of measurement, or by investigation and removal of the cause of the effect.

The recommendations in this International Standard are primarily for guidance. It is recognized that while the recommendations presented do form a valid approach to the evaluation of uncertainty for many purposes, it is also possible to adopt other suitable approaches.

In general, references to measurement results, methods and processes in this International Standard are normally understood to apply also to testing results, methods and processes.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply. In addition, reference is made to “intermediate precision conditions”, which are discussed in detail in ISO 5725-3:1994.

2.1

bias

difference between the expectation of a test result or measurement result and a true value

NOTE 1 Bias is the total systematic error as contrasted to random error. There may be one or more systematic error components contributing to the bias. A larger systematic difference from the true value is reflected by a larger bias value.

NOTE 2 The bias of a measuring instrument is normally estimated by averaging the error of indication over an appropriate number of repeated measurements. The error of indication is the “indication of a measuring instrument minus a true value of the corresponding input quantity”.

NOTE 3 In practice, the accepted reference value is substituted for the true value.

[ISO 3534-2:2006, definition 3.3.2]

ISO 21748:2010(E)

2.2
combined standard uncertainty
 $u(y)$
standard uncertainty of the result of a measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or covariances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[ISO/IEC Guide 98-3:2008, definition 2.3.4]

2.3
coverage factor
 k
numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

NOTE A coverage factor, k , is typically in the range 2 to 3.

[ISO/IEC Guide 98-3:2008, definition 2.3.6]

2.4
expanded uncertainty
 U
quantity defining an interval about a result of a measurement expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

NOTE 1 The fraction may be regarded as the coverage probability or level of confidence of the interval.

NOTE 2 To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterized by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions can be justified.

NOTE 3 Expanded uncertainty is termed overall uncertainty in paragraph 5 of Recommendation INC-1 (1980).

[ISO/IEC Guide 98-3:2008, definition 2.3.5]

2.5
precision
closeness of agreement between independent test/measurement results obtained under stipulated conditions

NOTE 1 Precision depends only on the distribution of random errors and does not relate to the true value or the specified value.

NOTE 2 The measure of precision is usually expressed in terms of imprecision and computed as a standard deviation of the test results or measurement results. Less precision is reflected by a larger standard deviation.

NOTE 3 Quantitative measures of precision depend critically on the stipulated conditions. Repeatability conditions and reproducibility conditions are particular sets of extreme stipulated conditions.

[ISO 3534-2:2006, definition 3.3.4]

2.6
repeatability
precision under repeatability conditions

NOTE Repeatability can be expressed quantitatively in terms of the dispersion characteristics of the results.

[ISO 3534-2:2006, definition 3.3.5]

2.7**repeatability conditions**

observation conditions where independent test/measurement results are obtained with the same method on identical test/measurement items in the same test or measuring facility by the same operator using the same equipment within short intervals of time

NOTE Repeatability conditions include:

- the same measurement procedure or test procedure;
- the same operator;
- the same measuring or test equipment used under the same conditions;
- the same location;
- repetition over a short period of time.

[ISO 3534-2:2006, definition 3.3.6]

2.8**repeatability standard deviation**

standard deviation of test results or measurement results obtained under repeatability conditions

NOTE 1 It is a measure of the dispersion of the distribution of test or measurement results under repeatability conditions.

NOTE 2 Similarly, “repeatability variance” and “repeatability coefficient of variation” can be defined and used as measures of the dispersion of test or measurement results under repeatability conditions.

[ISO 3534-2:2006, definition 3.3.7] (standards.iteh.ai)

2.9**reproducibility**

precision under reproducibility conditions

SIST ISO 21748:2014

<https://standards.iteh.ai/catalog/standards/sist/6bb76a8c-0401-4901-8ff4-1f0647b39c58/sist-iso-21748-2014>

NOTE 1 Reproducibility can be expressed quantitatively in terms of the dispersion characteristics of the results.

NOTE 2 Results are usually understood to be corrected results.

[ISO 3534-2:2006, definition 3.3.10]

2.10**reproducibility conditions**

observation conditions where independent test/measurement results are obtained with the same method on identical test/measurement items in different test or measurement facilities with different operators using different equipment

[ISO 3534-2:2006, definition 3.3.11]

2.11**reproducibility standard deviation**

standard deviation of test results or measurement results obtained under reproducibility conditions

NOTE 1 It is a measure of the dispersion of the distribution of test or measurement results under reproducibility conditions.

NOTE 2 Similarly, “reproducibility variance” and “reproducibility coefficient of variation” can be defined and used as measures of the dispersion of test or measurement results under reproducibility conditions.

[ISO 3534-2:2006, definition 3.3.12]

ISO 21748:2010(E)**2.12
standard uncertainty** $u(x_i)$

uncertainty of the result of a measurement expressed as a standard deviation

[ISO/IEC Guide 98-3:2008, definition 2.3.1]

**2.13
trueness**

closeness of agreement between the expectation of a test result or a measurement result and a true value

NOTE 1 The measure of trueness is usually expressed in terms of bias.

NOTE 2 Trueness is sometimes referred to as “accuracy of the mean”. This usage is not recommended.

NOTE 3 In practice, the accepted reference value is substituted for the true value.

[ISO 3534-2:2006, definition 3.3.3]

**2.14
uncertainty**

〈measurement〉 parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand

NOTE 1 The parameter may be, for example, a standard deviation (or a given multiple of it), or the half-width of an interval having a stated level of confidence.

NOTE 2 Uncertainty of measurement comprises, in general, many components. Some of these components may be evaluated from the statistical distribution of the results of a series of measurements and can be characterized by experimental standard deviations. Other components, which also can be characterized by standard deviations, are evaluated from assumed probability distributions based on experience or other information.

NOTE 3 It is understood that the result of the measurement is the best estimate of the value of the measurand, and that all components of uncertainty, including those arising from systematic effects such as components associated with corrections and reference standards, contribute to the dispersion.

[ISO/IEC Guide 98-3:2008, definition 2.2.3]

**2.15
uncertainty budget**

list of sources of uncertainty and their associated standard uncertainties, compiled with a view to evaluating a combined standard uncertainty associated with a measurement result

NOTE The list often includes additional information such as sensitivity coefficients (change of result with change in a quantity affecting the result), degrees of freedom for each standard uncertainty, and an identification of the means of evaluating each standard uncertainty in terms of a Type A or Type B evaluation (see ISO/IEC Guide 98-3:2008).

3 Symbols a coefficient indicating an intercept in the empirical relationship $\hat{s}_R = a + bm$ B laboratory component of bias b coefficient indicating a slope in the empirical relationship $\hat{s}_R = a + bm$ c coefficient in the empirical relationship $\hat{s}_R = cm^d$ c_i sensitivity coefficient $\partial y / \partial x_i$

d	coefficient indicating an exponent in the empirical relationship $\hat{s}_R = cm^d$
e	random error under repeatability conditions
k	numerical factor used as a multiplier of the combined standard uncertainty u in order to obtain an expanded uncertainty U
l	laboratory number
m	mean value of the measurements
N	number of contributions included in combined uncertainty calculations
n'	number of contributions incorporated in combined uncertainty calculations in addition to collaborative study data
n_l	number of replicates by laboratory l in the study of a certified reference material
n_r	number of replicate measurements
p	number of laboratories
Q	number of test items from a larger batch
q	number of assigned values by consensus during a collaborative study
r_{ij}	correlation coefficient between x_i and x_j , in the interval -1 to $+1$
s_b	between-group component of variance expressed as a standard deviation
s_b^2	between-group component of variance
s_D	estimated, or experimental, standard deviation of results obtained by repeated measurement on a reference material used for checking control of bias
s_{inh}	uncertainty associated with the inhomogeneity of the sample
s_{inh}^2	component of variance associated with the inhomogeneity of the sample
s_l	estimated repeatability standard deviation with ν_l degrees of freedom for laboratory l during verification of repeatability
s_L	experimental or estimated inter-laboratory standard deviation
\hat{s}_L	adjusted estimate of standard deviation associated with B where s_L is dependent on the response
s_L^2	estimated variance of B
s_r	estimate of intra-laboratory standard deviation; the estimated standard deviation for e
s'_r	adjusted estimate of intra-laboratory standard deviation, where the contribution is dependent on the response
s_r^2	estimated variance of e
s_R	estimated reproducibility standard deviation
s'_R	estimate of the reproducibility standard deviation adjusted for laboratory estimate of repeatability standard deviation