

ETSI EN 302 195 V2.1.1 (2016-06)



**Short Range Devices (SRD);
Ultra Low Power Active Medical Implants (ULP-AMI)
and accessories (ULP-AMI-P) operating in the
frequency range 9 kHz to 315 kHz
Harmonised Standard covering the essential requirements
of article 3.2 of the Directive 2014/53/EU**

PREVIEW
[https://standards.iso.org/standards-catalogue/standards-catalogue/302-195-v2.1.1-2016-06](https://standards.iso.org/standards/catalogue/standards-catalogue/302-195-v2.1.1-2016-06)
450c-a195-1ef5508e-8e8e-421f-21cd-

Reference

REN/ERM-TG30-310

Keywords

health, inductive, magnetic, mobile, radio,
regulation, short range, SRD, testing

ETSI

650 Route des Lucioles
F-06921 Sophia Antipolis Cedex - FRANCE

Tel.: +33 4 92 94 42 00 Fax: +33 4 93 65 47 16

Siret N° 348 623 562 00017 - NAF 742 C
Association à but non lucratif enregistrée à la
Sous-Préfecture de Grasse (06) N° 7803/88

Important notice

The present document can be downloaded from:
<http://www.etsi.org/standards-search>

The present document may be made available in electronic versions and/or in print. The content of any electronic and/or print versions of the present document shall not be modified without the prior written authorization of ETSI. In case of any existing or perceived difference in contents between such versions and/or in print, the only prevailing document is the print of the Portable Document Format (PDF) version kept on a specific network drive within ETSI Secretariat.

Users of the present document should be aware that the document may be subject to revision or change of status. Information on the current status of this and other ETSI documents is available at
<https://portal.etsi.org/TB/ETSIDeliverableStatus.aspx>

If you find errors in the present document, please send your comment to one of the following services:
<https://portal.etsi.org/People/CommiteeSupportStaff.aspx>

Copyright Notification

No part may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm except as authorized by written permission of ETSI.

The content of the PDF version shall not be modified without the written authorization of ETSI.
The copyright and the foregoing restriction extend to reproduction in all media.

© European Telecommunications Standards Institute 2016.
All rights reserved.

DECT™, **PLUGTESTS™**, **UMTS™** and the ETSI logo are Trade Marks of ETSI registered for the benefit of its Members.
3GPP™ and **LTE™** are Trade Marks of ETSI registered for the benefit of its Members and of the 3GPP Organizational Partners.
GSM® and the GSM logo are Trade Marks registered and owned by the GSM Association.

Contents

Intellectual Property Rights	6
Foreword.....	6
Modal verbs terminology.....	6
Introduction	6
1 Scope	8
2 References	8
2.1 Normative references	8
2.2 Informative references.....	9
3 Definitions, symbols and abbreviations	9
3.1 Definitions	9
3.2 Symbols.....	10
3.3 Abbreviations	10
4 Technical requirements specifications	11
4.1 Environmental profile.....	11
4.2 Transmitter requirements	11
4.2.1 Radiated field strength	11
4.2.1.0 General	11
4.2.1.1 Radiated H-field	11
4.2.1.1.0 General	11
4.2.1.1.1 Definition.....	11
4.2.1.1.2 Limits	12
4.2.1.1.3 Conformance	12
4.2.2 Permitted range of modulation bandwidth	12
4.2.2.0 General	12
4.2.2.1 Definition	12
4.2.2.2 Limits	12
4.2.2.3 Conformance	13
4.2.3 Spurious emissions	13
4.2.3.0 General	13
4.2.3.1 Definition	13
4.2.3.2 Limits	13
4.2.3.3 Conformance.....	13
4.2.4 Duty cycle.....	13
4.2.4.0 General	13
4.2.4.1 Definitions.....	13
4.2.4.2 Limits	13
4.2.4.3 Conformance	13
4.3 Receiver requirements	14
4.3.1 Receiver Classification	14
4.3.2 Blocking.....	14
4.3.2.0 General	14
4.3.2.1 Definition	14
4.3.2.2 Limits	14
4.3.2.3 Conformance.....	15
4.3.3 Receiver spurious radiations	15
4.3.3.0 General	15
4.3.3.1 Definition	15
4.3.3.2 Limits	15
4.3.3.3 Conformance.....	15
5 Testing for compliance with technical requirements.....	15
5.1 Environmental conditions for testing	15
5.1.0 General remarks	15
5.1.1 Presentation of equipment for testing purposes	16

5.1.2	Choice of model for testing	16
5.1.3	Presentation of equipment that does not have an external 50 Ω RF connector (integral antenna equipment)	16
5.1.3.0	General remarks	16
5.1.3.1	Equipment with an internal permanent or temporary antenna connector	16
5.1.3.2	Equipment with a temporary antenna connector	16
5.1.4	Controls	17
5.1.5	Transmitter shut-off facility	17
5.1.6	Receiver power save capability	17
5.1.7	Equipment intended to be implanted in a human body	17
5.1.8	Declarations by the Applicant	17
5.1.9	Auxiliary test equipment	17
5.1.10	Test conditions	17
5.1.10.1	Normal and extreme test-conditions	17
5.1.10.2	Test power source	17
5.1.10.2.0	General remarks	17
5.1.10.2.1	External test power source	18
5.1.10.2.2	Internal test power source	18
5.1.10.3	Normal test Condition	18
5.1.10.3.1	Normal temperature and humidity	18
5.1.10.3.2	Normal test power source	18
5.1.10.4	Extreme test conditions	19
5.1.10.4.1	Extreme temperatures	19
5.1.10.4.2	Extreme test source voltages	20
5.1.10.5	Normal test signals and test modulation	21
5.1.10.5.0	General remarks	21
5.1.10.5.1	Normal modulation test signals for data	21
5.1.10.6	Antenna	21
5.1.10.6.0	General remarks	21
5.1.10.6.1	Artificial Antenna	21
5.1.10.7	Test fixture	22
5.1.10.7.0	General remarks	22
5.1.10.7.1	Alternate test fixture for equipment intended to be implanted within a human body	22
5.1.10.8	Test sites and general arrangements for radiated measurements	22
5.1.10.9	Modes of operation of the transmitter	22
5.1.10.10	Measuring receiver	23
5.2	Interpretation of the measurement results	23
5.3	Method of Measurements	23
5.3.1	Radiated field strength	23
5.3.1.0	General remarks	23
5.3.1.1	Radiated Field Strength (H-field)	24
5.3.2	Permitted frequency range of the modulation bandwidth	24
5.3.3	Spurious emissions	25
5.3.4	Receiver requirement	25
5.3.4.1	Blocking	25
5.3.4.2	Receiver spurious radiation	26
Annex A (normative):	Relationship between the present document and the essential requirements of Directive 2014/53/EU	27
Annex B (normative):	Radiated measurements	28
B.1	Test sites and general arrangements for measurements involving the use of radiated fields	28
B.1.1	Outdoor test site	28
B.1.1.0	General remarks	28
B.1.1.1	Standard position	28
B.1.1.2	Equipment in close proximity to the human body but external to it	29
B.1.1.3	Active medical implant equipment	29
B.1.2	Test antenna	31
B.1.2.1	Below 30 MHz	31
B.1.3	Optional additional indoor site	31
B.2	Guidance on the use of radiation test sites	31

B.2.0	General remarks	31
B.2.1	Measuring distance	32
B.2.2	Auxiliary cables	32
Annex C (normative):	H-field measurements at other distances than 10 m	33
Annex D (informative):	Bibliography	35
History		36

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/ce5eb8df-21cd-450c-a195-1ef55d9c34e7/etsi-en-302-195-v2.1.1-2016-06>

Intellectual Property Rights

IPRs essential or potentially essential to the present document may have been declared to ETSI. The information pertaining to these essential IPRs, if any, is publicly available for **ETSI members and non-members**, and can be found in ETSI SR 000 314: "*Intellectual Property Rights (IPRs); Essential, or potentially Essential, IPRs notified to ETSI in respect of ETSI standards*", which is available from the ETSI Secretariat. Latest updates are available on the ETSI Web server (<https://ipr.etsi.org/>).

Pursuant to the ETSI IPR Policy, no investigation, including IPR searches, has been carried out by ETSI. No guarantee can be given as to the existence of other IPRs not referenced in ETSI SR 000 314 (or the updates on the ETSI Web server) which are, or may be, or may become, essential to the present document.

Foreword

This Harmonised European Standard (EN) has been produced by ETSI Technical Committee Electromagnetic compatibility and Radio spectrum Matters (ERM).

The present document has been prepared under the Commission's standardisation request C(2015) 5376 final [i.6] to provide one voluntary means of conforming to the essential requirements of Directive 2014/53/EU on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC [i.2].

Once the present document is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of the present document given in table A.1 confers, within the limits of the scope of the present document, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

National transposition dates	
Date of adoption of this EN:	14 June 2016
Date of latest announcement of this EN (doa):	30 September 2016
Date of latest publication of new National Standard or endorsement of this EN (dop/e):	31 March 2017
Date of withdrawal of any conflicting National Standard (dow):	31 March 2018

Modal verbs terminology

In the present document "**shall**", "**shall not**", "**should**", "**should not**", "**may**", "**need not**", "**will**", "**will not**", "**can**" and "**cannot**" are to be interpreted as described in clause 3.2 of the [ETSI Drafting Rules](#) (Verbal forms for the expression of provisions).

"**must**" and "**must not**" are **NOT** allowed in ETSI deliverables except when used in direct citation.

Introduction

The present document covers Ultra Low Power Active Medical Implant (ULP-AMI) and any associated non-implantable Peripherals (ULP-AMI-P) equipment incorporating low frequency technology, which is designed to operate in the frequency range of 9 kHz to 315 kHz for the purpose of providing a digital communication link.

The present document includes methods of measurement for ULP-AMI and ULP-AMI-P incorporating attachable/detachable antenna connector(s) and/or integral antenna(s). Equipment designed for use with an integral antenna may use a temporary or permanent internal connector for the purpose of testing, provided the characteristics being measured are representative of the final product placed on the market.

If equipment already placed on the market is required to be inspected it should be tested in accordance with the methods of measurement specified in the present document.

Clauses 1 through 3 provide a general description on the types of equipment covered by the present document and the definitions, symbols and abbreviations used.

Clause 4 provides the technical requirements specifications, limits and conformance relative to transmitter and receiver function.

Clauses 5.1 and 5.2 specify the conditions for testing of the equipment and interpretation of the measurement results with the maximum measurement uncertainty values.

Clause 5.3 specifies the required measurement methods.

Annex A (normative) provides the relationship between the present document and the essential requirements of Directive 2014/53/EU [i.2].

Annex B (normative) provides specifications concerning radiated measurements.

Annex C (normative) provides procedures for H-field measurements at other distances than 10 m.

Annex D (informative) bibliography; provides additional information.

ITeH STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/ce5eb8df-21cd-450c-a195-1ef55d9c34e7/etsi-en-302-195-v2.1.1-2016-06>

1 Scope

The present document applies to ULP-AMI equipment operating in the frequency range from 9 kHz to 315 kHz and any associated Peripherals (ULP-AMI-P) transmitters and receivers operating in the frequency range of 9 kHz to 315 kHz including external programmers and patient related telecommunication devices using digital modulation techniques such as, but not limited to, FSK or pulse position modulation. Analogue voice modulation is not within the scope of the present document.

The present document applies to ULP-AMI/ULP/AMI-P transmitters and receivers:

- transmitters operating in range from 9 kHz to 315 kHz with power levels ranging up to 30 dBuA/m at 10m;
- receivers operating in the range from 9 kHz to 315 kHz.

The present document applies to ULP-AMI devices:

- either with a Radio Frequency (RF) output connection and dedicated antenna, or with an integral antenna;
- for telecommand, telemetry etc. applications;
- for all types of digital modulation.

The present document covers ULP-AMI-P fixed stations (physician programmer/controllers), mobile stations (patient programmers, handheld or otherwise) and portable stations (implanted devices providing medical benefit to the implanted patient).

The present document contains the technical requirements for characteristics of ULP-AMI/ULP-AMI-P radio equipment which are aligned with annex 12 Sub-band (a) of CEPT/ERC Recommendation 70-03 [i.1].

The present document contains requirements to demonstrate that Ultra Low Power Active Medical Implants (ULP-AMI) and Peripherals (ULP-AMI-P) operating in the frequency range 9 kHz to 315 kHz *"shall be so constructed that it both effectively uses and supports the efficient use of radio spectrum in order to avoid harmful interference"* (article 3.2 of the Directive 2014/53/EU) [i.2]. It does not necessarily include all the characteristics, which may be required by a user, nor does it necessarily represent the optimum performance achievable.

2 References

2.1 Normative references

References are either specific (identified by date of publication and/or edition number or version number) or non-specific. For specific references, only the cited version applies. For non-specific references, the latest version of the referenced document (including any amendments) applies.

Referenced documents which are not found to be publicly available in the expected location might be found at <http://docbox.etsi.org/Reference>.

NOTE: While any hyperlinks included in this clause were valid at the time of publication, ETSI cannot guarantee their long term validity.

The following referenced documents are necessary for the application of the present document.

Not applicable.

2.2 Informative references

References are either specific (identified by date of publication and/or edition number or version number) or non-specific. For specific references, only the cited version applies. For non-specific references, the latest version of the referenced document (including any amendments) applies.

NOTE: While any hyperlinks included in this clause were valid at the time of publication, ETSI cannot guarantee their long term validity.

The following referenced documents are not necessary for the application of the present document but they assist the user with regard to a particular subject area.

- [i.1] CEPT/ERC Recommendation 70-03: "Relating to the use of Short Range Devices (SRD)".
- [i.2] Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC.
- [i.3] ETSI TR 100 028 (all parts) (V1.4.1): "Electromagnetic compatibility and Radio spectrum Matters (ERM); Uncertainties in the measurement of mobile radio equipment characteristics".
- [i.4] Recommendation ITU-T O.153: "Basic parameters for the measurement of error performance at bit rates below the primary rate".
- [i.5] Air Force Technical Report AL/OE-TR-1996-0037: "Compilation of the Dielectric Properties of Body Tissues at RF and Microwave Frequencies", Camelia Gabriel.
- [i.6] Commission Implementing Decision C(2015) 5376 final of 4.8.2015 on a standardization request to the European Committee for Electrotechnical Standardization and to the European Telecommunications Standards Institute as regards radio equipment in support of Directive 2014/53/EU of the European Parliament and of the Council.

3 Definitions, symbols and abbreviations

3.1 Definitions

For the purposes of the present document, the following terms and definitions apply:

active medical implant: diagnostic or therapeutic device designed to be implanted in a human body containing a power source and capable of generating radio frequency energy within the 9 kHz to 315 kHz frequency band for the purpose of providing a digital communications link

artificial antenna: tuned reduced-radiating dummy load equal to the nominal impedance specified by the applicant

assigned frequency: frequency within the applicable band on which the device is authorized to operate

conducted measurements: measurements which are made using a direct connection to the equipment under test

custom antenna: antenna built according to providers antenna design rules

dedicated antenna: removable antenna supplied and tested with the radio equipment that is designed as an indispensable part of the equipment

fixed station: equipment intended for use in a fixed location

H-field test antenna: electrically screened loop or equivalent antenna, with which the magnetic component of the field can be measured

integral antenna: permanent fixed antenna, which may be built-in, that is designed as an indispensable part of the equipment

magnetic dipole moment: product of (Number of coil turns) \times (coil area) \times (coil current)

NOTE: Air coils only.

medical implant device: apparatus that includes a transmitter with an integral receiver that operates in the ULP-AMI band that is placed inside the human body for the purpose of performing diagnostic functions and/or delivery of therapeutic treatment

medical implant programmer/control transmitter: transmitter, operating outside of a human body in the ULP-AMI frequency band that transfers information to/from the implant after a communications link is initiated

mobile station: equipment external to the body, normally used by a patient, to provide telecommand or telemetry communication functions to a medical implant device placed within the body

patient activator: equipment intended to be used by a patient to communicate with an implanted device

portable station: equipment intended to be carried, attached or implanted in a human body that is operated at a separation distance less than 20 cm from or internal to a human body

programmer/controller: ULP-AMI equipment used by a physician to communicate with an implanted device

radiated measurements: measurements which involve the absolute measurement of a radiated field

telecommand: use of radio communication for the transmission of signals to initiate, modify or terminate functions of equipment at a distance

telecommunications: use of radio communications for the transmission of data between various ULP-AMI devices

telemetry: use of radio communication for indicating or recording data at a distance

Ultra Low Power Active Medical Implant (ULP-AMI): active medical implant transmitter that is designed to radiate RF energy in accordance with the provisions of annex 12, sub-band (a) and to CEPT/ERC Recommendation 70-03 [i.1]

Ultra Low Power Active Medical Implant Peripheral (ULP-AMI-P): associated medical implant programmer/control radio part located outside the human body that communicates with an ULP-AMI in accordance with the provisions of annex 12, sub-band (a) of CEPT/ERC Recommendation 70-03 [i.1]

3.2 Symbols

For the purposes of the present document, the following symbols apply:

f	frequency
H	Magnetic field strength
H ₀	Reference magnetic field strength
m	magnetic dipole moment
P	Power
R	Distance
R ₀	Reference distance
t	time

3.3 Abbreviations

For the purposes of the present document, the following abbreviations apply:

CEPT	European Conference of Postal & Telecommunications Administrations
EMC	ElectroMagnetic Compatibility
ERC	European Research Council
EUT	Equipment Under Test
FSK	Frequency Shift Keying
RF	Radio Frequency
SRD	Short Range Device
ULP-AMI	Ultra Low Power Active Medical Implant
ULP-AMI-P	Ultra Low Power Active Medical Implant Peripheral

4 Technical requirements specifications

4.1 Environmental profile

The technical requirements of the present document apply under the environmental profile for operation of the equipment, which shall be declared by the supplier. The equipment shall comply with all the technical requirements of the present document at all times when operating within the boundary limits of the declared operational environmental profile.

4.2 Transmitter requirements

4.2.1 Radiated field strength

4.2.1.0 General

Transmitters covered by the present document are considered to have internal and/or dedicated external antennas. In this case, radiated field strength measurements are required. User defined antenna systems are not permitted.

4.2.1.1 Radiated H-field

4.2.1.1.0 General

The radiated H-field, as defined in clause 4.2.1.1.1, shall not exceed the limits of clause 4.2.1.1.2.

4.2.1.1.1 Definition

4.2.1.1.1.0 General

In the case of a transmitter with an integral or dedicated antenna, the H-field is measured in the direction of maximum field strength under specified conditions of measurement.

4.2.1.1.1.1 The inductive loop coil transmitters

These transmitters are characterized by:

- a) the loop coil antenna area A shall be $< 30 \text{ m}^2$;
- b) the length of any antenna loop element shall be $< \frac{\lambda}{4}$ ($< \frac{75}{f}$, where f is in MHz) or $< 30 \text{ m}$ whichever is shorter;
- c) antenna coil may have one or multiple turns.

4.2.1.1.1.2 Antenna types

This equipment is defined according to the antenna types referenced in CEPT/ERC Recommendation 70-03 [i.1].

Inductive loop coil transmitter, tested with an antenna as either:

- an integral antenna (antenna type 1); or
- a dedicated antenna supplied with the equipment (antenna type 2).

The following restrictions apply to these antenna types:

- 9 kHz to 315 kHz frequency range;