

SLOVENSKI STANDARD SIST EN 303 348 V1.1.2:2017

01-september-2017

Sistemi z indukcijsko zanko v frekvenčnem območju od 10 Hz do 9 kHz za pomoč slušno prizadetim - Harmonizirani standard, ki zajema bistvene zahteve člena 3.2 direktive 2014/53/EU

Induction loop systems intended to assist the hearing impaired in the frequency range 10 Hz to 9 kHz - Harmonised Standard covering the essential requirements of article 3.2 of the Directive 2014/53/EU

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 303 348 V1.1.2:2017

https://standards.iteh.ai/catalog/standards/sist/d15612e9-db44-4b3a-a0ab-a7d5a2c8940f/sist-en-303-348-v1-1-2-2017

Ta slovenski standard je istoveten z: ETSI EN 303 348 V1.1.2 (2017-07)

ICS:

11.180.15 Pripomočki za gluhe osebe in Aids for deaf and hearing

osebe z okvaro sluha impaired people

33.020 Telekomunikacije na splošno Telecommunications in

general

SIST EN 303 348 V1.1.2:2017 en

SIST EN 303 348 V1.1.2:2017

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 303 348 V1.1.2:2017 https://standards.iteh.ai/catalog/standards/sist/d15612e9-db44-4b3a-a0ab-a7d5a2c8940f/sist-en-303-348-v1-1-2-2017 SIST EN 303 348 V1.1.2:2017

ETSI EN 303 348 V1.1.2 (2017-07)



Induction loop systems
intended to assist the hearing impaired
in the frequency range 10 Hz to 9 kHz;
Harmonised Standard covering the essential requirements
of article 3.2 of Directive 2014/53/EU

a7d5a2c8940f/sist-en-303-348-v1-1-2-2017

Reference

DEN/ERM-TG17WG3-16

Keywords

harmonised standard, hearing aid, inductive

ETSI

650 Route des Lucioles F-06921 Sophia Antipolis Cedex - FRANCE

Tel.: +33 4 92 94 42 00 Fax: +33 4 93 65 47 16

Siret N° 348 623 562 00017 - NAF 742 C Association à but non lucratif enregistrée à la iTeh Sous-Préfecture de Grasse (06) N° 7803/88 (standards.iteh.ai)

SIST EN 303 348 V1.1.2:2017

https://standards.iteh.ai/catalog/standards/sist/dl 5612e9-db44-4b3a-a0ab-

The present document can be downloaded from: http://www.etsi.org/standards-search

The present document may be made available in electronic versions and/or in print. The content of any electronic and/or print versions of the present document shall not be modified without the prior written authorization of ETSI. In case of any existing or perceived difference in contents between such versions and/or in print, the only prevailing document is the print of the Portable Document Format (PDF) version kept on a specific network drive within ETSI Secretariat.

Users of the present document should be aware that the document may be subject to revision or change of status. Information on the current status of this and other ETSI documents is available at https://portal.etsi.org/TB/ETSIDeliverableStatus.aspx

If you find errors in the present document, please send your comment to one of the following services: https://portal.etsi.org/People/CommiteeSupportStaff.aspx

Copyright Notification

No part may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm except as authorized by written permission of ETSI.

The content of the PDF version shall not be modified without the written authorization of ETSI. The copyright and the foregoing restriction extend to reproduction in all media.

> © ETSI 2017. All rights reserved.

DECT[™], **PLUGTESTS**[™], **UMTS**[™] and the ETSI logo are Trade Marks of ETSI registered for the benefit of its Members. 3GPP™ and LTE™ are Trade Marks of ETSI registered for the benefit of its Members and of the 3GPP Organizational Partners.

oneM2M logo is protected for the benefit of its Members GSM® and the GSM logo are Trade Marks registered and owned by the GSM Association.

Contents

Intelle	ectual Property Rights	5
Forev	word	5
Moda	al verbs terminology	5
Introd	duction	6
1	Scope	7
1	•	
2 2.1	References	
2.1 2.2	Informative references.	
3 3.1	Definitions, symbols and abbreviations	
3.1	Symbols	
3.3	Abbreviations	
4	Technical requirements specifications	
4 4.1	Environmental conditions	
4.1.1	Environmental profile	
4.2	General requirements	
4.2.1	Presentation of induction loop amplifier equipment for testing purposes	11
4.2.2	Choice of model for testing	
4.2.3	ControlsAmplifier shut-off facility Annual Processing Strain An	11
4.2.4 4.2.5	Amplifier shut-off facility	12 12
4.2.6	Declarations by the manufacturer Auxiliary test equipment Standards item ai	12
5	Testing for compliance with technical requirements vi. 1.2.2017	
5.1 5.1.1	Environmental conditions for testing catalog/standards/sist/d15612e9-db44-4b3a-a0ab-	12 12
5.1.1	General requirements	12 12
5.1.3	Test power source	
5.1.3.1	*	
5.1.3.2	±	
5.1.3.3		
5.1.3.3 5.1.3.3		
5.1.3.3 5.1.3.3	e	
5.1.3.3	· · · · · · · · · · · · · · · · · · ·	
5.2	General conditions	
5.2.1	Normal test signals	
5.2.1.1		
5.2.2 5.2.2.1	Loop connections	
5.2.2.1 5.2.2.2		
5.2.3	Measuring receiver	
5.3	Induction loop amplifier	
5.3.1	General	
5.3.2	Induction loop amplifier definitions	
5.3.2.1		
5.3.2.2 5.3.3	2 Product Classes	
5.3.3.1	· ·	
5.3.3.2		
5.3.3.3		
5.3.4	Spurious domain emission limits	
5.3.4.1		
5.3.4.2	2 Methods of measurement	1 /

ETSI EN 303 348 V1.1.2 (2017-07)

5.3.4.3	Limits		17
5.4	Induction loop recei	ver	18
5.4.1	General		18
5.4.2	Headphone outp	ut	18
5.4.3	Induction loop re	eceiver definitions	18
5.4.4	Methods of mea	surement	18
5.4.5	Limits		18
5.5	Interpretation of the	measurement results	19
Annex	A (informative):	Relationship between the present document and the essential requirements of Directive 2014/53/EU	20
Annex	B (informative):	Technical overview	21
B.1 Iı	nduction loop amplif	fiers	21
B.1.1	General		21
B.1.2	Frequency response		21
B.1.3	Field strength		21
B.1.4	General performance	e criteria	21
B.2 Induction loop receivers and test equipment			
B.2.1	General		21
B.2.2	Sensitivity		22
Annex	C (normative):	Measurement	23
C.1 S	tandard test position	to be used with internal loop equipment or equipment having a loop	
			23
		Teh STANDARD PREVIEW	
	D (normative):	The relationship between loop current and H-Field strength	24
		(standards.iten.ai)	
Annex	E (informative):	Artificial loop	26
Annex	F (normative):	Induction loop amplifier emission levels and spectrum mask ://standards.teh.avcatalog/standards/sist/d15612e9-db44-4b3a-a0ab-	
	https	://standards.iteh.a/catalog/standards/sist/d15612e9-db44-4b3a-a0ab- measurements a / d5a2c89401/sist-en-303-348-v1-1-2-2017	28
Annev	G (informative):	Determination and use of the measurement bandwidth	
	,		
Annex	H (informative):	Bibliography	30
Annex I (informative): Change history			31
History			32

Intellectual Property Rights

IPRs essential or potentially essential to the present document may have been declared to ETSI. The information pertaining to these essential IPRs, if any, is publicly available for **ETSI members and non-members**, and can be found in ETSI SR 000 314: "Intellectual Property Rights (IPRs); Essential, or potentially Essential, IPRs notified to ETSI in respect of ETSI standards", which is available from the ETSI Secretariat. Latest updates are available on the ETSI Web server (https://ipr.etsi.org/).

Pursuant to the ETSI IPR Policy, no investigation, including IPR searches, has been carried out by ETSI. No guarantee can be given as to the existence of other IPRs not referenced in ETSI SR 000 314 (or the updates on the ETSI Web server) which are, or may be, or may become, essential to the present document.

Foreword

This Harmonised European Standard (EN) has been produced by ETSI Technical Committee Electromagnetic compatibility and Radio spectrum Matters (ERM).

The present document has been prepared under the Commission's standardisation request C(2015) 5376 final [i.8] to provide one voluntary means of conforming to the essential requirements of Directive 2014/53/EU on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC [i.1].

Once the present document is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of the present document given in table A.1 confers, within the limits of the scope of the present document, a presumption of conformity with the corresponding essential requirements of that Directive, and associated EFTA regulations.

(standards.iteh.ai)

(Standards.itch.ar)		
National transposition dates		
Date of adoption of this ENes://standards.iteh.ai/catalog/standards/sist/d15612e9-db44-thjuly02017		
Date of latest announcement of this EN (doa):	31 October 2017	
Date of latest publication of new National Standard		
or endorsement of this EN (dop/e):	30 April 2018	
Date of withdrawal of any conflicting National Standard (dow):	30 April 2019	

Modal verbs terminology

In the present document "shall", "shall not", "should", "should not", "may", "need not", "will", "will not", "can" and "cannot" are to be interpreted as described in clause 3.2 of the <u>ETSI Drafting Rules</u> (Verbal forms for the expression of provisions).

"must" and "must not" are NOT allowed in ETSI deliverables except when used in direct citation.

Introduction

Audio Frequency Induction Loop Systems (AFILS) have been on the market since the middle of the twentieth century, with the first recognizable patents appearing circa 1938. AFILS are primarily used to facilitate improved communication to people with impaired hearing and are an important tool in the reduction of discrimination against disabled people.

The present document represents the performance of equipment which is currently on the market, which has not previously been subjected to compliance to a "radio" directive.

AFILS are installed in places of worship, places of entertainment, places of education, ticket booths and service counters, etc., as well as in domestic situations, providing huge benefits to users with impaired hearing.

AFILS provide an audio frequency magnetic field that couples with a receiving coil (Telecoil) fitted in hearing aids, cochlear implants, loop listeners and testing devices. This magnetic field is generated in a wire loop that is fed by an audio frequency amplifier which is capable of driving current through the "induction loop" which, in turn, is fed from external signals such as those generated by microphones, audio-visual equipment and musical instruments.

AFILS operate below 9 kHz and have a very limited range (some few metres) and there is no known evidence of interference with radio equipment.

The market for AFILS is relatively small compared with technologies such as RFID, and is physically separated from most radio systems, so the opportunity for mutual interference problems is reduced compared to other users of the spectrum in this frequency range.

The present document has been developed in response to Directive 2014/53/EU [i.1], which is the first radio standard that has been produced for AFILS equipment and has been prepared to allow the assessment of audio frequency induction loop amplifiers and receivers for compliance with Directive 2014/53/EU [i.1].

(standards.iteh.ai)

SIST EN 303 348 V1.1.2:2017 https://standards.iteh.ai/catalog/standards/sist/d15612e9-db44-4b3a-a0ab-a7d5a2c8940f/sist-en-303-348-v1-1-2-2017

1 Scope

The present document specifies technical characteristics and methods of measurements for audio frequency induction loop amplifiers and receivers operating from 10 Hz to 9 kHz used in audio frequency induction loop systems (AFILS).

NOTE: The object of an AFILS is to transmit an audio signal to people with hearing difficulties. The receiver in this case is normally a hearing aid with a built in telecoil.

These radio equipment types are capable of operating in the frequency band within the 10 Hz to 9 kHz range:

- either with an output connection(s) and dedicated loop(s) or with an internal loop(s);
- for audio frequency baseband transmission (un-modulated and without the use of a carrier).

The present document covers fixed induction loop amplifiers, mobile induction loop amplifiers and portable induction loop amplifiers.

The present document covers the essential requirements of article 3.2 of Directive 2014/53/EU [i.1] under the conditions identified in annex A.

2 References

2.1 Normative references

References are specific, identified by date of publication and/or edition number or version number. Only the cited version applies.

(standards iteh ai)

Referenced documents which are not found to be publicly available in the expected location might be found at https://docbox.etsi.org/Reference/. SIST EN 303 348 V1.1.2:2017

NOTE: While any hyperlinks included in this clause were valid at the time of publication, ETSI cannot guarantee their long term validity.

Addiagonal formula for the first statement of the first statement o

The following referenced documents are necessary for the application of the present document.

- [1] Void.
- [2] CEPT/ERC/Recommendation 74-01E (Siófok 98, Nice 99, Sesimbra 02, Hradec Kralove 05, Cardiff 11): "Unwanted Emissions in the Spurious Domain".

2.2 Informative references

References are either specific (identified by date of publication and/or edition number or version number) or non-specific. For specific references, only the cited version applies. For non-specific references, the latest version of the referenced document (including any amendments) applies.

NOTE: While any hyperlinks included in this clause were valid at the time of publication, ETSI cannot guarantee their long term validity.

The following referenced documents are not necessary for the application of the present document but they assist the user with regard to a particular subject area.

- [i.1] Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC.
- [i.2] CEPT/ERC/REC 70-03: "Relating to the use of Short Range Devices (SRD)".

[i.3]	CISPR 16-2-3: "Specification for radio disturbance and immunity measuring apparatus and methods - Part 2-3: Methods of measurement of disturbances and immunity - Radiated disturbance measurements".
[i.4]	IEC 60118-4: "Electroacoustics - Hearing aids - Part 4: Induction loop systems for hearing aid purposes - Magnetic field strength".
[i.5]	IEC 62489-1 + Amd 1: "Electroacoustics - Audio-frequency induction loop systems for assisted hearing - Part 1: Methods of measuring and specifying the performance of system components".
[i.6]	IEC 61672-1: "Electroacoustics. Sound level meters. Specifications".
[i.7]	IEC 60268-10: "Sound system equipment. Methods for specifying and measuring the characteristics of peak programme level meters".
[i.8]	Commission Implementing Decision C(2015) 5376 final of 4.8.2015 on a standardisation request to the European Committee for Electrotechnical Standardisation and to the European Telecommunications Standards Institute as regards radio equipment in support of Directive 2014/53/EU of the European Parliament and of the Council.
[i.9]	ETSI TR 100 028 (all parts) (V1.4.1) (12-2001): "Electromagnetic compatibility and Radio spectrum Matters (ERM); Uncertainties in the measurement of mobile radio equipment characteristics".
[i.10]	ETSI TR 100 028-2 (V1.4.1) (12-2001): "Electromagnetic compatibility and Radio spectrum Matters (ERM); Uncertainties in the measurement of mobile radio equipment characteristics; Part 2".

iTeh STANDARD PREVIEW

3 Definitions, symbols and abbreviations

3.1 Definitions SIST EN 303 348 V1.1.2:2017 Definitions SIST EN 303 348 V1.1.2:2017 Definitions SIST EN 303 348 V1.1.2:2017

For the purposes of the present document, the terms and definitions given in Directive 2014/53/EU [i.1] and the following apply:

artificial loop: reduced size radiating dummy load equal to the nominal impedance of the loop specified by the manufacturer

Audio Frequency Induction Loop System (AFILS): system including induction loop amplifier(s), microphones and/or other signal sources, in which a magnetic field is created by the flow of audio-frequency current in a conductor arranged in the form of a loop or coil

conducted measurements: measurements that are made using a direct connection to the equipment under test

customized loop: loop built according to manufacturers' loop design rules inside tested limits

dedicated loop: removable loop supplied and type tested with the AFILS equipment, designed as an indispensable part of the equipment

NOTE: The loop has been designed or developed for one or more specific types of equipment. It is the combination of dedicated loop and induction loop amplifier that is expected to be compliant with the regulations.

hearing aid: personal amplification system, worn entirely on the listener, which is designed to enable a person with impaired hearing to hear more easily

hearing instrument: hearing aid or cochlear implant

hearing loop system: See AFILS.

in band emissions: emissions on a frequency or frequencies which are inside the occupied bandwidth (10 Hz to 9 kHz)

ETSI EN 303 348 V1.1.2 (2017-07)

9

induction loop: current carrying loop or coil of an AFILS used to create the magnetic field

NOTE: This is equivalent to the term "antenna" used in other ETSI documents and is used in the present document as it is the term commonly understood by the AFILS industry.

induction loop amplifier: audio amplifier designed to drive an induction loop

NOTE: This is equivalent to the term "transmitter" used in other ETSI documents and is used in the present document as it is the term commonly understood by the AFILS industry.

induction loop listener: portable listening device that is designed to give an audible output in response to signals produced by an AFILS

induction loop monitor receiver: equipment designed to verify the performance of an AFILS by audio and visual means:

- a) providing visible indication that it is powered and indicating when the strength of the magnetic field produced by the loop falls within a specified range; and
- b) providing an audio-frequency output by which the sound quality of the AFILS transmissions can be assessed

induction loop system: See AFILS.

internal loop: loop designed as a fixed part of the equipment, without the use of an external connector and as such which cannot be disconnected from the equipment by the user

loop: See induction loop.

loop listener: See induction loop listener.

iTeh STANDARD PREVIEW

magnetic dipole moment: product of (Number of loop turns) × (loop area) × (loop current) (standards.iteh.ai)

NOTE: Air loops only.

magnetic field strength level meter: instrument designed to measure magnetic field strength of audio frequency magnetic fields https://standards.iteh.ai/catalog/standards/sist/d15612e9-db44-4b3a-a0ab-

a7d5a2c8940f/sist-en-303-348-v1-1-2-2017

NOTE: Two types are in common use; a peak-programme meter (PPM) type having dynamic characteristics similar to those of the Type II meter as specified in IEC 60268-10 [i.7], and a true rms meter type that incorporates a true rms rectifier, and meets the relevant requirements for a Class 2 sound level meter as specified in IEC 61672-1 [i.6]. Full functional specifications for both types of meter can be found in

IEC 60118-4 [i.4].

mobile amplifier: equipment normally installed in a vehicle

out of band emissions: emissions on a frequency or frequencies which are outside the occupied bandwidth (10 Hz to 9 kHz), but for which the levels are in transition to a frequency where the level may be reduced without affecting the corresponding transmission of information

phased loop array: system of overlapping loops in which the currents are out of phase with each other

portable amplifier: amplifier intended to be carried or attached

radiated measurements: measurements which involve the absolute measurement of a radiated field

rated load: the load, stated by the manufacturer, to which the amplifier output is connected for measurement purposes

reference magnetic field strength level: 0 dB reference for magnetic field strength levels, which is 400 mAm⁻¹

spurious emissions: emissions on a frequency or frequencies which are outside the occupied bandwidth (10 Hz to 9 kHz) and the level of which may be reduced without affecting the corresponding transmission of information

telecoil: magnetic pickup coil intended to receive signals from an audio frequency induction loop system

NOTE: A telecoil can be part of a hearing aid, cochlear implant, or of any other device for receiving signals from an audio frequency induction loop system in accordance with IEC 60118-4 [i.4].

10

type designation: manufacturers' marking of the equipment

useful magnetic field volume: volume within which the AFILS provides a hearing-aid user with a signal of acceptable quality

NOTE: See IEC 60118-4 [i.4].

3.2 Symbols

For the purposes of the present document, the following symbols apply:

Ω	ohm
A	loop area
C	correction factor
E	electrical field strength
f	frequency
f_C	centre frequency in Hz
H	magnetic field strength
Hf	H-field-strength limit
H_{C}	H-field strength at the centre of the frequency of interest
H_S	H-field-strength limit for radiated spurious emissions
$I_{\mathbb{C}}$	audio frequency baseband output current
I_S	spurious output current
λ	wavelength
m	magnetic dipole moment
N	number of turns for a loop NDARD PREVIEW
P	power
t	time (standards.iteh.ai)

3.3 Abbreviations SIST EN 303 348 V1.1.2:2017 https://standards.iteh.ai/catalog/standards/sist/d15612e9-db44-4b3a-a0ab-

For the purposes of the present document, the following abbreviations apply:

AFILS Audio Frequency Induction Loop System, also known as a Hearing Loop

AMN Artificial Mains Network
CDN Coupling/Decoupling Network

CEPT Conférence Européenne des Postes et Télécommunications CISPR Comité International Spécial des Perturbations Radioélectriques

EC European Community
EFTA European Free Trade Area
EMC ElectroMagnetic Compatibility

ERC European Radiocommunications Committee

EU European Union

ISN Impedance Stabilization Network
LISN Line Impedance Stabilization Network

OOB Out Of Band RF Radio Frequency

RFID Radio Frequency Identification Device

rms root mean square
SRD Short Range Device
THD Total Harmonic Distortion

TR Technical Report