



Designation: E1009 – 95(Reapproved 2006)

Standard Practice for Evaluating an Optical Emission Vacuum Spectrometer to Analyze Carbon and Low-Alloy Steel¹

This standard is issued under the fixed designation E1009; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers evaluation of an optical emission vacuum spectrometer to analyze carbon and low-alloy steels. It covers instruments used for the analysis of solid samples taken from molten metal for production control or from products to confirm the composition. Both pre-installation and post-installation precision and accuracy are included in the evaluation.

1.2 While Tables 1–3 are specific for plain carbon and low-alloy steel, they could be supplemented by similar tables for other materials.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[E135 Terminology Relating to Analytical Chemistry for Metals, Ores, and Related Materials](#)

[E305 Practice for Establishing and Controlling Atomic Emission Spectrochemical Analytical Curves](#)

¹ This practice is under the jurisdiction of ASTM Committee E01 on Analytical Chemistry for Metals, Ores, and Related Materials and is the direct responsibility of Subcommittee E01.01 on Iron, Steel, and Ferroalloys.

Current edition approved June 1, 2006. Published June 2006. Originally approved in 1990. Last previous edition approved in 2000 as E1009 – 95 (2000). DOI: 10.1520/E1009-95R06.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

[E406 Practice for Using Controlled Atmospheres in Spectrochemical Analysis](#)

[E528 Practice for Grounding Basic Optical Emission Spectrochemical Equipment \(Withdrawn 1998\)](#)³

[E876 Practice for Use of Statistics in the Evaluation of Spectrometric Data \(Withdrawn 2003\)](#)³

3. Terminology

3.1 *Definitions*—For definitions of terms used in this practice, refer to Terminology [E135](#) and Practice [E876](#).

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *accuracy*—the closeness of a spectrochemical determination to an accepted reference; it is affected by imprecision and bias.

3.2.2 *standard error (SE)*—although primarily a calculation that measures how well a calibration has been defined, standard error (SE) is used in this practice as an indicator of accuracy. It is CRM-dependent and instrument-operator dependent. Some expected maximum SE values are listed, but comparisons between instrument calibrations can strictly be done only when identical suites of calibrants are used.

4. Summary of Practice

4.1 After the spectrometer is calibrated, use this practice to evaluate the instrument and its calibration. Certified reference materials are run as unknowns and precision is compared to [Table 1](#). Before comparing standard errors to those in [Table 2](#), ascertain that the calibration does not include unrealistic inflections. Values equal to or less than those in [Tables 1 and 2](#) indicate that the instrument is acceptable.

³ The last approved version of this historical standard is referenced on www.astm.org.

TABLE 1 Recommended Precision Requirements For Steel Using An Optical Emission Vacuum Spectrometer^A

Element	Approximate Concentration, %	Standard Deviation
C	0.06	±0.002
	0.15	0.004
	0.50	0.010
	0.90	0.015
Mn	0.35	0.007
	0.60	0.015
	1.00	0.02
	1.50	0.03
P	0.006	0.0007
	0.04	0.0015
S	0.005	0.0015
	0.04	0.002
	0.06	0.004
Si	0.02	0.004
	0.30	0.006
	0.50	0.010
Ni	0.03	0.002
	0.10	0.003
	0.70	0.007
	1.60	0.03
Cr	0.04	0.002
	0.30	0.007
	0.80	0.015
Sn	0.003	0.0006
	0.02	0.0015
V	0.05	0.002
	0.01	0.0015
	0.03	0.002
Mo	0.25	0.007
	0.03	0.003
	0.30	0.008
Cu	0.02	0.0007
	0.15	0.003
	0.02	0.002
Ti	0.20	0.008
	0.02	0.002
	0.02	0.003
Al	0.07	0.004
	0.006	0.002
	0.02	0.003
Nb	0.07	0.003
	0.02	0.0015
	0.07	0.003
B	0.001	0.00015
	0.07	0.003
Zr	0.05	0.002
Pb	0.01	0.0015
Se	0.02	0.002

^AThese precisions were generated from actual data in one laboratory; as such, they represent what *has* been done with proven, homogeneous materials.

TABLE 1(a) Revised Data^A

Element	Approximate Concentration, %	Standard Deviation
C	0.015	0.0009
C	1.03	0.012
Mn	0.067	0.0006
Mn	2.00	0.023
P	0.0032	0.0007
S	0.0024	0.0001
Si	0.01	0.0002
Ni	0.021	0.0004
Ni	2.00	0.005
CR	1.48	0.009
Mo	0.005	0.0004
Mo	0.50	0.002
Cu	0.015	0.0003
Cu	0.30	0.0007
Ti	0.0055	0.0004
Al	0.004	0.0004
Al	0.04	0.0007
Nb	0.01	0.0005
Nb	0.1	0.002
B	0.001	0.0001
B	0.005	0.0001
Zr	0.013	0.0007
Pb	0.002	0.0004
As	0.01	0.0003
As	0.055	0.002

^AThese precisions were generated from data that were collected on newer instruments than the original data.

TABLE 2 Elements, Concentration Ranges, and Recommended Acceptable Standard Error (SE) for Steel

Element	Approximate Concentration Range, %, as Covered by the Certified Standards	Max Allowable Standard Error, % ^A
C	0.02 to 1.00	0.009
Mn	0.01 to 1.50	0.012
P	0.001 to 0.07	0.002
S	0.003 to 0.06	0.002
Si	0.005 to 1.00	0.012
Ni	0.01 to 2.00	0.014
Cr	0.01 to 1.00	0.012
Sn	0.002 to 0.10	0.002
V	0.002 to 0.50	0.005
Mo	0.003 to 0.50	0.005
Cu	0.005 to 0.50	0.007
Ti	0.003 to 0.25	0.005
Al	0.005 to 0.25	0.005
Nb	0.002 to 0.30	^B
B	0.0001 to 0.04	0.001
Zr	0.01 to 1.00	^B
Pb	0.001 to 0.03	^B
Se	0.001 to 0.05	^B

^AThese values will depend on the standards used and the distribution of their compositions throughout the range of calibration.

^BAccuracy will not be assessed for these elements due to lack of CRM's.

5. Significance and Use

5.1 Periodically throughout the useful life of an optical emission spectrometer it becomes necessary to evaluate its performance. This is especially true at manufacture and during installation. The objective at this time is to establish whether the instrument meets design specifications and performs to customer specifications. A manufacturer's objective may be to compare production line instruments. With data on many instruments, such an evaluation procedure would be a valuable contribution to the manufacturer's quality control plan.

5.2 Use of this procedure at installation can tell the manufacturer or user whether there has been a significant change in performance due to faulty shipping or handling of the instrument. At this time, the procedure could be the beginning of a