
Reference

REN/ERM-TG30-314

KeywordsHarmonised standard, health, radio, regulation,
SRD, testing**ETSI**650 Route des Lucioles
F-06921 Sophia Antipolis Cedex - FRANCE

Tel.: +33 4 92 94 42 00 Fax: +33 4 93 65 47 16

Siret N° 348 623 562 00017 - NAF 742 C
Association à but non lucratif enregistrée à la
Sous-Préfecture de Grasse (06) N° 7803/88

Important notice

The present document can be downloaded from:

<http://www.etsi.org/standards-search>

The present document may be made available in electronic versions and/or in print. The content of any electronic and/or print versions of the present document shall not be modified without the prior written authorization of ETSI. In case of any existing or perceived difference in contents between such versions and/or in print, the only prevailing document is the print of the Portable Document Format (PDF) version kept on a specific network drive within ETSI Secretariat.

Users of the present document should be aware that the document may be subject to revision or change of status. Information on the current status of this and other ETSI documents is available at

<https://portal.etsi.org/TB/ETSIDeliverableStatus.aspx>

If you find errors in the present document, please send your comment to one of the following services:

<https://portal.etsi.org/People/CommiteeSupportStaff.aspx>

Copyright Notification

No part may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm except as authorized by written permission of ETSI.

The content of the PDF version shall not be modified without the written authorization of ETSI.

The copyright and the foregoing restriction extend to reproduction in all media.

© European Telecommunications Standards Institute 2016.

All rights reserved.

DECT™, **PLUGTESTS™**, **UMTS™** and the ETSI logo are Trade Marks of ETSI registered for the benefit of its Members. **3GPP™** and **LTE™** are Trade Marks of ETSI registered for the benefit of its Members and of the 3GPP Organizational Partners.

GSM® and the GSM logo are Trade Marks registered and owned by the GSM Association.

Contents

Intellectual Property Rights	7
Foreword.....	7
Modal verbs terminology.....	7
Introduction	8
1 Scope	9
2 References	9
2.1 Normative references	9
2.2 Informative references.....	10
3 Definitions, symbols and abbreviations	11
3.1 Definitions.....	11
3.2 Symbols.....	13
3.3 Abbreviations	13
4 Technical requirements and specifications.....	14
4.1 Environmental profile.....	14
4.2 Conformance requirements	14
4.2.1 Transmitter requirements	14
4.2.1.1 Frequency Error	14
4.2.1.1.1 Definition.....	14
4.2.1.1.2 Limits	14
4.2.1.1.3 Conformance	14
4.2.1.2 Emission bandwidth.....	14
4.2.1.2.1 Definition.....	14
4.2.1.2.2 Limits	15
4.2.1.2.3 Conformance	15
4.2.1.3 Effective isotropic radiated power of the fundamental emission	15
4.2.1.3.1 Definition.....	15
4.2.1.3.2 Limits	15
4.2.1.3.3 Conformance	15
4.2.1.4 Spurious emissions.....	15
4.2.1.4.1 Definition.....	15
4.2.1.4.2 Limits	15
4.2.1.4.3 Conformance	16
4.2.1.5 Out-of-band emissions	16
4.2.1.5.1 Definition.....	16
4.2.1.5.2 Limits	16
4.2.1.5.3 Conformance	16
4.2.1.6 Frequency stability under low voltage conditions.....	16
4.2.1.6.0 Applicability	16
4.2.1.6.1 Definition.....	16
4.2.1.6.2 Limits	16
4.2.1.6.3 Conformance	16
4.2.1.7 LP-AMI-P with restricted duty cycle	17
4.2.1.7.0 General remarks.....	17
4.2.1.7.1 Definition.....	17
4.2.1.7.2 Limits	17
4.2.1.7.3 Conformance	17
4.2.2 Receiver requirements	17
4.2.2.1 Spurious radiation	17
4.2.2.1.0 General provision	17
4.2.2.1.1 Definition.....	17
4.2.2.1.2 Limits	17
4.2.2.1.3 Conformance	17
4.2.2.2 Receiver blocking	18
4.2.2.2.1 Definition.....	18

4.2.2.2.2	Limits	18
4.2.2.2.3	Conformance	18
4.2.3	Spectrum access	18
4.2.3.0	General requirements	18
4.2.3.1	LBT threshold power level.....	18
4.2.3.1.1	Definition.....	18
4.2.3.1.2	Limits	19
4.2.3.1.3	Conformance	19
4.2.3.2	Monitoring system bandwidth.....	19
4.2.3.2.0	General remarks.....	19
4.2.3.2.1	Definition.....	19
4.2.3.2.2	Limits	19
4.2.3.2.3	Conformance	19
4.2.3.3	Minimum channel monitoring period	20
4.2.3.3.0	General remarks.....	20
4.2.3.3.1	Definition.....	20
4.2.3.3.2	Limits	20
4.2.3.3.3	Conformance	20
4.2.3.4	Channel access based on ambient levels relative to the calculated access LBT threshold level, P_{Th}	20
4.2.3.4.0	General requirements.....	20
4.2.3.4.1	Conformance	20
4.3	Mechanical and electrical design.....	20
4.3.1	General.....	20
4.3.2	Controls	20
4.3.3	Transmitter shut-off facility	21
4.3.4	Marking	21
4.3.5	Equipment identification.....	21
5	Testing for compliance with technical requirements.....	21
5.1	Environmental conditions for testing	21
5.1.0	General requirements.....	21
5.1.1	Presentation of equipment for testing purposes	21
5.1.1.0	General provisions	21
5.1.1.1	Choice of model for testing.....	21
5.1.1.2	Spurious emission testing for composite equipment.....	22
5.1.1.3	Testing of equipment with alternative power levels.....	22
5.1.1.4	Presentation of equipment that does not have an external RF connector (integral antenna equipment).....	22
5.1.1.4.1	Equipment with an internal permanent or temporary antenna connector	22
5.1.1.4.2	Equipment with a temporary antenna connector	22
5.1.1.4.3	Equipment intended to be implanted in a human body.....	23
5.1.2	Declarations by the applicant.....	23
5.1.3	Auxiliary test equipment.....	23
5.1.4	Test conditions.....	23
5.1.4.1	Normal and extreme test conditions.....	23
5.1.4.2	Test power source	23
5.1.4.2.0	General requirements.....	23
5.1.4.2.1	External test power source.....	23
5.1.4.2.2	Internal test power source	23
5.1.4.3	Normal test conditions	24
5.1.4.3.1	Normal temperature and humidity.....	24
5.1.4.3.2	Normal test power source	24
5.1.4.3.2.1	Mains voltage.....	24
5.1.4.3.2.2	Other power sources	24
5.1.4.4	Extreme test conditions	24
5.1.4.4.1	Extreme temperatures	24
5.1.4.4.2	Extreme test source voltages	26
5.1.4.5	Normal test signals and test modulation.....	26
5.1.4.5.0	General requirements.....	26
5.1.4.5.1	Normal modulation test signals for data.....	26
5.1.4.6	Antennas	26
5.1.4.7	Artificial Antennas	26

5.1.4.8	Artificial antenna for transmitters with 50 Ω impedance connector	26
5.1.4.9	Test fixture for LP-AMI-P	27
5.1.4.10	Test fixture for LP-AMI	27
5.1.4.11	Test sites and general arrangements for radiated measurements	27
5.1.4.12	Modes of operation of the transmitter	27
5.1.4.13	Measuring receiver	28
5.2	Interpretation of the measurement results	28
5.3	Methods of measurement	29
5.3.1	Methods of measurement for transmitters	29
5.3.1.0	General provision	29
5.3.1.1	Frequency error	30
5.3.1.1.0	General requirements	30
5.3.1.1.1	Method of measurement for systems with an unmodulated carrier frequency operating mode	30
5.3.1.1.2	Method of measurement for systems with a modulated output frequency	30
5.3.1.2	Emission bandwidth	30
5.3.1.3	Effective isotropic radiated power of the fundamental emission	31
5.3.1.4	Transmitter spurious emissions	32
5.3.1.5	Out-of-band emissions	33
5.3.1.6	Frequency stability under low voltage conditions	34
5.3.2	Methods of measurement for receivers	34
5.3.2.0	General provisions	34
5.3.2.1	Receiver spurious emissions	34
5.3.2.1	Receiver Blocking	34
5.3.2.1.0	General remarks	34
5.3.2.1.1	Measurement method using frequency administration commands	35
5.3.2.1.2	Results based on the above test method	35
5.3.3	Methods of measurement for Monitoring Systems	35
5.3.3.0	Purpose	35
5.3.3.1	General Remarks on the Measurement Configuration	36
5.3.3.2	LBT threshold power level	36
5.3.3.2.0	General Remarks	36
5.3.3.2.1	Measurement method using out-of-operating-region disturbance	36
5.3.3.2.2	Measurement method using frequency administration commands	37
5.3.3.2.3	Measurement method for LBT operation under interference condition	37
5.3.3.2.3	Results based on above test method	37
5.3.3.3	Monitoring system bandwidth	37
5.3.3.3.0	General Remarks	37
5.3.3.3.1	Measurement method using out-of-operating-region disturbance	38
5.3.3.3.2	Measurement method using frequency administration commands	38
5.3.3.3.3	Results based on above test method	38
5.3.3.4	Monitoring system scan cycle time and minimum channel monitoring period	39
5.3.3.4.0	General Remarks	39
5.3.3.4.1	Measurement method using out-of-operating-region disturbance	39

Annex A (normative):	Relationship between the present document and the essential requirements of Directive 2014/53/EU	44
-----------------------------	---	-----------

Annex B (normative):	Radiated measurements	45
-----------------------------	------------------------------------	-----------

B.1	Test sites and general arrangements for measurements involving the use of radiated fields	45
B.1.1	Outdoor test site	45
B.1.1.0	General requirement	45
B.1.1.1	Standard position	45
B.1.1.2	Equipment in close proximity to the human body but external to it	46
B.1.1.3	Applicative simulator	46
B.1.1.3.1	General matters	46
B.1.1.3.2	Vertical Human torso simulator for LP-AMI	46
B.1.1.3.3	Horizontal Human torso simulator for LP-AMI	47
B.1.2	Test antenna	48
B.1.3	Substitution antenna	48
B.1.4	Optional additional indoor site	49

B.2	Guidance on the use of radiation test sites	50
B.2.0	General requirement	50
B.2.1	Measuring distance	50
B.2.2	Test antenna.....	50
B.2.3	Substitution antenna	50
B.2.4	Artificial antenna.....	50
B.2.5	Auxiliary cables.....	50
B.3	Further optional alternative indoor test site using an anechoic chamber	51
B.3.0	General requirements	51
B.3.1	Example of the construction of a shielded anechoic chamber	51
B.3.2	Influence of parasitic reflections in anechoic chambers	51
B.3.3	Calibration of the shielded RF anechoic chamber	52
Annex C (informative):	Bibliography.....	54
History		55

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/90443b73-91ab-4f8-b5c3-5cd828205bd7/etsi-en-301-559-v2.1.1-2016-10>

Intellectual Property Rights

IPRs essential or potentially essential to the present document may have been declared to ETSI. The information pertaining to these essential IPRs, if any, is publicly available for **ETSI members and non-members**, and can be found in ETSI SR 000 314: "*Intellectual Property Rights (IPRs); Essential, or potentially Essential, IPRs notified to ETSI in respect of ETSI standards*", which is available from the ETSI Secretariat. Latest updates are available on the ETSI Web server (<https://ipr.etsi.org>).

Pursuant to the ETSI IPR Policy, no investigation, including IPR searches, has been carried out by ETSI. No guarantee can be given as to the existence of other IPRs not referenced in ETSI SR 000 314 (or the updates on the ETSI Web server) which are, or may be, or may become, essential to the present document.

Foreword

This draft Harmonised European Standard (EN) has been produced by ETSI Technical Committee Electromagnetic compatibility and Radio spectrum Matters (ERM), and is now submitted for the combined Public Enquiry and Vote phase of the ETSI standards EN Approval Procedure.

The present document has been prepared under the Commission's standardisation request C(2015) 5376 final [i.12] to provide one voluntary means of conforming to the essential requirements of Directive 2014/53/EU on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC [i.1].

Once the present document is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of the present document given in table A.1 confers, within the limits of the scope of the present document, a presumption of conformity with the corresponding essential requirements of that Directive, and associated EFTA regulations.

The present document covers Low Power Active Medical Implants (LP-AMI) and associated Peripherals (LP-AMI-P) operating in the frequency range 2 483,5 MHz to 2 500 MHz as described in the systems reference document for the equipment, ETSI TR 102 655 [i.2].

Proposed national transposition dates	
Date of latest announcement of this EN (doa):	3 months after ETSI publication
Date of latest publication of new National Standard or endorsement of this EN (dop/e):	6 months after doa
Date of withdrawal of any conflicting National Standard (dow):	18 months after doa

Modal verbs terminology

In the present document "**shall**", "**shall not**", "**should**", "**should not**", "**may**", "**need not**", "**will**", "**will not**", "**can**" and "**cannot**" are to be interpreted as described in clause 3.2 of the [ETSI Drafting Rules](#) (Verbal forms for the expression of provisions).

"**must**" and "**must not**" are **NOT** allowed in ETSI deliverables except when used in direct citation.

Introduction

The present document is part of a set of standards developed by ETSI and is designed to fit in a modular structure to cover all radio and telecommunications terminal equipment within the scope of the Radio Equipment Directive (RE-D) [i.1]. The modular structure is shown in ETSI EG 201 399 [i.4].

LP-AMI/LP-AMI-P equipment in the AMICS is a unique technology using the frequency band 2 483,5 MHz to 2 500 MHz, that will provide for example high speed communications capability between individuals with AIMDs and medical practitioners engaged in utilizing these AIMDs for the purposes of diagnosing and delivering therapy to individuals with various illnesses. Equipment in the AMICS consists of LP-AMI and/or LP-AMI-P that provide human therapeutic and diagnostic data storage and analysis capability.

The present document includes methods of measurement for Low Power Active Medical Implants (LP-AMI), and Peripherals (LP-AMI-P), fitted with antenna connector and/or integral antenna. Equipment designed for use with an integral antenna may be supplied with a temporary or permanent internal connector for the purpose of testing, providing the characteristics being measured are not expected to be affected.

The frequency usage conditions for the bands 2 483,5 MHz to 2 500 MHz are European wide harmonised for the SRD category "active medical implant devices" according to 2013/752/EU [i.13].

If equipment, which is available on the market, is required to be checked it should be tested in accordance with the methods of measurement specified in the present document.

The present document is structured as follows:

- Clauses 1 through 3 provide a general description of the types of equipment covered by the present document and the definitions of terms and symbols and abbreviations used.
- Clause 4 specifies the requirements and limits relative to transmitter, receiver, and spectrum access. The latter are primarily designed to minimize the possibility of disturbance between LP-AMI/LP-AMI-P equipment and other users of the 2 483,5 MHz to 2 500 MHz frequency range.
- Clauses 5.1 and 5.2 specify the test and general conditions for testing of the equipment.
- Clause 5.3 specifies the methods of measurement for the parameters specified in clause 4, related to transmitter, receiver, and spectrum access.
- Annex A (normative) provides an overview of the relationship between the present document and the essential requirements of the RE-D [i.1].
- Annex B (normative) gives the specifications concerning radiated measurements.
- Annex C (informative) provides the bibliography.

1 Scope

The present document covers, for Low Power Active Medical Implants (LP-AMI) using the band bands 2 483,5 MHz to 2 500 MHz, and associated Peripherals (LP-AMI-P) used in an Active Medical Implant Communications System (AMICS), the required characteristics considered necessary to efficiently use the available spectrum and serve the interests of patients with implanted devices. The specifications contained in the present document were developed to ensure that the health and safety of the patients that are using this equipment under the direction of medical practitioners is protected. Of particular importance is the inclusion of spectrum monitoring and access requirements designed to significantly reduce any interference potential between AMICS operating in the band or between AMICS and other primary or secondary users of the band. An AIMD is regulated under the AIMD Directive 90/385/EEC [i.5] radio parts contained therein (referred to herein as LP-AMI and LP-AMI-P for associated peripheral devices) are regulated under the Directive 2014/53/EU [i.1].

The frequency usage conditions for the bands 2 483,5 MHz to 2 500 MHz are EU wide harmonised for the SRD category "active medical implant devices" according to Commission Implementing Decision 2013/752/EU [i.13] with the following usage restrictions:

- "This set of usage conditions is only available to active implantable medical devices. Peripheral master units are for indoor use only."

The present document contains the technical characteristics for LP-AMI and associated peripherals LP-AMI-P radio equipment which is also addressed by CEPT/ERC/REC 70-03 [i.3] annex 12 sub-band e) to that document. It does not necessarily include all the characteristics, which may be required by a user, nor does it necessarily represent the optimum performance achievable.

The present document applies to LP-AMI and LP-AMI_P operating in the band 2 483,5 MHz to 2 500 MHz:

- for telecommand and telemetry between LP-AMI and LP-AMI-P;
- for telecommand and telemetry between LP-AMI to another LP-AMI;
- with or without an integral antenna; and/or
- with an antenna connection provided only for the purpose of connecting a dedicated antenna.

The present document contains required characteristics considered necessary for the radio devices used in AMICS to meet in order to efficiently use the available spectrum for the purpose of transferring data that is used in diagnosing and delivering therapies to individuals with various illnesses. Of particular importance is the inclusion of spectrum monitoring and access requirements (listen before talk protocol) designed to significantly reduce any interference potential between AMICS operating in the band or between an AMICS and the primary users of the band.

The present document is a specific product standard applicable to low power transmitters that are part of a system used in the AMICS operating in spectrum within the frequency band 2 483,5 MHz to 2 500 MHz.

The present document contains requirements to demonstrate that Low Power Active Medical Implants (LP-AMI) "...shall be so constructed that it both effectively uses and supports the efficient use of radio spectrum in order to avoid harmful interference" (article 3.2 of the Directive 2014/53/EU) [i.1]. The present document does not necessarily include all the requirements which may be required by a user, nor does it necessarily represent the optimum performance achievable.

2 References

2.1 Normative references

References are either specific (identified by date of publication and/or edition number or version number) or non-specific. For specific references, only the cited version applies. For non-specific references, the latest version of the referenced document (including any amendments) applies.

Referenced documents which are not found to be publicly available in the expected location might be found at <http://docbox.etsi.org/Reference>.

NOTE: While any hyperlinks included in this clause were valid at the time of publication, ETSI cannot guarantee their long term validity.

The following referenced documents are necessary for the application of the present document.

- [1] CISPR 16-2-3 (2010) +AMD1:2010+AMD2:2014: "Specification for radio disturbance and immunity measuring apparatus and methods - Part 2-3: Methods of measurement of disturbances and immunity - Radiated disturbance measurements".

2.2 Informative references

References are either specific (identified by date of publication and/or edition number or version number) or non-specific. For specific references, only the cited version applies. For non-specific references, the latest version of the referenced document (including any amendments) applies.

NOTE: While any hyperlinks included in this clause were valid at the time of publication, ETSI cannot guarantee their long term validity.

The following referenced documents are not necessary for the application of the present document but they assist the user with regard to a particular subject area.

- [i.1] Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC.
- [i.2] ETSI TR 102 655: "Electromagnetic compatibility and Radio spectrum Matters (ERM); System reference document; Short Range Devices (SRD); Low Power Active Medical Implants (LP-AMI) operating in a 20 MHz band within 2 360 MHz to 3 400 MHz".
- [i.3] CEPT/ERC/REC 70-03: "Relating to the use of Short Range Devices (SRD)".
- [i.4] ETSI EG 201 399 (V3.1.1): "Electromagnetic compatibility and Radio spectrum Matters (ERM); A guide to the production of Harmonized Standards for application under the Radio & Telecommunication Terminal Equipment Directive 1999/5/EC (R&TTE) and a first guide on the impact of the Radio Equipment Directive 2014/53/EU (RED) on Harmonized Standards".
- [i.5] Council Directive 90/385/EEC of 20 June 1990 on the approximation of the laws of the Member States relating to active implantable medical devices (AIMD Directive).
- [i.6] CEPT/ERC/REC 74-01: "Unwanted emissions in the spurious domain".
- [i.7] Recommendation ITU-T O.153: "Basic parameters for the measurement of error performance at bit rates below the primary rate".
- [i.8] ETSI TR 100 028 (all parts) (V1.4.1) (12-2001): "Electromagnetic compatibility and Radio spectrum Matters (ERM); Uncertainties in the measurement of mobile radio equipment characteristics".
- [i.9] Simulated Biological Materials for Electromagnetic Radiation Absorption Studies" Hartsgrove G, Kraszewski A, Surowiec A 1987.
- [i.10] Hartsgrove and Kraszewski: "Composition And Electrical Properties Of A Liquid That Has The Electrical Properties Of Tissue", 1984.
- [i.11] Carl H. Durney, Ph.D., Habib Massoudi, Ph.D., Magdy F. Iskander, USAFSAM-TR-85-73, Radiofrequency Radiation Dosimetry Handbook (Fourth Edition), in line document.
- [i.12] Commission Implementing Decision C(2015) 5376 final of 4.8.2015 on a standardisation request to the European Committee for Electrotechnical Standardisation and to the European Telecommunications Standards Institute as regards radio equipment in support of Directive 2014/53/EU of the European Parliament and of the Council.

- [i.13] Commission Implementing Decision 2013/752/EU of 11 December 2013 amending Decision 2006/771/EC on harmonisation of the radio spectrum for use by short-range devices and repealing Decision 2005/928/EC.

3 Definitions, symbols and abbreviations

3.1 Definitions

For the purposes of the present document, the following terms and definitions apply:

access protocol: specification for measuring natural and man-made ambient background levels for the purpose of providing a technique for spectrum access that reduces the potential for harmful interference to/from other users of the spectrum

Active Medical Implant Communication Channel (AMICC): any continuous segment of spectrum that is equal to the emission bandwidth of the device with the largest bandwidth that is to participate in an AMICL session

NOTE: The following type of devices for Active Medical Implant Communications Systems is covered by the present document: Frequency agile devices (i.e. having implemented LBT&AFA) designed to access a minimum one of 16 channels or 8 in the case of two channels dynamically aggregated for greater instantaneous bandwidth, that are evenly distributed across the 2 483,5 MHz to 2 500 MHz band and having Duty Cycle of less than 10 % for LP-AMI-P.

Active Medical Implant Communication Link (AMICL): collection of digitally modulated transmissions that may or may not be continuous, between LP-AMIs, and LP-AMI-Ps transferring information in a communications system

Active Medical Implant Communications System (AMICS): system specifically for the purpose of providing transmission of human therapeutic digital information between one or several LP-AMI and one or several LP-AMI-P

Active Medical Implant Communications System (AMICS) session: collection of transmissions that may or may not be continuous, between a co-operating LP-AMI and LP-AMI-P

NOTE: Under normal operational circumstances the AMICS are allowed to be triggered, set-up and maintained only by an LP-AMI-P acting as a master device. LP-AMI may attempt initiating the link only in cases of emergencies, described as "medical implant event".

Active Implantable Medical Device (AIMD): any active medical device (AMD) which is intended to be totally or partially introduced, surgically or medically, into the human body or by medical intervention into a natural orifice, and which is intended to remain there for a long period after the procedure

Adaptive Frequency Agility (AFA): ability to determine and change to an unoccupied or least interfered sub-band or channel of operation in order to maximize spectrum utilization

artificial antenna: reduced-radiation dummy load equal to the nominal impedance specified by the applicant

channel aggregation: combining two or more adjacent channels for greater bandwidth up to 2 MHz

composite equipment: any combined equipment made of two or more individual products or functions

NOTE: The individual products or functions in composite equipment might be subject to different technical standards.

conducted measurements: measurements made using a direct connection to the equipment under test

dedicated antenna: removable antenna supplied and tested with the radio equipment, designed as an indispensable part of the equipment

duty cycle: ratio, expressed as a percentage, of the maximum transmitter "on" time monitored over one hour per AMICS session, relative to a one hour period

NOTE: See clause 4.2.1.7.1.

effective radiated power: power radiated within the emission bandwidth of the EUT in the direction of the maximum level under specified conditions of measurements in the presence of modulation or without modulation as appropriate

effective isotropically radiated power (e.i.r.p.): product of the power supplied to the antenna and the antenna gain in a given direction relative to an isotropic antenna

NOTE: See clause 4.2.1.3.1.

emission bandwidth: measured as the width of the signal between the points on either side of centre frequency that are 20 dB down relative to the maximum level of the modulated signal

frequency error: difference between the nominal frequency as measured on the devices under test and under normal test conditions and the frequency under extreme conditions

NOTE: See also clause 4.2.1.1.1.

frequency range: range of operating frequencies over which the equipment can be adjusted

frequency stability under low voltage condition: ability of the equipment to remain on the nominal operating frequency when the battery voltage falls below the lower extreme voltage level

NOTE: See also clause 4.2.1.6.1.

integral antenna: permanent fixed antenna, which may be built-in, designed as an indispensable part of the equipment

LBT threshold power level: ambient signal power level above which the monitoring system selects spectrum for use in an AMICS communication session according to the next available channel

Least Interfered Channel (LIC): channel, among the available channels that has the lowest potential for causing disturbance to or receiving disturbance from other users of the band

Listen Before Talk (LBT): combination of the listen mode followed by the talk mode

listen mode: action taken by an interrogator to detect an unoccupied sub-band or channel

Low Power Active Medical Implant Peripheral (LP-AMI-P) device: radio transmitting/receiving part of an equipment that communicates indoor with one or more LP-AMI to establish an AMICL

NOTE: LP-AMI-P transmissions are allowed without limitation in cases of emergencies, described as "medical implant event".

Low Power Active Medical Implant (LP-AMI): radio transmitting/receiving part of an AIMD inside the human body

NOTE: LP-AMI transmissions are allowed without limitation in cases of emergencies, described as "medical implant event".

Medical Device (MD): any instrument, apparatus, appliance, material or other article, whether used alone or in combination, together with any accessories or software for its proper functioning, intended by the manufacturer to be used for human beings in the:

- diagnosis, prevention, monitoring, treatment or alleviation of disease or injury;
- investigation, replacement or modification of the anatomy or of a physiological process;
- control of conception.

medical implant event: occurrence recognized by a medical implant system device that requires the immediate transmission of data from a medical implant transmitter in order to protect the safety of the person in whom the medical implant system transmitter has been placed

NOTE: An example of medical implant event is when the LP-AMI identifies the imminent critical health condition of monitored patient. Cannot be used for scheduled data transmissions.

monitoring system: circuitry in an LP-AMI and/or LP-AMI-P that assures conformity with the spectrum access protocol requirements based on Listen before Talk, Adaptive Frequency Agility and selection of the least interfered channel for operation (LIC) or the unoccupied sub-band or channel

out-of-band emissions: emissions on a frequency or frequencies immediately outside the necessary **emission bandwidth**, which result from the modulation process, but excluding **emissions in the spurious domain**

spurious domain emissions: emissions at frequencies separated by more than 250 % of the occupied bandwidth from the centre of the AMICC

spurious radiations from the receiver: components at any frequency, generated and radiated by active receiver circuitry and the antenna

NOTE: See clause 4.2.2.1.1.

talk mode: transmission of intentional radiation by a transmitter

telecommand: use of radio communication for the transmission of signals to initiate, modify or terminate functions of equipment at a distance

telemetry: use of radio communication for indicating or recording data at a distance

time-critical data: data which if not transferred immediately may result in compromising the health and/or safety of the patient

transient power: power falling into adjacent spectrum due to switching the transmitter on and off during normal operation (e.g. cyclic keying during data transmission)

unwanted emissions: emissions in the spurious domain and out of band emissions

3.2 Symbols

For the purposes of the present document, the following symbols apply:

B	bandwidth
dB	decibel
dBm	absolute power level referred to one milliwatt
E	electrical field strength
e.i.r.p.	effective isotropically radiated power
f	frequency
f _c	channel centre frequency
f _e	frequency under extreme conditions
G	Antenna Gain
NaCl	sodium chloride
P	power
ppm	parts per million
R	distance
P _{Th}	maximum threshold power level (see clause 4.2.3.1)
T	temperature
t	time
λ	wavelength

3.3 Abbreviations

For the purposes of the present document, the following abbreviations apply:

AC	Alternating Current
AFA	Adaptive Frequency Agility
AIMD	Active Implantable Medical Device
AMICC	Active Medical Implant Communication Channel
AMICL	Active Medical Implant Communication Link
AMICS	Active Medical Implant Communications System
CW	Continuous Wave
DC	Direct Current
DUT	Device Under Test
EUT	Equipment Under Test