



# SLOVENSKI STANDARD SIST ISO 15819:2009

01-marec-2009

?cna Yh\_U!`5 bU]nbY`a YrcXY!`B]hcnUa ]b].`XYHY\_WY`U]b`Xc`c Yj Ub`Y`B!  
b]hfcncX]YhUbc`Ua ]bUfB89 @Łj`cna Yh] b]`]nXY`]`g`hY`c ]bg`c`\_fca Urc[ fU]t`  
j ]gc`Y`c` `]j cgh]`n`a Ugbc`gdY`\_fca Yff]`g`c`XYHY\_WY`c`fk D@`!A G!A GŁ

Cosmetics - Analytical methods - Nitrosamines: Detection and determination of N-nitrosodiethanolamine (NDELA) in cosmetics by HPLC-MS-MS

## iTeh STANDARD PREVIEW (standards.iteh.ai)

Cosmétiques - Méthodes analytiques - Nitrosamines: Recherche et dosage des N-nitrosodiéthanolamines (NDELA) dans les produits cosmétiques par CLHP-SM-SM

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>

Ta slovenski standard je istoveten z: **ISO 15819:2008**

### ICS:

71.100.70 S[ : { ^ã aãV[ a^çã Cosmetics. Toiletries  
]iã[[ [ \ã

SIST ISO 15819:2009

en,fr

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 15819:2009

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>

---

---

**Cosmetics — Analytical methods —  
Nitrosamines: Detection and  
determination of N-nitrosodiethanolamine  
(NDELA) in cosmetics by HPLC-MS-MS**

*Cosmétiques — Méthodes analytiques — Nitrosamines: Recherche et dosage des N-nitrosodiéthanolamines (NDELA) dans les produits cosmétiques par CLHP-SM-SM*

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST ISO 15819:2009](https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009)

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>



**ISO 15819:2008(E)****PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15819:2009](https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009)

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15819 was prepared by Technical Committee ISO/TC 217, *Cosmetics*.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15819:2009](https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009)

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>

## Introduction

Human exposure to N-nitrosamines can occur through diverse sources such as environment, food or personal care products. As a result of their perceived carcinogenic potential on several animal species, minimization of exposure to N-nitrosamines is recognised as important to the preservation of human health. Among N-nitrosamines, N-nitrosodiethanolamine (NDELA) has been recognised as a potential contaminant of cosmetics.

In this context, several analytical methods have been developed to detect and determine its presence in cosmetics – such as gas chromatography/thermal energy analysis, high performance liquid chromatography (HPLC) coupled either with photolysis and colorimetric quantification or with mass spectrometry (MS) determination. This latter method uses advanced technology to ensure the maximum specificity towards NDELA, to minimize the risk of artifactual formation of the analyte of interest and to allow precise quantification.

This analytical method uses high performance liquid chromatography coupled with mass spectrometry to separate and detect trace levels of NDELA from a cosmetic ingredient or product matrix with maximum specificity for NDELA.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15819:2009](https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009)

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>

# Cosmetics — Analytical methods — Nitrosamines: Detection and determination of N-nitrosodiethanolamine (NDELA) in cosmetics by HPLC-MS-MS

## 1 Scope

This International Standard describes a method for the detection and quantification of NDELA in cosmetics and raw materials used in cosmetics.

This method is not applicable to the detection and/or quantification of nitrosamines other than NDELA nor to the detection and/or quantification of NDELA in products other than cosmetics or raw materials used in cosmetics.

If a product has a possibility of either NDELA contamination from ingredients or NDELA formation by the composition of ingredients, the method shall be applied for quantitative determination of NDELA. Accordingly the method would not be applied in routine testing of cosmetic products. Because of the large variety of cosmetic products within this field of application, this method might need to be adapted for certain matrices.

Therefore, International Standards dedicated to alternative methods for testing nitrosamines in cosmetic products are being developed separately. Other methods can be employed provided that they are verified as to their detection of NDELA and validated in terms of recovery and quantification of the analyte.

<https://standards.iteh.ai/catalog/standards/sist/b67baf80-f89d-4cce-84d6-3add2002ec98/sist-iso-15819-2009>

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

## 3 Principle

Extraction of the nitrosamine NDELA in cosmetic samples is carried out with water in the presence of deuterated d8-NDELA used as internal standard. Clean-up is performed either using solid phase extraction (SPE clean-up, see 6.3.1) with a C18 cartridge or liquid-liquid extraction using dichloromethane (DCM clean-up, see 6.3.2) when the samples are not dispersible in water. The extracts are analysed by HPLC-MS-MS (high performance liquid chromatography coupled with tandem mass spectrometric detection).

NDELA quantification is done by comparing the ratio of the major fragment ions of NDELA and d8-NDELA with the calibration curve.

Confirmation of the presence of NDELA is carried out by using the molecular ion and two diagnostic ions.

## ISO 15819:2008(E)

## 4 Reagents

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and only distilled water or water of grade 1 in accordance with ISO 3696:1987. Solvent shall be of quality for HPLC analysis.

- 4.1 **Methanol** (MeOH), HPLC grade.
- 4.2 **Ethanol** (EtOH), HPLC grade.
- 4.3 **Dichloromethane**, HPLC grade.
- 4.4 **N-nitrosodiethanolamine**, with known purity greater than 95 %.
- 4.5 **d8-N-nitrosodiethanolamine**, with known purity greater than 95 %.
- 4.6 **Ammonium acetate** (NH<sub>4</sub>Ac), analytical grade.
- 4.7 **1 mol/l ammonium acetate solution**, formed by dissolving 77,08 g of NH<sub>4</sub>Ac in 1,0 l water.
- 4.8 **Eluent A: 2 mmol NH<sub>4</sub>Ac in water**, formed by taking 2 ml of 1 mol/l NH<sub>4</sub>Ac (4.7) and making up to 1 l with water.
- 4.9 **Eluent B: 2 mmol NH<sub>4</sub>Ac in 90 % MeOH/water**, formed by taking 2 ml of 1 mol/l NH<sub>4</sub>Ac (4.7) and adding 900 ml MeOH and 98 ml water.

iTech STANDARD PREVIEW  
(standards.iteh.ai)

## 5 Apparatus

Use standard laboratory glassware and equipment, with the addition of:

- 5.1 **Vortex mixer**.
- 5.2 **Sample processing station**, in SPE application (such as Vacmaster<sup>®1</sup>) sample processing station, IST).
- 5.3 **Centrifuge**, capable of reaching not less than 20 000 G.
- 5.4 **Solid phase extraction columns**, e.g. Bakerbond<sup>®1</sup> C18 – 6 ml, 500 mg reversed phase octadecylsilane bonded to silicagel, 40 APD, 60 Å.
- 5.5 **HPLC-MS-MS equipment** consisting of:
  - 5.5.1 **High performance liquid chromatography apparatus**, consisting of an eluent reservoir, a pump, an injection system, a data processor, e.g. an integrator with plotter, coupled with tandem mass spectrometry using electrospray ionization.

---

1) Vacmaster<sup>®</sup>, Bakerbond<sup>®</sup> and Spherisorb<sup>®</sup> are examples of suitable products available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.



**5.5.2 Analytical reversed phase HPLC separating column**, C18, e.g. Spherisorb® ODS II<sup>1)</sup> protected with a guard column, the dimensions of which are:

separating column

- length: 150 mm
- internal diameter: 4,6 mm
- size of spherical particles: 5 µm

guard column

- length: 10 mm
- internal diameter: 4,6 mm
- size of spherical particles: 5 µm

## 6 Sample preparation and conservation

### 6.1 General

**WARNING** — Most N-nitrosamines are potent carcinogens and every possible precaution shall be taken to avoid human exposure.

All operations involving handling of N-nitrosamines or their solutions should take place in an adequately ventilated fume hood or glove box.

**NOTE** Rubber surgical gloves, which are frequently employed, do not provide complete protection. They should be removed and disposed of immediately after use and not worn for long periods.

Thought should be given to safe disposal of any solution of material containing N-nitrosamines.

N-Nitrosodiethanolamine shall be stored in the absence of light between 2 °C and 8 °C.

UV degrades N-nitrosamines, so all solutions (standards/extracts) shall be stored in such a way that deterioration and change in composition are prevented.

Injection for analysis shall be made within 30 min of preparation of the extract sample.

### 6.2 Standards preparation

**6.2.1** Prepare stock solutions (A) of NDELA containing 1,016 mg/ml in ethanol and store them in the absence of light at –20 °C.

**6.2.2** Prepare stock solutions (d8A) of d8-NDELA containing 1,016 mg/ml in ethanol and store them in the absence of light at –20 °C.

**6.2.3** Prepare working solutions (B, C, D, E and F) by sequential 10-fold dilutions of the 1,016 mg/ml solution. All solutions shall be stored in the absence of light between 2 °C and 8 °C.