



DRAFT AMENDMENT ISO 7218:2007/DAM 1

ISO/TC 34/SC 9

Secretariat: **AFNOR**

Voting begins on
2011-10-13

Voting terminates on
2012-03-13

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Microbiology of food and animal feeding stuffs — General requirements and guidance for microbiological examinations

AMENDMENT 1

Microbiologie des aliments — Exigences générales et recommandations

AMENDEMENT 1

ICS 07.100.30

ISO/CEN PARALLEL PROCESSING

This draft has been developed within the International Organization for Standardization (ISO), and processed under the **ISO-lead** mode of collaboration as defined in the Vienna Agreement.

This draft is hereby submitted to the ISO member bodies and to the CEN member bodies for a parallel five-month enquiry.

Should this draft be accepted, a final draft, established on the basis of comments received, will be submitted to a parallel two-month approval vote in ISO and formal vote in CEN.

In accordance with the provisions of Council Resolution 15/1993 this document is circulated in the English language only.

Conformément aux dispositions de la Résolution du Conseil 15/1993, ce document est distribué en version anglaise seulement.

To expedite distribution, this document is circulated as received from the committee secretariat. ISO Central Secretariat work of editing and text composition will be undertaken at publication stage.

Pour accélérer la distribution, le présent document est distribué tel qu'il est parvenu du secrétariat du comité. Le travail de rédaction et de composition de texte sera effectué au Secrétariat central de l'ISO au stade de publication.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/69cd73fc-7220-4631-80c4-2f4c8e512f9a/iso-7218-2007-amd-1-2013>

Copyright notice

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

Contents

Page

Foreword	vii
5 Apparatus and equipment	1
5.1 General	1
5.2 Protective cabinets.....	1
5.2.1 Description	1
5.2.2 Use	2
5.2.3 Cleaning and disinfection.....	2
5.2.4 Maintenance and inspection	3
5.3 Balances and gravimetric diluters.....	3
5.3.1 Use and measurement uncertainty.....	3
5.3.2 Cleaning and disinfection.....	3
5.3.3 Performance verification and calibration.....	3
5.4 Homogenizers, blenders and mixers.....	4
5.4.1 Description	4
5.4.2 Use	4
5.4.3 Cleaning and disinfection.....	4
5.4.4 Maintenance	4
5.5 pH meter	5
5.5.1 Description	5
5.5.2 Use	5
5.5.3 Verification	6
5.5.4 Maintenance	6
5.6 Autoclave	6
5.6.1 Description	6
5.6.2 Use	7
5.6.3 Maintenance	7
5.6.4 Verification	7
5.7 Media preparator	7
5.7.1 Description	7
5.7.2 Use	7
5.7.3 Maintenance	7
5.7.4 Verification	8
5.8 Incubator	8
5.8.1 Description	8
5.8.2 Use	8
5.8.3 Cleaning and sanitization	8
5.8.4 Verification	8
5.9 Refrigerator, cold-storage room	9
5.9.1 Description	9
5.9.2 Use	9
5.9.3 Verification	9
5.9.4 Maintenance and cleaning.....	9
5.10 Freezer and deep freezer	9
5.10.1 Description	9
5.10.2 Use	10
5.10.3 Verification	10
5.10.4 Maintenance	10
5.11 Thermostatically controlled bath.....	10
5.11.1 Description	10
5.11.2 Use	11
5.11.3 Verification	11
5.11.4 Maintenance.....	11

5.12	Steamers, including boiling-water baths	12
5.12.1	Description	12
5.12.2	Use	12
5.12.3	Maintenance	12
5.13	Sterilizing oven	12
5.13.1	Description	12
5.13.2	Use	12
5.13.3	Verification	13
5.13.4	Maintenance	13
5.14	Microwave oven	13
5.14.1	Description	13
5.14.2	Use	13
5.14.3	Verification	13
5.14.4	Maintenance	13
5.15	Glass washer	14
5.15.1	Description	14
5.15.2	Use	14
5.15.3	Verification	14
5.15.4	Maintenance	14
5.16	Optical microscope.....	14
5.16.1	Description	14
5.16.2	Use	14
5.16.3	Maintenance	14
5.17	Gas burner or wire incinerator	15
5.17.1	Description	15
5.17.2	Use	15
5.17.3	Maintenance	15
5.18	Dispenser for culture media and reagents.....	15
5.18.1	Description	15
5.18.2	Use	15
5.18.3	Verification	15
5.18.4	Cleaning and maintenance	16
5.19	Vortex mixer	16
5.19.1	Description	16
5.19.2	Use	16
5.19.3	Verification	16
5.19.4	Maintenance	16
5.20	Colony-counting device	16
5.20.1	Description	16
5.20.2	Use	16
5.20.3	Verification	17
5.20.4	Maintenance	17
5.21	Equipment for culture in a modified atmosphere.....	17
5.21.1	Description	17
5.21.2	Use	17
5.21.3	Verification	17
5.21.4	Maintenance	17
5.22	Centrifuge	17
5.22.1	Description	17
5.22.2	Use	18
5.22.3	Verification	18
5.22.4	Maintenance	18
5.23	Hotplate and heating mantle.....	18
5.23.1	Description	18
5.23.2	Use	18
5.23.3	Maintenance	18
5.24	Spiral plater	18
5.24.1	Description	18
5.24.2	Use	18
5.24.3	Verification	19

5.24.4	Maintenance	19
5.25	Still, deionizers and reverse-osmosis units	19
5.25.1	Description	19
5.25.2	Use	19
5.25.3	Verification	19
5.25.4	Maintenance	20
5.26	Timers and timing devices	20
5.26.1	Description	20
5.26.2	Use	20
5.26.3	Verification	20
5.26.4	Maintenance	20
5.27	Pipettes and pipettors	20
5.27.1	Description	20
5.27.2	Use	20
5.27.3	Verification	21
5.27.4	Maintenance	21
5.28	Thermometers and temperature-monitoring devices, including automatic recorders	21
5.28.1	Description	21
5.28.2	Use	21
5.28.3	Verification	22
5.28.4	Maintenance	22
5.29	Immunomagnetic separator	22
5.29.1	Description	22
5.29.2	Use and verification	23
5.29.3	Maintenance	23
5.30	Filtration system	23
5.31	Other equipment and software	23
6	Preparation of glassware and other laboratory materials	23
6.1	Preparation	23
6.2	Sterilization/decontamination	23
6.2.1	General	23
6.2.2	Sterilization by dry heat	23
6.2.3	Sterilization by moist heat (steam)	24
6.2.4	Decontamination with chemical compounds	24
6.3	Disposable equipment and materials	24
6.4	Storage of clean glassware and materials	24
6.5	Management of sterile glassware and materials	24
6.6	Use of decontamination and disinfection	24
6.6.1	Decontamination of disposable equipment	24
6.6.2	Decontamination of glassware and materials prior to use	24
6.6.3	Decontamination of glassware and materials after use	24
6.7	Waste management	25
6.8	Washing	25
10	Enumeration	26
10.1	General	26
10.2	Enumeration using a solid medium	27
10.2.1	General	27
10.2.2	Number of Petri dishes per dilution	27
10.2.3	Pour plate techniques	27
10.2.4	Surface inoculation	28
10.2.5	Incubation	29
10.3	Calculation and expression of results obtained with solid media	30
10.3.1	Counting of colonies	30
10.3.2	Expression of results	30
10.4	Enumeration of yeasts and moulds	38
10.4.1	General	38
10.4.2	Counting of colonies for yeasts and moulds	38
10.5	Enumeration using a liquid medium	38
10.5.1	Principle	38

10.5.2	Inoculation	39
10.5.3	Choice of inoculation system	39
10.5.4	Incubation	40
10.5.5	Interpretation of results	40
10.5.6	Determination of MPN values	41
Annex A (informative) Properties of some disinfectants		46
Annex B (informative) Confidence intervals for colony count technique and choice of the method of calculation in special cases		47
B.1	Confidence intervals for colony count technique	47
B.2	Choice of the method of calculation in special cases, in the system "One plate per dilution"	49
B.2.1	Special cases not only with low numbers	49
B.2.2	Special cases with low numbers	50
Annex C (normative) Determination of most probable number (MPN)		51
Annex D (normative) Counting of colonies with two dishes per dilution		62
D.1	Counting of colonies	62
D.2	Method of calculation: General case (counting of total colonies or typical colonies)	63
D.3	Method of calculation: Case after identification or confirmation	64
D.4	Method of calculation: Estimated counts	65
D.4.1	Case of two dishes (test sample or initial suspension or first dilution) containing less than 10 colonies	65
D.4.2	Case of two dishes (test sample or initial suspension or first dilution) containing no colonies	65
D.5	Special cases (counting of typical or presumed suspect colonies)	66
D.6	Method of calculation: Special cases	67

iTeh STANDARDS PREVIEW
 (standards.iteh.ai)
 Full standard:
<https://standards.iteh.ai/catalog/standards/sist/7220-4631-80c4-2f4c8e512f92/iso-7218-2007>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 7218:2007 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 9, and by Technical Committee CEN/TC 275, *Food analysis - Horizontal methods* in collaboration.

ITeH STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sis/99cd74c7-7220-4631-80c4-2f4c8e512f9a/iso-7218-2007-amd-1-2011>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/69cd73fc-7220-4631-80c4-2f4c8e512f9a/iso-7218-2007-amd-1-2013>

Microbiology of food and animal feeding stuffs — General requirements and guidance for microbiological examinations

AMENDMENT 1

(excerpt from ISO 7218:2007)

page 6, clause 5

page 28, clause 6

page 34, clause 10

Delete the existing clauses and insert the following clauses.

5 Apparatus and equipment

5.1 General

In accordance with good laboratory practice, all apparatus and equipment should be kept clean and in good working condition. Before use, equipment should be verified as fit for the intended purpose and its performance monitored during use, where appropriate.

Where necessary, equipment and monitoring devices should be calibrated to traceable national standards, and recalibration and any necessary intermediate checks performed, and procedures and results documented.

Equipment should be regularly checked and maintained to ensure safety and fitness for use. Equipment should be monitored according to the working conditions and the accuracy demanded for the results.

The frequency of calibration and verification checks of each item of equipment is, in most cases, not specified in this International Standard, since it shall be determined by each laboratory, depending on the type of equipment and on the laboratory's level of activity, and in accordance with the manufacturer's instructions. In a limited number of cases, a frequency has been specified since it was considered to be essential.

Apparatus and equipment shall be constructed and installed to facilitate operation and to allow for ease of maintenance, cleaning, decontamination and calibration.

Any measurement uncertainties given in this clause relate to the apparatus and equipment concerned and not to the whole method of analysis.

Throughout this clause, requirements for accuracy of measuring equipment are given. These are based on the practical tolerance required to demonstrate suitable control of equipment in routine use. The accuracy stated is related to the metrological uncertainty of the device (see ISO Guide 99).

For temperature control equipment, check the stability and homogeneity of the temperature before initial use and after any repair or modification which might have an effect on the temperature control.

5.2 Protective cabinets

5.2.1 Description

A protective cabinet is a work station with horizontal or vertical laminar airflow to remove dust and other particles, such as microbes, from the air.

The maximum tolerable number of particles per cubic metre with a size greater than or equal to 0,5 µm represents the dust-spreading class of a safety cabinet. For cabinets used in food microbiology, the number of particles shall not exceed 4 000 per cubic metre.

Cabinets for use in food microbiology laboratories are of four types.

- a) Class I safety cabinets are open-fronted exhaust-protective cabinets that are intended to protect the operator and the environment but will not protect the product from extraneous contamination. Potentially infected aerosols will be contained within the cabinet and trapped by impaction on the filter. The filtered air is normally discharged to the atmosphere; if this is not done, the air shall pass through two HEPA filters mounted in series. They are not recommended for work with risk category 3 pathogens because of the difficulties in maintaining and ensuring appropriate operator protection.
- a) Class II safety cabinets protect the product, the operator and the environment. They recirculate some filtered air, exhaust some to the atmosphere and take in replacement air through the working aperture, thereby providing operator protection. They are suitable for work with risk category 3 pathogens.
- b) Horizontal laminar outflow cabinets protect the work from contamination, but blow any aerosols generated into the operator's face. Therefore they are not suitable for handling inoculated cultures or preparation of tissue culture.
- c) Vertical laminar airflow cabinets protect the product by the use of vertical laminar flow of HEPA-filtered air. They also protect the operator by the use of internally recirculated air. They are particularly suitable for providing an aseptic environment for handling sterile products and for protecting the operator when handling powders.

Use protective cabinets for all work involving the handling of pathogens and contaminated powders, if required by national regulations.

The use of a gas burner or wire incinerator is not recommended in protective cabinets. If it is necessary, the gas burner should have a small flame so that the airflow is not disturbed. The use of disposable equipment (loops, pipettes, etc.) is a suitable alternative.

5.2.2 Use

Cabinets should be kept as free of equipment as possible.

Where practicable, place everything needed inside the cabinet before starting work to minimize the number of arm movements into and out of the working aperture. Position equipment and materials so as to minimize disturbance to the airflow at the working aperture.

Operators should be adequately trained in the correct use of cabinets to ensure their safety and the integrity of the product or culture.

5.2.3 Cleaning and disinfection

Clean and disinfect the working area after use with appropriate and non-corrosive disinfectant in accordance with the manufacturer's instructions. Regularly examine wire grids protecting prefilters and wipe clean with a disinfectant-soaked cloth.

For laminar flow cabinets, the filter face should be vacuum cleaned regularly, taking care not to damage the filter medium.

Safety cabinets should be fumigated before filter changing or servicing.

After cleaning of the cabinets, UV lamps may be used for disinfection. UV lamps should be regularly cleaned and replaced in accordance with the manufacturer's instructions.

5.2.4 Maintenance and inspection

Use protective cabinets that are appropriate for the intended application and environmental conditions in the laboratory.

The efficiency of a protective cabinet shall be checked by a qualified person on receipt and thereafter at regular intervals as recommended by the manufacturer, as well as after any repair or modification.

Periodic verification of freedom from any microbial contamination should be carried out by a check of the working surface and walls of the cabinet.

A periodic verification of the number of airborne microorganisms present should be carried out during operation of the filters using the usual equipment. For example, expose several open Petri dishes containing a non-selective agar culture medium (e.g. PCA) in each cabinet for 30 min. Other methods may be used.

5.3 Balances and gravimetric diluters

5.3.1 Use and measurement uncertainty

Balances are mainly used for weighing the test portion of the sample to be examined and the components of the culture media and reagents. In addition, they may be used for carrying out measurements of dilution fluid volumes by mass.

Gravimetric diluters are electronic instruments consisting of a balance and programmable liquid dispenser and are used during the preparation of initial sample suspensions; they function by adding diluent to a subsample at a set ratio. The subsample is then weighed to the tolerance specified in the application, and the diluter set to dispense sufficient diluent for the ratio required (e.g. 9 to 1 for decimal dilutions).

A food microbiology laboratory shall be equipped with balances of the required range and measurement uncertainty for the different products to be weighed.

Unless otherwise stated, the resolution of the balance should achieve a tolerance of 1 % but shall be sufficient to achieve a maximum tolerance of 5 % of the weight.

EXAMPLE To weigh 10 g, the balance should be capable of being read to 0,1 g..

To weigh 1 g, the balance should be capable of being read to 0,01g.

Place the equipment on a stable horizontal surface, adjusted as necessary to ensure that it is level and protected from vibration and draughts.

5.3.2 Cleaning and disinfection

Equipment should be cleaned and disinfected after use or following spillage during weighing with an appropriate and non-corrosive disinfectant.

5.3.3 Performance verification and calibration

The performance of the balance system shall be regularly verified during use and after cleaning with check weights by a trained person. Calibration shall be checked across the entire range by a qualified person at a frequency dependent on use.

Check weights may also be verified immediately after calibration of the balance.

5.4 Homogenizers, blenders and mixers

5.4.1 Description

This equipment is used to prepare the initial suspension from the test sample of non-liquid products.

The following apparatus may be used:

- a peristaltic blender (stomacher) with sterile bags, possibly with a device for adjusting speed and time; or
- a rotary homogenizer (blender), the notional speed of which is between 8 000 r/min and 45 000 r/min inclusive, with sterilizable glass or metals bowls equipped with covers; or
- a vibrational mixer (pulsifier) with sterile bags; or
- another homogenizing system with equivalent efficiency.

In certain cases, manual mixing may be carried out using sterile glass beads having an appropriate diameter (approximately 6 mm; see ISO 6887-2 to ISO 6887-4 and ISO 8261).

5.4.2 Use

The usual operating time of a peristaltic homogenizer is 1 min to 3 min (see ISO 6887-2 to ISO 6887-4 and ISO 8261 for specific foods).

Do not use this type of apparatus for certain foodstuffs, such as:

- products which risk puncturing the bag (presence of sharp, hard or dry particles);
- products which are difficult to homogenize because of their texture (e.g. salami-type sausage).

The rotary homogenizer shall operate for a duration such that the total number of revolutions is between 15 000 r/min and 20 000 r/min inclusive. Even with the slowest homogenizer, this time shall not exceed 2,5 min.

The vibrational mixer may be used for most foodstuffs, including hard or dry products. The usual operating time is 0,5 min to 1 min. If microorganisms are likely to be encountered deep inside cohesive structures, the sample should be cut into small pieces prior to processing.

Glass beads can be used for the preparation, by shaking, of the initial suspensions of certain viscous or thick products, in particular certain dairy products (see specific standards).

5.4.3 Cleaning and disinfection

Clean and disinfect peristaltic homogenizers and vibrational mixers regularly and after any bag spillage or leakage.

For rotary homogenizers, clean and sterilize the glass or metal bowl after each use.

5.4.4 Maintenance

Inspect and maintain equipment in accordance with the manufacturer's instructions.

5.5 pH meter

5.5.1 Description

A pH meter is used to measure the potential difference, at a determined temperature, between a measuring electrode and a reference one, both electrodes being introduced into the product. It shall be capable of being read to the nearest 0,01 pH unit, enabling measurements to be made with a tolerance of $\pm 0,1$ pH unit. The pH meter shall be equipped with either manual or automatic temperature compensation.

NOTE The measuring electrode and the reference electrode are usually grouped together in a combined electrode system.

5.5.2 Use

A pH meter is used to measure the pH value of culture media and reagents to check if adjustment is needed during preparation and as a quality check after sterilization.

It may also be used to measure the pH value of samples and sample suspensions. The use of a pH meter is discussed in the standard specific to the product to be analysed, in which the conditions for the determination of the pH value and for adjustment of the pH value are specified.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/69cd73fc-7220-4631-80c4-2f4c8e512f9a/iso-7218-2007-amd-1-2013>