

---

---

**Soil quality — Gas chromatographic  
determination of the content of  
volatile aromatic hydrocarbons,  
naphthalene and volatile halogenated  
hydrocarbons — Purge-and-trap method  
with thermal desorption**

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

*Qualité du sol — Détermination par chromatographie en phase gazeuse  
des teneurs en hydrocarbures aromatiques volatils, en naphthalène et  
en hydrocarbures halogénés volatils — Méthode par purge et piégeage  
avec désorption thermique*

ISO 15009:2012

<https://standards.iteh.ai/catalog/standards/sist/ae11a7f8-213e-4bd9-88c8-318d0c337aa2/iso-15009-2012>



## iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 15009:2012

<https://standards.iteh.ai/catalog/standards/sist/ae11a7f8-213e-4bd9-88c8-318d0c337aa2/iso-15009-2012>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword .....	iv
<b>1 Scope .....</b>	<b>1</b>
<b>2 Normative references .....</b>	<b>1</b>
<b>3 Principle .....</b>	<b>2</b>
<b>4 Reagents .....</b>	<b>2</b>
<b>4.1 Water, free of volatile aromatic and volatile halogenated hydrocarbons .....</b>	<b>2</b>
<b>4.2 Internal standard compounds .....</b>	<b>2</b>
<b>4.3 Standard compounds .....</b>	<b>3</b>
<b>4.4 Methanol (CAS RN 67-56-1) .....</b>	<b>4</b>
<b>4.5 Adsorbing agent .....</b>	<b>4</b>
<b>4.6 Cooling water for purge and trap .....</b>	<b>4</b>
<b>4.7 Inert carrier gas for the gas chromatograph .....</b>	<b>4</b>
<b>4.8 Nitrogen or helium as inert gas for the purge equipment .....</b>	<b>4</b>
<b>4.9 Standard solutions .....</b>	<b>4</b>
<b>5 Apparatus .....</b>	<b>5</b>
<b>6 Sampling, preservation and sample pretreatment .....</b>	<b>7</b>
<b>6.1 General .....</b>	<b>7</b>
<b>6.2 Sampling using vials prefilled with methanol .....</b>	<b>7</b>
<b>6.3 Sampling using coring tube method .....</b>	<b>7</b>
<b>7 Procedure .....</b>	<b>7</b>
<b>7.1 Blank determination .....</b>	<b>7</b>
<b>7.2 Extraction .....</b>	<b>8</b>
<b>7.3 Purge and trap .....</b>	<b>8</b>
<b>7.4 Gas chromatographic analysis .....</b>	<b>8</b>
<b>8 Calculation .....</b>	<b>10</b>
<b>9 Expression of results .....</b>	<b>11</b>
<b>10 Precision .....</b>	<b>11</b>
<b>11 Test report .....</b>	<b>11</b>
<b>Annex A (informative) Relative retention time with respect to ethylbenzene-D10 of volatile aromatic hydrocarbons and volatile halogenated hydrocarbons on the following columns: CP-Sil 5 CB and CP-Sil 13 CB .....</b>	<b>12</b>
<b>Annex B (normative) Check on internal standards .....</b>	<b>13</b>
<b>Annex C (informative) Validation .....</b>	<b>14</b>
<b>Annex D (informative) Information on purge-and-trap instruments .....</b>	<b>17</b>
<b>Bibliography .....</b>	<b>18</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15009 was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

This second edition cancels and replaces the first edition (ISO 15009:2002), which has been technically revised.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 15009:2012

<https://standards.iteh.ai/catalog/standards/sist/ae11a7f8-213e-4bd9-88c8-318d0c337aa2/iso-15009-2012>

# Soil quality — Gas chromatographic determination of the content of volatile aromatic hydrocarbons, naphthalene and volatile halogenated hydrocarbons — Purge-and-trap method with thermal desorption

## 1 Scope

This International Standard specifies a method for quantitative gas chromatographic determination of volatile hydrocarbons, naphthalene and volatile halogenated hydrocarbons in soil.

This International Standard is applicable to all types of soil.

NOTE In the case of unsaturated peaty soils, absorption of the extraction solution may occur.

The lower limit of determination is dependent on the equipment used and the quality of the methanol grade used for the extraction of the soil sample.

Under the conditions specified in this International Standard, the following limits of determinations apply (expressed on a basis of dry matter):

a) Typical limit of determination when using gas chromatography/flame ionization detection (GC/FID):

— volatile aromatic hydrocarbons: 0,1 mg/kg.

b) Typical limit of determination when using gas chromatography/electron capture detector (GC/ECD):

— volatile halogenated hydrocarbons: 0,01 mg/kg.

Lower limits of determination for some compounds can be achieved by using mass spectrometry (MS) with selected ion detection.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4799, *Laboratory glassware — Condensers*

ISO 10381-1, *Soil quality — Sampling — Part 1: Guidance on the design of sampling programmes*

ISO 10381-2, *Soil quality — Sampling — Part 2: Guidance on sampling techniques*

ISO 10381-5, *Soil quality — Sampling — Part 5: Guidance on the procedure for the investigation of urban and industrial sites with regard to soil contamination*

ISO 11465, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*

ISO 11465:1993/Cor 1:1994, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method — Technical Corrigendum 1*

ISO 15680, *Water quality — Gas-chromatographic determination of a number of monocyclic aromatic hydrocarbons, naphthalene and several chlorinated compounds using purge-and-trap and thermal desorption*

ISO 18512, *Soil quality — Guidance on long and short term storage of soil samples*

ISO 22892, *Soil quality — Guidelines for the identification of target compounds by gas chromatography and mass spectrometry*

### 3 Principle

Test samples are taken from an untreated field-moist soil sample. To prevent losses of the volatiles, samples are taken as undisturbed as possible in the field with a tube corer or by adding methanol immediately in the field.

The test sample is extracted with methanol. After centrifugation, part of the methanol extract is brought into a purge vessel filled with water. The volatile compounds are purged with nitrogen or helium and adsorbed on a suitable adsorbing agent. The adsorbed compounds are desorbed thermally and by means of a carrier gas flow, whether or not via a cold trap, brought into a gas chromatograph. The various compounds are separated by using a capillary column with an immobile phase of low polarity. Volatile organic compounds are detected with appropriate detectors such as: mass spectrometric detector (MS), flame ionization detector (FID), electron capture detector (ECD), photo ionization detector (PID) or electrolytic conductivity detector (ELCD). Identification and quantification takes place by comparison of retention times and peak heights (or peak areas) towards an internal standard added with the corresponding variables of an external standard solution. The efficiency of the procedure depends on the composition of the soil that is investigated. The described procedure does not take into account incomplete extraction caused by structure and composition of the soil sample.

When using non-specific detectors such as FIDs and ECDs, the confirmation of the identity of the detected compounds and their concentrations should be done by repeating the gas chromatographic analysis using a column of different polarity. When using gas chromatography/mass spectrometry (GC/MS), the identity confirmation and the quantification can be done in a single run.

NOTE 1 This International Standard follows the description of an off-line purge-and-trap method. The use of commercial available online instruments is allowed, provided that equivalent results are obtained during validation of this equipment. With such an instrument, purge and trap occurs on line with gas chromatography and detection. Follow the manufacturer's manual, especially regarding the items composing the apparatus which are listed in 5.1.1 to 5.1.9.

NOTE 2 Other injection techniques, such as static headspace followed by thermal desorption (ISO 22155) or solid-phase micro-extraction (SPME), can be used, provided that their applicability is proven.

### 4 Reagents

All reagents shall be of recognized analytical grade. Verify whether the reagents are applicable for this specific purpose and free of interfering compounds.

#### 4.1 Water, free of volatile aromatic and volatile halogenated hydrocarbons

Usually boiler water with a temperature of at least 80 °C and 1 day old can be applied. Purging with an inert gas, e.g. a flow of 10 ml/min of nitrogen for 30 min, is another means of removing interfering compounds from water. A sufficient amount of water from the same batch should be available to complete each batch of analyses, including all preparations.

#### 4.2 Internal standard compounds

4.2.1 For the determination of volatile aromatic hydrocarbons, preferably two internal standard compounds shall be selected that do not interfere with compounds present in the sample extract.

Examples of suitable internal standards are

- toluene-D8 (CAS RN 2037-26-5),
- ethylbenzene-D10 (CAS RN 25837-05-2), and
- 2-bromofluorobenzene (CAS RN 1072-85-1).

**4.2.2** For the determination of volatile halogenated hydrocarbons, preferably two internal standard compounds shall be selected that do not interfere with compounds present in the sample extract.

Examples of suitable internal standards are:

- 1,4-dichlorobutane (CAS RN 110-56-5);
- $\alpha,\alpha,\alpha$ -trifluorotoluene (CAS RN 98-08-8);
- 2-bromofluorobenzene (CAS RN 1072-85-1).

### 4.3 Standard compounds

#### 4.3.1 Volatile aromatic hydrocarbons

- benzene (CAS RN 71-43-2);
- toluene (CAS RN 108-88-3);
- ethylbenzene (CAS RN 100-41-4);
- *o*-xylene (CAS RN 95-47-6);
- *m*-xylene (CAS RN 108-38-3);
- *p*-xylene (CAS RN 106-42-3);
- styrene (CAS RN 100-42-5);
- naphthalene (CAS RN 91-20-3).

#### 4.3.2 Volatile halogenated hydrocarbons

- dichloromethane (CAS RN 75-09-2);
- trichloromethane (CAS RN 67-66-3);
- tetrachloromethane (CAS RN 56-23-5);
- 1,1-dichloroethane (CAS RN 75-34-3);
- 1,2-dichloroethane (CAS RN 107-06-2);
- 1,1,1-trichloroethane (CAS RN 79-01-6);
- 1,1,2-trichloroethane (CAS RN 79-00-5);
- 1,2-dichloropropane (CAS RN 78-87-5);
- 1,2,3-trichloropropane (CAS RN 98-18-4);
- *cis*-1,3-dichloropropene (CAS RN 10061-01-5);
- *trans*-1,3-dichloropropene (CAS RN 10061-02-6);
- *cis*-1,2-dichloroethene (CAS RN 156-59-2);
- *trans*-1,2-dichloroethene (CAS RN 156-60-5);
- 3-chloropropene (CAS RN 107-05-1);
- trichloroethene (CAS RN 79-01-6);
- tetrachloroethene (CAS RN 127-18-4);

- monochlorobenzene (CAS RN 108-90-7);
- 1,2-dichlorobenzene (CAS RN 95-50-1).

#### 4.4 Methanol (CAS RN 67-56-1)

The methanol used shall not contain more than 100 µg/l of the individual volatile aromatic compounds and not more than 10 µg/l of the volatile halogenated hydrocarbons that are to be analysed.

#### 4.5 Adsorbing agent

Polymer of 2,6-diphenyl-*p*-phenoxide (40 mesh to 60 mesh) of a grade suitable for thermal desorption.

NOTE 1 2,6-diphenyl-*p*-phenoxide is commercially available as Tenax TA<sup>1)</sup>.

NOTE 2 Other adsorbing agents may be used provided that their suitability has been tested.

#### 4.6 Cooling water for purge and trap

The temperature of the cooling water depends on the dimensions of the purge-and-trap equipment (5.1). A temperature of about 10 °C is recommended. A cryostat shall be used if the temperature of the cooling water is too high.

#### 4.7 Inert carrier gas for the gas chromatograph

Helium, nitrogen or argon-methane mixture ultra-pure. Other gases for gas chromatography shall be used in accordance with the instrument manufacturer's instructions.

#### 4.8 Nitrogen or helium as inert gas for the purge equipment

ISO 15009:2012  
<https://standards.iteh.ai/catalog/standards/sist/ae11a7f8-213e-4bd9-88c8-318d0c337aa2/iso-15009-2012>

##### 4.9.1 Standard stock solutions for volatile aromatic and halogenated compounds in methanol, 4 g/l.

Weigh about 100 mg of the individual standard compounds (4.3) with an accuracy of 0,1 mg into a closed septum flask containing 25 ml of methanol. Transfer the standard compounds into the flask by using a syringe.

NOTE The stock solution is stable for about 6 months when stored at -18 °C.

##### 4.9.2 Internal standard solutions in methanol, 4 g/l.

Weigh about 100 mg of the individual internal standard compounds (4.2.1 and 4.2.2), accurate to 0,1 mg, into a closed septum flask containing 25 ml of methanol. Using a syringe, transfer the standard compounds into the flask.

The containers containing the solutions shall be marked or weighed so that any evaporation losses of the solvent may be recognized. The solutions shall be stored in the dark at a temperature of 4 °C ± 2 °C. Prior to use, they shall be brought to ambient temperature.

##### 4.9.3 Calibration solutions

Calibration solutions containing 0 mg/l to 200 mg/l of each standard (4.3) and the selected internal standard compounds (4.2), 200 mg/l. The calibration solutions are prepared in methanol.

Dilute the amounts indicated in Table 1 of the solutions obtained according to 6.2 and 6.3 with methanol (4.4) to 100 ml.

Other volumes of methanol may be used as long as suitability is proven.

1) Tenax TA is an example of a suitable product available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of this product.



Table 1 — Example for preparation of calibration solutions

Calibration solution	Internal standard solution (4.9.2) ml	Standard stock solution (4.9.1) ml	Concentration in the calibration solution mg/l	Quantity in µg/5 µl calibration solution in 100 ml (sample) water
1	5	0	0	0
2	5	1	40	0,2
3	5	2	80	0,4
4	5	3	120	0,6
5	5	4	160	0,8
6	5	5	200	1,0

The total volume of the methanol used for calibration shall be the same as that which will be taken for the methanol extract of the soil sample (see 7.2).

## 5 Apparatus

Usual laboratory glassware, free of interfering compounds.

All glassware shall be cleaned according to the usual procedures for this type of analysis.

### 5.1 Purge-and-trap apparatus

The instrument described here is for an off-line purge-and-trap method. As mentioned in Note 1 to Clause 3, commercially available automated systems are allowed provided the requirements of this International Standard are met. Annex D gives some considerations for the use of such systems.

<https://standards.iteh.ai/catalog/standards/sist/a611a7f8-213e-4bd9-88c8-333333333333/iso-15009-2012>

5.1.1 Round-bottom flask with three angled side necks; volume 100 ml.

5.1.2 Gas inlet tube with a tip of sintered glass.

5.1.3 Ball-and-cup stopcock with a polytetrafluoroethylene (PTFE) ring.

5.1.4 Flow adjustment; the nitrogen flow shall be 40 ml/min ± 2 ml/min.

5.1.5 Inlet tube for the thermocouple.

5.1.6 Allin- or Graham-type condenser (see ISO 4799).

5.1.7 Screw cap with cut-off ring made of silicone rubber with PTFE inlay.

5.1.8 Adsorption tubes

Tubes made of glass or stainless steel, filled with at least 240 mg of adsorbing agent (4.5).

The adsorbent is kept in place by using inert material, e.g. silanized glass fibre. The tubes shall be suitable for direct use in connection with the apparatus for thermal desorption. The tubes shall be marked on one side. The tubes shall be provided with caps of inert material, e.g. polyethylene or metal, with screw caps and a PTFE ring, that allow tight closing after purging.

Before use, the adsorbent shall be activated and purified by slowly heating the tubes to 250 °C and keeping them at that temperature for 3 h while a nitrogen flow of 10 ml/min is maintained. The adsorbent shall be cooled under

nitrogen and the tubes analysed. The result of a blank determination shall not exceed the equivalent of 1 ng of a compound to be analysed. When the result is higher than this, the adsorbent shall be desorbed once more.

NOTE The use of commercially available tubes is recommended.

Tubes that are used should not be used again, unless the blank determination meets the above-mentioned requirements.

Care should be taken to avoid cross-contamination. A heavily loaded tube can contaminate a lightly loaded tube in the sample change platform.

**5.1.9** Heating block with thermocouple, suitable for heating 100 ml flasks

**5.2 Centrifuge**, suitable for centrifuging tubes of 200 ml with such a rotation frequency that the radial acceleration is 2 000g to 3 000g.

**5.3 Centrifuge tubes**, with a volume of 200 ml.

**5.4 Capillary columns**

Fused silica capillary columns with a non-polar or semipolar stationary phase allowing sufficient separation of the compounds of interest. A thick film of stationary phase increases the efficiency of the more volatile compounds.

Examples are given in 7.4 and Annex A.

**5.5 Gas chromatograph**, equipped with one or two appropriate detectors. Detectors, such as flame ionization detectors (FIDs), electron capture detectors (ECDs), photo-ionization detectors (PIDs) or electrolytic conductivity detectors (ELCDs), and a mass spectrometer (MS) can be used, depending on the substances to be analysed and their target level of contamination. The mass spectrometer should be capable of operating over the total mass range of interest and being equipped with a data system capable of quantifying ions using selected  $m/z$  values.

**5.6 Apparatus for thermal desorption**

The apparatus used shall meet the following requirements:

- a primary desorption oven with adjustable desorption temperature up to 250 °C and adjustable desorption time;
- a cold trap/secondary desorption oven;
- a connecting tube to the gas chromatograph, with adjustable heating up to 150 °C;
- adjustable carrier-gas flow rate up to 40 ml/min.

NOTE Instruments for thermal desorption are commercially available.

**5.7 Electronic integrator or automatic recorder**

**5.8 Syringe**, of volume 5 µl, readable to 0,1 µl, and of volume 50 µl, readable to 1 µl.

**5.9 Horizontal shaking machine**

A shaking machine with horizontal movement (200 to 300 movements per minute).

## 6 Sampling, preservation and sample pretreatment

### 6.1 General

Sampling shall be carried out in accordance with ISO 10381-1 and ISO 10381-5, as appropriate, using equipment according to ISO 10381-2, after coordination with the analytical laboratory.

Samples shall be analysed as soon as possible upon their receipt in the laboratory. If it is necessary to store samples, they shall be stored in cool conditions in accordance with ISO 18512. Samples are not pretreated. Exposure of samples to air, even during sampling, shall be avoided.

Samples for the determination of volatile compounds can be obtained using several techniques. It is strongly recommended that one of the procedures described in 6.2 and 6.3 be used.

Determine the dry matter content of the field-moist sample in accordance with ISO 11465. In case the sampling method in 6.2 is used, a separate sample should be delivered to the laboratory for determination of the dry matter.

### 6.2 Sampling using vials prefilled with methanol

Transfer a defined volume of soil using an appropriate device into a pre-weighed vial which is filled with a defined volume of methanol (4.4). Prevent leakages by cleaning the top of the vessel before sealing.

The soil samples should be taken from undisturbed material using an appropriate sample cutter of known volume, e.g. a modified 20 ml disposable plastic syringe with the tip cut off. The soil sample should be collected immediately after exposing a fresh soil surface of the drilling core, e.g. of an open-window sampler or the trial pit wall. The incorporation of material-like roots or stones should be avoided as far as possible.

Make sure that the sample is completely covered with methanol (4.4). Then close the cap of the PTFE-coated septum. At least one blank sample on every site shall be prepared in the field by opening the prepared vial for the same time period as necessary for filling with the soil sample. Add methanol (4.4) and close the cap of the vial.

The sampling vials should be kept dark in a cooler (before and after sampling) throughout the whole transportation. For details, see ISO 18512.

### 6.3 Sampling using coring tube method

This method, which consists of taking an undisturbed sample, greatly reduces or eliminates common losses (e.g. due to evaporation, diffusion, sorption onto plastics). It involves a stainless-steel coring tube of minimal volume 200 ml which is filled *in situ*, retrieved and capped with a non-permeable material, e.g. stainless steel, aluminium foil. The tube should be filled totally.

NOTE This method is not suitable for very stony soils.

Store in cool conditions at a temperature of 2 °C to 8 °C for no longer than 4 d; see ISO 18512.

In the laboratory during sub-sampling, take care that no volatile compounds are lost. Start as soon as possible with the cooled sample. Use the whole content of the coring tube or take a sub-sample with a suitable instrument, e.g. an apple corer, and put it directly into the vial (see 7.2).

## 7 Procedure

### 7.1 Blank determination

For each series of samples, a blank determination shall be carried out by treating 50 ml of water (4.1), to which 5 µl to 50 µl of methanol (4.4) has been added, as a sample. Ensure that no contamination occurs from the laboratory atmosphere.