
**Water quality — Determination of
carbon 14 activity — Liquid scintillation
counting method**

*Qualité de l'eau — Détermination de l'activité volumique du
carbone 14 — Méthode par comptage des scintillations en milieu liquide*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 13162:2011](https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011)

[https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-
292baaa231e3/iso-13162-2011](https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011)



iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 13162:2011

<https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Symbols, definitions, units, and abbreviations	2
4 Principle	2
5 Reagents and equipment	3
5.1 Reagents	3
5.2 Equipment	4
6 Sampling and samples	5
6.1 Sampling	5
6.2 Sample storage	5
7 Procedure	5
7.1 Sample preparation	5
7.2 Preparation of the sources to be measured	5
7.3 Counting procedure	6
7.4 Calibration and verification	6
7.5 Measurement conditions	6
8 Expression of results	7
8.1 General	7
8.2 Calculation of activity concentration	7
8.3 Decision threshold	8
8.4 Detection limit	8
8.5 Confidence limits	9
8.6 Calculations using the activity per mass	9
9 Test report	9
Annex A (informative) Numerical applications	11
Annex B (informative) Internal standard method	13
Annex C (informative) Extraction of total carbon: precipitate counting	15
Annex D (informative) Extraction of total carbon: absorption counting	18
Bibliography	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 13162 was prepared by Technical Committee ISO/TC 147, *Water quality*.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 13162:2011

<https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011>

Introduction

The carbon 14 (^{14}C) present in the environment is of natural origin and man made. As a result of atmospheric nuclear weapon testing, emissions from nuclear engineering installations, and the application and processing of isotopes, relatively large amounts of ^{14}C have been released into the environment. Due to the substantial proportion of ^{14}C in the human internal dose contribution, monitoring of ^{14}C activity concentrations in the environment is necessary in order to follow its circulation in the hydrosphere and biosphere. ^{14}C is the second radionuclide ($\sim 3\,500\text{ Bq}$) to contribute to the human body natural radioactivity, behind ^{40}K ($\sim 6\,000\text{ Bq}$).

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 13162:2011

<https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 13162:2011

<https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011>

Water quality — Determination of carbon 14 activity — Liquid scintillation counting method

WARNING — Persons using this International Standard should be familiar with normal laboratory practice. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this International Standard be carried out by suitably trained staff.

1 Scope

This International Standard specifies the conditions for the determination of ^{14}C activity concentration in samples of environmental water or of ^{14}C -containing water using liquid scintillation counting.

The method is applicable to the analysis of any organic molecule soluble in water that is well mixed with the scintillation cocktail. It does not apply to micelles or “large” particles (lipids, fulvic acid, humic acid, etc.) that are inadequately mixed with the scintillation cocktail and the water. Some beta energy is lost without any excitation of the scintillation cocktail and the results are underestimated. The method is not applicable to the analysis of organically bound ^{14}C , whose determination requires additional chemical processing (such as chemical oxidation, combustion).

It is possible to determine ^{14}C activity concentrations below 10^6 Bq l^{-1} without any sample dilution.

<https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011>

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 11929, *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Symbols, definitions, units, and abbreviations

For the purposes of this document, the symbols, definitions, units, and abbreviations defined in ISO 80000-10, ISO 11929, ISO/IEC Guide 98-3 and the following apply.

A	Activity of the calibration source, in becquerels
c_A	Activity concentration, in becquerels per litre
c_A^*	Decision threshold, in becquerels per litre
$c_A^\#$	Detection limit, in becquerels per litre
$c_A^{<}, c_A^{>}$	Lower and upper limits of the confidence interval, in becquerels per litre
f_q	Quench factor
m	Mass of test sample, in kilograms
r_0	Background count rate, in reciprocal seconds
r_g	Sample count rate, in reciprocal seconds
r_s	Count rate of the calibration sample, in reciprocal seconds
t_0	Background counting time, in seconds
t_g	Sample counting time, in seconds
t_s	Counting time of the calibration sample, in seconds
U	Expanded uncertainty, calculated by $U = k u(c_A)$ with $k = 1, 2, \dots$, in becquerels per litre
$u(c_A)$	Standard uncertainty associated with the measurement result, in becquerels per litre
V	Volume of test sample, in litres
α	Activity per mass, in becquerels per kilogram
β_{\max}	Maximum energy for the beta emission, in kiloelectronvolts
ε	Detection efficiency
ρ	Mass density of the sample, in kilograms per litre

4 Principle

The scintillation phenomenon results from interaction of ionizing radiation with solvents and compounds exhibiting fluorescence (scintillators). Both solvents and scintillators constitute the scintillation cocktail. The scintillation mixture is achieved by adding the scintillation cocktail to the test sample in order to obtain a homogeneous mixture.

The test sample is mixed with the scintillation cocktail in a counting vial to obtain a homogeneous medium. Electrons emitted by ^{14}C transfer their energy to the scintillation medium. Molecules excited by this process return to their ground state by emitting photons that are detected by photodetectors.

The electric pulses emitted by the photodetectors are amplified, sorted (in order to remove random events) and analysed by the electronic systems and the data analysis software. The count rate of these electric pulses allows the determination of the test sample activity, after correcting for the background count rate and detection efficiency.

In order to determine the background, a blank sample is prepared in the same way as the test sample. The blank sample is prepared using a reference water of the lowest activity available, in accordance with the activities to be measured.

The detection efficiency is determined with a calibration sample that is prepared with a standard of aqueous ^{14}C , or a dilution of this standard with reference water, measured under the same conditions as the test sample.

The sample (blank, test, calibration) and the measurement conditions shall be:

- same type of counting vial;
- same filling geometry;
- same scintillation cocktail;
- same ratio between test sample and scintillation cocktail;
- temperature stability of the detection apparatus;
- value of quench-indicating parameter included in calibration curve.

A prerequisite for the direct determination of ^{14}C in a water sample is the absence of or a negligible contribution from other beta-emitting radionuclides, such as ^{90}Sr and Ra isotopes. When the radionuclide content of the sample is unknown, the method specified in this International Standard only provides a ^{14}C equivalent activity for the sample.

Examples of methods of sample pretreatment are described in Annexes C and D.

Concerning quench correction, if particular conditions of chemical quenching affect the measurement results, it is recommended that a quench curve be established. It is important to choose the chemical quenching agent in accordance with the supposed type of quenching observed in the sample.

ISO 13162:2011

<https://standards.iteh.ai/catalog/standards/sist/a2de48d0-6906-4373-ad55-292baaa231e3/iso-13162-2011>

5 Reagents and equipment

5.1 Reagents

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade.

5.1.1 Reference water for the blank

The reference water for the blank should be as free as possible of chemical or radioactive impurities.

The reference water may have a low ^{14}C activity concentration, in becquerels per litre, at the time t at which the samples are measured.

For example, obtain water with a ^{14}C activity concentration as low as possible, e.g. (deep) subterranean water. Distil the water. Keep the distillate in a well-sealed borosilicate glass bottle in the dark at a temperature as constant as possible; this reference water shall be kept physically remote from any ^{14}C -containing material (see next paragraph). Determine (see final paragraph) the ^{14}C activity concentration ($t = 0$), in becquerels per litre, of this water and note the date ($t = 0$) of this determination.

It is advisable to keep an adequate quantity of reference water in stock and to draw off small working volumes from it for immediate use, as required. Contamination with ^{14}C (e.g. from CO_2 in the air) or other radioactive species should be avoided.

For measurement of activity concentrations close to 1 Bq l^{-1} , water with a very low activity concentration is necessary as reference water.

5.1.2 Calibration source solution

In order to avoid cross-contamination, prepare the calibration source solution in a suitable location which is remote from the area where the ^{14}C analyses are to be carried out. Weigh and pour the requisite quantity of a ^{14}C aqueous standard solution into a weighed volumetric flask (e.g. of capacity 100 ml), so that the ^{14}C activity concentration generates sufficient counts to reach the required measurement uncertainty after dilution with water for the blank and thorough mixing. Calculate the ^{14}C activity concentration of the resulting internal standard solution ($t = 0$), in becquerels per litre. Note the date at which the standard solution was made up ($t = 0$).

5.1.3 Scintillation solution

The scintillation cocktail is chosen according to the characteristics of the sample to be analysed and according to the properties of the detection apparatus.

It is recommended that a good hydrophilic scintillation cocktail be used, especially for the measurement of low activity levels.

The scintillation cocktail shall be homogeneous and stable.

For the measurement of raw waters containing particles in suspension, it is recommended that a scintillation cocktail leading to a gel-type mixture for the added volume of water be used.

It is recommended that the scintillation solution be:

- stored in the dark and, particularly just before use, exposure to direct sunlight or fluorescent light avoided in order to prevent interfering luminescence,
- compliant with the storage conditions specified by the scintillation cocktail supplier.

The mixtures (scintillation cocktail and test sample) should be disposed of as chemical waste, and, depending on the radioactivity, may require disposal as radioactive waste.

5.1.4 Quenching agent

Examples of chemical quenching agents (non acid): organochloride compounds, nitromethane.

NOTE Some quenching agents are dangerous or toxic.

5.2 Equipment

Usual laboratory equipment and in particular the following.

5.2.1 Liquid scintillation counter, preferably with an automatic sample transfer and ability to measure or correct for sample quench.

Operation at constant temperature is recommended. Follow the manufacturer's instructions.

The method specified in this International Standard relates to the widely used liquid scintillation counters (LSCs) with vials that hold about 20 ml. When other vials are used with appropriate counters, the method specified shall be adapted accordingly.

5.2.2 Pipette, suitable for

- accurate transfer of the standard solution (e.g. a micropipette of capacity 100 μl);
- accurate transfer of the test sample.

5.2.3 Balance, e.g. capable of being read to the nearest 0,1 mg.

5.2.4 Counting vials.

Different types of scintillation vials exist, manufactured using a range of materials. The most common are glass vials and polyethylene (PE) vials. Glass vials allow visual inspection of the scintillation medium, but have an inherent background, due to the presence of ^{40}K . However, some organic solvents contained in scintillation cocktails diffuse through PE, accelerating the degradation of the mixture.

There are other types of vials.

- Glass vials with low level of ^{40}K , which exhibit a lower background than “normal” glass vials.
- Polytetrafluoroethylene vials (PTFE) or PE vials with a layer of PTFE on the inside wall, which are strongly recommended for the determination of very low ^{14}C concentrations. Diffusion of organic solvents is slower through PTFE than through PE. These vials are used for long counting times when very low-level activity is to be measured.

Generally, the vials are single use. If the vial is re-used, efficient cleaning is essential.

To prevent interfering luminescence, counting vials should be kept in the dark and should not be exposed to direct sunlight or fluorescent light, particularly just before use.

NOTE Toluene-based scintillation solutions can physically distort PE and counting vials made of that polymer are unsuitable for use with them. Diffusion of organic solvents into and through the walls is also a serious drawback of PE vials.

6 Sampling and samples

iTeh STANDARD PREVIEW
(standards.iteh.ai)

6.1 Sampling

Conditions of sampling shall conform to ISO 5667-1 and ISO 5667-3.

The samples shall not be acidified to avoid the destruction of the carbonic equilibrium (CO_3^{2-} , HCO_3^- , H_2CO_3), as specified in ISO 5667-3.

It is important that the laboratory receives a representative sample, unmodified during transport or storage and in an undamaged container. It is recommended that a glass flask filled to the maximum be used to minimize ^{14}C exchange with atmospheric CO_2 .

For low level activity measurements, it is important to avoid any contact between sample and atmosphere during the sampling.

6.2 Sample storage

If required, the sample shall be stored in compliance with ISO 5667-3 for carbon dioxide. If the storage duration exceeds that specified in ISO 5667-3, it is advisable to store the samples in glass flasks.

7 Procedure

7.1 Sample preparation

On a raw sample, measurement of the test sample is generally performed without removal of suspended matter, if the sample has low levels of such material. If the activity of a filtered or centrifuged sample is to be measured, the removal of suspended matter shall be performed as soon as possible after sampling.

7.2 Preparation of the sources to be measured

Known quantities of test sample and scintillation cocktail are introduced into the counting vial.

After closing the vial, it shall be thoroughly shaken to homogenize the mixture.