



Designation: D6129 – 97(Reapproved 2007)

Standard Test Method for Silicon in Engine Coolant Concentrates by Atomic Absorption Spectroscopy¹

This standard is issued under the fixed designation D6129; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of silicon in the range from 200 to 500 ppm in engine coolant concentrates by atomic absorption. This method is as accurate and precise as photometric methods, while requiring considerably less operator time and avoiding problems with reagent instability.

1.1.1 Coolants with silicon content outside of this range may be analyzed by this method by suitably adjusting the sample size. Care should be taken to ensure that the glycol content of the working standards corresponds to that of the sample solution being analyzed.

1.2 The values stated in SI units are to be regarded as the standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

<http://www.astm.org/standards/D1176> Practice for Sampling and Preparing Aqueous Solutions of Engine Coolants or Antirusts for Testing Purposes

3. Summary of Test Method

3.1 The sample is dissolved in water and the silicon content is determined from a standard curve established at the time the sample is run. An atomic absorption spectrometer with a nitrous oxide-acetylene flame is used.

4. Apparatus

4.1 *Atomic Absorption Spectrometer*, with nitrous oxide-acetylene flame.

¹ This specification is under the jurisdiction of ASTM Committee D15 on Engine Coolants and Related Fluids and is the direct responsibility of Subcommittee D15.04 on Chemical Properties.

Current edition approved April 1, 2007. Published April 2007. Originally approved in 1997. Last previous edition approved in 2002 as D6129 - 97(2002). DOI: 10.1520/D6129-97R07.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

4.2 *Hollow-Cathode Lamp*, for silicon.

5. Reagents

5.1 Reagent grade chemicals shall be used in all tests.³

5.2 References to water shall be understood to mean distilled water or water of equal quality.

5.3 *Standard Silicon Solution*, 1000 mg/L (see Note 1).

5.4 *Ethylene Glycol* (propylene glycol should be used when analyzing PG-based coolants).

5.5 *Sodium Hydroxide, 5 %* in water (see Note 1). When preparing from solid sodium hydroxide avoid prolonged exposure to the atmosphere. DO NOT STORE IN GLASS.

5.6 Working standards: 0, 20, 40, and 60 mg/L silicon.

5.6.1 Prepare these solutions by pipeting 2.0, 4.0, and 6.0 mL of the 1000 mg/L standard silicon solution into separate 100 mL volumetric flasks. Add 10.0 mL of ethylene glycol and 2.0 mL of 5 % sodium hydroxide solution to each flask and dilute to volume with water. Prepare the zero standard solution by adding 10.0 mL ethylene glycol and 2.0 mL 5 % sodium hydroxide solution to another 100 mL volumetric flask and diluting to volume with water.

NOTE 1—Standard silicon solutions and sodium hydroxide solutions are available from most laboratory supply companies.

6. Sampling

6.1 Sample material in accordance with Test Method D1176.

7. Procedure

7.1 To the nearest 0.001 g, weigh 5 g of sample into a 50 mL volumetric flask. Dissolve in and dilute to volume with water.

7.2 Using the nitrous oxide-acetylene flame, optimize the instrument operating conditions for silicon determination at the 251.6 nm spectral line according to the manufacturer's recommendations.

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC) Rockville, MD.