

---

---

**Microbiology of food and animal feeding  
stuffs — Protocol for the validation of  
alternative methods**

**AMENDMENT 1**

*Microbiologie des aliments — Protocole pour la validation des  
méthodes alternatives*  
**AMENDEMENT 1**  
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011>



**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 16140:2003/Amd 1:2011](https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011)  
<https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

Page

|  |           |
|--|-----------|
| Foreword .....   | iv        |
| <b>6.3 Interlaboratory study</b> .....   | <b>2</b>  |
| <b>6.3.1 General</b> .....   | <b>2</b>  |
| <b>6.3.2 Terms and definitions</b> .....   | <b>2</b>  |
| <b>6.3.3 Measurement protocol and samples</b> .....  | <b>4</b>  |
| <b>6.3.4 Determination of the trueness and precision characteristics</b> .....   | <b>5</b>  |
| <b>6.3.5 Scrutiny of the measurement results for consistency</b> .....   | <b>7</b>  |
| <b>6.3.6 Comparison of the trueness and precision characteristics of the reference method and the alternative method</b> ..... | <b>8</b>  |
| <b>Annex Q (normative) Calculation of Rousseeuw's scale estimator <math>Q_n</math></b> .....                                   | <b>9</b>  |
| <b>Annex V (informative) Indicators for Mandel's robust <math>h</math>- and <math>k</math>-statistics</b> .....                | <b>11</b> |
| <b>Annex W (informative) Example: Determination of <i>E. coli</i></b> .....  | <b>12</b> |
| <b>Bibliography additions</b> .....  | <b>18</b> |

## iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 16140:2003/Amd 1:2011  
<https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 16140:2003 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 9, *Microbiology*.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 16140:2003/Amd 1:2011](https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011)  
<https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011>

# Microbiology of food and animal feeding stuffs — Protocol for the validation of alternative methods

## AMENDMENT 1

*Page 22, 6.3*

Replace the existing text with that commencing on p. 2.

*Page 61, Annex Q*

Replace the existing text with that commencing on p. 9.

*Page 74, Bibliography*

Insert Annexes V and W commencing on p. 11 before the bibliography.

Additions to the bibliography are given on p. 18.

## 6.3 Interlaboratory study

### 6.3.1 General

An interlaboratory study aims to **determine the comparative performance characteristics** (trueness and precision characteristics) of the alternative method against the reference method.

Guidelines and requirements for organizing, dispatching and conducting the interlaboratory study are given in Annex H and in ISO 5725-2.

At least eight laboratories shall participate in an interlaboratory study.

The organizing laboratory is responsible for the preparation of the test protocol and a data sheet (see below) to be used by each laboratory for recording all measurement results and critical experimental conditions (see H.3).

The analyst in each collaborating laboratory shall demonstrate competence in the use of the alternative method and of the reference method prior to participating in the study.

In microbiology, the data  $\{y\}$  of repeated measurements do not always show a normal (Gaussian) distribution. Therefore, the distribution of these data should be checked for normality if more than 30 values are available at the same level. In order to get a more symmetric distribution, take logarithms of the counts.

The data obtained from interlaboratory studies often contain outliers, i.e. measurement values that deviate so much from comparable measurements that they are considered inconsistent. If they are retained in the data set, the trueness and precision characteristics (averages, standard deviations, etc.) obtained with classical methods of statistical analysis are unreliable. Therefore, ISO 5725-2 includes **outlier tests** (Cochran, Grubbs) in order to detect and eventually discard outliers and to obtain reliable trueness and precision characteristics. Often this causes disputes as to which of the outliers should be discarded from or retained in the statistical analysis. In order to avoid such disputes, **robust estimates** of the trueness and precision characteristics are used in this International Standard. Since they are insensitive to any extreme values, they always use the complete data set obtained from the interlaboratory study.

However, each extreme value should be checked for a clerical error in transcribing the measurement result, an error in computation, a slip in performing the measurement or an analysis of the wrong sample. If possible, such values should be replaced by the correct values. **Other extreme measurement results or laboratories reporting extreme values are not excluded from the statistical analysis unless exclusion is based on sound microbiological reasons.**

### 6.3.2 Terms and definitions

For the purpose of this document, the following terms and definitions apply.

#### 6.3.2.1

##### **accuracy**

closeness of agreement between a test result or measurement result and the true value

NOTE 1 In practice, the accepted reference value is substituted for the true value.

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.1]

NOTE 2 Accuracy refers to a combination of trueness and precision.

**6.3.2.2****trueness**

closeness of agreement between the expectation of a test result or a measurement result and the true value

NOTE 1 The measure of trueness is usually expressed in terms of bias.

NOTE 2 In practice, the accepted reference value is substituted for the true value.

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.3]

**6.3.2.3****precision**

closeness of agreement between independent test/measurement results obtained under stipulated conditions

NOTE Quantitative measures of precision depend critically on the stipulated conditions. Repeatability conditions and reproducibility conditions are particular sets of extreme stipulated conditions.

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.4]

**6.3.2.4****repeatability**

precision under repeatability conditions

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.5]

**6.3.2.5****repeatability conditions**

observation conditions where independent test/measurement results are obtained with the same method on identical test/measurement items in the same test or measuring facility by the same operator using the same equipment within short intervals of time

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.6]

[ISO 16140:2003/Amd 1:2011](https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011)

<https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-96c64b4d7427/iso-16140-2003-amd-1-2011>

**6.3.2.6****repeatability standard deviation**

standard deviation of test result or measurement results obtained under repeatability conditions

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.7]

**6.3.2.7****repeatability limit**

$r$

value less than or equal to which the absolute difference between two measurement results obtained under repeatability conditions is expected to be with a probability of 95 %

NOTE Adapted from ISO 3534-2:2006<sup>[13]</sup>, 3.3.9.

**6.3.2.8****reproducibility**

precision under reproducibility conditions

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.10]

**6.3.2.9****reproducibility conditions**

observation conditions where independent test/measurement results are obtained with the same method on identical test/measurement items in different test or measuring facilities with different operators using different equipment

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.11]

**6.3.2.10 reproducibility standard deviation**

standard deviation of test results or measurement results obtained under reproducibility conditions

[ISO 3534-2:2006<sup>[13]</sup>, 3.3.12]

**6.3.2.11 reproducibility limit**

*R*

value less than or equal to which the absolute difference between two measurement results obtained under reproducibility conditions is expected to be with a probability of 95 %

NOTE Adapted from ISO 3534-2:2006<sup>[13]</sup>, 3.3.14.

**6.3.3 Measurement protocol and samples**

The protocol is the following.

- One relevant food matrix is used (see Annex B).
- The analyte concentrations should be chosen to cover at least the lower, middle and upper levels of the entire range of the alternative method. The organizing laboratory shall assure that the samples are homogeneous (Reference [14]). A negative control should also be included.
- Artificial contamination of a food sample with the target analyte may be used.
- To compare the alternative method with the reference method the same samples shall be used for each method. Four sub-samples from each level (or two aliquots, each measured by both methods) are prepared for each laboratory. These are blind coded but labelled so that two are measured by the reference method and two are measured by the alternative method.
- Liquid samples (compared to solid samples) give greater assurance of homogeneity if prepared and dispatched without change in microbiological content and used correctly. In specific cases, it could be necessary to subdivide the samples immediately before measurement with both methods.
- The measurements shall be performed in each collaborative laboratory and the organizing laboratory at a stipulated date using common batches of media and kits.
- The rounding of results shall be set by the organizing laboratory.
- For each level *j*, the results of the interlaboratory study shall be presented as shown in Table 9.

**Table 9 — Presentation of the results of the interlaboratory study at level *j***

| Laboratory<br><i>i</i> | Reference method (coded) |             | Alternative method (coded) |             |
|------------------------|--------------------------|-------------|----------------------------|-------------|
|                        | Duplicate 1              | Duplicate 2 | Duplicate 1                | Duplicate 2 |
| 1                      |                          |             |                            |             |
| 2                      |                          |             |                            |             |
| ...                    |                          |             |                            |             |
| <i>p</i>               |                          |             |                            |             |



**6.3.4 Determination of the trueness and precision characteristics**

**6.3.4.1 General**

For each level *j*, the following trueness and precision characteristics are determined:

- the median of the laboratory means of the measurement results of the reference method and of the alternative method (for the estimation of bias);
- the repeatability standard deviation of the reference method and of the alternative method based on the calculation of Rousseeuw's  $Q_n$  (see Annex Q);
- the reproducibility standard deviation of the reference method and of the alternative method based on Rousseeuw's  $Q_n$  (see Annex Q).

**6.3.4.2 Calculations**

For each level *j*, perform the calculations according to the following steps. In the following text, the subscript *j* is omitted.

**Step 1.** With the measurement results for the reference method (Table 9, columns 2 and 3), prepare Table 10 for each level *j*.

**Table 10 — Measurement results, means and deviations at level *j***  
(for the reference method or for the alternative method)

| 1          | 2           | 3        | 4           | 5                       | 6        |
|------------|-------------|----------|-------------|-------------------------|----------|
| Laboratory | Measurement |          | Mean        | Deviation from the mean |          |
| 1          | $y_{11}$    | $y_{12}$ | $\bar{y}_1$ | $d_{11}$                | $d_{12}$ |
| 2          | $y_{21}$    | $y_{22}$ | $\bar{y}_2$ | $d_{21}$                | $d_{22}$ |
| ...        | ...         | ...      | ...         | ...                     | ...      |
| <i>i</i>   | $y_{i1}$    | $y_{i2}$ | $\bar{y}_i$ | $d_{i1}$                | $d_{i2}$ |
| ...        | ...         | ...      | ...         | ...                     | ...      |
| <i>p</i>   | $y_{p1}$    | $y_{p2}$ | $\bar{y}_p$ | $d_{p1}$                | $d_{p2}$ |

In Table 10,

$$\bar{y}_i = \frac{y_{i1} + y_{i2}}{2}$$

is the mean of the two measurements in laboratory *i* and

$$d_{i1} = y_{i1} - \bar{y}_i = \frac{y_{i1} - y_{i2}}{2}$$

$$d_{i2} = y_{i2} - \bar{y}_i = \frac{y_{i2} - y_{i1}}{2}$$

are the deviations of these two measurements from their mean.

**Step 2.** Use the deviations of Table 10, columns 5 and 6 as a series of  $n = 2p$  values and calculate Rousseeuw's bias corrected scale estimator  $Q_{\text{intra}}$ :

$$Q_{\text{intra}} = c_n Q_n \text{ with } n = 2p$$

as described in Annex Q.

**Step 3.** Calculate Rousseeuw's bias corrected scale estimator  $Q_{\text{inter}}$ :

$$Q_{\text{inter}} = c_n Q_n \text{ with } n = p$$

of the laboratory means of Table 10, column 4 as described in Annex Q.

**Step 4.** Calculate the median  $m$  of the laboratory means of the results of the reference method (Table 10, column 4).

**Step 5.** Calculate the repeatability standard deviation  $s_r$ :

$$s_r = \sqrt{2} Q_{\text{intra}}$$

Calculate the coefficient of variation<sup>1)</sup> of repeatability,  $C_{V,r}$ :

$$C_{V,r} = \frac{s_r}{m}$$

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

Calculate the repeatability limit  $r$ :

$$r = 2,8s_r$$

ISO 16140:2003/Amd 1:2011  
<https://standards.iteh.ai/catalog/standards/sist/9708df8d-89c4-4608-83be-966c6e4d77/c/iso-16140-2003-amd-1-2011>

**Step 6.** Calculate the between laboratories standard deviation  $s_L$ :

$$s_L = \sqrt{Q_{\text{inter}}^2 - Q_{\text{intra}}^2}$$

or 0, if  $Q_{\text{inter}}^2 - Q_{\text{intra}}^2 \leq 0$ .

Calculate the reproducibility standard deviation  $s_R$ :

$$s_R = \sqrt{s_L^2 + s_r^2}$$

Calculate the coefficient of variation<sup>1)</sup> of reproducibility,  $C_{V,R}$ :

$$C_{V,R} = \frac{s_R}{m}$$

Calculate the reproducibility limit,  $R$ :

$$R = 2,8s_R$$

---

1) The predecessor term to "coefficient of variation" used in ISO 16140:2003, "relative standard deviation" is deprecated in ISO 3534-1:2006, 2.38, Note 2. In this amendment, the symbol  $C_V$  is used in place of  $RSD$ .

**Step 7.** Repeat steps 1 to 6 for the alternative method.

After having carried out the calculations of steps 1 to 7 for each level  $j$ , combine the results of the determinations for all  $q$  levels, as in Table 11.

A functional relationship between the repeatability standard deviation or the reproducibility standard deviation and the median level  $m$  may exist. Methods for their establishment can be found in ISO 5725-2.

**Table 11 — Results of the statistical analysis**

| Level | Reference method |                    |                      | Alternative method |                    |                      |
|-------|------------------|--------------------|----------------------|--------------------|--------------------|----------------------|
|       | median           | repeatability s.d. | reproducibility s.d. | median             | repeatability s.d. | reproducibility s.d. |
| 1     | $m_{1,ref}$      | $s_{r1,ref}$       | $s_{R1,ref}$         | $m_{1,alt}$        | $s_{r1,alt}$       | $s_{R1,alt}$         |
| ...   | ...              | ...                | ...                  | ...                | ...                | ...                  |
| $j$   | $m_{j,ref}$      | $s_{rj,ref}$       | $s_{Rj,ref}$         | $m_{j,alt}$        | $s_{rj,alt}$       | $s_{Rj,alt}$         |
| ...   | ...              | ...                | ...                  | ...                | ...                | ...                  |
| $q$   | $m_{q,ref}$      | $s_{rq,ref}$       | $s_{Rq,ref}$         | $m_{q,alt}$        | $s_{rq,alt}$       | $s_{Rq,alt}$         |

**6.3.5 Scrutiny of the measurement results for consistency**

In order to identify measurement results or laboratories that are inconsistent with the other measurement results or laboratories, two graphical consistency techniques are applied, i.e. Mandel's  $h$ - and  $k$ - statistics in a robustified version.

**6.3.5.1** For the reference method, calculate the between-laboratory consistency statistic, Mandel's  $h_{ij}$ , for each of the  $p$  laboratories,  $i = 1 \dots p$ , and for each of the  $q$  levels  $j = 1 \dots q$  by dividing the mean  $\bar{y}_{ij}$  (Table 10, column 4, for each level  $j$ ) minus the median  $m_j$  (Table 11, column 2) for that level  $j$  by Rousseeuw's bias corrected scale estimator  $Q_{inter,j}$  for that level  $j$  (as determined in Step 3),

$$h_{ij} = \frac{\bar{y}_{ij} - m_j}{Q_{inter,j}}$$

Plot the  $p q$  values ( $h_{ij}$ ) sequentially in the order: laboratory 1, levels 1 ...  $q$ , laboratory 2, levels, 1 ...  $q$ , ... to laboratory  $p$ , levels 1 ...  $q$  (see Figure W.1).

Add the indicators for Mandel's  $h$  at the 5 % and 1 % significance levels (Table V.1) as horizontal lines to this plot. In case of between-laboratory consistency, only 5 % or 1 %, respectively, of the values  $h_{ij}$  are expected to lie above these horizontal lines.

**6.3.5.2** For the reference method, calculate the within-laboratory consistency statistic, Mandel's  $k_{ij}$ , for each of the  $p$  laboratories,  $i = 1 \dots p$  and for each of the  $q$  levels  $j = 1 \dots q$  by dividing the absolute difference between the two repeated measurements,  $|y_{ij1} - y_{ij2}|$  (Table 10, columns 2 and 3) by  $\sqrt{2}$  multiplied by the repeatability standard deviation  $s_{rj}$  for that level  $j$ ,

$$k_{ij} = \frac{|y_{ij1} - y_{ij2}|}{\sqrt{2}s_{rj}}$$

Plot the  $p q$  values ( $k_{ij}$ ) sequentially in the order: laboratory 1, levels 1 ...  $q$ , laboratory 2, levels, 1 ...  $q$ , to laboratory  $p$ , levels 1 ...  $q$  (see Figure W.2).