

ISO/TC 147/SC 6

Secretariat: BSI

Voting begins on:
2012-09-06

Voting terminates on:
2012-11-06

Water quality — Sampling — Part 3: Preservation and handling of water samples

Qualité de l'eau — Échantillonnage —

Partie 3: Conservation et la manipulation des échantillons d'eau

ITEH STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sis/391cfe76-62d0-46f0-bd04-d17ffe73a156/iso-5667-3-2012>

Please see the administrative notes on page iii

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 5667-3:2012(E)

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/391cfe76-62d0-46f0-bd04-d17ffe73a156/iso-5667-3-2012>

Copyright notice

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

ISO/CEN PARALLEL PROCESSING

This final draft has been developed within the International Organization for Standardization (ISO), and processed under the **ISO-lead** mode of collaboration as defined in the Vienna Agreement. The final draft was established on the basis of comments received during a parallel enquiry on the draft.

This final draft is hereby submitted to the ISO member bodies and to the CEN member bodies for a parallel two-month approval vote in ISO and formal vote in CEN.

Positive votes shall not be accompanied by comments.

Negative votes shall be accompanied by the relevant technical reasons.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/391cfe76-62d0-46f0-bd04-d17ffe73a156/iso-5667-3-2012>

Contents		Page
Foreword		v
Introduction		vii
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Sampling and chain of custody	2
5	Reagents and materials	2
5.1	Solids.....	3
5.2	Solutions.....	3
5.3	Materials.....	4
6	Containers	4
6.1	Container selection and preparation.....	4
6.2	Filtration on site.....	5
6.3	Filling the container.....	5
7	Sample handling and preservation	5
7.1	Sample handling and preservation for physical and chemical examination.....	5
7.2	Sample handling and preservation for biological examination.....	6
7.3	Sample handling and preservation for radiochemical analysis.....	6
8	Sample transport	7
9	Identification of samples	7
10	Sample reception	8
11	Sample storage	8
Annex A (informative) Techniques for sample preservation		9
Annex B (informative) Container preparation		35
Annex C (informative) Protocol as used in Dutch validation studies		36
Bibliography		38

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5667-3 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 6, *Sampling (general methods)*.

This fourth edition cancels and replaces the third edition (ISO 5667-3:2003), which has been technically revised.

ISO 5667 consists of the following parts, under the general title *Water quality — Sampling*:

- *Part 1: Guidance on the design of sampling programmes and sampling techniques*
- *Part 3: Preservation and handling of water samples*
- *Part 4: Guidance on sampling from lakes, natural and man-made*
- *Part 5: Guidance on sampling of drinking water from treatment works and piped distribution systems*
- *Part 6: Guidance on sampling of rivers and streams*
- *Part 7: Guidance on sampling of water and steam in boiler plants*
- *Part 8: Guidance on the sampling of wet deposition*
- *Part 9: Guidance on sampling from marine waters*
- *Part 10: Guidance on sampling of waste waters*
- *Part 11: Guidance on sampling of groundwaters*
- *Part 12: Guidance on sampling of bottom sediments*
- *Part 13: Guidance on sampling of sludges*
- *Part 14: Guidance on quality assurance of environmental water-sampling and handling*
- *Part 15: Guidance on the preservation and handling of sludge and sediment samples*
- *Part 16: Guidance on biotesting of samples*
- *Part 17: Guidance on sampling of bulk suspended solids*
- *Part 19: Guidance on sampling of marine sediments*

- *Part 20: Guidance on the use of sampling data for decision making — Compliance with thresholds and classification systems*
- *Part 21: Guidance on sampling of drinking water distributed by tankers or means other than distribution pipes*
- *Part 22: Guidance on the design and installation of groundwater monitoring points*
- *Part 23: Guidance on passive sampling in surface waters*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/391cfe76-62d0-46f0-bd04-d17ffe73a156/iso-5667-3-2012>

Introduction

This part of ISO 5667 is intended to be used in conjunction with ISO 5667-1, which deals with the design of sampling programmes and sampling techniques.

Where possible this part of ISO 5667 has been brought into line with current standards. Where new research or validation results have provided new insights, the latest knowledge has been used.

Guidance on validation protocols can be found in ISO Guide 34.^[63]

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/391cfe76-62d0-46f0-bd04-d17ffe73a156/iso-5667-3-2012>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/391cfe76-62d0-46f0-bd04-d17ffe73a156/iso-5667-3-2012>

Water quality — Sampling —

Part 3: Preservation and handling of water samples

NOTICE — This part of ISO 5667 and the analytical International Standards listed in Annex A are complementary. Where no analytical International Standard is applicable, the technique(s) described in Tables A.1 to A.3 take(s) normative status.

When new or revised analytical standards are developed with storage times or preservative techniques differing from those in Tables A.1 to A.3, then the storage times or preservative techniques should be validated and presented to ISO/TC 147/SC 6/WG 3 for incorporation into the next revision of this part of ISO 5667.

1 Scope

This part of ISO 5667 establishes general requirements for sampling, preservation, handling, transport and storage of all water samples including those for biological analyses. It is not applicable to water samples intended for microbiological analyses as specified in ISO 19458, ecotoxicological assays, biological assays, and passive sampling as specified in the scope of ISO 5667-23.

This part of ISO 5667 is particularly appropriate when spot or composite samples cannot be analysed on site and have to be transported to a laboratory for analysis.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667 (all parts), *Water quality — Sampling*

ISO 19458, *Water quality — Sampling for microbiological analysis*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

integrity

property that the parameter(s) of interest, information or content of the sample container has not been altered or lost in an unauthorized manner or subject to loss of representativeness

3.2

sample preservation

any procedure used to stabilize a sample in such a way that the properties under examination are maintained stable from the collection step until preparation for analysis

[ISO 11074:2005,^[29] 4.4.20]

NOTE Different analytes may require several samples from the same source that are stabilized by different procedures.

3.3

sample storage

process, and the result, of keeping a sample available under predefined conditions for a (usually) specified time interval between collection and further treatment of a sample

NOTE 1 Adapted from ISO 11074:2005,^[29] 4.4.22.

NOTE 2 Specified time is the maximum time interval.

3.4

storage time

period of time between filling of the sample container and further treatment of the sample in the laboratory, if stored under predefined conditions

NOTE 1 Sampling finishes as soon as the sample container has been filled with the sample. Storage time ends when the sample is taken by the analyst to start sample preparation prior to analysis.

NOTE 2 Further treatment is, for most analytes, a solvent extraction or acid destruction. The initial steps of sample preparation can be steps complementary to the storage conditions for the maintenance of analyte concentrations.

4 Sampling and chain of custody

If there is a need to take samples, this is done according to a sampling programme. The first step is to design a sampling programme. Guidance on this topic is given in ISO 5667-1.

Depending on the sample type and matrix, the guidelines found in the relevant part(s) of ISO 5667 and ISO 19458 shall be consulted.

The process of preservation and handling of water samples consists of several steps. During this process, the responsibility for the samples might change. To ensure the integrity of the samples, all steps involving the sample shall be documented.

All preparation procedures shall be checked to ensure positive or negative interferences do not occur. As a minimum, this shall include the analysis of blanks (e.g. field blank or sample container) or samples containing known levels of relevant analytes as specified in ISO 5667-14.

5 Reagents and materials

WARNING — Certain preservatives (e.g. acids, alkalis, formaldehyde) need to be used with caution. Sampling personnel should be warned of potential dangers, and appropriate safety procedures should be followed.

The following reagents are used for the sample preservation and shall only be prepared according to individual sampling requirements. All reagents used shall be of at least analytical reagent grade and water shall be of at least ISO 3696, grade 2. Acids referred to in this part of ISO 5667 are commercially available “concentrated” acids.

All reagents shall be labelled with a “shelf-life”. The shelf-life represents the period for which the reagent is suitable for use, if stored correctly. This shelf-life shall not be exceeded. Any reagents that are not completely used by the expiry of the shelf-life date shall be discarded.

NOTE Often the shelf-life of reagents is supplied by the receiving laboratory.

Check reagents periodically, e.g. by field blanks, and discard any reagent found to be unsuitable.

Between on-site visits, reagents shall be stored separately from sample containers and other equipment in a clean, secure cabinet in order to prevent contamination.

Each sample shall be labelled accordingly, after the addition of the preservative. Otherwise, there could be no visible indication as to which samples have been preserved, and which have not.

5.1 Solids

5.1.1 **Sodium thiosulfate pentahydrate**, $\text{Na}_2\text{S}_2\text{O}_3 \cdot 5\text{H}_2\text{O}$, $w(\text{Na}_2\text{S}_2\text{O}_3 \cdot 5\text{H}_2\text{O}) > 99 \%$.

5.1.2 **Ascorbic acid**, $\text{C}_6\text{H}_8\text{O}_6$, $w(\text{C}_6\text{H}_8\text{O}_6) > 99 \%$.

5.1.3 **Sodium hydroxide**, NaOH , $w(\text{NaOH}) > 99 \%$.

5.1.4 **Sodium tetraborate decahydrate**, $\text{Na}_2\text{B}_4\text{O}_7 \cdot 10\text{H}_2\text{O}$, $w(\text{Na}_2\text{B}_4\text{O}_7 \cdot 10\text{H}_2\text{O}) > 99 \%$.

CAUTION Sodium tetraborate decahydrate is known to be a carcinogen, mutagen and reproductive toxin (CMR).

5.1.5 **Hexamethylenetetramine (hexamine, urotropine)**, $\text{C}_6\text{H}_{12}\text{N}_4$, $w(\text{C}_6\text{H}_{12}\text{N}_4) > 99 \%$.

5.1.6 **Potassium iodide**, KI , $w(\text{KI}) > 99 \%$.

5.1.7 **Iodine**, I_2 , $w(\text{I}_2) > 99 \%$.

5.1.8 **Sodium acetate**, $\text{C}_2\text{H}_3\text{NaO}_2$, $w(\text{C}_2\text{H}_3\text{NaO}_2) > 99 \%$.

5.1.9 **Ethylenediamine**, $\text{C}_2\text{H}_8\text{N}_2$, $w(\text{C}_2\text{H}_8\text{N}_2) > 99 \%$.

5.2 Solutions

5.2.1 **Zinc acetate solution** $\text{C}_4\text{H}_6\text{O}_4\text{Zn}$ (10 g/l).

Dissolve 10,0 g of zinc acetate in ~100 ml of water. Dilute to the mark with water. Store the solution in a polypropylene or glass bottle for a maximum period of 1 a.

5.2.2 **Orthophosphoric acid** ($\rho \approx 1,7$ g/ml), H_3PO_4 , $w(\text{H}_3\text{PO}_4) > 85 \%$, $c(\text{H}_3\text{PO}_4) = 15$ mol/l.

5.2.3 **Hydrochloric acid** ($\rho \approx 1,2$ g/ml), HCl , $w(\text{HCl}) > 36 \%$, $c(\text{HCl}) = 12,0$ mol/l.

5.2.4 **Nitric acid** ($\rho \approx 1,42$ g/ml), HNO_3 , $w(\text{HNO}_3) > 65 \%$, $c(\text{HNO}_3) = 15,8$ mol/l.

5.2.5 **Sulfuric acid** ($\rho \approx 1,84$ g/ml), H_2SO_4 (freshly prepared).

Dilute concentrated sulfuric acid (H_2SO_4), $\rho \approx 1,84$ g/ml, $w(\text{H}_2\text{SO}_4) \approx 98 \%$ 1 + 1 by carefully adding the concentrated acid to an equal volume of water and mix.

WARNING — Adding the concentrated acid to the water can give violent reactions because of an exothermic reaction.

5.2.6 **Sodium hydroxide solution** ($\rho \approx 0,40$ g/ml), NaOH .

5.2.7 **Formaldehyde solution** (formalin), CH_2O , $\phi(\text{CH}_2\text{O}) = 37 \%$ to 40% (freshly prepared),

WARNING — Beware of formaldehyde vapours. Do not store large numbers of samples in small work areas.

5.2.8 Disodium salt of ethylenediaminetetraacetic acid (EDTA) ($\rho \approx 0,025$ g/ml), $C_{10}H_{14}N_2Na_2O_8 \cdot 2H_2O$, $w(C_{10}H_{14}N_2Na_2O_8 \cdot 2H_2O) > 99$ %.

Dissolve 25 g EDTA in 1 000 ml of water.

5.2.9 Ethanol C_2H_5OH , $\phi(C_2H_5OH) = 96$ %.

5.2.10 Alkaline Lugol's solution, 100 g potassium iodide (5.1.6), 50 g iodine (5.1.7), and 250 g sodium acetate (5.1.8) in 1 000 ml water to pH 10.

5.2.11 Acidic Lugol's solution, 100 g potassium iodide (5.1.6), 50 g iodine (5.1.7) and 100 ml glacial acetic acid (5.2.17) in 1 000 ml water to pH 2.

5.2.12 Neutralized formaldehyde solution, formaldehyde solution (5.2.7) neutralized with sodium tetraborate (5.1.4) or hexamethylenetetramine (5.1.5). Formalin solution at 100 g/l gives a final solution of $\phi(CH_2O) = 3,7$ % to 4,0 %.

WARNING — Beware of formaldehyde vapours. Do not store large numbers of samples in small work areas.

5.2.13 Ethanol preservative solution.

Ethanol (5.2.9), formaldehyde solution (5.2.7) and glycerol (5.2.18): (100 + 2 + 1 parts by volume, respectively).

5.2.14 Sodium hypochlorite $NaOCl$, $w(NaOCl) = 10$ %. Dissolve 100 g sodium hypochlorite ($NaOCl$) in 1 000 ml of water.

5.2.15 Potassium iodate KIO_3 , $w(KIO_3) = 10$ %. Dissolve 100 g potassium iodate (KIO_3) in 1 000 ml of water.

5.2.16 Methanoic acid (formic acid) CH_2O_2 , $\phi(CH_2O_2) > 98$ %.

5.2.17 Glacial acetic acid $C_2H_4O_2$, $w(C_2H_4O_2) > 99$ %.

5.2.18 Glycerol (glycerin, glycerine) $C_3H_5(OH)_3$.

5.3 Materials

5.3.1 Container and cap, types as specified in Tables A.1 to A.3.

5.3.2 Filter, pore size 0,40 μm to 0,45 μm , unless a different filter size is specified in the analytical International Standard.

6 Containers

6.1 Container selection and preparation

The choice of sample container (5.3.1) is of major importance and ISO 5667-1 provides some guidance on this subject.

Details of the type of container used for the collection and storage of samples are given in Tables A.1 to A.3. The same considerations given to this selection of suitable container material shall also be given to the selection of cap liner materials.

Sample containers shall be made of a material appropriate for preserving the natural properties of both the sample and the expected range of contaminants. Suitable types of containers for each analyte to be measured are given in Tables A.1 to A.3.

NOTE For very low concentrations of metals, containers prescribed can be different from those used for higher concentrations. Details can be found in Table A.1 or in the analytical International Standards.

If the samples are to be frozen, suitable containers, such as polyethylene (PE) or polytetrafluoroethylene (PTFE), shall be used to prevent breakage.

The use of disposables is preferred. Some manufacturers supply containers with a certificate of cleanliness. If such a certificate of cleanliness is supplied, it is not necessary to clean or rinse the containers before use.

6.2 Filtration on site

Filtration on site is required in some cases.

- Groundwaters shall be filtered on site if dissolved metals need to be analysed.
- Waters shall be filtered (5.3.2) on site, if this is required according to Annex A. Unless specified otherwise, a filter pore size 0,40 µm to 0,45 µm shall be used.

If immediate filtration on site is impossible, then the reason and the time between sampling and filtration shall be added to the test report.

6.3 Filling the container

The container (5.3.1) shall be filled completely unless prescribed differently in Tables A.1 to A.3 or the analytical International Standard used. If the samples are to be frozen as part of their preservation, sample containers shall not be completely filled. This is in order to prevent breakage which may arise from expansion of ice during the freezing and thawing process.

If no preservatives are present in the bottle, then prerinsing the bottle may be advisable. Guidance on prerinsing can be found in ISO 5667-14.

7 Sample handling and preservation

7.1 Sample handling and preservation for physical and chemical examination

Waters, particularly fresh waters, waste waters and groundwaters, are susceptible to changes as a result of physical, chemical or biological reactions which may take place between the time of sampling and the commencement of analysis. The nature and rate of these reactions are often such that, if precautions are not taken during sampling, transport and storage (for specific analytes), the concentrations determined are different to those existing at the time of sampling.

The extent of these changes is dependent on the chemical and biological nature of the sample, its temperature, its exposure to light, the type of the container in which it is placed, the time between sampling and analysis, and the conditions to which it is subjected, e.g. agitation during transport. Further specific causes of variation are listed in a) to f).

- a) The presence of bacteria, algae and other organisms can consume certain constituents of the samples. These organisms can also modify the nature of the constituents to produce new constituents. This biological activity affects, for example, the concentrations of dissolved oxygen, carbon dioxide, compounds of nitrogen, phosphorus and, sometimes, silicon.
- b) Certain compounds can be oxidized either by dissolved oxygen present in the samples, or by atmospheric oxygen [e.g. organic compounds, Fe(II) and sulfides].