
**Microbiology of food and animal
feed — Preparation of test samples,
initial suspension and decimal
dilutions for microbiological
examination —**

Part 6:
**Specific rules for the preparation
of samples taken at the primary
production stage**

[https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-](https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-2013-03-01)

*Microbiologie des aliments — Préparation des échantillons, de
la suspension mère et des dilutions décimales en vue de l'examen
microbiologique —*

*Partie 6: Règles spécifiques pour la préparation des échantillons
prélevés au stade de production primaire*



iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 6887-6:2013

<https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-25746956ba8b/iso-6887-6-2013>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Diluents and disinfectants	2
5.1 Diluents for special purposes	2
5.2 Disinfectants for use during laboratory examination	3
6 Apparatus and glassware	3
7 Types of sample that can be sent to the laboratory	3
8 Preparation of samples	3
8.1 General	3
8.2 Storage	4
9 Specific procedures	4
9.1 Procedures carried out on samples taken at farm	4
9.2 Procedures carried out on samples taken from slaughterhouse	6
9.3 Procedures for samples taken from poultry at hatchery or when transporting from hatchery to farm	7
10 Further decimal dilutions	8
Bibliography	9

[ISO 6887-6:2013](https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-25746956ba8b/iso-6887-6-2013)

<https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-25746956ba8b/iso-6887-6-2013>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 6887-6 was prepared by the European Committee for Standardization (CEN) Technical Committee CEN/TC 275, *Food analysis — Horizontal methods*, in collaboration with Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 9, *Microbiology*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

ISO 6887 consists of the following parts, under the general title *Microbiology of food and animal feed*¹⁾ — *Preparation of test samples, initial suspension and decimal dilutions for microbiological examination*:

- Part 1: *General rules for the preparation of the initial suspension and decimal dilutions*
- Part 2: *Specific rules for the preparation of meat and meat products*
- Part 3: *Specific rules for the preparation of fish and fishery products*
- Part 4: *Specific rules for the preparation of products other than milk and milk products, meat and meat products, and fish and fishery products*
- Part 5: *Specific rules for the preparation of milk and milk products*
- Part 6: *Specific rules for the preparation of samples taken at the primary production stage*

1) It is intended that, upon revision, the main element of Parts 2 to 5 will be aligned with the main element of the title of Part 6 (i.e. “*Microbiology of food and animal feed*”).

Microbiology of food and animal feed — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination —

Part 6:

Specific rules for the preparation of samples taken at the primary production stage

1 Scope

This part of ISO 6887 specifies rules for the preparation of samples taken at all stages from the farm to the slaughterhouse and their suspension for microbiological examination when the samples require different preparation from the methods described in ISO 6887-1. ISO 6887-1 defines the general rules for the preparation of the initial suspension and decimal dilutions for microbiological examination.

This part of ISO 6887 excludes the preparation of samples for both enumeration and detection test methods where preparation details are specified in the relevant International Standards.

This part of ISO 6887 is applicable to various samples taken from the hatchery, the farm, from the vehicle or the animals during transportation, or from animals or their carcasses in the slaughterhouse, to indicate the microbiological status of the animals in relation to zoonotic agents. This part of ISO 6887 does not apply to samples taken to assess the hygiene of meat. These are covered by ISO 6887-2.

This part of ISO 6887 does not consider samples taken from the aquatic environment (marine or freshwater) at the primary production stage. These are covered by ISO 6887-3.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6887-1, *Microbiology of food and animal feeding stuffs — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination — Part 1: General rules for the preparation of the initial suspension and decimal dilutions*¹⁾

ISO 7218:2007, *Microbiology of food and animal feeding stuffs — General requirements and guidance for microbiological examinations*

ISO 13307, *Microbiology of food and animal feed — Primary production stage — Sampling techniques*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

laboratory sample

sample prepared for sending to the laboratory and intended for inspection or testing

3.2

test portion

measured (volume or mass) representative sample taken from the *laboratory sample* (3.1) for use in the preparation of the *initial suspension* (3.4)

3.3

pooled sample

composite sample taken from a number of different animals or composites of separate environmental samples

3.4

initial suspension

primary dilution

suspension, solution or emulsion obtained after a weighed or measured quantity of the product under examination (or of a test sample prepared from the product) has been mixed with, normally, a nine-fold quantity of diluent, allowing large particles, if present, to settle

Note 1 to entry: Sometimes higher or lower initial dilutions are prepared, for example 1 in 5 or 1 in 100 ($V_1 \rightarrow V_2$).

3.5

further decimal dilutions

suspension or solution obtained by mixing a measured volume of the *initial suspension* (3.4) with a nine-fold volume of diluent and by repeating this operation with each dilution prepared in this way, until a decimal dilution series, suitable for the inoculation of culture media, is obtained

iTeh STANDARD PREVIEW
(standards.iteh.ai)

4 Principle

An initial suspension (3.4) is prepared to obtain as uniform a distribution as possible of the microorganisms contained in the test sample, without reducing their viability.

A suspension for pre-enrichment or enrichment is prepared in the same way, using the medium recommended by the method of analysis concerned, except in the special cases mentioned in each product subclause (section) of this part of ISO 6887.

If necessary, further dilutions (3.5) are prepared in order to reduce the number of microorganisms per unit volume to allow, after incubation, observation of any growth (in the case of liquid media) or enumeration of colonies (in or on agar plates), as stated in each specific standard.

In order to restrict, if required, the range of enumeration to a given interval, or if high numbers of microorganisms are foreseen, it is possible to inoculate only the appropriate dilutions (at least two successive dilutions) according to the calculation described in ISO 7218.

5 Diluents and disinfectants

Diluents for general use are described in ISO 6887-1.

5.1 Diluents for special purposes

5.1.1 Neutralizing agents

Neutralizing liquid, prepared in accordance with ISO 13307, is used if needed at 10 % volume fraction in diluent. Neutralizer is normally added when the sample is taken, before transportation to the laboratory.

NOTE If a high level of formalin is expected in the samples, L-histidine (0.9 %) can be also added to the pre-enrichment broth.

5.2 Disinfectants for use during laboratory examination

Disinfectants are given in ISO 7218:2007, 6.2.4.

6 Apparatus and glassware

Usual microbiological laboratory equipment for general use (see ISO 6887-1 and ISO 7218) and, in particular, the following:

6.1 Homogenizers

6.1.1 Peristaltic homogenizer (stomacher)

6.1.2 Rotary homogenizer (blender)

6.1.3 Vibrational mixer (pulsifier)

6.2 Sterile hammer or plastic mallet

6.3 Sterile sand, pestle and mortar

6.4 Sterile forceps, scissors, scalpels, spatulas, spoons

6.5 Sterile flasks or screw-cap bottles of appropriate capacities

6.6 Sterile total delivery graduated pipettes, pipette tips

[https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-](https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-25746956ba8b/iso-6887-6-2013)

[25746956ba8b/iso-6887-6-2013](https://standards.iteh.ai/catalog/standards/sist/2441d49f-79d2-46b0-8cb6-25746956ba8b/iso-6887-6-2013)

7 Types of sample that can be sent to the laboratory

The samples shall be taken and transported in accordance with ISO 7218 and ISO 13307.

The following list includes examples only. More detailed information on samples needed in different situations is given in [Clause 9](#):

- samples taken at the farm:
 - from the environment (e.g. swabs, litter, faeces, dust, water);
 - from animals (e.g. swabs);
- samples taken at slaughter (e.g. rectal or caecal content, mesenteric lymph nodes);
- samples taken at the hatchery (e.g. hatcher basket liners, broken egg-shells);
- samples taken on vehicles, modules and crates for animal transport (e.g. swabs).

8 Preparation of samples

8.1 General

All preparations and manipulations should be carried out using good aseptic techniques and with sterile equipment to prevent microbial contamination of samples from all external sources (see ISO 7218).

If the sample is transferred to another container and the whole laboratory sample is used, ensure that all of the sample is transferred (e.g. when transferring bootsocks from the original container to a new one, take care that no material from the bootsocks is left in the first container).

Indicate in the report which procedure has been used for analysis if it is different from the procedure described in this part of ISO 6887.

8.2 Storage

Samples shall be stored in conditions suitable for the optimal survival of the target microorganism(s) in accordance with ISO 7218.

9 Specific procedures

9.1 Procedures carried out on samples taken at farm

9.1.1 Samples from the environment or live animals

9.1.1.1 Fabric swabs

If possible, add the appropriate quantity/amount of medium directly to the sample in the transport container. The particular quantity to be added depends on the dimensions of the swab and the purpose of the test. Ensure that all parts of the swabs are submerged.

Take account of the effect of the quantity/amount of medium added on the detection limit of the test (see the note to 3.4).

For enumeration, add sufficient quantity/amount of medium to saturate the swab fully while providing the minimum amount of free liquid needed for enumeration, and for detection, ensure a sufficient amount of free liquid is present.

For example for enumeration from a 10 x 10 cm sponge swab add 100 ml of diluent. Squeeze the swab several times (e.g. by hand if the container is a bag) so that the microorganisms are released into the suspension, then shake well.

For detection, the swab is included in the culture.

9.1.1.2 Stick swabs

Transfer the swab to a suitable container. Break (or cut if necessary) the stick using sterile tools, add the appropriate quantity/amount of medium and mix. If the swab is already in a tube or other suitable container, add the medium to the same container, unless the stick swab container includes solid transport medium.

Where appropriate, stick swabs can be pooled, adding a suitable volume of medium.

9.1.1.3 Bootsock swabs, drag swabs, rope swabs

Where possible, add the appropriate quantity/amount of medium directly into the transport container.

Ensure that all parts of the swabs are submerged.

For example for *Salmonella* testing, add at least 225 ml of the appropriate diluent (see Annex D of ISO 6579:2002) per pair of bootsock swabs.

9.1.1.4 Moore's drain swab (tampon swab)

Handle these like bootsocks, (see 9.1.1.3), but because of the accumulation of high numbers of organisms over time, a dilution of 1/20 may be advantageous.

9.1.1.5 Litter samples and naturally pooled faeces

It is important to homogenize the laboratory sample by mixing dry material or by adding the sample to an equal quantity/amount of diluent, which may be in the same container used for transport. Mix to make a slurry. Allow to stand for 10 min to 15 min and mix again. Transfer 50 ml of the suspension (containing 25 g of initial sample) to an appropriate volume of diluent according to the specific standard for the target microorganism.

9.1.1.6 Faecal samples

Individual sample: take a portion or the whole dropping, if small. Mix it gently and add the appropriate quantity/amount of medium according to the specific ISO procedure being followed.

For samples which are difficult to mix, refer to [9.1.1.5](#).

To pool samples: mix each sample as thoroughly as possible, take an equal quantity of each sample, add the appropriate quantity/amount of medium according to the specific standard for the target microorganism and mix again thoroughly.

It is preferable to pool no more than 20 animal faeces.

9.1.1.7 Dust

This sample is usually relevant for *Salmonella* or other robust organisms, but not, for instance, for *Campylobacter*.

At least 10 g of dust shall be tested. A 1:20 ratio of sample to pre-enrichment medium is advantageous for detection of *Salmonella* in very dry absorbent samples such as dust. Large dust samples may be prepared in the laboratory by mixing 1:4 with diluent, and then taking a subsample, which is subsequently diluted 1:5 for pre-enrichment, making sure that at least 10 g of the original sample is included. Samples of up to 25 g may be cultured without subsampling.

To reduce handling of dust in the laboratory and the consequent risk of cross-contamination, it is advisable to collect the dust to be analysed in sufficiently large bags or vessels so that only the required quantity/amount of medium is added in the laboratory. Free dust should be handled in a laminar flow cabinet.

9.1.1.8 Water

Small volumes of water (such as 100 ml) can be added to an equal volume of double strength medium for culture. Put the water in a container of suitable size, add medium, according to the ratio stated in the specific international procedure and mix.

For larger volumes of water from water systems: filter the sample through a membrane filter of pore size 0,45 µm, (for *Campylobacter*, 0,20 µm), then culture the filter. The larger the volume of water filtered, the more sensitive the detection.

For further details, see ISO 8199.

9.1.2 Animals from the farm

9.1.2.1 General

A separate autopsy room should preferably be used or, if this is not possible, a cabinet or dedicated areas should be used.

9.1.2.2 Birds

Isolating bacteria from bird viscera needs special attention with regard to the risk of cross-contamination and the need for post-mortem dissection according to organs to be analysed. Post mortem dissection