
**Water quality — Enumeration of
Clostridium perfringens — Method
using membrane filtration**

*Qualité de l'eau — Dénombrement de Clostridium perfringens —
Méthode de filtration sur membrane*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 14189:2013](https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013)

[https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-
57638cb653f4/iso-14189-2013](https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013)



iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 14189:2013

<https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Principle.....	2
5 Apparatus and glassware.....	2
6 Culture media and reagents.....	3
6.1 Basic materials.....	3
6.2 Culture media.....	3
7 Sampling.....	3
8 Procedure.....	3
8.1 Transport and storage of the sample.....	3
8.2 Heat pre-treatment to select spores.....	3
8.3 Sample dilution.....	3
8.4 Filtration.....	4
8.5 Incubation and examination.....	4
8.6 Confirmation of <i>Clostridium perfringens</i>	4
9 Expression of results.....	5
10 Test report.....	5
11 Quality assurance.....	5
Annex A (normative) Composition and preparation of culture media and reagents.....	7
Annex B (informative) Performance data.....	9
Bibliography.....	12

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 147, *Water quality*, Subcommittee SC 4, *Microbiological methods*.

[ISO 14189:2013](https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013)

<https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013>

Introduction

Clostridium perfringens is widely recognized as a valuable indicator for faecal pollution. Within the intestinal tract of animals and man, these Gram-positive bacteria form spores which are resistant to heating compared with vegetative cells. *C. perfringens* in the intestine exists both as spores and vegetative cells, spores are also found in environmental samples. The spores of *C. perfringens* survive in water for months, much longer than vegetative faecal indicator bacteria and consequently their presence may indicate remote or intermittent faecal pollution. Monitoring of *C. perfringens* has proven useful for the assessment of the quality of water resources and to check the stages of water treatment to evaluate the treatment-works performance. The spores are not always inactivated by routine disinfection procedures (e.g. chlorination).

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 14189:2013](https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013)

<https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 14189:2013

<https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013>

Water quality — Enumeration of *Clostridium perfringens* — Method using membrane filtration

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

1 Scope

This International Standard specifies a method for the enumeration of vegetative cells and spores of *Clostridium perfringens* by the membrane filtration method in samples of water intended for human consumption. However, the method can be applied to all types of water samples provided they do not contain particulate or colloidal matter that interferes with filtration.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8199, *Water quality — General guidance on the enumeration of micro-organisms by culture*
<https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-17ad7361fe1992065>

ISO/TS 11133-1, *Microbiology of food and animal feeding stuffs — Guidelines on preparation and production of culture media — Part 1: General guidelines on quality assurance for the preparation of culture media in the laboratory*

ISO 19458, *Water quality — Sampling for microbiological analysis*

ISO/IEC Guide 2:2004, *Standardization and related activities — General vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 2 and the following apply:

3.1

presumptive *Clostridium perfringens*

bacteria which produce all shades of black or grey to yellow brown colonies on tryptose-sulfite-cycloserine agar, even if the colour is faint, after anaerobic incubation at $(44 \pm 1)^\circ\text{C}$ for (21 ± 3) h

Note 1 to entry: Unlike colonies growing directly on the agar medium, colonies on the membrane do not always display a distinct blackening, so faint colonies are included in the count.

3.2

confirmed *Clostridium perfringens*

bacteria that produce characteristic colonies on tryptose-sulfite-cycloserine agar and possess the enzyme acid phosphatase

4 Principle

A measured volume of the sample, or a dilution of it, is filtered through a membrane with a pore size of 0,45 µm sufficient to retain spores of clostridia. The membrane is incubated on a selective/differential agar (tryptose-sulfite-cycloserine agar) anaerobically at $(44 \pm 1) ^\circ\text{C}$ for (21 ± 3) h. *C. perfringens* usually produce black or grey to yellow brown colonies as a result of the reduction of sulfite to sulfide which reacts with a ferric salt in the medium. Characteristic colonies are counted and confirmatory tests are carried out. The result is calculated as the colony count per sample volume. If a count of spores alone is required the sample is first pre-treated at $(60 \pm 2) ^\circ\text{C}$ to inactivate vegetative bacteria.

NOTE 1 The medium contains cycloserine as a selective agent to inhibit *Bacillus* species.

NOTE 2 Incubation at 44 °C increases the selectivity of the test for *C. perfringens*.

5 Apparatus and glassware

Except for disposable glassware or plastics ware which is delivered sterile, sterilize glassware as specified in ISO 8199.

Usual microbiological equipment and particularly:

5.1 Membrane filtration apparatus, as specified in ISO 8199.

5.2 Sterile filter funnels

Use funnels either delivered sterile or sterilized as specified in ISO 8199. Alternatively flaming of funnels made of metal prior to their use is acceptable.

NOTE For this method it is insufficient to disinfect funnels by boiling.

5.3 Sterile membrane filters, nominal pore size 0,45 µm.

The quality of membrane filters may vary from brand to brand or even from batch to batch. It is therefore advisable to check the quality on a regular basis.

5.4 Incubators, capable of being maintained at $(36 \pm 2) ^\circ\text{C}$ and at $(44 \pm 1) ^\circ\text{C}$.

5.5 Water bath (optional), capable of being maintained at $(60 \pm 2) ^\circ\text{C}$ equipped with a means of circulating the water.

5.6 Autoclave, capable of being maintained at $(121 \pm 3) ^\circ\text{C}$.

5.7 Sterile forceps

5.8 Anaerobic jars, or similar equipment.

5.9 Anaerobic gas generating system, to generate an atmosphere of approximately 90 % hydrogen and 10 % carbon dioxide.

6 Culture media and reagents

6.1 Basic materials

For uniformity of results, in the preparation of media, use constituents of uniform quality and chemicals of recognized analytical grade. For preparation of the media use glass-distilled water or deionized water of equivalent purity, as specified for water grade 3 in ISO 3696^[1].

Alternatively, use commercially available dehydrated complete medium and reagents prepared and used according to the manufacturer's instructions.

Other grades of chemicals may be used provided they can be shown to lead to the same results.

6.2 Culture media

See [Annex A](#).

6.2.1 Tryptose sulfite cycloserine agar (TSC agar), References [6][11][12]

See [A.1](#).

6.2.2 Blood agar or Columbia agar base or another suitable nutrient-rich agar

See [A.2](#).

6.2.3 Acid phosphatase reagent

See [A.3](#).

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 14189:2013

<https://standards.iteh.ai/catalog/standards/sist/a29c62e8-a222-4951-a332-57638cb653f4/iso-14189-2013>

7 Sampling

Carry out sampling as specified in ISO 19458.

8 Procedure

8.1 Transport and storage of the sample

Start examination as soon as possible after the collection of the sample preferably within the same working day. Samples should be cooled during transport ideally at $(5 \pm 3)^\circ\text{C}$. The recommended maximum sample storage time including transport is for vegetative bacteria 12 h and for spores 24 h. The sample storage time including transport shall not exceed 18 h for vegetative bacteria and 72 h for spores.

8.2 Heat pre-treatment to select spores

If it is the intention to count only spores, mix the sample well and then heat it to $(60 \pm 2)^\circ\text{C}$ in a water bath for (15 ± 1) min. The volume heated should be greater than the volume to be analysed. The temperature should be monitored by placing an appropriate thermometer in a reference bottle of the same size as the sample bottle and containing the same volume of water at the same initial temperature as the sample being treated. The time taken to reach $(60 \pm 2)^\circ\text{C}$ shall not exceed 15 min and can be minimized by ensuring the water in the water bath is circulated to maximize heat exchange.

8.3 Sample dilution

A test volume of sample or dilution of it - after heat treatment if required - should be chosen to yield, if possible, between 10 and 80 colonies on a membrane 47 mm to 50 mm in diameter. Test volumes or dilutions should be prepared as described in ISO 8199.

8.4 Filtration

For a general description of the membrane filtration technique see ISO 8199.

Filter a measured volume of water. The volume should be appropriate to the water being examined. For water intended for human consumption, it is usual to filter a volume of 100 ml. Record the volume filtered.

After filtration, place the membrane grid face upwards on a TSC agar plate ensuring that no air bubbles are trapped under the filter.

NOTE Alternatively, a thin layer (about 5 ml to 10 ml related to a petri dish of 90 mm diameter) molten TSC agar (equilibrated in a water bath at (45 ± 1) °C) as an overlay on the filter can be used. Allow to solidify before anaerobic incubation. This procedure may enhance the blackening of the colonies. It is not necessary to add cycloserine in the TSC agar for the overlay. However, obtaining pure cultures for the confirmation test may be more laborious.

As the spores of *C. perfringens* are more heat resistant, sterile funnels shall be used (5.2). Placing in a boiling water bath between samples may not be sufficient to inactivate spores. Flaming of metal funnels is considered acceptable. For different volumes of the same sample, the funnel may be reused provided the smallest volumes of sample are filtered first.

8.5 Incubation and examination

The time between placing the membrane on the TSC agar and starting incubation should be as short as possible and shall not exceed 1 h.

Incubate the plates with the filters, anaerobically at (44 ± 1) °C for (21 ± 3) h inverted to avoid interference with condensing water.

After incubation, enumerate the presumptive *C. perfringens* by counting all colonies which show black or grey to yellow brown staining, even if the colour is faint, of the TSC medium when viewed from either above or below the membrane filter.

Since the black colour of the colonies rapidly fades and finally disappear, the plates have to be counted within 30 min after completion of the anaerobic incubation. If more anaerobic jars are used, the plates should be checked jar by jar or in portions if the incubation was performed in an anaerobic incubator.

8.6 Confirmation of *Clostridium perfringens*

8.6.1 General

It is recommended for membrane filtration methods for water analysis that for counts of 1 to 10 all colonies are subject to confirmation, and for counts above 10, at least 10 colonies taken at random are subject to confirmation.

For confirmation subculture all colonies for counts of 1 to 10 and at least 10 colonies for counts above 10 taken randomly onto blood agar plates. When this is impracticable, all typical colonies shall be examined from a sub-area of the membrane.

If blood agar is not available, Columbia agar base or another nutrient-rich agar (e.g. tryptone soya agar) may be used for subculture.

Incubate anaerobically in an incubator at (36 ± 2) °C for (21 ± 3) h.

8.6.2 Acid phosphatase test

Colonies grown anaerobically on blood or nutrient agar plates are spread on filter paper and 2 to 3 drops of the acid phosphatase reagent are placed onto the colonies. A purplish colour developed within 3 min to 4 min is considered as a positive reaction.

8.6.3 Interpretation

C. perfringens produces black or grey to yellow brown colonies on TSC agar, even if the colour is faint, and possesses acid phosphatase.

9 Expression of results

From the numbers of total and confirmed colonies, calculate the numbers of presumptive *C. perfringens* and *C. perfringens* and number of spores, if applicable, present in 100 ml of the filtered volume in accordance with ISO 8199.

Where required, the variability of the test results should be evaluated according to ISO 29201[3].

10 Test report

The test report shall contain at least the following information:

- a) the test method used, together with a reference to this International Standard (ISO 14189:2013);
- b) all details necessary for the complete identification of the sample;
- c) the number of colonies presumptive *C. perfringens* (optional, depending on the purpose of the investigation);
- d) the number of colonies confirmed as *C. perfringens*;
- e) information, if the result represents the total number of *C. perfringens* (vegetative cells and spores) or spores only;
- f) any particular occurrence(s) observed during the analysis and any operation(s) not specified in this method, which may have affected the results.

11 Quality assurance

The laboratory shall have a clearly defined quality control system to ensure that the apparatus, reagents and techniques are suitable for the test. The use of positive controls, negative controls and blanks is part of the test.

Media quality control is described in ISO/TS 11133-1. For quantitative process quality control a suspension of *C. perfringens* is used for this purpose. Select the volume filtered to contain between 10 cfu to 80 cfu and treat the control like a sample. Compare recovery to that on a non-selective agar such as blood agar. Alternatively to the suspension of *C. perfringens* use reference materials.

NOTE Since the medium should be used as fresh as possible it is acceptable to perform media and process control for samples simultaneously in parallel with the analysis of the samples.

Include a blank control in each batch of analyses. Filter 100 ml of sterile water or another appropriate diluent according to ISO 8199 and continue to treat it like a sample but without pasteurization. No colonies should be visible after incubation.

Include an appropriate control for correct anaerobic conditions (e.g. anaerobic indicator strip) whenever anaerobic incubation is performed (anaerobic jar or anaerobic incubator).

For the confirmation step performed by acid phosphatase test include known positive and negative control strains.

At least one of the *C. perfringens* strains given in [Table 1](#) shall be used as a positive control for media control and the confirmation test.