

ISO/TC 215

Secretariat: ANSI

ISO voting begins on:  
2009-10-22

ISO voting terminates on:  
2010-03-22

---

---

## Health informatics — Personal health device communication —

### Part 10404: Device specialization — Pulse oximeter

*Informatique de santé — Communication entre dispositifs médicaux sur  
le site des soins —*

*Partie 10404: Spécialisation des dispositifs — Oxymètre de pouls*

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION. SEE ADDITIONAL INFORMATION ABOUT PATENTS ON THE INSIDE BACK COVER. BECAUSE IT IS AN UNAPPROVED DRAFT, THIS DOCUMENT SHALL NOT BE USED FOR ANY CONFORMANCE/COMPLIANCE PURPOSES.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

Please see the administrative notes on page iii  
and on the inside back cover



Reference number  
ISO/IEEE FDIS 11073-10404:2009(E)

© ISO 2009  
© IEEE 2009

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor IEEE accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and IEEE members. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat or IEEE at the address given below.

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/0737aaec-d8ee-4fc4-9c4c-33a2d1afe0d6/iso-ieee-11073-10404-2010>

**Copyright notice**

This ISO/IEEE document is a Final Draft International Standard and is copyright-protected by ISO and IEEE. Except as permitted under the applicable laws of the user's country, neither this ISO/IEEE draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to IEEE.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Institute of Electrical and Electronics Engineers, Inc.  
3 Park Avenue, New York • NY 10016-5997, USA  
E-mail [stds.ipr@ieee.org](mailto:stds.ipr@ieee.org)  
Web [www.ieee.org](http://www.ieee.org)

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

# FAST-TRACK PROCEDURE

This document is submitted under the fast-track procedure in accordance with the Partner Standards Development Organization cooperation agreement between ISO and IEEE, as approved by Council Resolution 49/2007, and is submitted to all ISO member bodies for voting within 5 months.

Positive votes shall not be accompanied by comments.

Negative votes shall be accompanied by the relevant technical reasons.

In accordance with the provisions of Council Resolution 15/1993, this document is **circulated in the English language only**.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/0737aacc-d8ee-4fc4-9c4c-33a2d1afe0d6/iso-ieee-11073-10404-2010>

<b>Contents</b>	<b>Page</b>
Foreword.....	vi
Introduction.....	viii
1. Overview.....	1
1.1 Scope.....	1
1.2 Purpose.....	1
1.3 Context.....	2
2. Normative references.....	2
3. Definitions, acronyms, and abbreviations.....	2
3.1 Definitions.....	2
3.2 Acronyms and abbreviations.....	3
4. Introduction to ISO/IEEE 11073 personal health devices.....	3
4.1 General.....	3
4.2 Introduction to IEEE 11073-20601 modeling constructs.....	4
5. Pulse oximeter device concepts and modalities.....	4
5.1 General.....	4
5.2 Device types.....	5
5.3 General concepts.....	5
5.4 Collected data.....	6
5.5 Derived data.....	8
5.6 Stored data.....	8
5.7 Device configurations.....	8
6. Pulse oximeter DIM.....	9
6.1 Overview.....	9
6.2 Class extensions.....	9
6.3 Object instance diagram.....	9
6.4 Types of configuration.....	10
6.5 MDS object.....	11
6.6 Numeric objects.....	14
6.7 Real-time sample array (RT-SA) objects.....	24
6.8 Enumeration objects.....	25
6.9 PM-store objects.....	29
6.10 Scanner objects.....	33
6.11 Class extension objects.....	37
6.12 Pulse oximeter information model extensibility rules.....	37

7. Pulse oximeter service model .....	37
7.1 General.....	37
7.2 Object access services.....	37
7.3 Object access EVENT REPORT services.....	40
8. Pulse oximeter communication model.....	41
8.1 Overview.....	41
8.2 Communications characteristics.....	41
8.3 Association procedure.....	42
8.4 Configuring procedure.....	43
8.5 Operating procedure.....	45
8.6 Time synchronization.....	46
9. Test associations .....	46
9.1 Behavior with standard configuration.....	46
9.2 Behavior with extended configurations.....	46
10. Conformance.....	46
10.1 Applicability.....	46
10.2 Conformance specification.....	47
10.3 Levels of conformance.....	47
10.4 Implementation conformance statements (ICSS).....	48
Annex A (informative) Bibliography.....	52
Annex B (normative) Additional ASN.1 definitions.....	53
Annex C (normative) Allocation of identifiers.....	55
Annex D (informative) Message sequence examples.....	57
Annex E (informative) PDU examples.....	59

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is called to the possibility that implementation of this standard may require the use of subject matter covered by patent rights. By publication of this standard, no position is taken with respect to the existence or validity of any patent rights in connection therewith. ISO/IEEE is not responsible for identifying essential patents or patent claims for which a license may be required, for conducting inquiries into the legal validity or scope of patents or patent claims or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance or a Patent Statement and Licensing Declaration Form, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from ISO or the IEEE Standards Association.

ISO/IEEE 11073-10404 was prepared by the 11073 Committee of the Engineering in Medicine and Biology Society of the IEEE (as IEEE Std 11073-10404-2008). It was adopted by Technical Committee ISO/TC 215, *Health informatics*, in parallel with its approval by the ISO member bodies, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE. Both parties are responsible for the maintenance of this document.

ISO/IEEE 11073 consists of the following parts, under the general title *Health informatics — Personal health device communication* (text in parentheses gives a variant of subtitle):

- *Part 10101: (Point-of-care medical device communication) Nomenclature*
- *Part 10201: Domain information model*
- *Part 10404: Device specialization — Pulse oximeter*
- *Part 10407: Device specialization — Blood pressure monitor*

- Part 10408: (Point-of-care medical device communication) Device specialization — Thermometer
- Part 10415: (Point-of-care medical device communication) Device specialization — Weighing scale
- Part 10417: Device specialization — Glucose meter
- Part 10471: (Point-of-care medical device communication) Device specialization — Independent living activity hub
- Part 20101: (Point-of-care medical device communication) Application profiles — Base standard
- Part 20601: (Point-of-care medical device communication) Application profile — Optimized exchange protocol
- Part 30200: (Point-of-care medical device communication) Transport profile — Cable connected
- Part 30300: (Point-of-care medical device communication) Transport profile — Infrared wireless

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/0737a98c-d8ee-4fc4-9c4c-33a2d1afe0d6/iso-ieee-11073-10404-2009>

## Introduction

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. This standard uses the optimized framework created in IEEE Std 11073-20601™-2008<sup>a</sup> and describes a specific, interoperable communication approach for pulse oximeters. These standards align with, and draw upon, the existing clinically focused standards to provide support for communication of data from clinical or personal health devices.

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/0737aacc-d8ee-4fc4-9c4c-33a2d1afe0d6/iso-ieee-11073-10404-2010>

---

<sup>a</sup> For information on references, see Clause 2.



# Health informatics—Personal health device communication—

## Part 10404: Device specialization—Pulse oximeter

**IMPORTANT NOTICE:** This standard is not intended to ensure safety, security, health, or environmental protection in all circumstances. Implementers of the standard are responsible for determining appropriate safety, security, environmental, and health practices or regulatory requirements.

This IEEE document is made available for use subject to important notices and legal disclaimers. These notices and disclaimers appear in all publications containing this document and may be found under the heading “Important Notice” or “Important Notices and Disclaimers Concerning IEEE Documents.” They can also be obtained on request from IEEE or viewed at <http://standards.ieee.org/IPR/disclaimers.html>.

### 1. Overview

#### 1.1 Scope

Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of communication between personal telehealth pulse oximeter devices and compute engines (e.g., cell phones, personal computers, personal health appliances, set top boxes) in a manner that enables plug-and-play (PnP) interoperability. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for personal telehealth pulse oximeters.

#### 1.2 Purpose

This standard addresses a need for an openly defined, independent standard for controlling information exchange to and from personal health devices and compute engines (e.g., cell phones, personal computers, personal health appliances, set top boxes). Interoperability is key to growing the potential market for these devices and enabling people to be better informed participants in the management of their health.

## 1.3 Context

See IEEE Std 11073-20601-2008<sup>1</sup> for an overview of the environment within which this standard is written.

This standard, IEEE Std 11073-10404-2008, defines the device specialization for the pulse oximeter, being a specific agent type, and provides a description of the device concepts, its capabilities, and its implementation according to this standard.

This standard is based on IEEE Std 11073-20601-2008, which in turn draws information from both ISO/IEEE 11073-10201:2004 [B3]<sup>2</sup> and ISO/IEEE 11073-20101:2004 [B4]. The medical device encoding rules (MDER) used within this standard are fully described in IEEE Std 11073-20601-2008.

This standard reproduces relevant portions of the nomenclature found in ISO/IEEE 11073-10101:2004 [B2] and adds new nomenclature codes for the purposes of this standard. Between this standard and IEEE Std 11073-20601-2008, all required nomenclature codes for implementation are documented.

NOTE—In this standard, ISO/IEEE P11073-104zz is used to refer to the collection of device specialization standards that utilize IEEE Std 11073-20601-2008, where zz can be any number from 01 to 99, inclusive.<sup>3</sup>

## 2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used, so that each referenced document is cited in text and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 11073-20601-2008, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Profile.<sup>4</sup>

See Annex A for all informative material referenced by this standard.

## 3. Definitions, acronyms, and abbreviations

### 3.1 Definitions

For the purposes of this standard, the following terms and definitions apply. *The Authoritative Dictionary of IEEE Standards* [B1] should be referenced for terms not defined in this clause.

**3.1.1 agent:** A node that collects and transmits personal health data to an associated manager.

**3.1.2 class:** In object-oriented modeling, a class describes the attributes, methods, and events that objects instantiated from the class utilize.

**3.1.3 compute engine:** *See: manager.*

**3.1.4 device:** A physical apparatus implementing either an agent or manager role.

**3.1.5 handle:** An unsigned 16-bit number that is locally unique and identifies one of the object instances within an agent.

<sup>1</sup> Information on references can be found in Clause 2.

<sup>2</sup> The numbers in brackets correspond to the numbers in the bibliography in Annex A.

<sup>3</sup> Notes in text, tables, and figures are given for information only and do not contain requirements needed to implement the standard.

<sup>4</sup> IEEE publications are available from the Institute of Electrical and Electronics Engineers, 445 Hoes Lane, Piscataway, NJ 08854, USA (<http://standards.ieee.org/>).

**3.1.6 manager:** A node receiving data from one or more associated agent systems. Examples of managers include a cellular phone, health appliance, set top box, or a computer system.

**3.1.7 obj-handle:** *See:* **handle**.

**3.1.8 object:** In object-oriented modeling, a particular instantiation of a class. The instantiation realizes attributes, methods, and events from the class.

**3.1.9 personal health device:** A device used in personal health applications.

**3.1.10 personal telehealth device:** *See:* **personal health device**.

**3.1.11 plethysmogram, plethysmographic, or photoplethysmographic waveform:** Sequence of samples related to the sequential time-varying light absorption due to effects of pulsatile blood flow.

**3.1.12 SpO<sub>2</sub>:** Percentage oxygen saturation of haemoglobin as measured by a pulse oximeter, where this measurement is an estimate of the fraction of functional haemoglobin (or hemoglobin) in arterial blood that is saturated with oxygen.

NOTE—For more information about SpO<sub>2</sub>, see ISO 9919 [B6].

## 3.2 Acronyms and abbreviations

APDU	application protocol data unit
ASN.1	Abstract Syntax Notation One
DIM	domain information model
ECG	electrocardiograph
EUI-64	extended unique identifier (64 bits)
ICS	implementation conformance statement
ID	identifier
MDC	medical device communication
MDER	medical device encoding rules
MDS	medical device system
MOC	managed object class
OID	object identifier
PDU	protocol data unit
PHD	personal health device
PnP	plug-and-play
RT-SA	real-time sample array
SpO <sub>2</sub>	percentage oxygen saturation of haemoglobin
VMO	virtual medical object
VMS	virtual medical system

## 4. Introduction to ISO/IEEE 11073 personal health devices

### 4.1 General

This standard and the remainder of the series of ISO/IEEE 11073 personal health device standards fit in the larger context of the ISO/IEEE 11073 series of standards. The full suite of standards enables agents to interconnect and interoperate with managers and with computerized healthcare information systems. See IEEE Std 11073-20601-2008 for a description of the guiding principles for this series of ISO/IEEE 11073 personal health device standards.

IEEE Std 11073-20601-2008 supports the modeling and implementation of an extensive set of personal health devices. IEEE Std 11073-10404-2008 (this standard) defines aspects of the pulse oximeter device. It describes all aspects necessary to implement the application layer services and data exchange protocol between an ISO/IEEE 11073 personal health device pulse oximetry agent and a manager. This standard

defines a subset of the objects and functionality contained in IEEE Std 11073-20601-2008, extending and adding definitions where appropriate. All new definitions are given in Annex B in Abstract Syntax Notation One (ASN.1). Nomenclature codes referenced in this standard, which are not defined in IEEE Std 11073-20601-2008, are normatively defined in Annex C.

### 4.2 Introduction to IEEE 11073-20601 modeling constructs

#### 4.2.1 General

The ISO/IEEE 11073 series of standards, and in particular IEEE Std 11073-20601-2008, is based on an object-oriented systems management paradigm. The overall system model is divided into three principal components: the domain information model (DIM), the service model, and the communication model. See IEEE Std 11073-20601-2008 for a detailed description of the modeling constructs.

#### 4.2.2 Domain information model (DIM)

The DIM is a hierarchical model that describes an agent as a set of objects. These objects and their attributes represent the elements that control behavior and report on the status of the agent and data that an agent can communicate to a manager. Communication between the agent and manager is defined by the application protocol in IEEE Std 11073-20601-2008.

#### 4.2.3 Service model

The service model defines the conceptual mechanisms for the data exchange services. Such services are mapped to messages that are exchanged between the agent and manager. Protocol messages within the ISO/IEEE 11073 series of standards are defined in ASN.1. The messages defined in IEEE Std 11073-20601-2008 can coexist with messages defined in other standard application profiles defined in the ISO/IEEE 11073 series of standards.

#### 4.2.4 Communication model

In general, the communication model supports the topology of one or more agents communicating over logical point-to-point connections to a single manager. For each logical point-to-point connection, the dynamic system behavior is defined by a connection state machine as specified in IEEE Std 11073-20601-2008.

#### 4.2.5 Implementing the models

An agent implementing this standard shall implement all mandatory elements of the information, service, and communication models as well as all conditional elements where the condition is met. The agent should implement the recommended elements, and it may implement any combination of the optional elements. A manager implementing this standard shall utilize at least one of the mandatory, conditional, recommended, or optional elements. In this context, “utilize” means to use the element as part of the primary function of the manager device. For example, a manager whose primary function is to display data would need to display a piece of data in the element in order to utilize it.

## 5. Pulse oximeter device concepts and modalities

### 5.1 General

This clause presents the general concepts of pulse oximeter equipment. In the context of personal health devices in the ISO/IEEE 11073 family of standards, a pulse oximeter, also called an oximeter, provides a noninvasive estimate of functional oxygen of arterial haemoglobin (SpO<sub>2</sub>) from a light signal interacting

with tissue, by using the time-dependent changes in tissue optical properties that occur with pulsatile blood flow (see Draft Guidance for Industry and FDA Staff [B5]). Applying the Beer-Lambert law of light absorption through such an arterial network, the fraction of oxygenation of arterial haemoglobin can be estimated. This estimate, normally expressed as a percentage by multiplying that fraction by 100, is known as SpO<sub>2</sub>. Occasionally, this estimate may be referenced as %SpO<sub>2</sub>. ISO 9919 [B6] contains additional information applicable to pulse oximetry.

## 5.2 Device types

Pulse oximeter systems with applicability in the personal health space may take on a variety of configurations and sensor compositions, and their configurations have suitability in different personal health application spaces. Pulse oximeter equipment comprises a pulse oximeter monitor, a pulse oximeter probe, and a probe cable extender, if provided. Some oximeters are all-in-one assemblies, where the optical probe, processing, and display components are in a single package. Other oximeters may consist of separate sensor and processing/display components. Still others may place the sensor and signal processing in one component, and send that information into an external component for display and storage. In addition, other configurations may add storage capability into the system. This implies that different information models may be best suited for each particular device configuration.

## 5.3 General concepts

### 5.3.1 Noninvasive measurement

The scope of this specialization covers the intended use of pulse oximeter equipment, which includes, but is not limited to, the estimation of arterial oxygen haemoglobin saturation and pulse rate. This standard is not applicable to pulse oximeter equipment intended for use in laboratory research applications or to oximeters that require a blood sample (see ISO 9919 [B6]). This standard does not cover measurement of oxygenation via blood extraction. This standard is not applicable to pulse oximeter equipment solely intended for foetal use.

The sensing mechanism may use either transmissive or reflective methods to measure blood oxygenation. In addition, blood oxygenation is usually determined as a ratio of the absorbance of two different wavelengths of light, although more wavelengths may be used.

### 5.3.2 Acquisition modes

#### 5.3.2.1 General

Pulse oximeters are used to measure SpO<sub>2</sub> within a variety of use scenarios.

#### 5.3.2.2 Spot-check

In a spot-check scenario, a user may simply want to take a single, fully processed reading for transmission to a manager. For example, the user would attach the oximeter, whereupon the agent would take an oximetry and pulse rate reading. The agent would then begin communication with a manager and send that single reading. The manager may acknowledge the transmission so the agent can subsequently disassociate and return to its prior state.

#### 5.3.2.3 Continuous monitoring

A continuous monitoring situation involves the pulse oximeter device measuring the user's oxygenation for some period of time greater than that needed to acquire a single measurement. Multiple measurements may be taken to acquire trending information.