



Designation: E1775 – 07

Standard Guide for Evaluating Performance of On-Site Extraction and Field- Portable Electrochemical or Spectrophotometric Analysis for Lead¹

This standard is issued under the fixed designation E1775; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This standard provides guidelines for determining the performance of field-portable quantitative lead analysis instruments.

1.2 This guide applies to field-portable electroanalytical and spectrophotometric (including reflectance and colorimetric) analyzers.

1.3 Sample matrices of concern herein include paint, dust, soil, and airborne particles.

1.4 This guide addresses the desired performance characteristics of field-based sample extraction procedures for lead, as well as on-site extraction followed by field-portable analysis.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

D5438 Practice for Collection of Floor Dust for Chemical Analysis

D6785 Test Method for Determination of Lead in Workplace Air Using Flame or Graphite Furnace Atomic Absorption Spectrometry

D6966 Practice for Collection of Settled Dust Samples Using Wipe Sampling Methods for Subsequent Determination of Metals

D7035 Test Method for Determination of Metals and Metalloids in Airborne Particulate Matter by Inductively

Coupled Plasma Atomic Emission Spectrometry (ICP-AES)

D7144 Practice for Collection of Surface Dust by Microvacuum Sampling for Subsequent Metals Determination

E1605 Terminology Relating to Lead in Buildings

E1613 Test Method for Determination of Lead by Inductively Coupled Plasma Atomic Emission Spectrometry (ICP-AES), Flame Atomic Absorption Spectrometry (FAAS), or Graphite Furnace Atomic Absorption Spectrometry (GFAAS) Techniques

E1644 Practice for Hot Plate Digestion of Dust Wipe Samples for the Determination of Lead

E1645 Practice for Preparation of Dried Paint Samples by Hotplate or Microwave Digestion for Subsequent Lead Analysis

E1726 Practice for Preparation of Soil Samples by Hotplate Digestion for Subsequent Lead Analysis

E1727 Practice for Field Collection of Soil Samples for Subsequent Lead Determination

E1729 Practice for Field Collection of Dried Paint Samples for Subsequent Lead Determination

E1792 Specification for Wipe Sampling Materials for Lead in Surface Dust

E1864 Practice for Evaluating Quality Systems of Organizations Conducting Facility and Hazard Assessments for Lead in Paint, Dust, Airborne Particulate, and Soil in and around Buildings and Related Structures (Withdrawn 2011)³

2.2 U.S. EPA Documents:⁴

EPA 600/R-93/200 Standard Operating Procedure for the Field Analysis of Lead in Paint, Bulk Dust, and Soil by Ultrasonic, Acid Digestion and Colorimetric Measurement (1993)

EPA 747-R-92-001 Laboratory Accreditation Guidelines: Measurement of Lead in Paint, Dust, and Soil (1992)

¹ This guide is under the jurisdiction of ASTM Committee E06 on Performance of Buildings and is the direct responsibility of Subcommittee E06.23 on Lead Hazards Associated with Buildings.

Current edition approved Aug. 1, 2007. Published August 2007. Originally approved in 1996. Last previous edition approved in 2001 as E1775 – 01. DOI: 10.1520/E1775-07.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

⁴ Available from United States Environmental Protection Agency (EPA), Ariel Rios Bldg., 1200 Pennsylvania Ave., NW, Washington, DC 20460, http://www.epa.gov.