# INTERNATIONAL STANDARD

ISO 13164-2

First edition 2013-09-01

### Water quality — Radon-222 —

Part 2: Test method using gamma-ray spectrometry

Qualité de l'eau — Radon 222 —

iTeh STPartie 2: Méthode d'essai par spectrométrie gamma

### (standards.iteh.ai)

<u>ISO 13164-2:2013</u> https://standards.iteh.ai/catalog/standards/sist/68951fe7-b872-4966-b2e2-11b7ab9f92a1/iso-13164-2-2013



Reference number ISO 13164-2:2013(E)

## iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO 13164-2:2013</u> https://standards.iteh.ai/catalog/standards/sist/68951fe7-b872-4966-b2e2-11b7ab9f92a1/iso-13164-2-2013



#### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.org Web www.iso.org

Published in Switzerland

### Contents

Foreword				
Intro	oduction	v		
1	Scope			
2	Normative references			
3	Terms, definitions and symbols	2		
	<ul><li>3.1 Terms and definitions</li><li>3.2 Symbols</li></ul>			
4	Principle			
5	Sampling5.1General requirement5.2Sampling requirement5.3Sample volume5.4Container characteristics			
6	Transportation and storage	4		
7	Detection	4		
8	Measurement procedure	5		
9	Quality assurance and guality control programme_EVIEW9.1General9.2Influence quantities9.3Instrument verification9.4Method verification9.5Demonstration of analyst capability 2:2013https://standards.iteh.ai/catalog/standards/sist/68951fe7-b872-4966-b2e2-Expression of results10.1Activity concentration	5 		
10	https://standards.iteh.a/catalog/standards/sist/68951te/-b8/2-4966-b2e2- Expression of results			
	<ul> <li>10.1 Activity concentration</li> <li>10.2 Standard uncertainty of the activity concentration</li> <li>10.3 Decision threshold</li> <li>10.4 Detection limit</li> <li>10.5 Confidence limits</li> </ul>			
11	Test report			
Ann	nex A (informative) Sample spectra			
Annex B (informative) Sample accuracy and precision data				
Bibli	liography			

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2, www.iso.org/directives.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received, www.iso.org/patents.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

The committee responsible for this document is ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements*. **Teh STANDARD PREVIEW** 

ISO 13164 consists of the following parts, under the general title *Water quality — Radon-222*:

— Part 1: General principles

- <u>ISO 13164-2:2013</u>
- Part 2: Test method using gamma-ray spectrometry ards/sist/68951fe7-b872-4966-b2e2-
- Part 3: Test method using emanometry

The following part is under preparation:

— Part 4: Test method using two-phase liquid scintillation counting

### Introduction

Radioactivity from several naturally occurring and human-made sources is present throughout the environment. Thus, water bodies (surface waters, groundwaters, sea waters) can contain radionuclides of natural and human-made origin.

- Natural radionuclides, including potassium-40, and those of the thorium and uranium decay series, in particular radium-226, radium-228, uranium-234, uranium-238, lead-210, can be found in water for natural reasons (e.g. desorption from the soil and wash-off by rain water) or releases from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizer production and use).
- Human-made radionuclides such as transuranium elements (americium, plutonium, neptunium, curium), tritium, carbon-14, strontium-90 and gamma-emitting radionuclides can also be found in natural waters as they can be authorized to be routinely released into the environment in small quantities in the effluent discharged from nuclear fuel cycle facilities and following their used in unsealed form in medicine or industry. They are also found in water due to the past fallout of the explosion in the atmosphere of nuclear devices and the accidents at Chernobyl and Fukushima.

Drinking-water can thus contain radionuclides at activity concentration which could present a risk to human health. In order to assess the quality of drinking-water (including mineral waters and spring waters) with respect to its radionuclide content and to provide guidance on reducing health risks by taking measures to decrease radionuclide activity concentrations, water resources (groundwater, river, lake, sea, etc.) and drinking water are monitor for their radioactivity content as recommended by the World Health Organization (WHO).

Standard test methods for rad**on-222 activity concentrations** in water samples are needed by test laboratories carrying out such measurements in fulfillment of national authority requirements. Laboratories may have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

https://standards.iteh.ai/catalog/standards/sist/68951fe7-b872-4966-b2e2-

The radon activity concentration in surface water is very low, usually below 1 Bq l<sup>-1</sup>. In groundwater, the activity concentration varies from 1 Bq l<sup>-1</sup> up to 50 Bq l<sup>-1</sup> in sedimentary rock aquifers, from 10 Bq l<sup>-1</sup> up to 300 Bq l<sup>-1</sup> in wells, and from 100 Bq l<sup>-1</sup> up to 1 000 Bq l<sup>-1</sup> in crystalline rocks. The highest activity concentrations are normally measured in rocks with high concentration of uranium (Reference [7]).

High variations in the activity concentrations of radon in aquifers have been observed. Even in a region with relatively uniform rock types, some well water may exhibit radon activity concentration greatly higher than the average value for the same region. Significant seasonal variations have also been recorded (see <u>Annex A</u>).

Water may dissolve chemical substances as it passes from the soil surface to an aquifer or spring waters. The water may pass through or remain for some time in rock, some formations of which may contain a high concentration of natural radionuclides. Under favourable geochemical conditions, the water may selectively dissolve some of these natural radionuclides.

Guidance on radon in drinking-water supplies provided by WHO in 2008 suggests that controls should be implemented if the radon concentration of drinking-water for public water supplies exceeds 100 Bq l<sup>-1</sup>. It also recommended that any new, especially public, drinking-water supply using groundwater should be tested prior to being used for general consumption and that if the radon concentration exceeds 100 Bq l<sup>-1</sup>, treatment of the water source should be undertaken to reduce the radon levels to well below that level (Reference [8]).

This International Standard is one of a series dealing with the measurement of the activity concentration of radionuclides in water samples.

The origin of radon-222 and its short-lived decay products in water and other measurement methods are described generally in ISO 13164-1.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO 13164-2:2013</u> https://standards.iteh.ai/catalog/standards/sist/68951fe7-b872-4966-b2e2-11b7ab9f92a1/iso-13164-2-2013

### Water quality — Radon-222 —

### Part 2: Test method using gamma-ray spectrometry

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

**IMPORTANT** — It is absolutely essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

#### 1 Scope

This part of ISO 13164 specifies a test method for the determination of radon-222 activity concentration in a sample of water following the measurement of its short-lived decay products by direct gamma-spectrometry of the water sample (see <u>Annex A</u>).

The radon-222 activity concentrations, which can be measured by this test method utilizing currently available gamma-ray instruments, range from a few becquerels per litre to several hundred thousand becquerels per litre for a 1 l test sample dards.iteh.ai)

This test method can be used successfully with drinking water samples. The laboratory is responsible for ensuring the validity of this test method for water samples of untested matrices. https://standards.iteh.ai/catalog/standards/sist/68951fe7-b872-4966-b2e2-

<u>Annex B</u> gives indication on the necessary acounting conditions to meet the required sensitivity for drinking water monitoring.

#### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5667-1, Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques

ISO 5667-3, Water quality — Sampling — Part 3: Preservation and handling of water samples

ISO 10703, Water quality — Determination of the activity concentration of radionuclides — Method by high resolution gamma-ray spectrometry

ISO 13164-1, Water quality — Radon-222 — Part 1: General principles

ISO/IEC 17025, General requirements for the competence of testing and calibration laboratories

ISO 80000-10, Quantities and units — Part 10: Atomic and nuclear physics

IEC 60973, Test procedures for germanium gamma-ray detectors

IEC 61151, Nuclear instrumentation — Amplifiers and preamplifiers used with detectors of ionizing radiation — Test procedures

#### ISO 13164-2:2013(E)

IEC 61452, Nuclear instrumentation — Measurement of gamma-ray emission rates of radionuclides — Calibration and use of germanium spectrometers

#### 3 Terms, definitions and symbols

#### 3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 80000-10 and ISO 13164-1 apply.

#### 3.2 Symbols

For the purposes of this document, the symbols defined in ISO 80000-10, ISO 13164-1, and the following apply.

Α	activity of each radionuclide in calibration source, at the calibration time, in becquerels
CA	activity concentration of radon in water, in becquerels per litre
$c_A^*$	decision threshold, in becquerels per litre
<i>c</i> <sup>#</sup> <sub><i>A</i></sub>	detection limit, in becquerels per litre
$c_A^{\triangleleft}, c_A^{\triangleright}$	<b>iTeh STANDARD PREVIEW</b> lower and upper limits of the confidence interval, in becquerels per litre (standards.iten.ai)
fd	correction factor for the decay of radon during time interval between sampling and measuring and for the decay of the radionuclide being measured ( <sup>214</sup> Bi or <sup>214</sup> Pb) over the counting time, dimensionless <sup>67</sup> -b872-4966-b2e2- 11b7ab9(92a1/so-13164-2-2013
$k_p, k_q$	quantiles of the standardized normal distribution for the probabilities, $p$ , $q$ , respectively
$n_{\mathrm{b},E}, n_{\mathrm{b}0,E}, n_{\mathrm{b}\mathrm{s},E}$	number of counts in the background of the peak, at energy, <i>E</i> , in the sample spectrum, in the background spectrum and in the calibration spectrum, respectively
$n_{\mathrm{g},E}, n_{\mathrm{g}0,E}, n_{\mathrm{g}\mathrm{s},E}$	number of counts in the gross area of the peak, at energy, <i>E</i> , in the sample spec- trum, in the background spectrum and in the calibration spectrum, respectively
$n_{\mathrm{N},E}, n_{\mathrm{N0},E}, n_{\mathrm{Ns},E}$	number of counts in the net area of the peak, at energy, <i>E</i> , in the sample spec- trum, in the background spectrum and in the calibration spectrum, respectively
$P_E$	probability of the emission of a gamma-ray with energy, <i>E</i> , of each radionuclide, per decay
$t_0$	background spectrum counting time, in seconds
tg	sample spectrum counting time, in seconds
ts	calibration spectrum counting time, in seconds
U	expanded uncertainty calculated by $U = ku(c_A)$ with $k = 2$
$u(c_A)$	standard uncertainty associated with the measurement result
V <sub>H2</sub> O	volume of test sample, in litres

α, β	probability of the error of the first and second kind, respectively
γ	probability for the confidence interval of the activity concentration
$\varepsilon_E$	efficiency of the detector at energy, <i>E</i>
$\lambda_{222}$ Rn	decay constant of radon-222, in reciprocal seconds
Φ	distribution function of the standardized normal distribution

#### 4 Principle

The determination of radon activity concentration in water using direct gamma-ray spectrometry is based on the:

- collection and storage in a suitable container of a representative sample of the water at time, *t*;
- detection and quantification of gamma-radiation emitted by the short-lived decay products of the radon present in the water sample (see <u>Table 1</u>).

# Table 1 — Energies and emission probability of the main X- and gamma-ray emission of short-<br/>lived <sup>222</sup>Rn decay products (Reference [5])

Radionuclideh	Half-life	A R Energy FV	Probability of emission
Rautonuchue	min	keV	%
214Ph	(standa	<b>rds.</b> 5£,93.a1)	35,60
214PD	26,8	295,22	18,50
https://standard	ls.iteh.ai/catalog/st	andards/sist/08931fe7-b872	-4966-b2e2 <sup>4</sup> 5,49
214Bi	11b7ab9f92	a1/iso-13 <b>16764,49</b> 13	15,28
217DI	19,9	1 120,29	14,907
		1 238,11	5,827

The activity concentration is computed from the decay product activity measured and the volume of the sample having the radon in equilibrium with its decay products.

#### 5 Sampling

#### 5.1 General requirement

The sample shall be representative of the environment to be analysed at a given time.

#### 5.2 Sampling requirement

The sampling shall be carried out in compliance with the conditions and techniques specified in ISO 5667-1, ISO 5667-3, and ISO 13164-1.

The sample shall be transferred directly into the container to be used for the gamma-spectrometric analysis.

The container is filled completely in order to avoid the presence of air in contact with the water sampled.

The container shall be filled in such a way as to avoid degassing the radon in the container sample. The sampling techniques to be used vary according to the actual situation.

When the analytical laboratory is not in charge of sampling, the laboratory shall supply the container for measurement and specify the sampling procedure to the person carrying out the sampling.

The temperature of the water shall be measured during sampling.

It is recommended that the container and its cap be weighed before and after sampling in order to estimate the quantity of suspended material and particles in the container.

#### 5.3 Sample volume

Experience shows that a sample volume of at least 1 l is needed for the sample to be representative of the environment to be analysed.

#### 5.4 Container characteristics

The selection and preparation of a suitable container are important. The container shall be in accordance with the requirements of ISO 5667-3.

The container and cap used to contain the sample shall comply with the following requirements.

- They shall be made from inert materials with low gamma-ray absorption. They shall be impermeable to radon, non-hydrophobic, and conductive (in order not to adsorb radon and its decay products from the surrounding atmosphere).
- They shall not themselves emit gamma-radiation interfering with the measurements.
- They shall be shock-proof.

6

# Transportation and storage

**Standards.iteh.ai**) During transportation and storage, the sample shall be maintained at a temperature below that of the original water (but above 0 °C) until its measurement. The container shall be protected and tightly sealed. The container shall be packed in an appropriate manner in order to prevent any leakage.

The period of transportation and storage prior to analysis shall be as short as possible given the halflife of radon-222, the expected activity concentration and the detection limit of the measurement equipment to be used.

On arrival at the laboratory, the sample shall be maintained at a temperature below that of the original water (but above 0 °C) if it cannot be analysed immediately. The sample shall be analysed as soon as possible after radioactive equilibrium has been achieved.

Experience indicates that it is essential the time between sampling and analysis not exceed 48 h.

#### 7 Detection

The activity concentration of the <sup>222</sup>Rn is determined from the gamma-ray emissions from <sup>214</sup>Bi and/or <sup>214</sup>Pb measured by direct gamma-spectrometry of the water sample.

The determination of the activity concentration of radon-222 in water samples shall be determined by high resolution gamma-ray spectrometry as specified in ISO 10703.

The measurement period shall start when the radon is at radioactive equilibrium with its decay products, i.e. at least 3 h after the end of the sampling.

The selection of the measurement equipment shall take into account the specifications given in IEC 60973 and IEC 61151.

The calibration shall follow the requirements of IEC 61452 using, for example, a radium-226 calibration source.

#### 8 Measurement procedure

The measurement procedure is as follows:

- a) choice and location of the sampling point;
- b) taking one or more water samples;
- c) recording of the location, date, and time of the sampling;
- d) establishment of a radioactive equilibrium between <sup>222</sup>Rn and its short-lived decay products (<sup>214</sup>Bi, <sup>214</sup>Pb) in the sample by waiting, for an optimum count, 3 h after the sampling;
- e) calibration of the spectrometer;
- f) determinination of the background of the detector;
- g) recording of a spectrum of the sample;
- h) analysis of the spectrum and determination of the activity concentration by calculation.

#### 9 Quality assurance and quality control programme

#### 9.1 General

Quality control operations shall meet the requirements of ISO/IEC 17025.

## 9.2 Influence quantities (standards.iteh.ai)

The determination of radon-222 activity <u>concentration s</u>hall take into consideration the correction of its decay during the storage time prior to the measurement/as well as the duration of the measurement process.

11b7ab9f92a1/iso-13164-2-2013 Special precautions shall be taken to minimize the influence of quantities that can affect the measurement results.

During the sampling, it is recommended that the influence be minimized of the:

- water temperature;
- turbulence in the water;
- volume of air in the container.

During measurement, take into account the influence of the:

- <sup>222</sup>Rn present in the atmosphere at the site where the measurements are performed;
- <sup>226</sup>Ra present in the water to be analysed;
- presence of suspended material in the sample.

Wide temporal variations of the radon activity concentration can be observed in the atmosphere of the laboratory. The continuous monitoring of the <sup>222</sup>Rn activity concentration in the laboratory using a radon-measuring device is recommended, since even in well-ventilated rooms, considerable variations are possible in special weather conditions (see ISO 11665-1<sup>[1]</sup> and ISO 11665-5<sup>[2]</sup>).

The air shall be continuously renewed through an air-conditioning system to maintain a constant temperature and relative humidity.

The walls of the laboratory should be painted in a suitable material in order to limit the release of radon from material within the walls (e.g. epoxy paints) (see Reference  $[\underline{6}]$ ).