



# SLOVENSKI STANDARD

## SIST ISO 10315:2014

01-januar-2014

Nadomešča:

SIST ISO 10315:2000

SIST ISO 10315:2000/Pop. 1:2002

---

**Cigarette - Določevanje nikotina in dimnih kondenzatov - Plinsko-kromatografska metoda**

Cigarettes - Determination of nicotine in smoke condensates - Gas-chromatographic method

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

Cigarettes - Dosage de la nicotine dans les condensats de fumée - Méthode par chromatographie en phase gazeuse

**Ta slovenski standard je istoveten z: ISO 10315:2013**

---

**ICS:**

65.160	Tobak, tobačni izdelki in oprema	Tobacco, tobacco products and related equipment
--------	----------------------------------	---

**SIST ISO 10315:2014**

**en,fr**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 10315:2014](#)

<https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014>

INTERNATIONAL  
STANDARD

ISO  
10315

Third edition  
2013-03-01

---

---

**Cigarettes — Determination of  
nicotine in smoke condensates — Gas-  
chromatographic method**

*Cigarettes — Dosage de la nicotine dans les condensats de fumée —  
Méthode par chromatographie en phase gazeuse*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 10315:2014](https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014)

[https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-  
8af5bb1e9924/sist-iso-10315-2014](https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014)



Reference number  
ISO 10315:2013(E)

© ISO 2013

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 10315:2014

<https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword.....	iv
Introduction.....	v
<b>1 Scope.....</b>	<b>1</b>
<b>2 Normative references.....</b>	<b>1</b>
<b>3 Principle.....</b>	<b>1</b>
<b>4 Reagents.....</b>	<b>1</b>
<b>5 Apparatus.....</b>	<b>2</b>
<b>6 Procedure.....</b>	<b>2</b>
6.1 Test portion.....	2
6.2 Setting up the apparatus.....	3
6.3 Calibration of the gas chromatograph.....	3
6.4 Determination.....	3
<b>7 Expression of results.....</b>	<b>3</b>
<b>8 Repeatability and reproducibility.....</b>	<b>3</b>
<b>9 Alternative gas-chromatographic procedures and analysis precautions.....</b>	<b>4</b>
9.1 General.....	4
9.2 Alternative columns.....	4
9.3 Injection systems.....	5
9.4 Alternative internal standards.....	5
<b>10 Test report.....</b>	<b>5</b>
<b>Annex A (informative) Use of this method with the gas-chromatographic determination of water.....</b>	<b>6</b>
<b>Bibliography.....</b>	<b>7</b>

**ISO 10315:2013(E)****Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10315 was prepared by Technical Committee ISO/TC 126, *Tobacco and tobacco products*.

This third edition cancels and replaces the second edition (ISO 10315:2000), which has been technically revised. It also incorporates the Amendment ISO 10315:2000/Amd 1:2011 and the Technical Corrigendum ISO 10315:2000/Cor 1:2000.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 10315:2014](https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014)

<https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014>

## Introduction

This International Standard may be considered as part of a set produced by ISO/TC 126 which describes the determination of total and nicotine-free dry particulate matter (NFDPM) in cigarette smoke condensates. The set comprises:

ISO 3308, ISO 3402, ISO 4387, ISO 8243, ISO 10315, and ISO 10362-1.

A related International Standard, ISO 3400, determines total alkaloids, whereas this International Standard determines only nicotine by virtue of the gas-chromatographic separation. Occasionally, differences can occur because of minor amounts of alkaloids other than nicotine in some types of tobacco.

[Annex A](#) provides information about the use of this method in conjunction with or simultaneously with the gas-chromatographic method of water determination specified in ISO 10362-1.

A bibliography is provided.

No machine smoking regime can represent all human smoking behaviour:

- it is recommended that cigarettes also be tested under conditions of a different intensity of machine smoking than those specified in this International Standard;
- machine smoking testing is useful to characterize cigarette emissions for design and regulatory purposes, but communication of machine measurements to smokers can result in misunderstandings about differences in exposure and risk across brands;
- smoke emission data from machine measurements may be used as inputs for product hazard assessment, but they are not intended to be nor are they valid as measures of human exposure or risks. Communicating differences between products in machine measurements as differences in exposure or risk is a misuse of testing using ISO standards.

<https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 10315:2014](#)

<https://standards.iteh.ai/catalog/standards/sist/703403ba-a129-4f97-8b45-8af5bb1e9924/sist-iso-10315-2014>



# Cigarettes — Determination of nicotine in smoke condensates — Gas-chromatographic method

**WARNING** — The use of this International Standard can involve hazardous materials, operations, and equipment. This International Standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

## 1 Scope

This International Standard specifies a method for the gas-chromatographic determination of nicotine in cigarette smoke condensates. The smoking of cigarettes and the collection of mainstream smoke are normally carried out in accordance with ISO 4387.

NOTE 1 The method specified in this International Standard is also applicable to the determination of nicotine in cigarette smoke condensates obtained by non-standard smoking.

NOTE 2 In countries not in a position to use the gas-chromatographic method, reference should be made to ISO 3400 for the determination of total nicotine alkaloids. In such cases, values obtained using the method described in ISO 3400 may be used with the addition of a note in the expression of results.

## 2 Normative references (standards.iteh.ai)

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4387, *Cigarettes — Determination of total and nicotine-free dry particulate matter using a routine analytical smoking machine*

## 3 Principle

The smoke condensate from the mainstream smoke is dissolved in a solvent containing an internal standard. The nicotine content of an aliquot of the solution is determined by gas chromatography, and the nicotine content of the whole of the smoke condensate is calculated.

## 4 Reagents

Use only reagents of recognized analytical reagent grade.

**4.1 Carrier gas:** helium (CAS: 7440-59-7) or nitrogen (CAS: 7727-37-9) of high purity (at least 99,999 %).

**4.2 Auxiliary gases:** air and hydrogen (CAS: 1333-74-0) of high purity (at least 99,999 %) for the flame ionization detector.

**4.3 Propan-2-ol** (CAS: 67-63-0), with maximum water content of 1,0 mg/ml.