



Designation: E287-94 (Reapproved 1998) Designation: E 287 - 02 (Reapproved 2007)

Standard Specification for Laboratory Glass Graduated Burets¹

This standard is issued under the fixed designation E 287; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers requirements for glass burets of precision and general-purpose grades suitable for laboratory use.

1.1.1 *Class A*—Each buret of precision grade may be marked with the letter “A” to signify compliance with applicable construction and accuracy requirements in Table 1 and Table 2. Each buret may be marked with a permanent serial number at the option of the manufacturer.

1.1.2 *Class B*—General purpose burets are of the same basic design as Class A burets. However, volumetric tolerances for Class B burets shall be within twice the specified range allowed for Class A burets. These burets need not be marked with their class designation or serial number. See Table 1 and Table 3 for tolerances and dimensional requirements.

NOTE 1—The Twelfth General (International) Conference on Weights and Measures redefined the litre as a “special name for the cubic decimetre,” but agreed to permit continuance of the terms litre, millilitre, and mL, except in association with measurements of the highest precision. For volumetric glassware the difference between the old and new meanings of litre is negligible. Therefore, either mL or cm³ may be marked on ware covered by this standard.

1.1.3 Product with a stated capacity not listed in this standard may be specified Class A Tolerance when product conforms to the tolerance range of the next smaller volumetric standard product listed in Table 1.

2. Referenced Documents

2.1 *ASTM Standards:*²

D 664 Test Method for Neutralization Number by Potentiometric Titration

E 438 Specification for Glasses in Laboratory Apparatus

E 542 Practice for Calibration of Volumetric Ware³ Test Method for Acid Number of Petroleum Products by Potentiometric Titration

E 675 Specification for Interchangeable Taper-Ground Stopcocks and Stoppers³

E 694 Specification for Volumetric Ware³ Specification for Interchangeable Taper-Ground Stopcocks And Stoppers

E 911 Specification for Glass Stopcocks with Polytetrafluoroethylene (PTFE) Plugs

E 920 Specification for Commercially Packaged Laboratory Apparatus

E 921 Specification for Export Packaged Laboratory Apparatus

E 1133 Practice for Performance Testing of Packaged Laboratory Apparatus for United States Government Procurements

E 1157 Specification for Sampling and Testing of Reusable Laboratory Glassware

3. General Requirements

3.1 *Calibration*—Burets shall be calibrated to deliver the intended volume at 20°C. No after-drainage period is allowed. Volumetric tolerances shall be within the limits specified in Table 1.

4. Design

4.1 *Shape*—This specification covers 10, 25, 50 and 100-mL straight burets with single-bore stopcock; 10, 25, 50 and 100-mL straight burets with 3-way stopcock; 100-mL burets with 50-mL bulb (intermediate bulb style); and 10, 25, 50 and 100-mL self-zeroing burets.

4.1.1 Self-zeroing burets shall have an ungraduated, tapered, interior extension that terminates with an overflow orifice. The exterior of the buret shall extend reasonably beyond the terminal point of the interior overflow orifice. An exterior outlet tube

¹ This specification is under the jurisdiction of ASTM Committee E-41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Glass Apparatus.

Current edition approved Feb. 15, 1994; Nov. 1, 2007. Published March 1994; December 2007. Originally published as E287-65T, approved in 1965. Last previous edition E287-83 (1988) in 2002 as E 287 - 02

² Annual Book of ASTM Standards, Vol 05.01.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.