

DRAFT INTERNATIONAL STANDARD

ISO/DIS 16782

ISO/TC 212

Secretariat: ANSI

Voting begins on:
2014-06-26

Voting terminates on:
2014-11-26

Clinical laboratory testing - Criteria for acceptable lots of dehydrated Mueller-Hinton agar and broth for antimicrobial susceptibility testing

Critères pour des lots acceptables d'Agar déshydraté Mueller-Hinton et de bouillon pour essai de susceptibilité antimicrobienne

ICS: 11.100.20

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/96be36c1-2d71-4f25-a40e-e23a3ecfeb1f/iso-prf-16782>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

ISO/CEN PARALLEL PROCESSING

This draft has been developed within the International Organization for Standardization (ISO), and processed under the **ISO lead** mode of collaboration as defined in the Vienna Agreement.

This draft is hereby submitted to the ISO member bodies and to the CEN member bodies for a parallel five month enquiry.

Should this draft be accepted, a final draft, established on the basis of comments received, will be submitted to a parallel two-month approval vote in ISO and formal vote in CEN.

To expedite distribution, this document is circulated as received from the committee secretariat. ISO Central Secretariat work of editing and text composition will be undertaken at publication stage.



Reference number
ISO/DIS 16782:2014(E)

© ISO 2014

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/96be36c1-2d71-4f25-a40e-e23a3ecfeb1f/iso-prf-16782>

Copyright notice

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

Contents

Page

Foreword	iv
Introduction.....	iv
1 Scope	1
2 Normative reference	1
3 Terms and definitions	1
4 Requirements for Mueller-Hinton broth	3
4.1 Components of Mueller-Hinton broth.....	3
4.2 Physical and chemical characteristics.....	3
4.2.1 Dehydrated powder or granules	3
4.2.2 Prepared broth medium	3
4.2.3 Cation supplementation and content for MHB	4
4.2.4 Other medium components.....	4
4.2.5 Specific adjustments required by the manufacturer	4
4.3 Manufacturers protocol for testing production lots of Dehydrated Mueller-Hinton broth	5
4.4 Interpreting the results	5
4.5 Evaluating the results	6
5 Requirements for Muller-Hinton agar	6
5.1 Components of Mueller-Hinton agar.....	6
5.2 Physical and chemical characteristics.....	6
5.2.1 Dehydrated powder or granules	6
5.2.2 Prepared agar medium.....	6
5.2.3 Cation supplementation and content for MHA	7
5.2.4 Other medium components.....	7
5.2.5 Specific adjustments required by the manufacturer	7
5.3 Manufacturer's protocol for testing production lots of dehydrated Mueller-Hinton agar	7
5.4 Interpreting the results	8
5.5 Evaluating the results	9
6 Testing new antimicrobial agents with production lots of Dehydrated Mueller-Hinton broth or agar	9
Annex A (informative) Mueller-Hinton medium.....	11
A.1 Broth	11
A.2 Agar.....	12
Annex B (informative) Preparing control cultures.....	13
B.1 Stock cultures	13
B.1.1 Preparing the initial stock cultures	14
B.1.2 Preparing frozen stock cultures	14
B.1.3 Preparing the stock culture test inoculum	14
Annex C (informative) Suggested data sheet for testing of production lots	15
C.1 Suggested data sheet for testing of production lots of dMHB	15
C.2 Suggested data sheet for testing of production lots of dMHA	16
Annex D (informative) Label statement	18
D.1 Dehydrated Mueller-Hinton broth	18
D.2 Dehydrated Mueller-Hinton agar	18
D.3 Label statement	18
Bibliography.....	19

Foreword

ISO 16782 was prepared by Technical Committee ISO/TC 212 Working Group 4 Project Team in collaboration with, the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 16782 was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*, Working Group 4, *Antimicrobial susceptibility testing*.

iTeh STANDARDS PREVIEW
(standards.itih.com)
Full standard:
<https://standards.itih.com/catalogue/standards/1966/16782-2d71-4f25-a40e-e23a3ecfe040-pr14-2>

Introduction

Historically, although various media have been recommended for susceptibility testing, Mueller-Hinton broth (MHB) has been selected as the medium for the reference broth microdilution minimum inhibitory concentration (MIC) method (ISO 20776-1) and Mueller-Hinton agar (MHA) is most widely used for disc diffusion testing of rapidly growing bacteria. Mueller-Hinton medium provides satisfactory growth of most non-fastidious pathogens, acceptable batch-to-batch reproducibility, low sulfonamide, trimethoprim, and tetracycline inhibitors, and a large amount of data has been collected from antimicrobial susceptibility tests with this medium over several decades. This standard is the result of an effort to establish a standard description and protocol by which manufacturers of dehydrated Mueller-Hinton agar (dMHA) and broth (dMHB) may determine its acceptable performance characteristics. This document describes methods for evaluation of production lots of the dehydrated product. Performance characteristics of production lots of agar and broth prepared with the dehydrated product are determined by testing defined microorganism-antimicrobial agent combinations. The results of testing must conform to defined quality control limit ranges for each combination of antimicrobial agent and quality control strains. Each production lot shall be tested at least against these combinations of antimicrobial agents and quality control strains. This standard does not address supplements (e.g., blood or blood products) that may be added to the medium to support growth of fastidious bacteria [1,2,3,4]. Those additives are provided after the dehydrated medium is prepared in its liquid state as a final product and fall outside of the scope of this document. Although dMHA can be used for determination of MICs using the agar dilution method [2,4] or the gradient diffusion method, this document only includes performance testing of dMHA using disc diffusion methodology as described by the Clinical and Laboratory Standards Institute (CLSI) [3] and European Committee on Antimicrobial Susceptibility Testing (EUCAST) [1]. Manufacturers may choose to test additional antimicrobial agents and strains, as well as Mueller-Hinton media supplemented for growth of fastidious strains. This is at the discretion of manufacturers, but expected performance limits must be validated appropriately.

This standard has been developed in part based upon two previous Clinical and Laboratory Standards Institute documents, CLSI M6-A2 (Protocols for evaluating dehydrated Mueller-Hinton agar) and CLSI M32-P (Evaluation of lots of dehydrated Mueller-Hinton broth for antimicrobial susceptibility testing) with permission. This standard supersedes and replaces the previous CLSI publications.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/96be36c1-2d71-4d25-a40e-e23a3ecfeb1f/iso-prf-16782>

Antimicrobial susceptibility testing — Criteria for acceptable lots of dehydrated Mueller-Hinton agar and broth for antimicrobial susceptibility testing

1 Scope

This International Standard provides a standard description of the physical properties of dehydrated Mueller-Hinton broth (dMHB) and Mueller-Hinton agar (dMHA) and performance criteria by which manufacturers can assess the performance characteristics of their production lots of dMHA and dMHB. Production lots of broth or agar may then be utilized by all users, including *in vitro* susceptibility testing device manufacturers, as the test medium for performance of antimicrobial susceptibility testing.

2 Normative reference

The following referenced document is indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20776-1, Clinical laboratory testing and *in vitro* diagnostic test systems – Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices – Part 1: Reference method for testing the *in vitro* activity of antimicrobial agents against rapidly growing aerobic bacteria involved in infectious diseases.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

antimicrobial agent

substance of biological, semi-synthetic or synthetic origin that inhibits the growth of or kills bacteria and is thus of potential use in the treatment of infections

NOTE 1 to entry: Disinfectants, antiseptics and preservatives are not included in this definition.

[SOURCE: ISO 20776-1; 2006, 2.1]

3.2

antimicrobial disc

small paper disc containing known amounts of antimicrobial agents used for *in vitro* susceptibility testing.

3.3

concentration

amount of an antimicrobial agent in a defined volume of liquid

NOTE 1 to entry: The concentration is expressed as mg/l.

NOTE 2 to entry: mg/l = µg/ml but it is not recommended to use the unit µg/ml.

[SOURCE: ISO 20776-1;2006, 2.2.2]

**3.4
stock solution**

initial solution used for further dilutions

[SOURCE: ISO 20776-1; 2006, 2.3]

**3.5
minimum inhibitory concentration
MIC**

lowest concentration of antimicrobial agent that, under defined *in vitro* conditions, prevents visible growth of bacteria within a defined period of time

NOTE 1 to entry: The MIC is expressed in mg/l.

[SOURCE: ISO 20776-1; 2006, 2.4 modified — “lowest concentration that” has been modified to “lowest concentration of antimicrobial agent that”.]

**3.6
reference strain**

catalogued, characterized microorganism with stable, defined antimicrobial susceptibility phenotype and/or genotype

NOTE 1 to entry: Reference strains are kept as stock cultures, from which working cultures are derived. They are obtained from recognised national culture collections and used for quality control.

[SOURCE: ISO 20776-1; 2006, 2.7 modified — “characterized bacteria” has been modified to “characterized microorganism” and “culture collections” in NOTE has been modified to “recognised national culture collections”.]

**3.7
Susceptibility testing method**

**3.7.1
broth dilution**

technique in which containers are filled with appropriate volumes of broth containing an antimicrobial agent in incrementally (usually two-fold) increasing concentrations and a defined inoculum

NOTE 1 to entry: The aim of this method is the determination of the MIC.

[SOURCE: ISO 20776-1; 2006, 2.8.1 modified — “an antimicrobial solution, employing incrementally (usually two-fold) increasing concentrations of the antimicrobial agent and appropriate volumes of broth with” has been modified to “broth containing an antimicrobial agent in incrementally (usually two-fold) increasing concentrations and”.]

**3.7.2
microdilution**

performance of broth dilution in microdilution trays with a capacity of 200 µl per well

[SOURCE: ISO 20776-1; 2006, 2.8.2 modified — “a capacity of ≤ 200 µl per well” has been modified to “a capacity of 200 µl per well”.]

**3.7.3
disc diffusion**

technique in which antimicrobial discs are applied to the surface of an agar medium that has been evenly inoculated with a defined inoculum and, following incubation under defined conditions, the resulting size of zones of growth inhibition of the microorganism correspond to the susceptibility/resistance of the microorganism to the antimicrobial agent

3.7.4**zone diameter**

the diameter (in mm) of the zone of growth inhibition around a paper disc containing an antimicrobial agent of specified amount used in a disc diffusion test

3.8**broth**

liquid medium used for the *in vitro* growth of bacteria

[SOURCE: ISO 20776-1; 2006, 2.9 modified — “fluid medium” has been modified to “liquid medium”.]

3.9**inoculum**

number of viable bacteria in a suspension, calculated with respect to the final volume

NOTE 1 to entry: The inoculum is expressed as colony-forming units per millilitre (CFU/ml).

[SOURCE: ISO 20776-1; 2006, 2.10 modified — “number of bacteria” has been modified to “number of viable bacteria”.]

3.10**dehydrated Mueller-Hinton broth****dMHB**

a dried bacteriological medium which is used to prepare liquid medium for broth dilution antimicrobial susceptibility tests

3.11**dehydrated Mueller-Hinton agar****dMHA**

a dried bacteriological medium which is used to prepare antimicrobial susceptibility testing agar plates for disc diffusion, gradient diffusion MIC and agar dilution MIC methods

4 Requirements for Mueller-Hinton broth**4.1 Components of Mueller-Hinton broth**

Historically, Mueller-Hinton broth medium for antimicrobial susceptibility testing contains approximately the following components per litre of purified water (adjustments may be needed to meet performance criteria) [5]:

Dehydrated infusion from 300 g beef (i.e., 2 g of beef extract powder)

Acid digest of casein 17,5 g

Starch 1,5 g

4.2 Physical and chemical characteristics**4.2.1 Dehydrated powder or granules**

Colour - beige to light beige

Uniform, free-flowing, homogeneous and free of extraneous material

4.2.2 Prepared broth medium

Once hydrated the final pH measured after autoclaving shall be 7,2-7,4 at 25 °C

The liquid is light straw coloured and clear with no visible precipitate

4.2.3 Cation supplementation and content for MHB

The broth shall contain sufficient concentrations of cations to provide adequate growth and to permit the user to determine MIC values (e.g., aminoglycosides and quinolones) for quality control strains within ranges identified in ISO 20776-1 Table 4 (check the latest version of CLSI and EUCAST documents for QC ranges). New lots of MHB may require testing for acceptable cation content. For standard production lots of dMHB, the broth prepared from the dehydrated product shall contain no greater than 25 mg/l of total calcium and 12,5 mg/l of total magnesium. Manufacturers may choose to provide commercial lots of dMHB with required concentrations of cations or actual levels less than 20 mg/l of calcium and 10 mg/l of magnesium. In the latter case, the final label shall specify the actual amounts contained in the lot of broth. For final testing, the prepared MHB shall contain 20-25 mg/l of Ca^{2+} and 10-12,5 mg/l of Mg^{2+} .

While trace amounts of manganese are required for growth, the concentration shall be below 8 mg/l to avoid false resistant interpretations with glycolcyclines [6]. This shall be determined by an MIC value within the acceptable range obtained by testing *Escherichia coli* WDCM 00013 with tigecycline.

While trace amounts of zinc are required for growth, the concentration of zinc shall be below 3 mg/l to avoid false resistance interpretations with imipenem [7] and potentially with other carbapenems. This shall be determined by an MIC value within the acceptable range obtained by testing *Pseudomonas aeruginosa* WDCM 00025 with imipenem.

Cation concentrations of calcium, magnesium, manganese, and zinc shall be determined by inductively coupled plasma mass spectrometry (ICP-MS) or flame atomic absorption spectroscopy (FAAS)[8].

Although ion effects known to affect susceptibility test results for other antimicrobial agents are not included in this document, they shall be considered for MHB dilution susceptibility tests by manufacturers at their discretion. Affected agents include daptomycin [9] and polymyxin[10]. When testing daptomycin, MHB shall be supplemented to a final concentration of 50 mg/l total Ca^{2+} . Refer to ISO 20776-1 for appropriate instructions on preparation of media and antimicrobial susceptibility testing.

4.2.4 Other medium components

The medium shall have a thymidine mass concentration of less than 0,03 mg/l as indicated by an MIC value of $\leq 0,5/9,5$ mg/l obtained by testing *Enterococcus faecalis* WDCM 00087 with trimethoprim-sulfamethoxazole [11].

4.2.5 Specific adjustments required by the manufacturer

For antimicrobial agents included in Table 1:

- a) Incorporation of sodium chloride (2% w/v NaCl) at a final concentration of 20 g/l in the broth is required for the detection of methicillin resistance in *Staphylococcus* spp. when testing with oxacillin.
- b) For broth microdilution testing of tigecycline, when MIC panels are prepared, the medium must be prepared fresh on the day of use. The medium must be no more than 12 hours old at the time the panels are made; however, the panels may then be frozen for later use. For further details refer to ISO 20776-1.

For organisms not included in Table 1 (i.e., for extended testing at the discretion of the manufacturer):

- a) Testing of fastidious organisms such as streptococci and *Haemophilus* spp. requires the addition of growth supplements (for example blood or blood components). If a Mueller-Hinton agar or broth lot that is found to perform acceptably according to the criteria in this document is to be used for testing fastidious organisms, the resulting MICs or zone diameters after addition of supplements must fall within the acceptable quality control ranges published in 20776-1 for the specific medium and organism tested.