
**Tissue paper and tissue products —
Part 7:
Determination of optical properties —
Measurement of brightness and colour
with D65/10° (outdoor daylight)**

iTeh STANDARD PREVIEW

(standards.iteh.ai)

Papier tissu et produits tissu —

*Partie 7: Détermination des propriétés optiques — Mesurage du
degré de blancheur et de la couleur avec l'illuminant D65/10°
(lumière du jour extérieure)*

ISO 12625-7:2014

<https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014>



iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 12625-7:2014](https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014)

<https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Apparatus	3
6 Calibration	4
7 Sampling	4
8 Conditioning	5
9 Preparation of test pieces	5
10 Procedure	5
10.1 General.....	5
10.2 Measurement of D65 brightness.....	5
10.3 Measurement of colour (D65/10°).....	5
11 Calculation	6
11.1 D65 brightness.....	6
11.2 Colour (D65/10°).....	6
12 Test report	7
Annex A (informative) Precision	8
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

ISO 12625-7 was prepared by European Committee for Standardization (CEN) Technical Committee CEN/TC 172 *Pulp, paper and board*, in collaboration with Technical Committee ISO/TC 6, *Paper, board and pulps*, Subcommittee SC 2, *Test methods and quality specifications for paper and board*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 12625-7:2007), which has been technically revised.

ISO 12625 consists of the following parts, under the general title *Tissue paper and tissue products*:

- Part 1: *General guidance on terms*;
- Part 3: *Determination of thickness, bulking thickness, apparent bulk density and bulk*;
- Part 4: *Determination of tensile strength, stretch at break and tensile energy absorption*;
- Part 5: *Determination of wet tensile strength*;
- Part 6: *Determination of grammage*;
- Part 7: *Determination of optical properties — Measurement of brightness and colour with D65/10° (outdoor daylight)*;
- Part 8: *Water-absorption time and water-absorption capacity; basket-immersion test method*;
- Part 9: *Determination of ball burst strength*;
- Part 11: *Determination of wet ball burst strength*;
- Part 12: *Determination of tensile strength of perforated lines — Calculation of perforation efficiency*;
- Part 15: *Determination of optical properties — Measurement of brightness and colour with C/2° (indoor daylight)*;
- Part 16: *Determination of optical properties — Opacity (paper backing) — Diffuse reflectance method*

Introduction

Brightness and colour measurement may be performed under various illumination and observation conditions. This part of ISO 12625 deals with D65/10° conditions, which refer to an outdoor daylight.

C/2° conditions (indoor daylight) are considered in ISO 12625-15. Although both international standards deal with brightness and colour, results obtained are usually different and do not correlate.

Optical measurement are affected by the geometry of the instruments used and by the texture of the material. The design of the instrument to be used according to this part of ISO 12625, and the routine to be adopted for its calibration, are specified in ISO 2469 and ISO 11475.

The optical properties are related to the visual appearance of the material. Therefore, although optical properties are intrinsic properties of tissue paper, they are not functional properties.

Brightness shall not be confused with the optical property called CIE-whiteness that is based on reflectance data obtained over the full visible spectral range (VIS) in contrast to the measurement of brightness which is limited to the blue region of VIS.

Due to the importance for some countries three different test methods for the determination of optical properties were developed:

- Part 7: *Determination of optical properties — Measurement of brightness and colour with D65/10° (outdoor daylight);*
- Part 15: *Determination of optical properties — Measurement of brightness and colour with C/2° (indoor) daylight;*
- Part 16: *Determination of optical properties — Opacity (paper backing) — Diffuse reflectance method;*

[ISO 12625-7:2014](https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014)

<https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 12625-7:2014](https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014)

<https://standards.iteh.ai/catalog/standards/sist/52a5f2da-01f3-48c8-8c69-d3049a9d94bf/iso-12625-7-2014>

Tissue paper and tissue products —

Part 7:

Determination of optical properties — Measurement of brightness and colour with D65/10° (outdoor daylight)

1 Scope

This part of ISO 12625 specifies testing procedures for the instrumental determination of brightness and colour of tissue paper and tissue products viewed under outdoor daylight conditions. It also gives specific instructions for the preparation of test pieces (single-ply, multi-ply products) and for the optical measurements of products, where special precautions may be necessary.

NOTE The properties called C/2° brightness and colour are measured with an instrument adjusted to a much lower UV content than that specified in this part of ISO 12625. The measurements of C/2° brightness and colour are described in ISO 12625-15.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 186, *Paper and board — Sampling to determine average quality*
[https://standards.iteh.ai/catalog/standards/sist/52a512da-0113-48c8-8c69-](https://standards.iteh.ai/catalog/standards/sist/52a512da-0113-48c8-8c69-d3049a9d94bf/iso-12625-7-2014)

ISO 187, *Paper, board and pulps — Standard atmosphere for conditioning and testing and procedure for monitoring the atmosphere and conditioning of samples*

ISO 2469:2007, *Paper, board and pulps — Measurement of diffuse radiance factor*

ISO 5631-2:2008, *Paper and board — Determination of colour by diffuse reflectance — Part 2: Outdoor daylight conditions (D65/10 degrees)*

ISO 11475, *Paper and board — Determination of CIE whiteness, D65/10 degrees (outdoor daylight)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

reflectance factor

R

ratio of the radiation reflected by a body to that reflected by the perfect diffuser under the same conditions of illumination and detection

Note 1 to entry: The reflectance factor is expressed as a percentage.

Note 2 to entry: The reflectance factor is influenced by the backing if the body is translucent.

**3.2
diffuse (radiance) reflectance factor**

R

ratio of the reflection from a body to that from the perfect reflecting diffuser under the same conditions of diffuse illumination and normal detection

Note 1 to entry: The ratio is often expressed as a percentage.

Note 2 to entry: Adapted from ISO 2469:2007, definition 3.5.

**3.3
intrinsic reflectance factor**

R_{∞}

diffuse reflectance factor of a layer or pad of material thick enough to be opaque, i.e. such that increasing the thickness of the pad by doubling the number of sheets results in no change in the measured reflectance factor

[SOURCE: ISO 2469:2007, definition 3.6]

**3.4
D65 brightness**

intrinsic reflectance factor measured with a reflectometer having the characteristics described in ISO 2469, equipped with a filter or corresponding function having an effective wavelength of 457 nm (and a half bandwidth of 44 nm), and adjusted so that the UV content of the irradiation incident upon the test piece corresponds to that of the CIE standard illuminant D65

Note 1 to entry: The filter function is described more fully by the weighing function factors given in ISO 2470-2[Z].

**3.5
tristimulus values**

X_{10}, Y_{10}, Z_{10}

amounts of the three reference colour stimuli, in a given chromatic system, required to match the stimulus considered

Note 1 to entry: In ISO 5631-2 the CIE standard illuminant D65 and the CIE 1964 (10°) standard observer are used to define the trichromatic system.

Note 2 to entry: The subscript 10 is applied to conform to the CIE convention that tristimulus units have the subscript 10 when the CIE 1964 (10°) standard observer is used.

[SOURCE: ISO 5631-2:2008, definition 3.5]

**3.6
colour (D65/10°)**

L^* , a^* and b^* values of the sample according to the CIELAB 1976 system, corresponding to the CIE standard illuminant D65, described in ISO 11664-2 and the CIE 1964 supplementary standard colorimetric observer, described in ISO 11664-1,[2] determined by measurement under the conditions specified in ISO 5631-2

Note 1 to entry: The quantity L^* is a measure of the lightness of the test piece, where $L^* = 0$ corresponds to black and $L^* = 100$ is defined by the perfect reflecting diffuser. Visually, the quantities a^* and b^* represent respectively the red-green and yellow-blue axes in colour space, such that

- $+a^*$ is a measure of the degree of redness of the test piece;
- $-a^*$ is a measure of the degree of greenness of the test piece;
- $+b^*$ is a measure of the degree of yellowness of the test piece;
- $-b^*$ is a measure of the degree of blueness of the test piece;
- if both a^* and b^* are equal to zero, the test piece is grey.

4 Principle

A test piece is illuminated diffusely in a standardized instrument and the light reflected normal to the surface is either allowed to pass through a defined optical filter and then measured by a photodetector or measured by an array of photosensitive diodes, where each diode responds to a different effective wavelength. The brightness is then determined directly from the output from the photodetector or by calculation from the photosensitive diode outputs using the appropriate weighting function and colour coordinates are calculated for D65/10° conditions.

5 Apparatus

5.1 Reflectometer or spectrophotometer, having the geometric, spectral and photometric characteristics described in ISO 2469 and calibrated in accordance with the provisions of ISO 2469, and equipped for the measurement of blue reflectance factor

5.1.1 In the case of a filter reflectometer, the radiation falling upon the test piece shall have a UV content corresponding to that of the CIE standard illuminant D65, adjusted or verified by the help of the fluorescent reference standard (5.2.2).

5.1.2 In the case of an abridged spectrophotometer, the instrument shall have an adjustable filter with a cut-off wavelength of 395 nm or some other system for adjustment and control, and this filter shall be adjusted or the system shall be calibrated with the help of the fluorescence reference standard (5.2.2), so that the UV content of the illumination falling upon the sample corresponds to that of the CIE standard illuminant D65.

(standards.iteh.ai)

5.2 Reference standard for calibration of the instrument

ISO 12625-7:2014

5.2.1 Non-fluorescent reference standard for photometric calibration, issued by an ISO/TC 6 authorized laboratory in accordance with the provisions of ISO 2469.

5.2.2 Fluorescent reference standard for use in adjusting the UV content of the radiation incident upon the sample, having a CIE whiteness (D65/10°) unit assigned by an ISO/TC 6 authorized laboratory as prescribed in ISO 11475:2004, Annex B.

NOTE 1 Greater precision in the D65 brightness measurement would be attained if a fluorescent reference standard having an assigned D65 brightness unit were used. It is, however, important for the industry to have only one UV-filter adjustment for all measurements under CIE illuminant D65 conditions. For this reason, a reference standard having an assigned CIE whiteness (D65/10°) values as prescribed in ISO 11475 is preferred.

NOTE 2 To be used sufficiently frequently to ensure satisfactory calibration and UV adjustment.

5.3 Working standards

5.3.1 Two plates of flat opal glass, ceramic or other suitable non-fluorescent material, cleaned and calibrated as described in ISO 2469.

NOTE In some instruments, the function of the primary working standard may be taken over by a built-in internal standard.

5.3.2 Stable plastic or other tablet, incorporating a fluorescent whitening agent.

5.3.3 Black cavity, having a reflectance factor which does not differ from its nominal value by more than 0,2 %, at all wavelengths. The black cavity should be stored upside down in a dust-free environment or with a protective cover.

NOTE The condition of the black cavity can be checked by reference to the instrument maker.