

SLOVENSKI STANDARD

SIST EN ISO 11737-1:2018

01-april-2018

Nadomešča:

SIST EN ISO 11737-1:2006

SIST EN ISO 11737-1:2006/AC:2009

**Sterilizacija izdelkov za zdravstveno nego - Mikrobiološke metode - 1. del:
Določevanje populacije mikroorganizmov na izdelku (ISO 11737-1:2018)**

Sterilization of health care products - Microbiological methods - Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018)

iTeh STANDARD PREVIEW

Sterilisation von Produkten für die Gesundheitsfürsorge - Mikrobiologische Verfahren - Teil 1: Bestimmung der Population von Mikroorganismen auf Produkten (ISO 11737-1:2018)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

[https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

[f290ab6f7dfb/sist-en-iso-11737-1-2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

Stérilisation des produits de santé - Méthodes microbiologiques - Partie 1: Détermination d'une population de micro-organismes sur des produits (ISO 11737-1:2018)

Ta slovenski standard je istoveten z: EN ISO 11737-1:2018

ICS:

07.100.10	Medicinska mikrobiologija	Medical microbiology
11.080.01	Sterilizacija in dezinfekcija na splošno	Sterilization and disinfection in general

SIST EN ISO 11737-1:2018

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 11737-1

January 2018

ICS 07.100.10; 11.080.01

Supersedes EN ISO 11737-1:2006

English Version

Sterilization of health care products - Microbiological methods - Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018)

Stérilisation des produits de santé - Méthodes microbiologiques - Partie 1: Détermination d'une population de micro-organismes sur des produits (ISO 11737-1:2018)

Sterilisation von Produkten für die Gesundheitsfürsorge - Mikrobiologische Verfahren - Teil 1: Bestimmung der Population von Mikroorganismen auf Produkten (ISO 11737-1:2018)

This European Standard was approved by CEN on 6 December 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	Error! Bookmark not defined.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)
<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

European foreword

This document (EN ISO 11737-1:2018) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2018, and conflicting national standards shall be withdrawn at the latest by July 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11737-1:2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom. (standards.iteh.ai)

Endorsement notice

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

The text of ISO 11737-1:2018 has been approved by CEN as EN ISO 11737-1:2018 without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

INTERNATIONAL
STANDARD

ISO
11737-1

Third edition
2018-01

**Sterilization of health care products —
Microbiological methods —**

Part 1:
**Determination of a population of
microorganisms on products**

iTeh STANDARD PREVIEW
*Stérilisation des produits de santé — Méthodes microbiologiques —
Partie 1: Détermination d'une population de microorganismes sur des
produits*
(standards.iteh.ai)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>



Reference number
ISO 11737-1:2018(E)

© ISO 2018

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 11737-1:2018

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements	4
4.1 Documentation.....	4
4.2 Management responsibility.....	5
4.3 Product realization.....	5
4.4 Measurement, analysis and improvement.....	5
5 Selection of products	5
5.1 General.....	5
5.2 Sample item portion (SIP).....	6
6 Methods of determination and microbial characterization of bioburden	6
6.1 Determination of bioburden.....	6
6.1.1 Selection of an appropriate method.....	6
6.1.2 Neutralization of inhibitory substances.....	7
6.1.3 Removal of microorganisms.....	7
6.1.4 Culturing of microorganisms.....	7
6.1.5 Enumeration of microorganisms.....	7
6.2 Microbial characterization of bioburden.....	7
7 Validation of the method for determining bioburden	8
7.1 General.....	8
7.2 Validation.....	8
8 Routine determination of bioburden and interpretation of data	8
8.1 General.....	8
8.2 Limits of detection and plate counting.....	8
8.3 Microbial characterization.....	8
8.4 Bioburden data for extent of treatment.....	9
8.5 Bioburden spikes.....	9
8.6 Bioburden levels.....	9
8.7 Data analysis.....	9
8.8 Statistical methods.....	9
9 Maintenance of the method for determining bioburden	9
9.1 Changes to the product and/or manufacturing process.....	9
9.2 Changes to the method for determining bioburden.....	9
9.3 Requalification of the method for determining bioburden.....	9
Annex A (informative) Guidance on the determination of a population of microorganisms on products	10
Annex B (informative) Guidance on methods to determine bioburden	27
Annex C (informative) Validation of bioburden recovery efficiency	37
Annex D (informative) Typical assignment of responsibilities	45
Bibliography	47

ISO 11737-1:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html. (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This third edition cancels and replaces the second edition (ISO 11737-1:2006), which has been technically revised. It also incorporates the Technical Corrigendum ISO 11737-1:2006/Cor.1:2007.

The main changes compared to the previous edition are as follows:

- the term “bioburden spikes” has been introduced as a normal and consistent part of the bioburden, and examples of data have been provided;
- clarification has been added that package testing is not typically done except when it is an integral part of the product;
- more information has been provided on the most probable number (MPN) technique and its applications;
- details have been provided on ways to improve limit of detection (LOD) and correct use of the data;
- some discussion has been deleted of statistical methods for the evaluation of bioburden data where information was not typical or not required;
- a table has been added with criteria for selection of a bioburden recovery efficiency approach, the use of the correction factor (CF) has been explained, and the bioburden recovery efficiency value of < 50 % mentioned for technique modifications has been eliminated;
- more information has been provided on the application and performance of a bioburden method suitability test;
- a section has been added to detail rules for direct plate counts, estimated counts and counts beyond the ideal range;
- a table has been added to clarify where typical responsibilities reside for the manufacturer or the laboratory;

- the focus on a risk-based approach has been increased, including the purpose for which bioburden data will be used.

A list of all parts in the ISO 11737 series can be found on the ISO website.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

ISO 11737-1:2018(E)

Introduction

A sterile health care product is one that is free of viable microorganisms. International Standards that specify requirements for the validation and routine control of sterilization processes require, when it is necessary to supply a sterile health care product, that adventitious microbiological contamination of a health care product prior to sterilization be minimized. Such products are non-sterile. The purpose of sterilization is to inactivate the microbiological contaminants and thereby transform the non-sterile products into sterile ones.

The kinetics of inactivation of a pure culture of microorganisms by physical and/or chemical agents used to sterilize health care products can generally best be described by an exponential relationship between the numbers of microorganisms surviving and the extent of treatment with the sterilizing agent. Inevitably, this means there is always a finite probability that a microorganism can survive regardless of the extent of treatment applied. For a given treatment, the probability of survival is determined by the number and resistance of microorganisms and by the environment in which the microorganisms exist during treatment. It follows that the sterility of any one product in a population subjected to sterilization processing cannot be guaranteed and the sterility of a processed population is defined in terms of the probability of there being a viable microorganism present on a product item.

Generic requirements of the quality management system for design and development, production, installation and servicing are given in ISO 9001 and particular requirements for quality management systems for medical device production are given in ISO 13485. The standards for quality management systems recognize that, for certain processes used in manufacturing, the effectiveness of the process cannot be fully verified by subsequent inspection and testing of the product. Sterilization is an example of such a process. For this reason, sterilization processes are validated for use, the performance of the sterilization process is monitored routinely and the equipment is maintained.

International Standards specifying procedures for the validation and routine control of the processes used for the sterilization of health care products have been prepared (see, for example, ISO 14937, ISO 11135, the ISO 11137 series, the ISO 17665 series and ISO 14160). However, it is important to be aware that exposure to a properly validated and accurately controlled sterilization process is not the only factor associated with the provision of assurance that the product is sterile and, in this respect, suitable for its intended use. Furthermore, for the effective validation and routine control of a sterilization process, it is important to be aware of the microbiological challenge that is presented in the process, in terms of number, characteristics and properties of microorganisms.

The term “bioburden” is used to describe the population of viable microorganisms present on or in a product and/or a sterile barrier system. A knowledge of bioburden can be used in a number of situations as part of the following:

- validation and requalification of sterilization processes;
- routine monitoring for control of manufacturing processes;
- monitoring of raw materials, components or packaging;
- assessment of the efficiency of cleaning processes;
- an overall environmental monitoring programme.

Bioburden is the sum of the microbial contributions from a number of sources, including raw materials, manufacturing of components, assembly processes, manufacturing environment, assembly/manufacturing aids (e.g. compressed gases, water, lubricants), cleaning processes and packaging of finished products. To control bioburden, attention should be given to the microbiological status of these sources.

It is not possible to enumerate bioburden exactly and, in practice, a determination of bioburden is made using a defined method. Definition of a single method for use in determining bioburden in all situations is not practicable because of the wide variety of designs and materials of construction of health care products. Nor is it possible to define a single technique to be used in all situations for the removal of

microorganisms in preparation for enumeration. Furthermore, the selection of culture conditions for enumeration of microorganisms will be influenced by the types of microorganism likely to be present on or in health care products.

This document specifies the requirements to be met for the determination of bioburden. In addition, it gives guidance in the annexes to provide explanations and methods that are deemed suitable to conform with the requirements. Methods other than those given in the guidance may be used, if they are effective in achieving conformity with the requirements of this document.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 11737-1:2018](https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018)

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 11737-1:2018

<https://standards.iteh.ai/catalog/standards/sist/9ef0d18f-ed18-4d3b-94d1-f290ab6f7dfb/sist-en-iso-11737-1-2018>

Sterilization of health care products — Microbiological methods —

Part 1: Determination of a population of microorganisms on products

1 Scope

This document specifies requirements and provides guidance on the enumeration and microbial characterization of the population of viable microorganisms on or in a health care product, component, raw material or package.

NOTE 1 The nature and extent of microbial characterization is dependent on the intended use of bioburden data.

NOTE 2 See [Annex A](#) for guidance on [Clauses 1](#) to [9](#).

This document does not apply to the enumeration or identification of viral, prion or protozoan contaminants. This includes the removal and detection of the causative agents of spongiform encephalopathies, such as scrapie, bovine spongiform encephalopathy and Creutzfeldt-Jakob disease.

NOTE 3 Guidance on inactivating viruses and prions can be found in ISO 22442-3, ICH Q5A(R1) and ISO 13022.

This document does not apply to the microbiological monitoring of the environment in which health care products are manufactured.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10012, *Measurement management systems — Requirements for measurement processes and measuring equipment*

ISO 13485, *Medical devices — Quality management systems — Requirements for regulatory purposes*

ISO 15189, *Medical laboratories — Requirements for quality and competence*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at <http://www.electropedia.org/>

— ISO Online browsing platform: available at <http://www.iso.org/obp>