

ISO/TC 168

Date: 2014-07-15

ISO 22675:2006/FDAM 1:2014(E)

ISO/TC 168/WG 3

Secretariat: DIN

Prosthetics — Testing of ankle-foot devices and foot units — Requirements and test methods

Prothèses — Essais des mécanismes de cheville-pied et des unités de pied — Exigences et méthodes d'essai

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 22675:2006/FDAmD 1](https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1)

<https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1>

Document type: International Standard
Document subtype: Amendment
Document stage: (50) Approval
Document language: E

Copyright notice

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office

Case postale 56 ♦♦ CH-1211 Geneva 20

Tel. + 41 22 749 01 11

Fax + 41 22 749 09 47

E-mail copyright@iso.org

Web ~~www.iso.org~~ www.iso.org

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

ITeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 22675:2006/FDAmD 1](https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1)

<https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

~~International Standards~~The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules given in the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

~~The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.~~

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

ISO 22675:2006/FDAmd 1

~~Amendment 1~~Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement. www.iso.org/standard/68888/iso-22675-2006-fdamd-1

~~For an explanation on the meaning of ISO specific terms and expressions related to ISO 22675:2006 was prepared by conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)~~Committee

~~The committee responsible for this document is ISO/TC 168, Prosthetics and orthotics and by Technical Committee CEN/TC 293, Assistive products for persons with disability in collaboration.~~

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 22675:2006/FDAmd 1](https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1)

<https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1>

Prosthetics — Testing of ankle-foot devices and foot units — Requirements and test methods

AMENDMENT 1

Page 2, Clause 2

Update the normative reference to ISO 10328 to the 2015 version with the following text:

ISO 10328:2015, Prosthetics — Structural testing of lower limb prostheses — Requirements and test methods

Page 8, 7.2.3

Replace 7.2.3 with the following text:

Test loading levels: P3, P4, P5, P6, P7, and P8

NOTE 1 Field experience has shown that there is a need for lower limb prostheses which sustain loads above the level covered by test loading level P5. In order to allow the structural testing of such prostheses on a uniform basis, test loading levels P6, P7, and P8 have been developed for the principal structural tests and the separate structural tests on ankle-foot devices and foot units (see Annex C).

NOTE 2 The values of the dimensions and loads of test loading levels P3, P4, and P5 are specified in separate tables in Clause 8. It is suggested that the values of the dimensions and loads specified in C.3 and Table C.2 are appropriate for test loading level P6, and for P7 and P8 (pending validation), as an interim measure. Further test loading levels will be defined, if necessary.

Page 10, Table 5

Delete the NOTE and replace by the following:

For the additional test loading levels P6, P7, and P8, the test forces are specified in Table C.1.

Page 11, Table 6

Delete the NOTE and replace by the following:

The total length and the segmental lengths also apply to the additional test loading levels P6, P7, and P8 as specified in Annex C [see C.3 a)].

Page 12, Table 7

Delete the NOTE and replace by the following:

The specified dimensions also apply to the additional test loading levels P6, P7, and P8 as specified in Annex C [see C.3 a)].

Page 12, Table 8

Delete the NOTE and replace by the following:

The specified dimensions also apply to the additional test loading levels P6, P7, and P8 as specified in Annex C [see C.3 a)].

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 22675:2006/FDAmd 1](https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdam-1)

<https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdam-1>

Page 13, Table-9

Replace Table-9 by the following:

Test procedure and test force			Unit	Test loading level (P _x) ^a and test loading condition (F _{1x} ; F _{2x})					
				P5		P4		P3	
				Heel loading, F _{1x}	Forefoot loading, F _{2x}	Heel loading, F _{1x}	Forefoot loading, F _{2x}	Heel loading, F _{1x}	Forefoot loading, F _{2x}
Static test procedure	Static proof test force	F _{1sp}	N	2 227	—	2 053	—	1 601	—
		F _{2sp}		—	2 198	—	2 026	—	1 580
	Static ultimate test force	F _{1su} , lower level	N	3 340	—	3 079	—	2 401	—
		F _{2su} , lower level		—	3 297	—	3 039	—	2 369
		F _{1su} , upper level	N	4 454	—	4 106	—	3 201	—
		F _{2su} , upper level		—	4 396	—	4 052	—	3 159
Cyclic test procedure	1st maximum value of pulsating test force	F _{1cmax}	N	1 273	—	1 173	—	915	—
	Intermediate minimum value of pulsating test force	F _{1cmin}	N	850		783		611	
	2nd maximum value of pulsating test force	F _{2cmax}	N	—	1 256	—	1 158	—	903
	Final static test force	F _{1fin} (= F _{1sp})	N	2 227	—	2 053	—	1 601	—
		F _{2fin} (= F _{2sp})		—	2 198	—	2 026	—	1 580
Prescribed number of cycles			1	2 × 10 ⁶					

NOTE The specific values of the different test forces are based on reference values described in A.2.3 and specified in Table A.1.

^a For the additional test loading level P6, P7, and P8, the values of the test forces and the prescribed number of cycles are specified in Table C.2.

Page 15, Table-10

Replace Table-10 by the following:

Reference point	Threshold															
	Instant (Time after heel contact) ms	Interval of time ms	Rate of loading/unloading (Value relevant to test loading level) kN/s						Test force $F_c(t)$ at reference point N							
			P8	P7	P6	P5	P4	P3	Symbol	P8	P7	P6	P5	P4	P3	
a	0									$F_c(t_a)$	0	0	0	0	0	0
		115	17,8	15,4	13,3	11,1	10,2	8,0								
b	115									F_{1cmax}	2 038	1 760	1 521	1 273	1 173	915
		51														
c	166									F_{1cmax}	2 038	1 760	1 521	1 273	1 173	915
		103	-6,5	-5,6	-4,9	-4,1	-3,8	-2,9								
d	269									F_{cmin}	1 361	1 175	1 016	850	783	611
		62														
e	331									F_{cmin}	1 361	1 175	1 016	850	783	611
		102	6,4	5,6	4,8	4,0	3,7	2,9								
f	433									F_{2cmax}	2 011	1 737	1 501	1 256	1 158	903
		51														
g	484									F_{2cmax}	2 011	1 737	1 501	1 256	1 158	903
		116	-17,4	-15,0	-13,0	-10,8	-9,9	-7,9								
h	600									$F_c(t_h)$	0	0	0	0	0	0

NOTE The loading period of 600 ms corresponds to the average stance phase time of a typical walking cycle of 1 s duration (the remaining time of 400 ms of the walking cycle corresponds to the swing phase). Simulating this stance phase time-swing phase time-relationship in a cyclic test, a loading period of 600 ms corresponds to a test frequency $f = 1$ Hz. For other test frequencies, preferably between 0,5 Hz and 3 Hz (see 16.4.1.6 and 16.4.1.7), the time intervals between each instant after heel contact for which rates of loading/unloading or test force are specified in this table can easily be adapted by linear scaling.

Page 44, 16.2.1.2

Replace the last sentence in the third paragraph by the following:

Make specific reference if the additional test loading level P6, P7, or P8 specified in Annex C are to be applied.

Page 47, 16.3.1.2

Replace the last sentence in the fourth paragraph by the following:

Make specific reference if the application of the additional test loading levels P6, P7, or P8 specified in Annex C is to be applied.

Page 48, 16.3.1.5

In the second sentence of the fourth paragraph replace [12.3.3 a)] by [12.3.4 a)]

Page 48, 16.3.1.8

Replace the second sentence in the fourth paragraph by the following:

Make specific reference if the application of the additional test loading levels P6, P7, or P8 specified in Annex C is to be applied.

Page 52, 16.4.1.3

Replace the second sentence in the second paragraph by the following:

Make specific reference if the application of the additional test loading levels P6, P7, or P8 specified in Annex C is to be applied.

Page 59, 18.2.1

Replace the second sentence in the first paragraph by the following:

This particularly applies to tests that are conducted at the additional test loading level P6, P7, or P8 according to Annex C (see 16.2.1.2, 16.3.1.2, and 16.4.1.3) and to the alternative static ultimate strength test according to Annex B (see 16.3.1.1, 16.3.1.5, and 16.3.1.11).

Page 63, A.1

Replace the second and third paragraph by the following:

The test loading level P5 is based on data from all amputees including a few whose body mass exceeded 100 kg. The test loading ~~level~~levels P6 and P7 is based on locomotion data from amputees of whose body mass is less than 125 kg and 150 kg, respectively, and obtained from simulations and field observations; P8 is extrapolated from these two levels to amputees whose body mass is less than 175 kg. The test loading levels P4 and P3 are based on locomotion data from amputees whose body mass is less than 80 kg and 60 kg, respectively.

For the proposed additional test loading levels P6, P7, and P8, see Annex C.

Page 65, ~~Equation-Formula~~ (A.4) [ISO 22675:2006/FDAmD 1](https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1)

Replace by the following: <https://standards.iteh.ai/catalog/standards/sist/4dcfc380-753d-4031-88c3-ed66dfdacc8e/iso-22675-2006-fdamd-1>

$$D_{PT} = \sqrt{\frac{f_{T,26}^2 + u_{T,26}^2}{26}}$$

(A.4)