

# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 8828

ISO/TC 150

Secretariat: DIN

Voting begins on:  
2014-01-19

Voting terminates on:  
2014-04-19

---

---

## Implants for surgery — Guidance on care and handling of orthopaedic implants

*Implants chirurgicaux — Principes directeurs pour l'entretien et la manipulation des implants orthopédiques*

[Revision of first edition (ISO 8828:1988)]

ICS: 11.040.40

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/9749c11a-4856-4290-a567-0894589fd242/iso-8828-2014>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.



Reference number  
ISO/DIS 8828:2013(E)

© ISO 2013

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/9749c11a-4856-4290-a567-0894589fd242/iso-8828-2014>

### Copyright notice

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

# Contents

Page

Foreword .....	iv
Introduction.....	v
1 Scope .....	1
2 Terms and definitions .....	1
3 General guidance .....	1
3.1 Manufacturer's instructions .....	1
3.2 On receipt.....	1
3.3 Transport.....	2
3.4 Stock records.....	2
3.5 Storage .....	3
3.6 Stock rotation .....	3
3.7 Cleaning and sterilization of non-sterile implants .....	3
3.8 Appearance .....	4
3.9 Contouring and modifying implants.....	4
3.10 Re-use.....	4
4 Additional guidance on polymeric implants and materials .....	4
4.1 Sterilization .....	4
4.2 Acrylic bone cement .....	4
4.3 Silicone implants .....	4
4.4 Biodegradable implants .....	5
5 Additional guidance on ceramic components.....	5
5.1 Sterilization and handling.....	5
5.2 Dropping of ceramic components .....	5
5.3 Manufacturer's instructions .....	5
6 Additional guidance on implants or components of implants with rough surfaces or surfaces with intrinsic porosity .....	5
6.1 Sterile implants.....	5
6.2 Subsequent cleaning of implants .....	5
6.3 Non-sterile implants .....	5
Bibliography.....	6

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8828 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*.

This second edition cancels and replaces the first edition (ISO 8828:1988), which has been technically revised.

**PREVIEW**  
iTech STANDARD  
(standards.iteh.ai)  
Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/49c11a-4856-4290-a567-0894589fd242/iso-8828-2013>

## Introduction

The guidance given in this International Standard on the care and handling of orthopaedic implants after delivery to the purchaser is intended to help ensure that implants remain free from contamination or damage prior to insertion into the patient. Guidance is given on the procedures for receiving, storing, transporting, handling, cleaning and sterilizing implants. Guidance on procedures for preparing the implants for use, as well as handling during the surgery, are also outlined. This guidance is aimed at all personnel involved in receiving and handling implants including surgeons. It is important that all personnel should be familiar with recommended procedures in order to minimize the risk and occurrence of damage to implants.

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/9749c11a-4856-4290-a567-0894589fd242/iso-8828-2014>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

Full standard:  
<https://standards.iteh.ai/catalog/standards/sist/9749c11a-4856-4290-a567-0894589fd242/iso-8828-2014>

# Implants for surgery — Guidance on care and handling of orthopaedic implants

## 1 Scope

This International Standard specifies the recommended procedures for handling orthopaedic implants, hereafter referred to as implants, from receipt at the hospital until they are implanted or discarded.

This guidance applies to implants (such as currently used metal, ceramic or polymeric implants) and also to acrylic resin and other bone cements.

This guidance does not apply to the implant manufacturer. However, it contains references to the stocking of implants that can be useful for manufacturers and especially for third-party suppliers.

## 2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 2.1 orthopaedic implant implant

device implanted surgically, wholly or partially, in the body, either temporarily or permanently, and used either as an aid in the repair of bone or related tissues, or as a temporary or permanent replacement for these tissues

NOTE Acrylic resin cement, used for fixing certain devices, is deemed to be an "implant".

## 3 General guidance

### 3.1 Manufacturer's instructions

All of the manufacturer's instructions should be followed and take precedence over the guidance provided in this standard.

### 3.2 On receipt

#### 3.2.1 General

Packaged implants can arrive either:

- a) pre-sterilized (see 3.2.2), or
- b) non-sterilized (see 3.2.3).

### 3.2.2 Products supplied sterile

The packaging of products supplied sterile shall be left intact until the time of use. The packaging shall be inspected for damage. If damage is found, the implant shall be considered non-sterile. The implant shall then either:

- a) be returned to the manufacturer for reprocessing; or,
- b) if appropriate and not prohibited by the device manufacturer, be taken out of the damaged packaging and re-sterilized in the user facility following the directions for an applicable method of sterilization provided in the instructions for use (for guidance see ISO 17664).

### 3.2.3 Non-sterile implants

Some non-sterile implants can be received in special packaging that is suitable for sterilization using the method(s) specified by the manufacturer in the instructions for use. The implant shall not be removed from this packaging prior to sterilization. Non-sterile implants not packaged in this way should only be unwrapped immediately prior to sterilization so as to preserve the surface finish and configuration intact, and they should be handled as infrequently as possible. The implant shall be sterilized following the directions for an applicable method of sterilization provided in the instructions for use.

### 3.2.4 Usability of implants

Any implant that has been dropped or mishandled and which is suspected of having suffered damage shall not be used. The implant should be disposed of or returned to the manufacturer as directed in the instructions for use. However, the final judgement as to the suitability of the implant shall always lie with the surgeon who uses the implant provided there are no restrictions in the instructions for use. If such an implant is used, the patient record shall include a description of the mishandling and any methods used to mitigate the possible effects of the mishandling on the safety and efficacy of the device.

## 3.3 Transport

Care shall be exercised during transport and handling of the implants so as to preclude any damage or alteration to the condition of the implant and its packaging as received. Attention shall be paid to the handling conditions specified by the manufacturer on the label of the outermost layer of packaging.

## 3.4 Stock records

### 3.4.1 General

Stock records are required to facilitate inventories, stock rotation, traceability to the manufacturer and, in some instances, for transfer to patient's records.

### 3.4.2 Lot or batch code or serial number

The label of the implant package should bear the model designation of the device and in most cases a lot, batch code or serial number of the implant. Also, some implants are marked on their surface with the lot or batch code or serial number. The lot, batch code or serial number of the implant shall be transferred to the patient's record.

### 3.4.3 Records to be compiled

The following information shall be recorded:

- a) the name of the manufacturer;
- b) a description of the implant including as applicable:



- the model designation;
  - the implant material(s); and
  - the characteristic dimensions;
- c) the lot or batch code or serial number of the implant;
- d) the number of implants in a package unit;
- e) the "use by" date or date of manufacture as appropriate; and
- f) the date of receipt by the hospital.

### 3.5 Storage

#### 3.5.1 General

In all storage areas, implants shall be stored prior to use so as to maintain the configuration and surface finish of the implant and to avoid damage to its packaging, particularly to the sterile packaging. Implants should be stored separately from instruments. Non-sterile implants shall be stored separately from those that have been sterilized.

#### 3.5.2 Storage conditions

The implant shall be stored in a suitable area following the conditions specified by the manufacturer on the label of the outermost layer of packaging or in the instructions for use (e.g. temperature, humidity and ambient pressure). If there are no such instructions, implants shall be stored in dry conditions and shall not be exposed to direct sunlight, ionizing radiation, extremes of temperature or particulate contamination.

### 3.6 Stock rotation

The principle of "first in, first out" is recommended. The practice of stock rotation should be adopted for all implants, sterile and non-sterile, in all storage areas.

### 3.7 Cleaning and sterilization of non-sterile implants

**3.7.1** Non-sterile implants may be sterilized without prior cleaning if the manufacturer's packaging has been removed immediately prior to sterilization.

**3.7.2** After each surgical procedure, all implants that may be subjected to a resterilization procedure shall be thoroughly and carefully cleaned according to the manufacturer's instructions for use. Ultrasonic cleaning, mechanized washing or scrubbing by hand are suitable methods provided that they are carried out carefully. The method used shall prevent impact, scratching, bending or surface contact with any materials that might affect the implant surface or configuration.

**3.7.3** The manufacturer's recommendations on cleaning shall be closely complied with. If scrubbing by hand is used, soft brushes shall be used and harsh chemicals or harsh cleaning solutions shall be avoided.

**3.7.4** After cleaning, the implants shall be rinsed completely free of all residues, soap, detergent or cleaning solutions. After rinsing, the implants shall be thoroughly dried. Special attention shall be paid to recesses since both chemicals and rinse water may be entrapped in them. If a cloth is used for the final implant cleaning or drying, this cloth shall be made of antistatic material so that dirt and dust from the environment are not attracted by the implant surface because of an electrostatic charge.