



SLOVENSKI STANDARD SIST EN ISO 16140-4:2020

01-november-2020

Mikrobiologija v prehranski verigi - Validacija metode - 4. del: Protokol za validacijo metode v posameznem laboratoriju (ISO 16140-4:2020)

Microbiology of the food chain - Method validation - Part 4: Protocol for method validation in a single laboratory (ISO 16140-4:2020)

Mikrobiologie der Lebensmittelkette - Verfahrensvalidierung - Teil 4: Arbeitsvorschrift für Einzel-Labor-Verfahrensvalidierung (ISO 16140-4:2020)

Microbiologie de la chaîne alimentaire - Validation des méthodes - Partie 4: Protocole pour la validation de méthodes dans un seul laboratoire (ISO 16140-4:2020)

Ta slovenski standard je istoveten z: EN ISO 16140-4:2020

[SIST EN ISO 16140-4:2020](https://standards.slovenski-standard.si/standards/sist/en/2020/01/01/16140-4:2020-383/standard/sist-en-iso-16140-4-2020)

<https://standards.slovenski-standard.si/standards/sist/en/2020/01/01/16140-4:2020-383/standard/sist-en-iso-16140-4-2020>

ICS:

07.100.30 Mikrobiologija živil Food microbiology

SIST EN ISO 16140-4:2020 **en**

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 16140-4

August 2020

ICS 07.100.30

English Version

**Microbiology of the food chain - Method validation - Part 4:
Protocol for method validation in a single laboratory (ISO
16140-4:2020)**

Microbiologie de la chaîne alimentaire - Validation des méthodes - Partie 4: Protocole pour la validation de méthodes dans un seul laboratoire (ISO 16140-4:2020)

Mikrobiologie der Lebensmittelkette -
Verfahrensvalidierung - Teil 4: Arbeitsvorschrift für
Einzel-Labor-Verfahrensvalidierung (ISO 16140-
4:2020)

This European Standard was approved by CEN on 25 May 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[SIST EN ISO 16140-4:2020](https://standards.itih.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020)

<https://standards.itih.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020>

European foreword

This document (EN ISO 16140-4:2020) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 463 "Microbiology of the food chain" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2021, and conflicting national standards shall be withdrawn at the latest by February 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 16140-4:2020 has been approved by CEN as EN ISO 16140-4:2020 without any modification.

[SIST EN ISO 16140-4:2020](https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020)

<https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020>

INTERNATIONAL STANDARD

ISO 16140-4

First edition
2020-07

Microbiology of the food chain — Method validation —

Part 4: Protocol for method validation in a single laboratory

*Microbiologie de la chaîne alimentaire — Validation des méthodes —
Partie 4: Protocole pour la validation de méthodes dans un seul
laboratoire*

*ITeH Standards
(<https://standards.iteh.ai>)
Document Preview*

[SIST EN ISO 16140-4:2020](https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020)

<https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020>



Reference number
ISO 16140-4:2020(E)

© ISO 2020

ISO 16140-4:2020(E)

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN ISO 16140-4:2020](https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020)

<https://standards.iteh.ai/catalog/standards/sist/c52ea26c-bef4-4a6d-ad22-38e95daaea71/sist-en-iso-16140-4-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 General principles of the single-laboratory detection or quantification method validation.....	4
4.1 General.....	4
4.2 Principles of the factorial approach.....	5
4.3 Principles of the conventional approach.....	5
5 Technical protocol for validation — Factorial approach.....	7
5.1 Qualitative methods.....	7
5.1.1 Single-laboratory method validation study against a reference method.....	7
5.1.2 Single-laboratory method validation study without a reference method.....	13
5.2 Quantitative methods.....	15
5.2.1 Single-laboratory method validation study against a reference method.....	15
5.2.2 Single-laboratory method validation study without a reference method.....	18
6 Technical protocol for validation — Conventional approach.....	19
6.1 Qualitative methods.....	19
6.1.1 Single-laboratory method validation study against a reference method.....	19
6.1.2 Single-laboratory method validation study without a reference method.....	20
6.2 Quantitative methods.....	21
6.2.1 Single-laboratory method validation study against a reference method.....	21
6.2.2 Single-laboratory method validation study without a reference method.....	23
7 Summary of acceptability limits.....	26
Annex A (informative) List of factors and factor levels for factorial method validation.....	27
Annex B (informative) Calculation of in-house reproducibility for qualitative methods based on the LOD₅₀ study described in 6.1.2.3.....	29
Annex C (informative) Example of a factorial single-laboratory method validation study for a quantitative method against a reference method.....	30
Annex D (informative) Example of a factorial single-laboratory method validation study for a qualitative method against a reference method.....	36
Annex E (informative) Example of a factorial single-laboratory method validation study for the variability of the LOD₅₀ for a qualitative method without a reference method.....	40
Annex F (informative) Determination of precision if the artificially contaminated samples are unstable.....	43
Annex G (informative) Protocol for single-laboratory validation of alternative methods for microbiological confirmation and typing procedures.....	45
Bibliography.....	46

ISO 16140-4:2020(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 9, *Microbiology*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 463, *Microbiology of the food chain*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 16140 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

0.1 The ISO 16140 series

The ISO 16140 series has been expanded in response to the need for various ways to validate or verify test methods. It is the successor to ISO 16140:2003. The ISO 16140 series consists of six parts with the general title, *Microbiology of the food chain — Method validation*:

- *Part 1: Vocabulary;*
- *Part 2: Protocol for the validation of alternative (proprietary) methods against a reference method;*
- *Part 3: Protocol for the verification of reference methods and validated alternative methods in a single laboratory;*
- *Part 4: Protocol for method validation in a single laboratory;*
- *Part 5: Protocol for factorial interlaboratory validation for non-proprietary methods;*
- *Part 6: Protocol for the validation of alternative (proprietary) methods for microbiological confirmation and typing procedures.*

ISO 17468 is a closely linked International Standard, which establishes technical rules for the development and validation of standardized methods.

In general, two stages are needed before a method can be used in a laboratory.

- The first stage is the validation of the method. Validation is conducted using a study in a single laboratory followed by an interlaboratory study (see ISO 16140-2, ISO 16140-5 and ISO 16140-6). In the case when a method is validated within one laboratory (as described in this document), no interlaboratory study is conducted.
- The second stage is method verification, where a laboratory demonstrates that it can satisfactorily perform a validated method. This is described in ISO 16140-3. Verification is only applicable to methods that have been validated using an interlaboratory study.

In general, two types of methods are distinguished: reference methods and alternative methods.

A reference method is defined in ISO 16140-1:2016, 2.59, as an “internationally recognized and widely accepted method”. The note to entry clarifies that “these are ISO standards and standards jointly published by ISO and CEN or other regional/national standards of equivalent standing”.


In the ISO 16140 series, reference methods include standardized reference (ISO and CEN) methods as defined in ISO 17468:2016, 3.5, as a “reference method described in a standard”.

An alternative method (method submitted for validation) is defined in ISO 16140-1:2016, 2.4, as a “method of analysis that detects or quantifies, for a given category of products, the same analyte as is detected or quantified using the corresponding reference method”. The note to entry clarifies that: “The method can be proprietary. The term ‘alternative’ is used to refer to the entire ‘test procedure and reaction system’. This term includes all ingredients, whether material or otherwise, required for implementing the method.”

This document, ISO 16140-4, addresses validation within a single laboratory. The results are therefore only valid for the laboratory that conducted the study. In this case, verification (as described in ISO 16140-3) is not applicable. ISO 16140-5 describes protocols for non-proprietary methods where a more rapid validation is required or when the method to be validated is highly specialized and the number of participating laboratories required by ISO 16140-2 cannot be reached. This document and ISO 16140-5 can be used for validation against a reference method. This document (regarding qualitative and quantitative methods) and ISO 16140-5 (regarding quantitative methods only) can also be used for validation without a reference method.

ISO 16140-4:2020(E)

The flow chart in [Figure 1](#) gives an overview of the links between the different parts mentioned above. It also guides the user in selecting the right part of the ISO 16140 series, taking into account the purpose of the study and the remarks given above.





SIST EN ISO 16140-4:2020

<https://standards.iteh.org/standards/iso/16140-4/2020> **Figure 1 — Flow chart for application of the ISO 16140 series** [71/sist-en-iso-16140-4-2020](https://standards.iteh.org/standards/iso/16140-4/2020)

NOTE In this document, the words “category”, “type” and/or “item” are sometimes combined with “(food)” to improve readability. However, the word “(food)” is interchangeable with “(feed)” and other areas of the food chain as mentioned in [Clause 1](#).

ISO 16140-6 is somewhat different from the other parts in the ISO 16140 series in that it relates to a very specific situation where only the confirmation procedure of a method is to be validated [e.g. the biochemical confirmation of *Enterobacteriaceae* (see ISO 21528-2)]. The confirmation procedure advances a suspected (presumptive) result to a confirmed positive result. The validation of alternative typing techniques (e.g. serotyping of *Salmonella*) is also covered by ISO 16140-6. The validation study in ISO 16140-6 clearly defines the selective agar(s) from which strains can be confirmed using the alternative confirmation method. If successfully validated, the alternative confirmation method can only be used if strains are recovered on an agar that was used and shown to be acceptable within the validation study. [Figure 2](#) shows the possibilities where an alternative confirmation method validated in accordance with ISO 16140-6 can be applied (see text in the boxes).

Reference method**Alternative method validated in accordance with ISO 16140-2****Figure 2 — Use of validated alternative confirmation methods (see ISO 16140-6)**

EXAMPLE An example application of a validated alternative confirmation method is as follows.

An alternative confirmation method based on ELISA has been validated (in accordance with ISO 16140-6) to replace the biochemical confirmation for *Salmonella* as described in ISO 6579-1. In the validation study, XLD (mandatory agar in accordance with ISO 6579-1) plus BGA and a specified chromogenic agar (two optional agars for second plating in accordance with ISO 6579-1) were used as the agars to start the confirmation. The validated confirmation method can be used to replace the biochemical confirmation under the following conditions:

- by laboratories using the ISO 6579-1; or
- by laboratories using an ISO 16140-2 validated alternative method that refers to ISO 6579-1 for confirmation; or
- by laboratories using an ISO 16140-2 validated alternative method that starts the confirmation from XLD and/or BGA agar and/or the specified chromogenic agar.

The validated confirmation method cannot be used under the following conditions:

- by laboratories using an ISO 16140-2 validated alternative method that refers only to agars other than those included in the validation to start the confirmation (e.g. Hektoen agar and SS agar only); or
- by laboratories using an ISO 16140-2 validated alternative method that refers only to a confirmation procedure that does not require isolation on agar.

0.2 Validation protocols in the ISO 16140 series

An interlaboratory validation study, in accordance with ISO 16140-2, requires at least eight laboratories for quantitative methods and at least ten laboratories for qualitative methods. ISO 16140-5 is intended to be used for interlaboratory studies comprising four to seven laboratories for quantitative methods and four to nine laboratories for qualitative methods. ISO 16140-5 can only be used for non-proprietary methods. [Table 1](#) provides an overview of the different protocols.

ISO 16140-4:2020(E)

Table 1 — Overview of different validation protocols described in the ISO 16140 series

Number of participating laboratories	With reference method	Without reference method
1	This document: — factorial (see 5.1.1 and 5.2.1), or — conventional (see 6.1.1 and 6.2.1)	This document: — factorial (see 5.1.2 and 5.2.2), or — conventional (see 6.1.2 and 6.2.2)
4 to 7 (quantitative method)/ 4 to 9 (qualitative method)	ISO 16140-5: for non-proprietary methods only	ISO 16140-5: for non-proprietary quantitative methods only
≥ 8 (quantitative method)/ ≥ 10 (qualitative method)	ISO 16140-2: for the interlaboratory study part	Not applicable

The aim of this document is to assess the performance of detection or quantification methods within a single laboratory, typically across a number of (food) categories and (food) types. Single-laboratory validation of alternative methods for microbiological confirmation and typing procedures can also be performed under certain conditions: the general principles are the same as those described in ISO 16140-6 for the validation of alternative (proprietary) methods for microbiological confirmation and typing procedures (except there is no interlaboratory study). Further information is given in [Annex G](#).

The protocols in this document only validate the method for the particular laboratory. A generalization to other laboratories is not within the scope of these protocols. However, extension to other laboratories is possible if this document is used as the first phase of validation of a reference method, to be followed by an interlaboratory study as described in ISO 17468.

If a reference method is available, the validation of a method is conducted by comparing the alternative method to the reference method. This allows inclusion of naturally contaminated samples in the validation process and thus provides a more realistic picture of the performance of the method. If no reference method is available, the validation process is based on samples with known contamination levels only. This document provides protocols for both situations.

The general principles for single-laboratory validations of detection and quantification methods are the same as those described in ISO 16140-2 for the validation of alternative (proprietary) methods against a reference method. This document cannot be used without ISO 16140-1 or ISO 16140-2, as many definitions and procedures are given in these International Standards. In addition to the validation parameters described in ISO 16140-2, this document describes the calculation of in-house repeatability and in-house reproducibility. Calculation of these parameters is not required if an interlaboratory study is to be conducted after the single-laboratory validation (i.e. if the single-laboratory validation is only the first phase of validation). Reliability of performance parameters obtained with this document is comparable to ISO 16140-2. This also means that the workload associated with the technical protocols for the single laboratory is comparable with the method comparison study of ISO 16140-2.

This document provides two strategies for the single-laboratory method validation of detection and quantification methods. The first strategy is based on a factorial approach while the second strategy uses the conventional approach derived from the protocols of ISO 16140-2. In addition, protocols for the determination of the in-house reproducibility for quantitative methods are described.

The advantages of using a factorial approach, over the conventional approach, are that it takes into account specific conditions that the laboratory encounters during routine testing and provides more information on the factors (technicians, culture media, etc.) that vary within the laboratory across relevant (food) items, while using fewer samples to assess the performance of the method. The factorial approach offers assessment of the precision of quantitative methods. It allows computation of reliable and representative single-laboratory method validation parameters such as in-house reproducibility standard deviation, LOD₅₀ or RLOD values because it provides information on the variability of these values under different measurement conditions. The factorial approach requires fewer test results in order to obtain similar or higher levels of reliability compared to the conventional approach.