

---

**Infuzijska oprema za uporabo v medicini - 14. del: Sponke in regulatorji pretoka transfuzijskih in infuzijskih naprav brez stika s tekočino (ISO 8536-14:2016)**

Infusion equipment for medical use - Part 14: Clamps and flow regulators for transfusion and infusion equipment without fluid contact (ISO 8536-14:2016)

Infusionsgeräte zur medizinischen Verwendung - Teil 14: Klemmen und Durchflussregler für Transfusions- und Infusionsgeräte ohne Flüssigkeitskontakt (ISO 8536-14:2016)

Matériel de perfusion à usage médical - Partie 14: Clamps et limiteurs de débit pour appareils de transfusion et de perfusion sans contact à fluide (ISO 8536-14:2016)

[https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-](https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018)

[4eb2b51e6d8d/sist-en-iso-8536-14-2018](https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018)

**Ta slovenski standard je istoveten z: EN ISO 8536-14:2018**

---

**ICS:**

11.040.20	Transfuzijska, infuzijska in injekcijska oprema	Transfusion, infusion and injection equipment
-----------	---	---

**SIST EN ISO 8536-14:2018**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN ISO 8536-14:2018

<https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018>

EUROPEAN STANDARD

**EN ISO 8536-14**

NORME EUROPÉENNE

EUROPÄISCHE NORM

February 2018

ICS 11.040.20

English Version

## Infusion equipment for medical use - Part 14: Clamps and flow regulators for transfusion and infusion equipment without fluid contact (ISO 8536-14:2016)

Matériel de perfusion à usage médical - Partie 14:  
Clamps et limiteurs de débit pour appareils de  
transfusion et de perfusion sans contact à fluide (ISO  
8536-14:2016)

Infusionsgeräte zur medizinischen Verwendung - Teil  
14: Klemmen und Durchflussregler für Transfusions-  
und Infusionsgeräte ohne Flüssigkeitskontakt (ISO  
8536-14:2016)

This European Standard was approved by CEN on 19 December 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	Page
<b>European foreword</b> .....	<b>3</b>
<b>Annex ZA (informative) Relationship between this European standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered</b> .....	<b>5</b>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-14:2018](https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018)  
<https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018>

## European foreword

This document (EN ISO 8536-14:2018) has been prepared by Technical Committee ISO/TC 76 “Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use” in collaboration with Technical Committee CEN/TC 205 “Non-active medical devices” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2018, and conflicting national standards shall be withdrawn at the latest by February 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN ISO 8536-14:2018](https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-46297cc0d318/iso-8536-14-2018)

<https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-46297cc0d318/iso-8536-14-2018>

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard “within the meaning of Annex ZA”, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE The way in which these references documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

**Table — Correlations between normative references and dated EN and ISO standards**

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 10993-1	EN ISO 10993-1:2009	ISO 10993-1:2009

EN ISO 8536-14:2018 (E)

## Endorsement notice

The text of ISO 8536-14:2016 has been approved by CEN as EN ISO 8536-14:2018 without any modification.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-14:2018](https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018)

<https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018>

**Annex ZA**  
(informative)  
**Relationship between this European standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered**

This European Standard has been prepared under a Commission's standardization request [M/295 concerning the development of European Standards related to medical devices] to provide one voluntary means of conforming to essential requirements of Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169]

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC. This means that risks have to be reduced "as far as possible", "to a minimum", "to the lowest possible level", "minimized" or "removed", according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard. <https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018>

**Table ZA.1 — Correspondence between this European standard and Annex I of Directive 93/42/EEC [OJ L 169]**

Essential Requirements of Directive 93/42/EEC	Clause(s)/subclause(s) of this EN	Remarks/Notes
7.2	4, 5, 6	<p>Clause 4 covers ER 7.2 in respect of controlling fluid flow and avoiding puncturing of flexible tubing only.</p> <p>Clause 5 covers ER 7.2 in respect of material physical performance as follows:</p> <ul style="list-style-type: none"> <li>• Tubing specifications</li> <li>• Operating temperature</li> <li>• Construction</li> <li>• Flow rates</li> </ul> <p>Clause 6 covers ER 7.2 in respect of the following items only:</p> <ul style="list-style-type: none"> <li>• Tubing specifications</li> <li>• Operating temperature</li> <li>• Construction</li> <li>• Flow rates</li> </ul> <p>The part of ER 7.2 relating to packaging is not addressed.</p>
7.5	6.3, A.1	<p>Clause 6.3 covers ER 7.5, first para, first sentence in respect of controlling the flow rate of substances being delivered via the infusion set only.</p> <p>A.1 covers ER 7.5, first para, first sentence in respect of controlling the flow rate of substances being delivered via the infusion set only.</p> <p>ER 7.5, first para, second sentence and ER 7.5, paragraphs two and three are not covered.</p>
7.6	6.3, A.1	<p>ER 7.6 is covered by Clause 6.3 but only to the extent that the clamp must ensure complete occlusion of the fluid pathway.</p> <p>ER 7.6 is covered by A.1 but only to the extent that the clamp must ensure complete occlusion of the fluid pathway.</p>



9.1	4	ER 9.1 is covered by Clause 4 in respect of the design of the clamp so that accidental operation of the clamp is avoided and flexible tubing is not damaged during operation only.
12.7.1	4	ER 12.7.1 is covered by Clause 4 in respect of the design of the clamp so that accidental operation of the clamp is avoided and flexible tubing is not damaged during operation only.

**WARNING 1** — Presumption of conformity stays valid only as long as a reference to the European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

**WARNING 2** — Other Union legislation may be applicable to the products falling within the scope of this standard.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8536-14:2018](https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018)

<https://standards.iteh.ai/catalog/standards/sist/3cbbbd6a-d415-44e6-a2e5-4eb2b51e6d8d/sist-en-iso-8536-14-2018>